

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM030723\
 Data File : BM038889.D
 Acq On : 07 Mar 2023 10:27
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020102

Quant Time: Mar 07 11:25:45 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM022823.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Mar 06 23:57:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.004	152	214862	20.000	ng/ul	0.00
20) Naphthalene-d8	10.822	136	1006104	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.639	164	644132	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.380	188	1401660	20.000	ng/ul	0.00
79) Chrysene-d12	21.562	240	1455842	20.000	ng/ul	0.00
88) Perylene-d12	24.015	264	1663030	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.375	96	44075	7.660	ng/uL	0.00
4) Pyridine-d5	3.799	84	322761	20.401	ng/ul	0.00
7) Phenol-d5	7.151	99	387344	19.229	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.328	67	250641	19.474	ng/ul	0.00
11) 2-Chlorophenol-d4	7.528	132	301166	20.064	ng/ul	0.00
15) 4-Methylphenol-d8	8.710	113	306516	19.315	ng/ul	0.00
21) Nitrobenzene-d5	9.169	128	153084	21.674	ng/ul	0.00
24) 2-Nitrophenol-d4	9.892	143	147480	22.283	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.433	165	303031	19.704	ng/ul	0.00
31) 4-Chloroaniline-d4	10.951	131	475791	18.561	ng/ul	0.00
46) Dimethylphthalate-d6	14.051	166	961662	19.276	ng/ul	0.00
49) Acenaphthylene-d8	14.333	160	1126354	19.316	ng/ul	0.00
54) 4-Nitrophenol-d4	14.821	143	190293	19.054	ng/ul	0.00
60) Fluorene-d10	15.627	176	856990	19.477	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.739	200	135540	18.600	ng/ul	0.00
73) Anthracene-d10	17.480	188	1341914	19.415	ng/ul	0.00
81) Pyrene-d10	19.768	212	1555719	19.429	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.850	264	1668088	19.904	ng/ul	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.416	88	48641	7.866	ng/uL	99
5) Pyridine	3.822	79	331228	19.448	ng/ul	97
6) Benzaldehyde	7.134	77	231852	23.765	ng/ul	98
8) Phenol	7.181	94	417842	19.380	ng/ul	98
10) Bis(2-Chloroethyl)ether	7.422	93	334992	19.516	ng/ul	99
12) 2-Chlorophenol	7.563	128	311839	20.171	ng/ul	98
13) 2-Methylphenol	8.445	108	308741	19.157	ng/ul	100
14) 2,2'-oxybis(1-Chloropr...	8.539	45	418441	19.419	ng/ul	99
16) Acetophenone	8.828	105	536725	19.460	ng/ul	99
17) N-Nitroso-di-n-propyla...	8.816	70	259088	20.453	ng/ul	99
18) 4-Methylphenol	8.769	108	344023	19.225	ng/ul	99
19) Hexachloroethane	9.092	117	126268	20.386	ng/ul	97
22) Nitrobenzene	9.210	77	384551	19.901	ng/ul	98
23) Isophorone	9.739	82	742391	18.888	ng/ul	98
25) 2-Nitrophenol	9.928	139	170121	21.676	ng/ul	99
26) 2,4-Dimethylphenol	9.981	107	374304	19.412	ng/ul	98
27) Bis(2-Chloroethoxy)met...	10.228	93	468407	19.420	ng/ul	98
29) 2,4-Dichlorophenol	10.457	162	306412	19.654	ng/ul	98
30) Naphthalene	10.869	128	1106743	19.278	ng/ul	100
32) 4-Chloroaniline	10.975	127	484532	18.668	ng/ul	99
33) Hexachlorobutadiene	11.157	225	187335	19.312	ng/ul	99
34) Caprolactam	11.733	113	94516	17.680	ng/ul	98
35) 4-Chloro-3-methylphenol	12.092	107	346316	19.649	ng/ul	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM030723\
 Data File : BM038889.D
 Acq On : 07 Mar 2023 10:27
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020102

Quant Time: Mar 07 11:25:45 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM022823.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Mar 06 23:57:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
36) 2-Methylnaphthalene	12.475	142	730943	19.506	ng/ul	100
37) 1-Methylnaphthalene	12.692	142	727230	19.261	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.839	216	376139	19.522	ng/ul	99
40) Hexachlorocyclopentadiene	12.822	237	207828	19.993	ng/ul	98
41) 2,4,6-Trichlorophenol	13.074	196	239541	19.710	ng/ul	97
42) 2,4,5-Trichlorophenol	13.139	196	267351	19.988	ng/ul	96
43) 1,1'-Biphenyl	13.480	154	997723	19.638	ng/ul	99
44) 2-Chloronaphthalene	13.522	162	779637	19.455	ng/ul	99
45) 2-Nitroaniline	13.721	65	202864	22.433	ng/ul	99
47) Dimethylphthalate	14.098	163	971329	19.212	ng/ul	99
48) 2,6-Dinitrotoluene	14.216	165	168500	21.061	ng/ul	98
50) Acenaphthylene	14.363	152	1248524	19.306	ng/ul	99
51) 3-Nitroaniline	14.539	138	190722	18.994	ng/ul	95
52) Acenaphthene	14.704	153	856385	19.238	ng/ul	99
53) 2,4-Dinitrophenol	14.739	184	78958	17.360	ng/ul	96
55) 4-Nitrophenol	14.833	109	155104	19.443	ng/ul	97
56) Dibenzofuran	15.039	168	1207750	19.543	ng/ul	98
57) 2,4-Dinitrotoluene	14.992	165	257473	20.818	ng/ul	99
58) 2,3,4,6-Tetrachlorophenol	15.257	232	233263	20.031	ng/ul	98
59) Diethylphthalate	15.457	149	980734	19.603	ng/ul	99
61) Fluorene	15.686	166	1005894	19.413	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.680	204	486058	19.431	ng/ul	99
63) 4-Nitroaniline	15.698	138	212798	20.489	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.757	198	142023	18.833	ng/ul	98
67) N-Nitrosodiphenylamine	15.892	169	831172	19.606	ng/ul	99
68) 4-Bromophenyl-phenylether	16.574	248	277174	19.212	ng/ul	98
69) Hexachlorobenzene	16.686	284	333408	19.080	ng/ul	100
70) Atrazine	16.839	200	270430	18.856	ng/ul	99
71) Pentachlorophenol	17.027	266	196255	18.472	ng/ul	99
72) Phenanthrene	17.427	178	1569210	19.377	ng/ul	99
74) Anthracene	17.515	178	1613826	19.571	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.445	216	384940	19.487	ng/uL	99
76) Pentachlorobenzene	14.957	250	395626	19.492	ng/uL	98
77) Carbazole	17.786	167	1470669	19.206	ng/ul	99
78) Di-n-butylphthalate	18.356	149	1615400	20.736	ng/ul	100
80) Fluoranthene	19.433	202	1802599	19.574	ng/ul	97
82) Pyrene	19.798	202	1968616	19.494	ng/ul	100
83) Butylbenzylphthalate	20.698	149	698433	26.029	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.480	252	576483	20.208	ng/ul	99
85) Benzo(a)anthracene	21.545	228	1991682	19.800	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.480	149	1082563	24.269	ng/ul	98
87) Chrysene	21.597	228	1911308	19.430	ng/ul	100
89) Di-n-octyl phthalate	22.427	149	1777848	21.261	ng/ul	100
90) Benzo(b)fluoranthene	23.262	252	2049170	19.845	ng/ul	99
91) Benzo(k)fluoranthene	23.315	252	2145281	20.225	ng/ul	99
93) Benzo(a)pyrene	23.903	252	1900044	19.864	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.568	276	2433994	19.250	ng/ul	100
95) Dibenzo(a,h)anthracene	26.591	278	2054045	19.256	ng/ul	99
96) Benzo(g,h,i)perylene	27.350	276	1963918	18.969	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :

BNA_M


ClientSampleId :


SSTD020102

```
Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM030723\  
Data File : BM038889.D  
Acq On    : 07 Mar 2023 10:27  
Operator   : CG/JU  
Sample     : SSTDCCC020  
Misc      :  
ALS Vial   : 2      Sample Multiplier: 1
```

Instrument :
BNA_M
ClientSampleId :
SSTD020102

Quant Time: Mar 07 11:25:45 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM022823.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Mar 06 23:57:30 2023
Response via : Initial Calibration









1,4-Dioxane
 Concen: 7.866 ng/uL
 RT: 3.416 min Scan# 21
 Delta R.T. -0.000 min
 Lab File: BM038889.D
 Acq: 07 Mar 2023 10:27




Instrument :
 BNA_M
 ClientSampleId :
 SST020102


Tgt Ion: 88 Resp: 48641
 Ion Ratio Lower Upper
 88 100
 43 27.1 21.5 32.3
 58 69.6 56.4 84.6



Pyridine
 Concen: 19.448 ng/uL
 RT: 3.822 min Scan# 91
 Delta R.T. -0.000 min
 Lab File: BM038889.D
 Acq: 07 Mar 2023 10:27

Tgt Ion: 79 Resp: 331228
 Ion Ratio Lower Upper
 79 100
 52 69.1 57.8 86.6
 51 32.9 27.8 41.6









Benzaldehyde
 Concen: 23.765 ng/ul
 RT: 7.134 min Scan# 655
 Delta R.T. -0.006 min
 Lab File: BM038889.D
 Acq: 07 Mar 2023 10:27




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020102


Tgt Ion: 77 Resp: 231852
 Ion Ratio Lower Upper
 77 100
 105 98.8 77.9 116.9
 106 94.9 78.3 117.5



Phenol
 Concen: 19.380 ng/ul
 RT: 7.181 min Scan# 662
 Delta R.T. -0.000 min
 Lab File: BM038889.D
 Acq: 07 Mar 2023 10:27

Tgt Ion: 94 Resp: 417842
 Ion Ratio Lower Upper
 94 100
 65 27.9 22.1 33.1
 66 34.0 28.6 43.0









#10
Bis(2-Chloroethyl)ether
Concen: 19.516 ng/ul
RT: 7.422 min Scan#
Delta R.T. -0.000 min
Lab File: BM038889.D
Acq: 07 Mar 2023 10:27




Instrument :
BNA_M
ClientSampleId :
SSTD020102


Tgt Ion:	93	Resp:	334992
Ion	Ratio	Lower	Upper
93	100		
63	74.0	60.3	90.5
95	32.1	25.8	38.6



#12
2-Chlorophenol
Concen: 20.171 ng/ul
RT: 7.563 min Scan# 727
Delta R.T. -0.000 min
Lab File: BM038889.D
Acq: 07 Mar 2023 10:27

Tgt Ion:	128	Resp:	311839
Ion	Ratio	Lower	Upper
128	100		
64	46.7	38.9	58.3
130	31.8	25.1	37.7









#13
2-Methylphenol
Concen: 19.157 ng/ul
RT: 8.445 min Scan# 877
Delta R.T. -0.000 min
Lab File: BM038889.D
Acq: 07 Mar 2023 10:27




Instrument :
BNA_M
ClientSampleId :
SSTD020102


Tgt Ion:108 Resp: 308741
Ion Ratio Lower Upper
108 100
107 86.0 69.0 103.6



#14
2,2'-oxybis(1-Chloropropane)
Concen: 19.419 ng/ul
RT: 8.539 min Scan# 893
Delta R.T. 0.006 min
Lab File: BM038889.D
Acq: 07 Mar 2023 10:27

Tgt Ion: 45 Resp: 418441
Ion Ratio Lower Upper
45 100
77 15.0 12.4 18.6
79 11.7 9.4 14.0









#16
Acetophenone
Concen: 19.460 ng/ul
RT: 8.828 min Scan# 942
Delta R.T. -0.006 min
Lab File: BM038889.D
Acq: 07 Mar 2023 10:27




Instrument :
BNA_M
ClientSampleId :
SSTD020102


Tgt Ion:	105	Resp:	536725
Ion	Ratio	Lower	Upper
105	100		
77	79.2	62.6	93.8
51	26.5	20.8	31.2



#17
N-Nitroso-di-n-propylamine
Concen: 20.453 ng/ul
RT: 8.816 min Scan# 940
Delta R.T. -0.000 min
Lab File: BM038889.D
Acq: 07 Mar 2023 10:27

Tgt Ion:	70	Resp:	259088
Ion	Ratio	Lower	Upper
70	100		
42	53.8	43.5	65.3
101	9.5	7.2	10.8
130	20.2	16.0	24.0





4-Methylphenol

Concen: 19.225 ng/ul

RT: 8.769 min Scan# 933

Delta R.T. -0.006 min

Lab File: BM038889.D

Acq: 07 Mar 2023 10:27

Instrument :

BNA_M

ClientSampleId :





SSTD020102

Tgt Ion:108 Resp: 344023

Ion Ratio Lower Upper

108 100

107 110.2 89.0 133.6



Hexachloroethane

Concen: 20.386 ng/ul

RT: 9.092 min Scan# 987

Delta R.T. -0.006 min

Lab File: BM038889.D

Acq: 07 Mar 2023 10:27




Tgt Ion:117 Resp: 126268


Ion Ratio Lower Upper

117 100

201 94.6 76.6 115.0

199 62.5 46.6 70.0





Nitrobenzene

Concen: 19.901 ng/ul

RT: 9.210 min Scan# 1007

Delta R.T. -0.000 min

Lab File: BM038889.D

Acq: 07 Mar 2023 10:27

Instrument :

BNA_M

ClientSampleId :

SSTD020102





Tgt Ion: 77 Resp: 384551

Ion Ratio Lower Upper

77 100

123 45.7 35.3 52.9

65 13.6 10.6 15.8



Isophorone

Concen: 18.888 ng/ul

RT: 9.739 min Scan# 1097

Delta R.T. -0.000 min

Lab File: BM038889.D

Acq: 07 Mar 2023 10:27




Tgt Ion: 82 Resp: 742391


Ion Ratio Lower Upper

82 100

95 6.8 5.2 7.8

138 17.2 13.1 19.7









#25(-)
2-Nitrophenol
Concen: 21.676 ng/ul
RT: 9.928 min Scan# 1129
Delta R.T. -0.000 min
Lab File: BM038889.D
Acq: 07 Mar 2023 10:27




Instrument :
BNA_M
ClientSampleId :
SSTD020102

Tgt Ion:	139	Resp:	170121
Ion	Ratio	Lower	Upper
139	100		
65	52.2	42.6	63.8
109	29.0	23.1	34.7



#26(-)
2,4-Dimethylphenol
Concen: 19.412 ng/ul
RT: 9.981 min Scan# 1138
Delta R.T. -0.006 min
Lab File: BM038889.D
Acq: 07 Mar 2023 10:27

Tgt Ion:	107	Resp:	374304
Ion	Ratio	Lower	Upper
107	100		
121	48.4	39.8	59.8
122	84.4	69.0	103.4



Abundance Scan 1180 (10.228 min): BM038872.D\data.ms (-1#23) (-)

Bis(2-Chloroethoxy)methane

Concen: 19.420 ng/ul

RT: 10.228 min Scan# 1180

Delta R.T. -0.000 min

Lab File: BM038889.D

Acq: 07 Mar 2023 10:27

Instrument :

BNA_M

ClientSampleId :

SSTD020102

Tgt Ion: 93 Resp: 468407

Ion Ratio Lower Upper

93 100

95 33.0 25.3 37.9

123 11.6 9.3 13.9

Abundance Scan 1180 (10.228 min): BM038889.D\data.ms

Ref

50

0

49.0 93.0 125.0 170.9 206.9

m/z-->

50 100 150 200 250

Abundance Scan 1180 (10.228 min): BM038889.D\data.ms

Raw

50

0

49.0 93.0 125.0 170.9 207.0 281.1

m/z-->

50 100 150 200 250

Abundance Scan 1180 (10.228 min): BM038889.D\data.ms (-1146) (-)

Sub

50

0

49.0 93.0 125.0 170.9 207.9 281.1

m/z-->

50 100 150 200 250

Abundance

300000

200000

100000

0

10.20

Time-->

Abundance Scan 1219 (10.457 min): BM038872.D\data.ms (-1#29) (-)

2,4-Dichlorophenol

Concen: 19.654 ng/ul

RT: 10.457 min Scan# 1219

Delta R.T. -0.000 min

Lab File: BM038889.D

Acq: 07 Mar 2023 10:27

Tgt Ion:162 Resp: 306412

Ion Ratio Lower Upper

162 100

164 64.4 52.6 78.8

98 37.7 29.0 43.6

Abundance Scan 1219 (10.457 min): BM038889.D\data.ms

Ref

50

0

63.0 98.0 161.9 207.8

m/z-->

50 100 150 200 250

Abundance Scan 1219 (10.457 min): BM038889.D\data.ms

Raw

50

0

63.0 97.9 129.9 161.9 206.9 280.8

m/z-->

50 100 150 200 250

Abundance Scan 1219 (10.457 min): BM038889.D\data.ms (-1185) (-)

Sub

50

0

63.0 97.9 129.9 161.9 208.0 280.8

m/z-->

50 100 150 200 250

Abundance

150000


100000

50000

0

10.457





Time-->



Naphthalene
Concen: 19.278 ng/ul
RT: 10.869 min Scan# 1289
Delta R.T. -0.000 min
Lab File: BM038889.D
Acq: 07 Mar 2023 10:27




Instrument :
BNA_M
ClientSampleId :
SSTD020102

Tgt Ion:128 Resp: 1106743
Ion Ratio Lower Upper
128 100
129 10.7 8.6 12.8
127 12.8 10.2 15.4




4-Chloroaniline
Concen: 18.668 ng/ul
RT: 10.975 min Scan# 1307
Delta R.T. -0.000 min
Lab File: BM038889.D
Acq: 07 Mar 2023 10:27

Tgt Ion:127 Resp: 484532
Ion Ratio Lower Upper
127 100
129 32.3 26.3 39.5





Abundance Scan 1339 (11.163 min): BM038872.D\data.ms (-1332) (-)




Hexachlorobutadiene
Concen: 19.312 ng/ul
RT: 11.157 min Scan# 1339
Delta R.T. -0.006 min
Lab File: BM038889.D
Acq: 07 Mar 2023 10:27

Instrument :
BNA_M
ClientSampleId :
SSTD020102


Tgt Ion:	225	Resp:	187335
Ion	Ratio	Lower	Upper
225	100		
223	62.9	50.9	76.3
227	62.6	50.4	75.6



Abundance Scan 1338 (11.157 min): BM038889.D\data.ms (-1305) (-)





Abundance Scan 1437 (11.739 min): BM038872.D\data.ms (-1434) (-)





Caprolactam
Concen: 17.680 ng/ul
RT: 11.733 min Scan# 1436
Delta R.T. -0.006 min
Lab File: BM038889.D
Acq: 07 Mar 2023 10:27

Tgt Ion:	113	Resp:	94516
Ion	Ratio	Lower	Upper
113	100		
55	167.2	137.0	205.6
56	125.7	99.9	149.9



Abundance Scan 1436 (11.733 min): BM038889.D\data.ms (-1403) (-)









4-Chloro-3-methylphenol
Concen: 19.649 ng/ul
RT: 12.092 min Scan# 1497
Delta R.T. -0.000 min
Lab File: BM038889.D
Acq: 07 Mar 2023 10:27




Instrument :
BNA_M
ClientSampleId :
SSTD020102


Tgt Ion:	107	Resp:	346316
Ion	Ratio	Lower	Upper
107	100		
144	25.3	21.1	31.7
142	80.9	64.7	97.1



2-Methylnaphthalene
Concen: 19.506 ng/ul
RT: 12.475 min Scan# 1562
Delta R.T. -0.000 min
Lab File: BM038889.D
Acq: 07 Mar 2023 10:27

Tgt Ion:	142	Resp:	730943
Ion	Ratio	Lower	Upper
142	100		
141	88.6	71.0	106.6









1-Methylnaphthalene
 Concen: 19.261 ng/ul
 RT: 12.692 min Scan# 1599
 Delta R.T. -0.000 min
 Lab File: BM038889.D
 Acq: 07 Mar 2023 10:27




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020102

Tgt Ion:	142	Resp:	727230
Ion	Ratio	Lower	Upper
142	100		
141	89.8	73.4	110.0
116	3.7	3.2	4.8



1,2,4,5-Tetrachlorobenzene
 Concen: 19.522 ng/ul
 RT: 12.839 min Scan# 1624
 Delta R.T. -0.000 min
 Lab File: BM038889.D
 Acq: 07 Mar 2023 10:27

Tgt Ion:	216	Resp:	376139
Ion	Ratio	Lower	Upper
216	100		
214	77.8	62.2	93.4
179	19.9	16.4	24.6
108	18.3	15.3	22.9



Abundance Scan 1621 (12.822 min): BM038872.D\data.ms (-1549) (-)

Hexachlorocyclopentadiene

Concen: 19.993 ng/ul

RT: 12.822 min Scan# 1621

Delta R.T. -0.000 min

Lab File: BM038889.D

Acq: 07 Mar 2023 10:27

Instrument :

BNA_M

ClientSampleId :

SSTD020102

Tgt Ion:237 Resp: 207828



Ion Ratio Lower Upper

237 100


235 62.0 48.1 72.1

272 12.1 10.0 15.0

Abundance Scan 1621 (12.822 min): BM038889.D\data.ms



Abundance Scan 1621 (12.822 min): BM038889.D\data.ms (-1587) (-)



Abundance Scan 1664 (13.075 min): BM038872.D\data.ms (-1651) (-)

2,4,6-Trichlorophenol

Concen: 19.710 ng/ul

RT: 13.074 min Scan# 1664

Delta R.T. -0.000 min

Lab File: BM038889.D

Acq: 07 Mar 2023 10:27

Tgt Ion:196 Resp: 239541



Ion Ratio Lower Upper

196 100


198 97.6 74.8 112.2

200 30.9 24.3 36.5


Abundance Scan 1664 (13.074 min): BM038889.D\data.ms



Abundance Scan 1664 (13.074 min): BM038889.D\data.ms (-1630) (-)



Abundance Scan 1676 (13.145 min): BM038872.D\data.ms (-1640) (-)



2,4,5-Trichlorophenol
Concen: 19.988 ng/ul
RT: 13.139 min Scan# 1676
Delta R.T. -0.006 min
Lab File: BM038889.D
Acq: 07 Mar 2023 10:27

Instrument :

BNA_M

ClientSampleId :

SSTD020102

Tgt Ion:196 Resp: 267351



Ion Ratio Lower Upper

196 100


198 93.3 77.5 116.3

200 29.3 24.8 37.2


Abundance Scan 1675 (13.139 min): BM038889.D\data.ms



Abundance Scan 1675 (13.139 min): BM038889.D\data.ms (-1659) (-)



Abundance Scan 1733 (13.480 min): BM038872.D\data.ms (-1743) (-)



1,1'-Biphenyl
Concen: 19.638 ng/ul
RT: 13.480 min Scan# 1733
Delta R.T. -0.000 min
Lab File: BM038889.D
Acq: 07 Mar 2023 10:27

Tgt Ion:154 Resp: 997723



Ion Ratio Lower Upper

154 100


153 38.3 31.3 46.9


76 13.3 11.0 16.6

Abundance Scan 1733 (13.480 min): BM038889.D\data.ms



Abundance Scan 1733 (13.480 min): BM038889.D\data.ms (-1699) (-)









2-Chloronaphthalene
Concen: 19.455 ng/ul
RT: 13.522 min Scan# 1740
Delta R.T. -0.000 min
Lab File: BM038889.D
Acq: 07 Mar 2023 10:27




Instrument :
BNA_M
ClientSampleId :
SSTD020102


Tgt Ion:	162	Resp:	779637
Ion	Ratio	Lower	Upper
162	100		
164	32.5	26.6	40.0
127	38.9	31.5	47.3



2-Nitroaniline
Concen: 22.433 ng/ul
RT: 13.721 min Scan# 1774
Delta R.T. -0.000 min
Lab File: BM038889.D
Acq: 07 Mar 2023 10:27

Tgt Ion:	65	Resp:	202864
Ion	Ratio	Lower	Upper
65	100		
92	67.3	54.2	81.2
138	106.8	84.3	126.5









Dimethylphthalate
Concen: 19.212 ng/ul
RT: 14.098 min Scan# 1838
Delta R.T. -0.000 min
Lab File: BM038889.D
Acq: 07 Mar 2023 10:27




Instrument :
BNA_M
ClientSampleId :
SSTD020102


Tgt Ion:	163	Resp:	971329
Ion	Ratio	Lower	Upper
163	100		
194	3.9	3.0	4.6
164	9.9	8.4	12.6



2,6-Dinitrotoluene
Concen: 21.061 ng/ul
RT: 14.216 min Scan# 1858
Delta R.T. -0.000 min
Lab File: BM038889.D
Acq: 07 Mar 2023 10:27

Tgt Ion:	165	Resp:	168500
Ion	Ratio	Lower	Upper
165	100		
89	62.8	51.4	77.2
121	23.3	19.4	29.2









Acenaphthylene
Concen: 19.306 ng/ul
RT: 14.363 min Scan# 1883
Delta R.T. -0.000 min
Lab File: BM038889.D
Acq: 07 Mar 2023 10:27




Instrument :
BNA_M
ClientSampleId :
SSTD020102


Tgt Ion:152 Resp: 1248524
Ion Ratio Lower Upper
152 100
151 19.8 15.9 23.9
153 13.6 10.4 15.6



3-Nitroaniline
Concen: 18.994 ng/ul
RT: 14.539 min Scan# 1913
Delta R.T. -0.000 min
Lab File: BM038889.D
Acq: 07 Mar 2023 10:27

Tgt Ion:138 Resp: 190722
Ion Ratio Lower Upper
138 100
108 11.9 10.3 15.5
92 116.4 97.7 146.5









Acenaphthene
Concen: 19.238 ng/ul
RT: 14.704 min Scan# 1941
Delta R.T. -0.000 min
Lab File: BM038889.D
Acq: 07 Mar 2023 10:27




Instrument :
BNA_M
ClientSampleId :
SSTD020102


Tgt Ion:153 Resp: 856385
Ion Ratio Lower Upper
153 100
152 48.1 39.1 58.7
154 89.5 70.4 105.6



2,4-Dinitrophenol
Concen: 17.360 ng/ul
RT: 14.739 min Scan# 1947
Delta R.T. -0.000 min
Lab File: BM038889.D
Acq: 07 Mar 2023 10:27

Tgt Ion:184 Resp: 78958
Ion Ratio Lower Upper
184 100
63 73.8 61.9 92.9
154 60.6 50.7 76.1









4-Nitrophenol
 Concen: 19.443 ng/ul
 RT: 14.833 min Scan# 1963
 Delta R.T. -0.000 min
 Lab File: BM038889.D
 Acq: 07 Mar 2023 10:27




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020102


Tgt Ion:109 Resp: 155104
 Ion Ratio Lower Upper
 109 100
 139 131.1 108.8 163.2
 65 132.2 107.8 161.6



Dibenzofuran
 Concen: 19.543 ng/ul
 RT: 15.039 min Scan# 1998
 Delta R.T. -0.000 min
 Lab File: BM038889.D
 Acq: 07 Mar 2023 10:27

Tgt Ion:168 Resp: 1207750
 Ion Ratio Lower Upper
 168 100
 139 37.3 30.9 46.3









2,4-Dinitrotoluene
Concen: 20.818 ng/ul
RT: 14.992 min Scan# 1991
Delta R.T. -0.006 min
Lab File: BM038889.D
Acq: 07 Mar 2023 10:27




Instrument :
BNA_M
ClientSampleId :
SSTD020102


Tgt Ion:	165	Resp:	257473
Ion	Ratio	Lower	Upper
165	100		
63	47.6	38.9	58.3
182	2.5	2.2	3.2



2,3,4,6-Tetrachlorophenol
Concen: 20.031 ng/ul
RT: 15.257 min Scan# 2035
Delta R.T. -0.000 min
Lab File: BM038889.D
Acq: 07 Mar 2023 10:27

Tgt Ion:	232	Resp:	233263
Ion	Ratio	Lower	Upper
232	100		
131	48.8	40.6	60.8
130	2.5	2.0	3.0
166	32.2	25.9	38.9









Diethylphthalate
 Concen: 19.603 ng/ul
 RT: 15.457 min Scan# 2069
 Delta R.T. -0.006 min
 Lab File: BM038889.D
 Acq: 07 Mar 2023 10:27




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020102

Tgt Ion:149 Resp: 980734
 Ion Ratio Lower Upper
 149 100
 177 21.2 17.2 25.8
 150 12.2 10.1 15.1



Fluorene
 Concen: 19.413 ng/ul
 RT: 15.686 min Scan# 2108
 Delta R.T. -0.000 min
 Lab File: BM038889.D
 Acq: 07 Mar 2023 10:27

Tgt Ion:166 Resp: 1005894
 Ion Ratio Lower Upper
 166 100
 165 97.3 78.2 117.4
 167 13.5 10.5 15.7



Abundance Scan 2107 (15.680 min): BM038872.D\data.ms (-2107) (-)

4-Chlorophenyl-phenylether

Concen: 19.431 ng/ul

RT: 15.680 min Scan# 2107

Delta R.T. -0.000 min

Lab File: BM038889.D

Acq: 07 Mar 2023 10:27

Instrument :

BNA_M

ClientSampleId :

SSTD020102

Tgt Ion:204 Resp: 486058

Ion Ratio Lower Upper

204 100

206 33.1 26.3 39.5

141 57.4 45.5 68.3

Abundance Scan 2107 (15.680 min): BM038889.D\data.ms

Ref

166.0

204.0

77.0

115.0

39.0

281.0

m/z-->

Abundance Scan 2107 (15.680 min): BM038889.D\data.ms (-2073) (-)

Ref

166.0

204.0

77.0

115.0

39.0

m/z-->

Abundance

15.680

300000

200000

100000

0

Time-->

Abundance Scan 2110 (15.698 min): BM038872.D\data.ms (-2110) (-)

4-Nitroaniline

Concen: 20.489 ng/ul

RT: 15.698 min Scan# 2110

Delta R.T. -0.000 min

Lab File: BM038889.D

Acq: 07 Mar 2023 10:27

Tgt Ion:138 Resp: 212798

Ion Ratio Lower Upper

138 100

92 51.5 41.8 62.8

108 81.8 65.0 97.6

Abundance Scan 2110 (15.698 min): BM038889.D\data.ms

Ref

166.0

204.0

65.0

108.0

280.9

m/z-->

Abundance Scan 2110 (15.698 min): BM038889.D\data.ms (-2076) (-)

Ref

166.0

204.0

65.0

108.0

280.9

m/z-->

Abundance

15.698

150000


100000

50000

0

Time-->

Abundance Scan 2120 (15.757 min): BM038872.D\data.ms (-2150) (-)





4,6-Dinitro-2-methylphenol
Concen: 18.833 ng/ul
RT: 15.757 min Scan# 2120
Delta R.T. -0.000 min
Lab File: BM038889.D
Acq: 07 Mar 2023 10:27


Instrument :
BNA_M
ClientSampleId :
SSTD020102

Tgt Ion:198 Resp: 142023
Ion Ratio Lower Upper
198 100
182 5.2 3.5 5.3
77 27.1 21.0 31.6


Abundance Scan 2120 (15.757 min): BM038889.D\data.ms



Abundance Scan 2120 (15.757 min): BM038889.D\data.ms (-2086) (-)





Abundance Scan 2143 (15.892 min): BM038872.D\data.ms (-2167) (-)




N-Nitrosodiphenylamine
Concen: 19.606 ng/ul
RT: 15.892 min Scan# 2143
Delta R.T. -0.000 min
Lab File: BM038889.D
Acq: 07 Mar 2023 10:27

Tgt Ion:169 Resp: 831172
Ion Ratio Lower Upper
169 100
168 66.4 53.5 80.3
167 36.7 29.1 43.7

Abundance Scan 2143 (15.892 min): BM038889.D\data.ms




Abundance Scan 2143 (15.892 min): BM038889.D\data.ms (-2109) (-)






Abundance Scan 2259 (16.574 min): BM038872.D\data.ms (-2259) (-)

4-Bromophenyl-phenylether
Concen: 19.212 ng/ul
RT: 16.574 min Scan# 2259
Delta R.T. -0.000 min
Lab File: BM038889.D
Acq: 07 Mar 2023 10:27

Instrument :
BNA_M
ClientSampleId :
SSTD020102




Tgt Ion:	248	Resp:	277174
Ion	Ratio	Lower	Upper
248	100		
250	97.0	77.6	116.4
141	76.2	63.7	95.5







Abundance Scan 2278 (16.686 min): BM038872.D\data.ms (-2278) (-)

Hexachlorobenzene
Concen: 19.080 ng/ul
RT: 16.686 min Scan# 2278
Delta R.T. -0.000 min
Lab File: BM038889.D
Acq: 07 Mar 2023 10:27



Tgt Ion:	284	Resp:	333408
Ion	Ratio	Lower	Upper
284	100		
142	37.0	29.4	44.2
249	30.2	24.0	36.0





Atrazine

Concen: 18.856 ng/ul

RT: 16.839 min Scan# 2304

Delta R.T. -0.006 min

Lab File: BM038889.D

Acq: 07 Mar 2023 10:27

Instrument :

BNA_M

ClientSampleId :

SSTD020102





Tgt Ion:200 Resp: 270430

Ion Ratio Lower Upper

200 100

173 27.3 21.3 31.9

215 47.4 38.0 57.0



Pentachlorophenol

Concen: 18.472 ng/ul

RT: 17.027 min Scan# 2336

Delta R.T. -0.000 min

Lab File: BM038889.D

Acq: 07 Mar 2023 10:27




Tgt Ion:266 Resp: 196255


Ion Ratio Lower Upper

266 100

264 63.5 51.0 76.6

268 65.6 51.7 77.5









Scan 2404 (17.427 min): BM038872.D\data.ms (-2372) (-)
 Phenanthrene
 Concen: 19.377 ng/ul
 RT: 17.427 min Scan# 2404
 Delta R.T. -0.000 min
 Lab File: BM038889.D
 Acq: 07 Mar 2023 10:27




Instrument :
 BNA_M
 ClientSampleId :
 SST020102

Tgt Ion:	178	Resp:	1569210
Ion	Ratio	Lower	Upper
178	100		
179	15.0	12.3	18.5
176	18.9	15.7	23.5




Scan 2419 (17.515 min): BM038872.D\data.ms (-2374) (-)
 Anthracene
 Concen: 19.571 ng/ul
 RT: 17.515 min Scan# 2419
 Delta R.T. -0.000 min
 Lab File: BM038889.D
 Acq: 07 Mar 2023 10:27

Tgt Ion:	178	Resp:	1613826
Ion	Ratio	Lower	Upper
178	100		
179	15.7	12.5	18.7
176	19.3	14.9	22.3






Abundance Scan 1727 (13.445 min): BM038872.D\data.ms (-1725) (-)




1,2,3,4-Tetrachlorobenzene
Concen: 19.487 ng/uL
RT: 13.445 min Scan# 1727
Delta R.T. -0.000 min
Lab File: BM038889.D
Acq: 07 Mar 2023 10:27

Instrument :
BNA_M
ClientSampleId :
SSTD020102

Tgt Ion:	216	Resp:	384940
Ion	Ratio	Lower	Upper
216	100		
214	78.1	62.4	93.6
218	47.1	38.8	58.2







Abundance Scan 1984 (14.957 min): BM038872.D\data.ms (-1976) (-)



Pentachlorobenzene
Concen: 19.492 ng/uL
RT: 14.957 min Scan# 1984
Delta R.T. -0.000 min
Lab File: BM038889.D
Acq: 07 Mar 2023 10:27

Tgt Ion:	250	Resp:	395626
Ion	Ratio	Lower	Upper
250	100		
248	60.8	49.6	74.4
252	62.7	51.2	76.8









Carbazole
Concen: 19.206 ng/ul
RT: 17.786 min Scan# 2465
Delta R.T. -0.000 min
Lab File: BM038889.D
Acq: 07 Mar 2023 10:27




Instrument :
BNA_M
ClientSampleId :
SSTD020102


Tgt Ion:167 Resp: 1470669
Ion Ratio Lower Upper
167 100
166 21.8 17.2 25.8
139 13.4 10.6 16.0



Di-n-butylphthalate
Concen: 20.736 ng/ul
RT: 18.356 min Scan# 2562
Delta R.T. -0.000 min
Lab File: BM038889.D
Acq: 07 Mar 2023 10:27

Tgt Ion:149 Resp: 1615400
Ion Ratio Lower Upper
149 100
150 9.0 7.3 10.9
104 5.1 4.2 6.4





#39 (-)

Fluoranthene

Concen: 19.574 ng/ul

RT: 19.433 min Scan# 2746

Delta R.T. -0.006 min

Lab File: BM038889.D

Acq: 07 Mar 2023 10:27

Instrument :

BNA_M

ClientSampleId :

SSTD020102





Tgt Ion:202 Resp: 1802599

Ion Ratio Lower Upper

202 100

101 12.6 9.0 13.4

100 9.5 6.9 10.3



#82 (-)

Pyrene

Concen: 19.494 ng/ul

RT: 19.798 min Scan# 2807

Delta R.T. -0.000 min

Lab File: BM038889.D

Acq: 07 Mar 2023 10:27




Tgt Ion:202 Resp: 1968616


Ion Ratio Lower Upper

202 100

101 14.0 11.2 16.8

100 11.5 9.1 13.7









Scan 2960 (20.698 min): BM038872.D\data.ms (-3089) (-)
 Butylbenzylphthalate
 Concen: 26.029 ng/ul
 RT: 20.698 min Scan# 2960
 Delta R.T. -0.000 min
 Lab File: BM038889.D
 Acq: 07 Mar 2023 10:27




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020102

Tgt Ion:	149	Resp:	698433
Ion	Ratio	Lower	Upper
149	100		
91	73.1	60.5	90.7
206	19.8	15.8	23.8




Scan 3092 (21.474 min): BM038872.D\data.ms (-3089) (-)
 3,3'-Dichlorobenzidine
 Concen: 20.208 ng/ul
 RT: 21.480 min Scan# 3093
 Delta R.T. 0.006 min
 Lab File: BM038889.D
 Acq: 07 Mar 2023 10:27

Tgt Ion:	252	Resp:	576483
Ion	Ratio	Lower	Upper
252	100		
254	65.3	51.8	77.6
126	11.3	9.0	13.4



Abundance Scan 3104 (21.545 min): BM038872.D\data.ms (-3085) (-)





Benzo(a)anthracene
Concen: 19.800 ng/ul
RT: 21.545 min Scan# 3104
Delta R.T. -0.000 min
Lab File: BM038889.D
Acq: 07 Mar 2023 10:27


Instrument :
BNA_M
ClientSampleId :
SSTD020102

Tgt Ion:	228	Resp:	1991682
Ion	Ratio	Lower	Upper
228	100		
229	19.6	15.8	23.6
226	26.8	21.4	32.2


Abundance Scan 3104 (21.545 min): BM038889.D\data.ms



Abundance Scan 3104 (21.545 min): BM038889.D\data.ms (-3071000) (-)





Abundance Scan 3093 (21.480 min): BM038872.D\data.ms (-3086) (-)




Bis(2-ethylhexyl)phthalate
Concen: 24.269 ng/ul
RT: 21.480 min Scan# 3093
Delta R.T. -0.000 min
Lab File: BM038889.D
Acq: 07 Mar 2023 10:27


Tgt Ion:	149	Resp:	1082563
Ion	Ratio	Lower	Upper
149	100		
167	27.9	21.4	32.2
279	4.0	3.1	4.7

Abundance Scan 3093 (21.480 min): BM038889.D\data.ms



Abundance Scan 3093 (21.480 min): BM038889.D\data.ms (-3059) (-)









Scan 3113 (21.597 min): BM038872.D\data.ms (-3109) (-)
 Chrysene
 Concen: 19.430 ng/ul
 RT: 21.597 min Scan# 3113
 Delta R.T. -0.000 min
 Lab File: BM038889.D
 Acq: 07 Mar 2023 10:27

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020102




Tgt Ion:228 Resp: 1911308


Ion	Ratio	Lower	Upper
228	100		
226	29.8	23.6	35.4
229	19.7	15.8	23.6



Scan 3254 (22.427 min): BM038872.D\data.ms (-3249) (-)
 Di-n-octyl phthalate
 Concen: 21.261 ng/ul
 RT: 22.427 min Scan# 3254
 Delta R.T. -0.000 min
 Lab File: BM038889.D
 Acq: 07 Mar 2023 10:27

Tgt Ion:149 Resp: 1777848









Benzo(b)fluoranthene
 Concen: 19.845 ng/ul
 RT: 23.262 min Scan# 3396
 Delta R.T. -0.000 min
 Lab File: BM038889.D
 Acq: 07 Mar 2023 10:27

Instrument :
 BNA_M
 ClientSampleId :
 SST020102

Tgt Ion:252 Resp: 2049170




Ion	Ratio	Lower	Upper
252	100		
253	21.9	17.2	25.8
125	10.8	8.6	13.0




Benzo(k)fluoranthene
 Concen: 20.225 ng/ul
 RT: 23.315 min Scan# 3405
 Delta R.T. -0.000 min
 Lab File: BM038889.D
 Acq: 07 Mar 2023 10:27

Tgt Ion:252 Resp: 2145281

Ion	Ratio	Lower	Upper
252	100		
253	21.7	17.7	26.5
125	10.4	8.4	12.6









Benzo(a)pyrene
Concen: 19.864 ng/ul
RT: 23.903 min Scan# 3499
Delta R.T. -0.000 min
Lab File: BM038889.D
Acq: 07 Mar 2023 10:27

Instrument :
BNA_M
ClientSampleId :
SSTD020102

Tgt Ion:252 Resp: 1900044




Ion	Ratio	Lower	Upper
252	100		
253	21.4	17.8	26.8
125	11.9	9.4	14.2



Indeno(1,2,3-cd)pyrene
Concen: 19.250 ng/ul
RT: 26.568 min Scan# 3958
Delta R.T. -0.006 min
Lab File: BM038889.D
Acq: 07 Mar 2023 10:27

Tgt Ion:276 Resp: 2433994


Ion	Ratio	Lower	Upper
276	100		
138	25.2	20.2	30.4
277	24.6	19.8	29.8






Abundance Scan 3962 (26.591 min): BM038872.D\data.ms (-3995) (-)

Dibenzo(a,h)anthracene
Concen: 19.256 ng/ul
RT: 26.591 min Scan# 3995
Delta R.T. -0.000 min
Lab File: BM038889.D
Acq: 07 Mar 2023 10:27

Instrument :
BNA_M
ClientSampleId :
SSTD020102



Tgt Ion:	278	Resp:	2054045
Ion	Ratio	Lower	Upper
278	100		
139	17.1	13.4	20.2
279	23.6	19.1	28.7



Abundance Scan 4092 (27.356 min): BM038872.D\data.ms (-4096) (-)

Benzo(g,h,i)perylene
Concen: 18.969 ng/ul
RT: 27.350 min Scan# 4091
Delta R.T. -0.006 min
Lab File: BM038889.D
Acq: 07 Mar 2023 10:27

Tgt Ion:	276	Resp:	1963918
Ion	Ratio	Lower	Upper
276	100		
138	23.5	19.4	29.0
277	23.7	18.7	28.1

