

Data Path : Z:\svoasrv\HPCHEM1\BNA\_M\Data\BM030723\  
 Data File : BM038921.D  
 Acq On : 08 Mar 2023 07:30  
 Operator : CG/JU  
 Sample : 01746-15  
 Misc :  
 ALS Vial : 34 Sample Multiplier: 1

Instrument :  
 BNA\_M  
 ClientSampleId :

Quant Time: Mar 08 11:55:57 2023  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_M\Methods\SFAM-EPA-BM022823.M  
 Quant Title : SVOA CALIBRATION  
 QLast Update : Wed Mar 08 04:25:35 2023  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |
|-------------------------------|--------|------|----------|--------|-------|----------|
| -----                         |        |      |          |        |       |          |
| Internal Standards            |        |      |          |        |       |          |
| 1) 1,4-Dichlorobenzene-d4     | 7.998  | 152  | 299848   | 20.000 | ng/ul | 0.00     |
| 20) Naphthalene-d8            | 10.816 | 136  | 1372017  | 20.000 | ng/ul | 0.00     |
| 38) Acenaphthene-d10          | 14.633 | 164  | 862739   | 20.000 | ng/ul | 0.00     |
| 64) Phenanthrene-d10          | 17.380 | 188  | 1852482  | 20.000 | ng/ul | 0.00     |
| 79) Chrysene-d12              | 21.556 | 240  | 1783234  | 20.000 | ng/ul | 0.00     |
| 88) Perylene-d12              | 24.009 | 264  | 2060129  | 20.000 | ng/ul | 0.00     |
| System Monitoring Compounds   |        |      |          |        |       |          |
| 3) 1,4-Dioxane-d8             | 3.375  | 96   | 34503    | 4.297  | ng/uL | 0.00     |
| 4) Pyridine-d5                | 3.804  | 84   | 108944   | 4.934  | ng/ul | 0.00     |
| 7) Phenol-d5                  | 7.151  | 99   | 160453   | 5.708  | ng/ul | 0.00     |
| 9) Bis-(2-Chloroethyl)eth...  | 7.322  | 67   | 541434   | 30.144 | ng/ul | 0.00     |
| 11) 2-Chlorophenol-d4         | 7.528  | 132  | 495385   | 23.649 | ng/ul | 0.00     |
| 15) 4-Methylphenol-d8         | 8.704  | 113  | 305453   | 13.792 | ng/ul | 0.00     |
| 21) Nitrobenzene-d5           | 9.163  | 128  | 341819   | 35.488 | ng/ul | 0.00     |
| 24) 2-Nitrophenol-d4          | 9.886  | 143  | 335000   | 37.117 | ng/ul | 0.00     |
| 28) 2,4-Dichlorophenol-d3     | 10.428 | 165  | 559643   | 26.685 | ng/ul | 0.00     |
| 31) 4-Chloroaniline-d4        | 10.945 | 131  | 792871   | 22.681 | ng/ul | 0.00     |
| 46) Dimethylphthalate-d6      | 14.045 | 166  | 2170975  | 32.489 | ng/ul | 0.00     |
| 49) Acenaphthylene-d8         | 14.327 | 160  | 2452761  | 31.405 | ng/ul | 0.00     |
| 54) 4-Nitrophenol-d4          | 14.815 | 143  | 62208    | 4.650  | ng/ul | 0.00     |
| 60) Fluorene-d10              | 15.627 | 176  | 1950821  | 33.102 | ng/ul | 0.00     |
| 65) 4,6-Dinitro-2-methylph... | 15.733 | 200  | 283025   | 29.388 | ng/ul | 0.00     |
| 73) Anthracene-d10            | 17.474 | 188  | 3340123  | 36.566 | ng/ul | 0.00     |
| 81) Pyrene-d10                | 19.768 | 212  | 3905589  | 39.821 | ng/ul | 0.00     |
| 92) Benzo(a)pyrene-d12        | 23.850 | 264  | 4016215  | 38.685 | ng/ul | 0.00     |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_M\Data\BM030723\  
Data File : BM038921.D  
Acq On : 08 Mar 2023 07:30  
Operator : CG/JU  
Sample : 01746-15  
Misc :  
ALS Vial : 34 Sample Multiplier: 1

Instrument :  
BNA\_M  
ClientSampleId :

Quant Time: Mar 08 11:55:57 2023  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_M\Methods\SFAM-EPA-BM022823.M  
Quant Title : SVOA CALIBRATION  
QLast Update : Wed Mar 08 04:25:35 2023  
Response via : Initial Calibration

