

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM030723\
 Data File : BM038925.D
 Acq On : 08 Mar 2023 09:59
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :

Quant Time: Mar 08 11:56:37 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM022823.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Mar 08 04:25:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.998	152	336935	20.000	ng/ul	0.00
20) Naphthalene-d8	10.816	136	1485733	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.633	164	935049	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.380	188	1968059	20.000	ng/ul	0.00
79) Chrysene-d12	21.556	240	1916764	20.000	ng/ul	0.00
88) Perylene-d12	24.009	264	2112251	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.375	96	69421	7.693	ng/uL	0.00
4) Pyridine-d5	3.798	84	495428	19.970	ng/ul	0.00
7) Phenol-d5	7.151	99	576353	18.246	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.328	67	366757	18.171	ng/ul	0.00
11) 2-Chlorophenol-d4	7.528	132	457149	19.421	ng/ul	0.00
15) 4-Methylphenol-d8	8.704	113	454255	18.254	ng/ul	0.00
21) Nitrobenzene-d5	9.169	128	226625	21.728	ng/ul	0.00
24) 2-Nitrophenol-d4	9.892	143	234639	24.008	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.427	165	445567	19.619	ng/ul	0.00
31) 4-Chloroaniline-d4	10.951	131	682111	18.019	ng/ul	0.00
46) Dimethylphthalate-d6	14.045	166	1331969	18.392	ng/ul	0.00
49) Acenaphthylene-d8	14.333	160	1592929	18.819	ng/ul	0.00
54) 4-Nitrophenol-d4	14.821	143	273518	18.866	ng/ul	0.00
60) Fluorene-d10	15.627	176	1194078	18.695	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.739	200	195878	19.144	ng/ul	0.00
73) Anthracene-d10	17.480	188	1837820	18.938	ng/ul	0.00
81) Pyrene-d10	19.768	212	2050435	19.450	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.850	264	2092950	19.662	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.416	88	76002	7.837	ng/uL	98
5) Pyridine	3.822	79	533393	19.972	ng/ul	95
6) Benzaldehyde	7.134	77	335964	21.960	ng/ul	98
8) Phenol	7.181	94	613270	18.139	ng/ul	99
10) Bis(2-Chloroethyl)ether	7.422	93	491958	18.277	ng/ul	99
12) 2-Chlorophenol	7.557	128	468531	19.326	ng/ul	98
13) 2-Methylphenol	8.439	108	453641	17.950	ng/ul	99
14) 2,2'-oxybis(1-Chloropr...	8.533	45	608488	18.008	ng/ul	100
16) Acetophenone	8.828	105	774153	17.899	ng/ul	100
17) N-Nitroso-di-n-propyla...	8.816	70	377140	18.985	ng/ul	98
18) 4-Methylphenol	8.769	108	510216	18.182	ng/ul	99
19) Hexachloroethane	9.086	117	193551	19.928	ng/ul	97
22) Nitrobenzene	9.210	77	568693	19.930	ng/ul	99
23) Isophorone	9.739	82	1079049	18.591	ng/ul	98
25) 2-Nitrophenol	9.922	139	258283	22.285	ng/ul	98
26) 2,4-Dimethylphenol	9.980	107	545478	19.157	ng/ul	100
27) Bis(2-Chloroethoxy)met...	10.222	93	675486	18.965	ng/ul	100
29) 2,4-Dichlorophenol	10.457	162	449928	19.543	ng/ul	99
30) Naphthalene	10.869	128	1593712	18.799	ng/ul	100
32) 4-Chloroaniline	10.974	127	696334	18.168	ng/ul	98
33) Hexachlorobutadiene	11.151	225	271456	18.950	ng/ul	99
34) Caprolactam	11.739	113	138879	17.592	ng/ul	98
35) 4-Chloro-3-methylphenol	12.086	107	494219	18.989	ng/ul	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM030723\
 Data File : BM038925.D
 Acq On : 08 Mar 2023 09:59
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 38 Sample Multiplier: 1


Instrument :
 BNA_M
 ClientSampleId :

Quant Time: Mar 08 11:56:37 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM022823.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Mar 08 04:25:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
36) 2-Methylnaphthalene	12.469	142	1044572	18.876	ng/ul	99
37) 1-Methylnaphthalene	12.686	142	1043374	18.713	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.833	216	527078	18.845	ng/ul	99
40) Hexachlorocyclopentadiene	12.816	237	295439	19.578	ng/ul	98
41) 2,4,6-Trichlorophenol	13.068	196	349639	19.819	ng/ul	98
42) 2,4,5-Trichlorophenol	13.139	196	387515	19.958	ng/ul	98
43) 1,1'-Biphenyl	13.474	154	1408660	19.100	ng/ul	98
44) 2-Chloronaphthalene	13.515	162	1106749	19.026	ng/ul	98
45) 2-Nitroaniline	13.715	65	304304	23.181	ng/ul	99
47) Dimethylphthalate	14.092	163	1342157	18.287	ng/ul	99
48) 2,6-Dinitrotoluene	14.210	165	247429	21.305	ng/ul	99
50) Acenaphthylene	14.357	152	1762947	18.779	ng/ul	100
51) 3-Nitroaniline	14.539	138	287931	19.753	ng/ul	95
52) Acenaphthene	14.698	153	1203241	18.620	ng/ul	98
53) 2,4-Dinitrophenol	14.739	184	118900	18.008	ng/ul	95
55) 4-Nitrophenol	14.833	109	216876	18.728	ng/ul	97
56) Dibenzofuran	15.033	168	1682462	18.754	ng/ul	99
57) 2,4-Dinitrotoluene	14.992	165	370374	20.630	ng/ul	97
58) 2,3,4,6-Tetrachlorophenol	15.257	232	326289	19.302	ng/ul	97
59) Diethylphthalate	15.457	149	1342648	18.487	ng/ul	99
61) Fluorene	15.680	166	1395891	18.558	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.674	204	671055	18.480	ng/ul	98
63) 4-Nitroaniline	15.698	138	299832	19.888	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.751	198	201376	19.018	ng/ul	96
67) N-Nitrosodiphenylamine	15.886	169	1131567	19.010	ng/ul	99
68) 4-Bromophenyl-phenylether	16.568	248	370415	18.286	ng/ul	99
69) Hexachlorobenzene	16.680	284	447252	18.229	ng/ul	98
70) Atrazine	16.839	200	373144	18.530	ng/ul	98
71) Pentachlorophenol	17.027	266	272979	18.299	ng/ul	99
72) Phenanthrene	17.421	178	2115870	18.608	ng/ul	100
74) Anthracene	17.515	178	2176988	18.803	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.439	216	536540	19.345	ng/uL	100
76) Pentachlorobenzene	14.951	250	541950	19.017	ng/uL	98
77) Carbazole	17.780	167	1985972	18.472	ng/ul	99
78) Di-n-butylphthalate	18.350	149	2196340	20.079	ng/ul	100
80) Fluoranthene	19.433	202	2382633	19.651	ng/ul	99
82) Pyrene	19.797	202	2590698	19.485	ng/ul	98
83) Butylbenzylphthalate	20.692	149	972214	27.519	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.474	252	769294	20.482	ng/ul	99
85) Benzo(a)anthracene	21.544	228	2566322	19.377	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.474	149	1458076	24.827	ng/ul	98
87) Chrysene	21.597	228	2481784	19.162	ng/ul	100
89) Di-n-octyl phthalate	22.415	149	2424079	22.824	ng/ul	100
90) Benzo(b)fluoranthene	23.262	252	2657455	20.263	ng/ul	100
91) Benzo(k)fluoranthene	23.309	252	2602483	19.318	ng/ul	99
93) Benzo(a)pyrene	23.897	252	2353289	19.370	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.562	276	2825672	17.595	ng/ul	99
95) Dibenzo(a,h)anthracene	26.579	278	2383418	17.592	ng/ul	99
96) Benzo(g,h,i)perylene	27.344	276	2238536	17.023	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed





Instrument :
BNA_M
ClientSampleId :



1,4-Dioxane
 Concen: 7.837 ng/uL
 RT: 3.416 min Scan# 2
 Delta R.T. 0.006 min
 Lab File: BM038925.D
 Acq: 08 Mar 2023 09:59




Instrument :
 BNA_M
 ClientSampleId :


Tgt Ion:	88	Resp:	76002
Ion	Ratio	Lower	Upper
88	100		
43	26.3	21.5	32.3
58	68.4	56.4	84.6



Pyridine
 Concen: 19.972 ng/uL
 RT: 3.822 min Scan# 91
 Delta R.T. -0.000 min
 Lab File: BM038925.D
 Acq: 08 Mar 2023 09:59

Tgt Ion:	79	Resp:	533393
Ion	Ratio	Lower	Upper
79	100		
52	67.8	57.8	86.6
51	32.4	27.8	41.6









Benzaldehyde
 Concen: 21.960 ng/ul
 RT: 7.134 min Scan# 654
 Delta R.T. -0.000 min
 Lab File: BM038925.D
 Acq: 08 Mar 2023 09:59




Instrument :
 BNA_M
 ClientSampleId :


Tgt Ion: 77 Resp: 335964
 Ion Ratio Lower Upper
 77 100
 105 101.6 77.9 116.9
 106 98.1 78.3 117.5



Phenol
 Concen: 18.139 ng/ul
 RT: 7.181 min Scan# 662
 Delta R.T. 0.006 min
 Lab File: BM038925.D
 Acq: 08 Mar 2023 09:59

Tgt Ion: 94 Resp: 613270
 Ion Ratio Lower Upper
 94 100
 65 27.1 22.1 33.1
 66 34.8 28.6 43.0









#10
Bis(2-Chloroethyl)ether
Concen: 18.277 ng/ul
RT: 7.422 min Scan# 726
Delta R.T. 0.006 min
Lab File: BM038925.D
Acq: 08 Mar 2023 09:59




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:	93	Resp:	491958
Ion	Ratio	Lower	Upper
93	100		
63	74.2	60.3	90.5
95	32.9	25.8	38.6



#12
2-Chlorophenol
Concen: 19.326 ng/ul
RT: 7.557 min Scan# 726
Delta R.T. -0.000 min
Lab File: BM038925.D
Acq: 08 Mar 2023 09:59

Tgt Ion:	128	Resp:	468531
Ion	Ratio	Lower	Upper
128	100		
64	46.6	38.9	58.3
130	32.0	25.1	37.7









Scan 877 (8.445 min): BM038889.D\data.ms (-870) (-)
 #13
 2-Methylphenol
 Concen: 17.950 ng/u1
 RT: 8.439 min Scan# 876
 Delta R.T. -0.000 min
 Lab File: BM038925.D
 Acq: 08 Mar 2023 09:59




Instrument :
 BNA_M
 ClientSampleId :


Tgt Ion:108 Resp: 453641
 Ion Ratio Lower Upper
 108 100
 107 87.0 69.0 103.6



Scan 893 (8.539 min): BM038889.D\data.ms (-884) (-)
 #14
 2,2'-oxybis(1-Chloropropane)
 Concen: 18.008 ng/u1
 RT: 8.533 min Scan# 892
 Delta R.T. -0.006 min
 Lab File: BM038925.D
 Acq: 08 Mar 2023 09:59

Tgt Ion: 45 Resp: 608488
 Ion Ratio Lower Upper
 45 100
 77 15.5 12.4 18.6
 79 11.9 9.4 14.0





Acetophenone

Concen: 17.899 ng/ul

RT: 8.828 min Scan# 942

Delta R.T. -0.000 min

Lab File: BM038925.D

Acq: 08 Mar 2023 09:59

Instrument :

BNA_M

ClientSampleId :





Tgt Ion:105 Resp: 774153

Ion Ratio Lower Upper

105 100

77 78.5 62.6 93.8

51 25.6 20.8 31.2



N-Nitroso-di-n-propylamine

Concen: 18.985 ng/ul

RT: 8.816 min Scan# 940

Delta R.T. 0.006 min

Lab File: BM038925.D

Acq: 08 Mar 2023 09:59

Tgt Ion: 70 Resp: 377140




Ion Ratio Lower Upper


70 100

42 53.8 43.5 65.3

101 10.0 7.2 10.8

130 22.6 16.0 24.0









#18
4-Methylphenol
Concen: 18.182 ng/ul
RT: 8.769 min Scan# 932
Delta R.T. -0.000 min
Lab File: BM038925.D
Acq: 08 Mar 2023 09:59




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:108 Resp: 510216
Ion Ratio Lower Upper
108 100
107 112.1 89.0 133.6



#19
Hexachloroethane
Concen: 19.928 ng/ul
RT: 9.086 min Scan# 986
Delta R.T. -0.006 min
Lab File: BM038925.D
Acq: 08 Mar 2023 09:59

Tgt Ion:117 Resp: 193551
Ion Ratio Lower Upper
117 100
201 92.6 76.6 115.0
199 56.0 46.6 70.0





Nitrobenzene

Concen: 19.930 ng/ul

RT: 9.210 min Scan#

Delta R.T. -0.000 min

Lab File: BM038925.D

Acq: 08 Mar 2023 09:59

Instrument :

BNA_M

ClientSampleId :





Tgt Ion: 77 Resp: 568693

Ion Ratio Lower Upper

77 100

123 44.6 35.3 52.9

65 13.1 10.6 15.8



Isophorone

Concen: 18.591 ng/ul

RT: 9.739 min Scan# 1097

Delta R.T. 0.006 min

Lab File: BM038925.D

Acq: 08 Mar 2023 09:59




Tgt Ion: 82 Resp: 1079049


Ion Ratio Lower Upper

82 100

95 6.7 5.2 7.8

138 17.5 13.1 19.7









2-Nitrophenol
Concen: 22.285 ng/ul
RT: 9.922 min Scan# 1125
Delta R.T. -0.000 min
Lab File: BM038925.D
Acq: 08 Mar 2023 09:59




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:	139	Resp:	258283
Ion	Ratio	Lower	Upper
139	100		
65	53.8	42.6	63.8
109	30.4	23.1	34.7



2,4-Dimethylphenol
Concen: 19.157 ng/ul
RT: 9.980 min Scan# 1138
Delta R.T. -0.000 min
Lab File: BM038925.D
Acq: 08 Mar 2023 09:59

Tgt Ion:	107	Resp:	545478
Ion	Ratio	Lower	Upper
107	100		
121	50.1	39.8	59.8
122	86.1	69.0	103.4









Bis(2-Chloroethoxy)methane
 Concen: 18.965 ng/u1
 RT: 10.222 min Scan# 1179
 Delta R.T. -0.000 min
 Lab File: BM038925.D
 Acq: 08 Mar 2023 09:59




Instrument :
 BNA_M
 ClientSampleId :


Tgt Ion: 93 Resp: 675486
 Ion Ratio Lower Upper
 93 100
 95 31.8 25.3 37.9
 123 11.5 9.3 13.9



2,4-Dichlorophenol
 Concen: 19.543 ng/u1
 RT: 10.457 min Scan# 1219
 Delta R.T. -0.000 min
 Lab File: BM038925.D
 Acq: 08 Mar 2023 09:59

Tgt Ion: 162 Resp: 449928
 Ion Ratio Lower Upper
 162 100
 164 66.7 52.6 78.8
 98 37.0 29.0 43.6









Naphthalene
Concen: 18.799 ng/ul
RT: 10.869 min Scan# 1289
Delta R.T. 0.006 min
Lab File: BM038925.D
Acq: 08 Mar 2023 09:59




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:128 Resp: 1593712
Ion Ratio Lower Upper
128 100
129 10.7 8.6 12.8
127 13.0 10.2 15.4



4-Chloroaniline
Concen: 18.168 ng/ul
RT: 10.974 min Scan# 1307
Delta R.T. -0.000 min
Lab File: BM038925.D
Acq: 08 Mar 2023 09:59

Tgt Ion:127 Resp: 696334
Ion Ratio Lower Upper
127 100
129 32.0 26.3 39.5









Hexachlorobutadiene
Concen: 18.950 ng/ul
RT: 11.151 min Scan# 1339
Delta R.T. -0.006 min
Lab File: BM038925.D
Acq: 08 Mar 2023 09:59




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:	225	Resp:	271456
Ion	Ratio	Lower	Upper
225	100		
223	63.0	50.9	76.3
227	63.3	50.4	75.6



Caprolactam
Concen: 17.592 ng/ul
RT: 11.739 min Scan# 1437
Delta R.T. 0.006 min
Lab File: BM038925.D
Acq: 08 Mar 2023 09:59

Tgt Ion:	113	Resp:	138879
Ion	Ratio	Lower	Upper
113	100		
55	168.3	137.0	205.6
56	122.1	99.9	149.9









4-Chloro-3-methylphenol
Concen: 18.989 ng/ul
RT: 12.086 min Scan# 1496
Delta R.T. -0.000 min
Lab File: BM038925.D
Acq: 08 Mar 2023 09:59




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:	107	Resp:	494219
Ion	Ratio	Lower	Upper
107	100		
144	25.7	21.1	31.7
142	81.6	64.7	97.1



2-Methylnaphthalene
Concen: 18.876 ng/ul
RT: 12.469 min Scan# 1561
Delta R.T. -0.000 min
Lab File: BM038925.D
Acq: 08 Mar 2023 09:59

Tgt Ion:	142	Resp:	1044572
Ion	Ratio	Lower	Upper
142	100		
141	87.7	71.0	106.6









1-Methylnaphthalene
 Concen: 18.713 ng/ul
 RT: 12.686 min Scan# 1598
 Delta R.T. -0.000 min
 Lab File: BM038925.D
 Acq: 08 Mar 2023 09:59




Instrument :
 BNA_M
 ClientSampleId :

Tgt Ion:142 Resp: 1043374
 Ion Ratio Lower Upper
 142 100
 141 91.1 73.4 110.0
 116 3.8 3.2 4.8



1,2,4,5-Tetrachlorobenzene
 Concen: 18.845 ng/ul
 RT: 12.833 min Scan# 1623
 Delta R.T. -0.000 min
 Lab File: BM038925.D
 Acq: 08 Mar 2023 09:59


Tgt Ion:216 Resp: 527078
 Ion Ratio Lower Upper
 216 100
 214 77.7 62.2 93.4
 179 21.3 16.4 24.6
 108 19.6 15.3 22.9






Abundance Scan 1621 (12.822 min): BM038889.D\data.ms (-1549) (-)

Hexachlorocyclopentadiene
Concen: 19.578 ng/ul
RT: 12.816 min Scan# 1619
Delta R.T. -0.000 min
Lab File: BM038925.D
Acq: 08 Mar 2023 09:59

Instrument :
BNA_M
ClientSampleId :







Tgt Ion	Ratio	Lower	Upper
237	100		
235	61.7	48.1	72.1
272	13.3	10.0	15.0



Abundance Scan 1664 (13.074 min): BM038889.D\data.ms (-1541) (-)

2,4,6-Trichlorophenol
Concen: 19.819 ng/ul
RT: 13.068 min Scan# 1663
Delta R.T. -0.000 min
Lab File: BM038925.D
Acq: 08 Mar 2023 09:59


Tgt Ion	Ratio	Lower	Upper
196	100		
198	95.2	74.8	112.2
200	31.2	24.3	36.5






Abundance Scan 1675 (13.139 min): BM038889.D\data.ms (-1640) (-)

2,4,5-Trichlorophenol
Concen: 19.958 ng/ul
RT: 13.139 min Scan# 1640
Delta R.T. -0.000 min
Lab File: BM038925.D
Acq: 08 Mar 2023 09:59

Instrument :
BNA_M
ClientSampleId :







Tgt Ion:	196	Resp:	387515
Ion	Ratio	Lower	Upper
196	100		
198	94.8	77.5	116.3
200	30.8	24.8	37.2




Abundance Scan 1733 (13.480 min): BM038889.D\data.ms (-1742) (-)

1,1'-Biphenyl
Concen: 19.100 ng/ul
RT: 13.474 min Scan# 1732
Delta R.T. -0.000 min
Lab File: BM038925.D
Acq: 08 Mar 2023 09:59

Tgt Ion:	154	Resp:	1408660
Ion	Ratio	Lower	Upper
154	100		
153	40.3	31.3	46.9
76	13.6	11.0	16.6









2-Chloronaphthalene
Concen: 19.026 ng/ul
RT: 13.515 min Scan# 1739
Delta R.T. -0.000 min
Lab File: BM038925.D
Acq: 08 Mar 2023 09:59




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:162 Resp: 1106749
Ion Ratio Lower Upper
162 100
164 32.3 26.6 40.0
127 38.4 31.5 47.3



2-Nitroaniline
Concen: 23.181 ng/ul
RT: 13.715 min Scan# 1773
Delta R.T. -0.000 min
Lab File: BM038925.D
Acq: 08 Mar 2023 09:59

Tgt Ion: 65 Resp: 304304
Ion Ratio Lower Upper
65 100
92 68.3 54.2 81.2
138 105.9 84.3 126.5









Dimethylphthalate
Concen: 18.287 ng/u1
RT: 14.092 min Scan# 1837
Delta R.T. -0.000 min
Lab File: BM038925.D
Acq: 08 Mar 2023 09:59




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:163 Resp: 1342157
Ion Ratio Lower Upper
163 100
194 4.0 3.0 4.6
164 9.8 8.4 12.6



2,6-Dinitrotoluene
Concen: 21.305 ng/u1
RT: 14.210 min Scan# 1857
Delta R.T. -0.000 min
Lab File: BM038925.D
Acq: 08 Mar 2023 09:59

Tgt Ion:165 Resp: 247429
Ion Ratio Lower Upper
165 100
89 64.0 51.4 77.2
121 23.4 19.4 29.2





(-1850) (-)

Acenaphthylene

Concen: 18.779 ng/ul

RT: 14.357 min Scan#

Delta R.T. -0.000 min

Lab File: BM038925.D

Acq: 08 Mar 2023 09:59

Instrument :

BNA_M

ClientSampleId :





Tgt Ion:152 Resp: 1762947

Ion Ratio Lower Upper

152 100

151 19.8 15.9 23.9

153 13.1 10.4 15.6



(-1851) (-)

3-Nitroaniline

Concen: 19.753 ng/ul

RT: 14.539 min Scan# 1913

Delta R.T. 0.006 min

Lab File: BM038925.D

Acq: 08 Mar 2023 09:59




Tgt Ion:138 Resp: 287931


Ion Ratio Lower Upper

138 100

108 11.3 10.3 15.5

92 116.0 97.7 146.5









Acenaphthene
 Concen: 18.620 ng/ul
 RT: 14.698 min Scan# 1940
 Delta R.T. -0.000 min
 Lab File: BM038925.D
 Acq: 08 Mar 2023 09:59




Instrument :
 BNA_M
 ClientSampleId :


Tgt Ion:153 Resp: 1203241
 Ion Ratio Lower Upper
 153 100
 152 48.5 39.1 58.7
 154 90.0 70.4 105.6



2,4-Dinitrophenol
 Concen: 18.008 ng/ul
 RT: 14.739 min Scan# 1947
 Delta R.T. -0.000 min
 Lab File: BM038925.D
 Acq: 08 Mar 2023 09:59

Tgt Ion:184 Resp: 118900
 Ion Ratio Lower Upper
 184 100
 63 70.4 61.9 92.9
 154 62.1 50.7 76.1









4-Nitrophenol
 Concen: 18.728 ng/ul
 RT: 14.833 min Scan# 1963
 Delta R.T. -0.000 min
 Lab File: BM038925.D
 Acq: 08 Mar 2023 09:59




Instrument :
 BNA_M
 ClientSampleId :


Tgt Ion:109 Resp: 216876
 Ion Ratio Lower Upper
 109 100
 139 132.3 108.8 163.2
 65 132.1 107.8 161.6



Dibenzofuran
 Concen: 18.754 ng/ul
 RT: 15.033 min Scan# 1997
 Delta R.T. -0.000 min
 Lab File: BM038925.D
 Acq: 08 Mar 2023 09:59

Tgt Ion:168 Resp: 1682462
 Ion Ratio Lower Upper
 168 100
 139 37.8 30.9 46.3









2,4-Dinitrotoluene
Concen: 20.630 ng/u1
RT: 14.992 min Scan# 1990
Delta R.T. -0.000 min
Lab File: BM038925.D
Acq: 08 Mar 2023 09:59




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:	165	Resp:	370374
Ion	Ratio	Lower	Upper
165	100		
63	46.8	38.9	58.3
182	2.6	2.2	3.2



2,3,4,6-Tetrachlorophenol
Concen: 19.302 ng/u1
RT: 15.257 min Scan# 2035
Delta R.T. -0.000 min
Lab File: BM038925.D
Acq: 08 Mar 2023 09:59

Tgt Ion:	232	Resp:	326289
Ion	Ratio	Lower	Upper
232	100		
131	48.4	40.6	60.8
130	2.2	2.0	3.0
166	30.4	25.9	38.9









Diethylphthalate
 Concen: 18.487 ng/ul
 RT: 15.457 min Scan# 2069
 Delta R.T. -0.000 min
 Lab File: BM038925.D
 Acq: 08 Mar 2023 09:59

Instrument :
 BNA_M
 ClientSampleId :

Tgt Ion:149 Resp: 1342648




Ion	Ratio	Lower	Upper
149	100		
177	21.9	17.2	25.8
150	12.2	10.1	15.1



Fluorene
 Concen: 18.558 ng/ul
 RT: 15.680 min Scan# 2107
 Delta R.T. -0.000 min
 Lab File: BM038925.D
 Acq: 08 Mar 2023 09:59

Tgt Ion:166 Resp: 1395891

Ion	Ratio	Lower	Upper
166	100		
165	96.4	78.2	117.4
167	13.4	10.5	15.7



Abundance Scan 2107 (15.680 min): BM038889.D\data.ms (-2#62) (-)

4-Chlorophenyl-phenylether

Concen: 18.480 ng/ul

RT: 15.674 min Scan# 2106

Delta R.T. -0.000 min


Lab File: BM038925.D

Acq: 08 Mar 2023 09:59

Instrument :

BNA_M

ClientSampleId :






Tgt Ion:204 Resp: 671055

Ion Ratio Lower Upper

204 100

206 32.8 26.3 39.5

141 59.4 45.5 68.3



Abundance Scan 2110 (15.698 min): BM038889.D\data.ms (-2#62) (-)

4-Nitroaniline

Concen: 19.888 ng/ul

RT: 15.698 min Scan# 2110

Delta R.T. 0.006 min

Lab File: BM038925.D

Acq: 08 Mar 2023 09:59




Tgt Ion:138 Resp: 299832

Ion Ratio Lower Upper


138 100

92 51.4 41.8 62.8

108 80.2 65.0 97.6



Abundance Scan 2120 (15.757 min): BM038889.D\data.ms (-2169) (-)





4,6-Dinitro-2-methylphenol
Concen: 19.018 ng/ul
RT: 15.751 min Scan# 2119
Delta R.T. -0.000 min
Lab File: BM038925.D
Acq: 08 Mar 2023 09:59


Instrument :
BNA_M
ClientSampleId :

Tgt Ion	Ratio	Lower	Upper
198	100		
182	4.8	3.5	5.3
77	28.4	21.0	31.6


Abundance Scan 2119 (15.751 min): BM038925.D\data.ms



Abundance Scan 2119 (15.751 min): BM038925.D\data.ms (-2085) (-)





Abundance Scan 2143 (15.892 min): BM038889.D\data.ms (-2167) (-)




N-Nitrosodiphenylamine
Concen: 19.010 ng/ul
RT: 15.886 min Scan# 2142
Delta R.T. -0.000 min
Lab File: BM038925.D
Acq: 08 Mar 2023 09:59

Tgt Ion	Ratio	Lower	Upper
169	100		
168	66.5	53.5	80.3
167	35.8	29.1	43.7

Abundance Scan 2142 (15.886 min): BM038925.D\data.ms



Abundance Scan 2142 (15.886 min): BM038925.D\data.ms (-2108) (-)



Abundance Scan 2259 (16.574 min): BM038889.D\data.ms (-2259) (-)

4-Bromophenyl-phenylether

Concen: 18.286 ng/ul

RT: 16.568 min Scan# 2258

Delta R.T. -0.000 min


Lab File: BM038925.D

Acq: 08 Mar 2023 09:59

Instrument :

BNA_M

ClientSampleId :






Tgt Ion:248 Resp: 370415

Ion Ratio Lower Upper

248 100

250 98.0 77.6 116.4

141 79.3 63.7 95.5



Abundance Scan 2278 (16.686 min): BM038889.D\data.ms (-2278) (-)

Hexachlorobenzene

Concen: 18.229 ng/ul

RT: 16.680 min Scan# 2277

Delta R.T. -0.000 min

Lab File: BM038925.D

Acq: 08 Mar 2023 09:59




Tgt Ion:284 Resp: 447252


Ion Ratio Lower Upper

284 100

142 38.3 29.4 44.2

249 30.4 24.0 36.0









Scan 2304 (16.839 min): BM038889.D\data.ms (-2770) (-)
 Atrazine
 Concen: 18.530 ng/ul
 RT: 16.839 min Scan# 2304
 Delta R.T. -0.000 min
 Lab File: BM038925.D
 Acq: 08 Mar 2023 09:59




Instrument :
 BNA_M
 ClientSampleId :


Tgt Ion:	200	Resp:	373144
Ion	Ratio	Lower	Upper
200	100		
173	27.3	21.3	31.9
215	48.7	38.0	57.0



Scan 2336 (17.027 min): BM038889.D\data.ms (-2770) (-)
 Pentachlorophenol
 Concen: 18.299 ng/ul
 RT: 17.027 min Scan# 2336
 Delta R.T. -0.000 min
 Lab File: BM038925.D
 Acq: 08 Mar 2023 09:59

Tgt Ion:	266	Resp:	272979
Ion	Ratio	Lower	Upper
266	100		
264	62.3	51.0	76.6
268	64.6	51.7	77.5









Scan 2404 (17.427 min): BM038889.D\data.ms (-2339) (-)
 Phenanthrene
 Concen: 18.608 ng/u1
 RT: 17.421 min Scan# 2403
 Delta R.T. -0.000 min
 Lab File: BM038925.D
 Acq: 08 Mar 2023 09:59




Instrument :
 BNA_M
 ClientSampleId :

Tgt Ion:	178	Resp:	2115870
Ion	Ratio	Lower	Upper
178	100		
179	15.2	12.3	18.5
176	19.4	15.7	23.5




Scan 2419 (17.515 min): BM038889.D\data.ms (-2339) (-)
 Anthracene
 Concen: 18.803 ng/u1
 RT: 17.515 min Scan# 2419
 Delta R.T. 0.006 min
 Lab File: BM038925.D
 Acq: 08 Mar 2023 09:59

Tgt Ion:	178	Resp:	2176988
Ion	Ratio	Lower	Upper
178	100		
179	15.5	12.5	18.7
176	18.8	14.9	22.3






Abundance Scan 1727 (13.445 min): BM038889.D\data.ms (-175) (-)




1,2,3,4-Tetrachlorobenzene
Concen: 19.345 ng/uL
RT: 13.439 min Scan# 1726
Delta R.T. -0.000 min
Lab File: BM038925.D
Acq: 08 Mar 2023 09:59

Instrument :
BNA_M
ClientSampleId :

Tgt Ion:	216	Resp:	536540
Ion	Ratio	Lower	Upper
216	100		
214	77.8	62.4	93.6
218	48.7	38.8	58.2







Abundance Scan 1984 (14.957 min): BM038889.D\data.ms (-177) (-)



Pentachlorobenzene
Concen: 19.017 ng/uL
RT: 14.951 min Scan# 1983
Delta R.T. -0.000 min
Lab File: BM038925.D
Acq: 08 Mar 2023 09:59

Tgt Ion:	250	Resp:	541950
Ion	Ratio	Lower	Upper
250	100		
248	63.3	49.6	74.4
252	62.5	51.2	76.8









Scan 2465 (17.786 min): BM038889.D\data.ms (-2430) (-)
 Carbazole
 Concen: 18.472 ng/ul
 RT: 17.780 min Scan# 2464
 Delta R.T. -0.000 min
 Lab File: BM038925.D
 Acq: 08 Mar 2023 09:59

Instrument :
 BNA_M
 ClientSampleId :

Tgt Ion:167 Resp: 1985972




Ion	Ratio	Lower	Upper
167	100		
166	21.0	17.2	25.8
139	13.0	10.6	16.0




Scan 2562 (18.356 min): BM038889.D\data.ms (-2570) (-)
 Di-n-butylphthalate
 Concen: 20.079 ng/ul
 RT: 18.350 min Scan# 2561
 Delta R.T. -0.000 min
 Lab File: BM038925.D
 Acq: 08 Mar 2023 09:59

Tgt Ion:149 Resp: 2196340

Ion	Ratio	Lower	Upper
149	100		
150	9.0	7.3	10.9
104	5.2	4.2	6.4



Abundance Scan 2745 (19.433 min): BM038889.D\data.ms (-2739) (-)



Fluoranthene

Concen: 19.651 ng/ul

RT: 19.433 min Scan# 2745

Delta R.T. -0.000 min

Lab File: BM038925.D

Acq: 08 Mar 2023 09:59

Instrument :

BNA_M

ClientSampleId :



Tgt Ion:202 Resp: 2382633

Ion Ratio Lower Upper


202 100

101 11.9 9.0 13.4


100 8.9 6.9 10.3



Abundance Scan 2745 (19.433 min): BM038925.D\data.ms (-2711) (-)



Abundance Scan 2807 (19.798 min): BM038889.D\data.ms (-2782) (-)



Pyrene

Concen: 19.485 ng/ul

RT: 19.797 min Scan# 2807

Delta R.T. 0.006 min

Lab File: BM038925.D

Acq: 08 Mar 2023 09:59



Tgt Ion:202 Resp: 2590698

Ion Ratio Lower Upper


202 100


101 13.4 11.2 16.8

100 10.8 9.1 13.7



Abundance Scan 2807 (19.797 min): BM038925.D\data.ms (-2772) (-)









Butylbenzylphthalate
 Concen: 27.519 ng/ul
 RT: 20.692 min Scan# 2959
 Delta R.T. -0.000 min
 Lab File: BM038925.D
 Acq: 08 Mar 2023 09:59




Instrument :
 BNA_M
 ClientSampleId :


Tgt Ion:	149	Resp:	972214
Ion	Ratio	Lower	Upper
149	100		
91	73.9	60.5	90.7
206	19.2	15.8	23.8



3,3'-Dichlorobenzidine
 Concen: 20.482 ng/ul
 RT: 21.474 min Scan# 3092
 Delta R.T. -0.000 min
 Lab File: BM038925.D
 Acq: 08 Mar 2023 09:59

Tgt Ion:	252	Resp:	769294
Ion	Ratio	Lower	Upper
252	100		
254	64.0	51.8	77.6
126	11.8	9.0	13.4









Scan 3104 (21.545 min): BM038889.D\data.ms (-3085) (-)
 Benzo(a)anthracene
 Concen: 19.377 ng/u1
 RT: 21.544 min Scan# 3104
 Delta R.T. 0.006 min
 Lab File: BM038925.D
 Acq: 08 Mar 2023 09:59




Instrument :
 BNA_M
 ClientSampleId :


Tgt Ion:	228	Resp:	2566322
Ion	Ratio	Lower	Upper
228	100		
229	19.4	15.8	23.6
226	26.6	21.4	32.2



Scan 3093 (21.480 min): BM038889.D\data.ms (-3086) (-)
 Bis(2-ethylhexyl)phthalate
 Concen: 24.827 ng/u1
 RT: 21.474 min Scan# 3092
 Delta R.T. 0.006 min
 Lab File: BM038925.D
 Acq: 08 Mar 2023 09:59

Tgt Ion:	149	Resp:	1458076
Ion	Ratio	Lower	Upper
149	100		
167	28.0	21.4	32.2
279	4.1	3.1	4.7









Scan 3113 (21.597 min): BM038889.D\data.ms (-3109) (-)
 Chrysene
 Concen: 19.162 ng/ul
 RT: 21.597 min Scan# 3113
 Delta R.T. 0.006 min
 Lab File: BM038925.D
 Acq: 08 Mar 2023 09:59

Instrument :
 BNA_M
 ClientSampleId :




Tgt Ion:228 Resp: 2481784


Ion	Ratio	Lower	Upper
228	100		
226	29.8	23.6	35.4
229	19.5	15.8	23.6



Scan 3254 (22.427 min): BM038889.D\data.ms (-3249) (-)
 Di-n-octyl phthalate
 Concen: 22.824 ng/ul
 RT: 22.415 min Scan# 3252
 Delta R.T. -0.000 min
 Lab File: BM038925.D
 Acq: 08 Mar 2023 09:59

Tgt Ion:149 Resp: 2424079









Benzo(b)fluoranthene
Concen: 20.263 ng/u1
RT: 23.262 min Scan# 3396
Delta R.T. 0.006 min
Lab File: BM038925.D
Acq: 08 Mar 2023 09:59




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:	252	Resp:	2657455
Ion	Ratio	Lower	Upper
252	100		
253	21.7	17.2	25.8
125	10.6	8.6	13.0



Benzo(k)fluoranthene
Concen: 19.318 ng/u1
RT: 23.309 min Scan# 3404
Delta R.T. -0.000 min
Lab File: BM038925.D
Acq: 08 Mar 2023 09:59

Tgt Ion:	252	Resp:	2602483
Ion	Ratio	Lower	Upper
252	100		
253	21.7	17.7	26.5
125	10.6	8.4	12.6









Benzo(a)pyrene
 Concen: 19.370 ng/u1
 RT: 23.897 min Scan# 3493
 Delta R.T. -0.000 min
 Lab File: BM038925.D
 Acq: 08 Mar 2023 09:59




Instrument :
 BNA_M
 ClientSampleId :

Tgt Ion:	252	Resp:	2353289
Ion	Ratio	Lower	Upper
252	100		
253	22.1	17.8	26.8
125	12.3	9.4	14.2




Indeno(1,2,3-cd)pyrene
 Concen: 17.595 ng/u1
 RT: 26.562 min Scan# 3957
 Delta R.T. -0.000 min
 Lab File: BM038925.D
 Acq: 08 Mar 2023 09:59

Tgt Ion:	276	Resp:	2825672
Ion	Ratio	Lower	Upper
276	100		
138	25.8	20.2	30.4
277	24.7	19.8	29.8





Abundance Scan 3962 (26.591 min): BM038889.D\data.ms (-3995) (-)




Dibenzo(a,h)anthracene
Concen: 17.592 ng/ul
RT: 26.579 min Scan# 4095
Delta R.T. -0.000 min
Lab File: BM038925.D
Acq: 08 Mar 2023 09:59

Instrument :
BNA_M
ClientSampleId :


Tgt Ion:	278	Resp:	2383418
Ion	Ratio	Lower	Upper
278	100		
139	17.4	13.4	20.2
279	23.9	19.1	28.7



Abundance Scan 3960 (26.579 min): BM038925.D\data.ms (-3926) (-)





Abundance Scan 4091 (27.350 min): BM038889.D\data.ms (-4096) (-)



Benzo(g,h,i)perylene
Concen: 17.023 ng/ul
RT: 27.344 min Scan# 4090
Delta R.T. -0.000 min
Lab File: BM038925.D
Acq: 08 Mar 2023 09:59

Tgt Ion:	276	Resp:	2238536
Ion	Ratio	Lower	Upper
276	100		
138	23.4	19.4	29.0
277	23.7	18.7	28.1



Abundance Scan 4090 (27.344 min): BM038925.D\data.ms (-4056) (-)

