

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM030723\
 Data File : BM038903.D
 Acq On : 07 Mar 2023 19:46
 Operator : CG/JU
 Sample : 01745-03
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 COLFO

Quant Time: Mar 07 21:41:00 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM022823.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Mar 06 23:57:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc Units	Dev(Min)

Internal Standards					
1) 1,4-Dichlorobenzene-d4	8.004	152	249644	20.000 ng/u1	0.00
20) Naphthalene-d8	10.816	136	1120550	20.000 ng/u1	0.00
38) Acenaphthene-d10	14.639	164	704622	20.000 ng/u1	0.00
64) Phenanthrene-d10	17.380	188	1532614	20.000 ng/u1	0.00
79) Chrysene-d12	21.556	240	1484157	20.000 ng/u1	0.00
88) Perylene-d12	24.009	264	1745819	20.000 ng/u1	0.00
System Monitoring Compounds					
3) 1,4-Dioxane-d8	3.381	96	35090	5.249 ng/uL	0.00
4) Pyridine-d5	3.804	84	181719	9.886 ng/u1	0.00
7) Phenol-d5	7.151	99	159548	6.817 ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.328	67	470746	31.479 ng/u1	0.00
11) 2-Chlorophenol-d4	7.528	132	447645	25.667 ng/u1	0.00
15) 4-Methylphenol-d8	8.704	113	293596	15.923 ng/u1	0.00
21) Nitrobenzene-d5	9.169	128	282552	35.918 ng/u1	0.00
24) 2-Nitrophenol-d4	9.892	143	262704	35.639 ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.427	165	489187	28.560 ng/u1	0.00
31) 4-Chloroaniline-d4	10.951	131	474038	16.603 ng/u1	0.00
46) Dimethylphthalate-d6	14.051	166	1892447	34.676 ng/u1	0.00
49) Acenaphthylene-d8	14.333	160	2096346	32.865 ng/u1	0.00
54) 4-Nitrophenol-d4	14.815	143	53179	4.868 ng/u1	0.00
60) Fluorene-d10	15.627	176	1728623	35.914 ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.739	200	207583	26.053 ng/u1	0.00
73) Anthracene-d10	17.480	188	3009766	39.826 ng/u1	0.00
81) Pyrene-d10	19.768	212	3524962	43.182 ng/u1	0.00
92) Benzo(a)pyrene-d12	23.856	264	3658646	41.585 ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM030723\
Data File : BM038903.D
Acq On : 07 Mar 2023 19:46
Operator : CG/JU
Sample : 01745-03
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
COLFO

Quant Time: Mar 07 21:41:00 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM022823.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Mar 06 23:57:30 2023
Response via : Initial Calibration

