

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM030723\
 Data File : BM038919.D
 Acq On : 08 Mar 2023 06:16
 Operator : CG/JU
 Sample : 01746-06
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DCAE0

Quant Time: Mar 08 07:07:28 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM022823.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Mar 08 04:25:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.998	152	280033	20.000	ng/u1	0.00
20) Naphthalene-d8	10.816	136	1274120	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.633	164	818616	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.380	188	1824223	20.000	ng/u1	0.00
79) Chrysene-d12	21.556	240	1810942	20.000	ng/u1	0.00
88) Perylene-d12	24.009	264	2095598	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.375	96	29580	3.944	ng/uL	0.00
4) Pyridine-d5	3.804	84	201984	9.796	ng/u1	0.00
7) Phenol-d5	7.151	99	142328	5.421	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.322	67	458680	27.344	ng/u1	0.00
11) 2-Chlorophenol-d4	7.528	132	428177	21.887	ng/u1	0.00
15) 4-Methylphenol-d8	8.704	113	272705	13.185	ng/u1	0.00
21) Nitrobenzene-d5	9.163	128	288170	32.217	ng/u1	0.00
24) 2-Nitrophenol-d4	9.892	143	278699	33.252	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.428	165	493161	25.321	ng/u1	0.00
31) 4-Chloroaniline-d4	10.945	131	745581	22.967	ng/u1	0.00
46) Dimethylphthalate-d6	14.045	166	1906702	30.072	ng/u1	0.00
49) Acenaphthylene-d8	14.327	160	2098724	28.320	ng/u1	0.00
54) 4-Nitrophenol-d4	14.815	143	56315	4.437	ng/u1	0.00
60) Fluorene-d10	15.627	176	1714410	30.659	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.733	200	242240	25.542	ng/u1	0.00
73) Anthracene-d10	17.480	188	2961387	32.922	ng/u1	0.00
81) Pyrene-d10	19.768	212	3685005	36.997	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.850	264	3841511	36.376	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM030723\
Data File : BM038919.D
Acq On : 08 Mar 2023 06:16
Operator : CG/JU
Sample : 01746-06
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DCAE0

Quant Time: Mar 08 07:07:28 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM022823.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Mar 08 04:25:35 2023
Response via : Initial Calibration

