

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM030816\
 Data File : BM004579.D
 Acq On : 08 Mar 2016 16:32
 Operator : SJ/UM
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM030816

Quant Time: Mar 08 18:24:19 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BM030816.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 08 18:22:06 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	114	-0.02
2	1,4-Dioxane	0.556	0.526	5.4	114	0.00
3	n-Nitrosodimethylamine	0.360	0.395	-9.7	125	-0.02
4 S	2-Fluorophenol	0.877	0.861	1.8	112	-0.03
5 S	Phenol-d6	1.013	1.042	-2.9	123	-0.02
6 I	Naphthalene-d8	1.000	1.000	0.0	111	-0.02
7 S	Nitrobenzene-d5	0.215	0.220	-2.3	126	-0.01
8	Naphthalene	1.383	1.448	-4.7	114	-0.02
9	2-Methylnaphthalene	0.864	0.906	-4.9	119	-0.01
10 I	Acenaphthene-d10	1.000	1.000	0.0	115	0.00
11 S	2,4,6-Tribromophenol	0.125	0.133	-6.4	132	0.00
12 S	2-Fluorobiphenyl	1.448	1.543	-6.6	119	-0.01
13	Acenaphthylene	2.631	2.849	-8.3	121	0.00
14	Acenaphthene	1.829	1.826	0.2	118	0.00
15	Fluorene	2.109	2.261	-7.2	123	-0.02
16 I	Phenanthrene-d10	1.000	1.000	0.0	124	0.00
17	Hexachlorobenzene	0.335	0.276	17.6	115	0.00
18	Pentachlorophenol	0.019	0.010	47.4#	132	-0.03
19	Phenanthrene	1.577	1.345	14.7	122	0.00
20	Anthracene	1.399	1.202	14.1	130	0.00
21	Fluoranthene	1.677	1.595	4.9	137	-0.02
22 I	Chrysene-d12	1.000	1.000	0.0	146	0.00
23	Pyrene	1.991	1.901	4.5	137	0.00
24 S	Terphenyl-d14	0.635	0.617	2.8	139	0.00
25	Benzo(a)anthracene	1.653	1.610	2.6	150	0.00
26	Chrysene	1.937	1.847	4.6	133	0.00
27	Indeno(1,2,3-cd)pyrene	1.439	1.301	9.6	126	0.00
28 I	Perylene-d12	1.000	1.000	0.0	137	0.01
29	Benzo(b)fluoranthene	1.897	1.913	-0.8	136	0.00
30	Benzo(k)fluoranthene	1.834	1.812	1.2	145	0.00
31 C	Benzo(a)pyrene	1.610	1.702	-5.7	137	0.00
32	Dibenzo(a,h)anthracene	1.301	1.188	8.7	125	0.00
33	Benzo(g,h,i)perylene	1.493	1.378	7.7	124	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0