

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM030817\
 Data File : BM009127.D
 Acq On : 09 Mar 2017 03:59
 Operator : SJ/MA
 Sample : I2062-11
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C0AD0

Manual Integrations
 APPROVED

mohammad
 3/9/2017 2:02:15 PM

Quant Time: Mar 09 05:55:39 2017
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM030317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 09 02:44:03 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.88	152	323	0.40	ng/ul	0.00
2) Naphthalene-d8	10.68	136	1070	0.40	ng/ul	-0.01
6) Acenaphthene-d10	14.52	164	652	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.26	188	1551m	0.40	ng/ul	0.00
16) Chrysene-d12	21.44	240	1389	0.40	ng/ul	0.00
20) Perylene-d12	23.78	264	1145m	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.27	152	347	0.20	ng/ul	0.00
14) Fluoranthene-d10	19.30	212	891	0.18	ng/ul	0.00
Target Compounds						
						Qvalue
3) Naphthalene	10.73	128	361	0.12	ng/ul#	65
5) 2-Methylnaphthalene	12.34	142	278	0.13	ng/ul	97
7) Acenaphthylene	14.23	152	300	0.09	ng/ul#	51
9) Fluorene	15.58	166	255	0.08	ng/ul#	90
12) Phenanthrene	17.31	178	2814m	0.58	ng/ul	
13) Anthracene	17.40	178	773	0.16	ng/ul#	90
15) Fluoranthene	19.32	202	5463m	0.86	ng/ul	
17) Pyrene	19.68	202	4946	1.04	ng/ul	96
18) Benzo(a)anthracene	21.42	228	2861	0.62	ng/ul	97
19) Chrysene	21.47	228	2837	0.65	ng/ul	93
21) Benzo(b)fluoranthene	23.07	252	3950m	0.88	ng/ul	
22) Benzo(k)fluoranthene	23.11	252	1155m	0.29	ng/ul	
23) Benzo(a)pyrene	23.67	252	2273m	0.56	ng/ul	
24) Indeno(1,2,3-cd)pyrene	26.15	276	1771m	0.34	ng/ul	
25) Dibenzo(a,h)anthracene	26.17	278	461m	0.11	ng/ul	
26) Benzo(g,h,i)perylene	26.89	276	1606m	0.37	ng/ul	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM030817\
Data File : BM009127.D
Acq On : 09 Mar 2017 03:59
Operator : SJ/MA
Sample : I2062-11
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampled :
C0AD0

Manual Integrations
APPROVED

mohammad
3/9/2017 2:02:15 PM

Quant Time: Mar 09 05:55:39 2017
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM030317.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 09 02:44:03 2017
Response via : Initial Calibration

