

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM030817\
 Data File : BM009121.D
 Acq On : 08 Mar 2017 22:17
 Operator : SJ/MA
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.449

Quant Time: Mar 09 01:20:59 2017
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM030317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 09 01:19:41 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	179#	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	198#	0.00
3	Naphthalene	1.153	1.107	4.0	193#	0.00
4 SURR	2-Methylnaphthalene-d10	0.664	0.612	7.8	187#	0.00
5	2-Methylnaphthalene	0.829	0.764	7.8	188#	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	186#	0.00
7	Acenaphthylene	2.082	2.060	1.1	192#	0.00
8 C	Acenaphthene	1.598	1.514	5.3	185#	0.00
9	Fluorene	1.869	1.800	3.7	177#	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	184#	0.00
11	Pentachlorophenol	0.194	0.179	7.7	160#	0.00
12	Phenanthrene	1.256	1.196	4.8	179#	0.00
13	Anthracene	1.241	1.152	7.2	171#	0.00
14 SURR	Fluoranthene-d10	1.251	1.172	6.3	172#	0.00
15 C	Fluoranthene	1.638	1.459	10.9	163#	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	153#	0.00
17	Pyrene	1.376	1.425	-3.6	167#	0.00
18	Benzo(a)anthracene	1.329	1.344	-1.1	153#	0.00
19	Chrysene	1.261	1.256	0.4	155#	0.00
20 I	Perylene-d12	1.000	1.000	0.0	130	0.00
21	Benzo(b)fluoranthene	1.571	1.539	2.0	133	0.00
22	Benzo(k)fluoranthene	1.389	1.509	-8.6	149	0.00
23 C	Benzo(a)pyrene	1.414	1.487	-5.2	139	0.00
24	Indeno(1,2,3-cd)pyrene	1.798	1.662	7.6	117	0.00
25	Dibenzo(a,h)anthracene	1.510	1.406	6.9	116	0.00
26	Benzo(g,h,i)perylene	1.534	1.372	10.6	112	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0