


Data Path : Z:\HPCHEM1\BNA_M\DATA\BM030817\
 Data File : BM009105.D
 Acq On : 08 Mar 2017 12:40
 Operator : SJ/MA
 Sample : DFTPP45
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP45

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM121416.M
 Title : SVOA CALIBRATION
 Last Update : Mon Dec 19 15:00:56 2016



AutoFind: Scans 2482, 2483, 2484; Background Corrected with Scan 2476

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	57.3	104821	PASS
68	69	0.00	2	2.0	1958	PASS
69	198	0.00	100	53.8	98560	PASS
70	69	0.00	2	0.6	640	PASS
127	198	10	80	45.8	83928	PASS
197	198	0.00	2	0.7	1367	PASS
198	198	100	100	100.0	183082	PASS
199	198	5	9	7.0	12777	PASS
275	198	10	60	28.4	52034	PASS
365	198	1	100	4.3	7787	PASS
441	443	0.01	100	84.4	19995	PASS
442	198	50	100	63.8	116741	PASS
443	442	15	24	20.3	23681	PASS

m/z	Abundance
37.00	762.0
38.10	1133.0
39.10	7645.0
40.00	1506.0
41.00	737.0
44.00	608.0
49.10	699.0
50.00	25968.0
51.00	94368.0
52.00	5550.0
56.00	3185.0
57.00	7179.0
59.00	321.0
61.00	1175.0
62.00	1331.0
63.00	4305.0
64.10	347.0
65.10	1582.0
68.10	1875.0
69.00	91848.0
69.90	517.0
73.00	1899.0
74.10	8924.0
75.00	15535.0
76.00	5795.0
77.10	100088.0
78.10	7623.0
78.90	5905.0
80.00	6745.0
81.00	9137.0
81.90	2180.0
82.90	1360.0
84.90	1173.0
85.90	2943.0
87.00	1036.0
90.80	578.0
91.20	866.0
92.10	1909.0
93.00	13543.0
93.80	974.0
96.00	525.0
98.00	12020.0
99.00	9460.0
100.00	917.0
101.00	3378.0
102.90	809.0
104.10	1562.0
105.00	2076.0
105.90	653.0
107.00	29080.0
108.00	4403.0
108.80	817.0
110.00	39072.0
111.00	6381.0
111.80	402.0
116.00	1098.0
117.00	30360.0
117.90	2734.0
122.00	1705.0
123.00	2504.0
124.10	1257.0
125.10	1062.0
126.10	380.0
127.00	72144.0
128.00	6364.0
129.00	33592.0
130.00	2980.0
131.20	429.0
132.70	387.0
133.90	1017.0
135.00	2336.0
136.00	973.0
136.90	1777.0
140.10	428.0
140.90	4266.0
142.00	1308.0
142.90	1606.0
143.90	498.0
145.80	717.0
147.00	2453.0
147.90	6234.0
148.80	1457.0
149.90	732.0
151.00	729.0
153.00	1405.0
153.80	1271.0
155.00	2337.0
156.00	3571.0
157.20	539.0
157.90	567.0
158.90	614.0

Instrument :
BNA_M
ClientSampleId :
DFTPP45

160.00	1309.0
161.00	2518.0
162.00	404.0
162.60	336.0
163.90	361.0
164.90	1556.0
166.00	1268.0
167.00	11465.0
168.00	6787.0
169.00	718.0
170.70	417.0
171.80	841.0
172.90	1257.0
174.00	1836.0
175.00	3664.0
175.90	782.0
176.90	1362.0
177.90	795.0
179.00	6662.0
179.90	4806.0
180.90	1466.0
181.90	381.0
183.90	612.0
185.00	2948.0
186.00	22768.0
187.00	6883.0
188.00	1066.0
189.00	1689.0
190.70	352.0
191.10	309.0
191.80	1712.0
193.00	2862.0
193.90	677.0
196.00	3512.0
197.00	779.0
197.90	136192.0
198.90	9477.0
199.90	1231.0
201.30	428.0
203.00	1318.0
204.00	7459.0
205.00	11395.0
206.00	40904.0
207.00	9404.0
208.00	2379.0
208.90	1138.0
209.80	836.0
211.30	2157.0
214.80	657.0
216.10	1280.0
216.90	12495.0
217.90	1701.0
220.20	334.0
221.10	3650.0
221.80	1451.0
223.00	3406.0
224.00	21456.0
225.00	5791.0
226.00	1242.0
226.90	12098.0
228.00	1760.0
228.90	2110.0
230.00	966.0
230.90	847.0
232.00	387.0
232.90	551.0
233.80	687.0
234.90	720.0
235.90	1340.0
237.00	2051.0
238.00	333.0
239.00	888.0
240.10	455.0
240.90	370.0
242.00	1074.0
242.90	1299.0
244.00	18112.0
245.00	2182.0
245.90	4871.0
247.20	899.0
249.00	359.0
253.90	680.0
255.00	90928.0
255.90	12611.0
257.00	1292.0
258.00	7601.0
258.80	1276.0
265.00	2776.0
265.90	739.0
267.10	388.0
267.80	305.0
270.90	333.0
272.90	2842.0
274.00	6492.0
275.00	39328.0

Instrument :
BNA_M
ClientSampleId :
DFTPP45

276.00	6477.0
276.90	3874.0
277.90	727.0
281.00	307.0
284.20	518.0
285.00	341.0
293.10	983.0
294.00	459.0
295.10	343.0
296.00	14288.0
296.90	1761.0
303.00	865.0
304.10	440.0
304.90	381.0
314.00	867.0
314.80	1154.0
316.00	1180.0
321.00	489.0
323.00	3099.0
324.10	976.0
326.80	592.0
328.00	352.0
331.70	305.0
334.00	2582.0
335.00	1163.0
341.00	1421.0
345.80	663.0
351.90	487.0
352.90	731.0
353.90	648.0
355.00	734.0
365.00	4052.0
365.90	848.0
372.00	2093.0
372.90	351.0
403.00	891.0
405.10	303.0
421.00	925.0
421.80	486.0
423.00	3136.0
424.00	1737.0
441.00	11026.0
442.00	61072.0
443.10	14180.0
443.80	1193.0

Instrument :

BNA_M

ClientSampleId :

DFTPP45

m/z	Abundance
37.00	593.0
38.10	1837.0
39.00	9120.0
40.10	940.0
41.00	358.0
44.10	598.0
48.80	1153.0
50.10	33040.0
51.10	126968.0
52.00	7258.0
53.10	351.0
55.00	698.0
56.00	3652.0
57.00	9151.0
57.90	512.0
61.00	1497.0
62.10	1510.0
63.00	5871.0
64.10	1045.0
65.10	1733.0
66.30	881.0
68.00	2762.0
69.00	118856.0
70.00	950.0
73.00	1400.0
74.00	13831.0
75.00	22536.0
76.10	7232.0
77.10	135232.0
78.00	9557.0
79.00	10678.0
80.00	6774.0
81.00	10654.0
82.00	2030.0
82.90	1925.0
85.00	2504.0
86.00	2920.0
87.00	1912.0
88.00	414.0
91.00	2545.0
91.90	2035.0
93.00	19248.0
94.10	1174.0
96.20	578.0
96.70	329.0
98.00	15566.0
99.00	11993.0
99.90	1044.0
101.00	4780.0
103.10	2105.0
104.00	3750.0
105.00	2908.0
105.90	1221.0
107.00	37624.0
108.00	4761.0
108.90	1473.0
110.00	57400.0
110.90	9783.0
112.00	1426.0
112.80	556.0
116.00	2067.0
117.00	45840.0
117.90	3002.0
119.90	563.0
122.10	1691.0
123.00	3926.0
124.00	1650.0
124.90	1521.0
126.10	584.0
127.00	103192.0
128.00	7010.0
129.00	45632.0
130.00	3805.0
131.00	951.0
133.10	333.0
134.00	2184.0
134.90	4456.0
135.80	1784.0
137.00	2236.0
139.90	1076.0
141.00	6310.0
141.90	2431.0
143.10	937.0
144.90	341.0
146.00	928.0
147.00	4111.0
147.90	10406.0
148.90	1962.0
150.10	679.0
151.00	779.0
153.00	2056.0

Instrument :
BNA_M
ClientSampleId :
DFTPP45

153.90	1250.0
155.00	3506.0
156.00	5745.0
157.10	572.0
157.90	1633.0
158.90	1597.0
159.90	2160.0
160.90	2595.0
161.90	710.0
162.90	942.0
164.80	2321.0
165.90	1857.0
167.00	15457.0
168.00	8612.0
168.90	1268.0
169.80	458.0
170.70	528.0
171.90	1040.0
172.90	1344.0
173.90	2781.0
175.00	7002.0
175.90	748.0
176.90	2381.0
177.90	988.0
179.00	9733.0
179.90	8920.0
181.10	2506.0
182.00	360.0
182.40	340.0
183.50	458.0
185.20	2987.0
186.00	35560.0
187.00	10048.0
187.90	2003.0
189.00	3075.0
190.80	1483.0
192.00	3457.0
193.00	4016.0
194.00	486.0
195.20	402.0
196.00	5657.0
196.90	1701.0
197.90	226368.0
198.90	16632.0
199.90	1567.0
201.30	457.0
202.00	335.0
202.90	2229.0
204.00	11229.0
205.00	15729.0
206.00	63696.0
207.00	11905.0
207.90	3215.0
209.10	1655.0
210.00	1014.0
210.80	2764.0
213.00	453.0
215.00	788.0
215.70	808.0
216.10	760.0
217.00	21528.0
217.90	2605.0
219.00	407.0
220.00	535.0
221.10	7339.0
221.90	2992.0
222.90	4627.0
224.00	41104.0
225.10	10414.0
226.00	1352.0
227.00	19416.0
228.00	3371.0
228.90	2743.0
229.90	340.0
231.10	2252.0
231.70	355.0
234.10	1283.0
235.00	2093.0
236.00	1235.0
236.90	1988.0
237.90	489.0
239.00	667.0
239.90	827.0
241.10	1244.0
242.00	2113.0
243.10	2380.0
244.10	29808.0
245.00	3421.0
246.10	7171.0
247.00	1780.0
248.00	348.0
248.70	868.0
250.90	382.0
251.90	715.0
253.10	1240.0

Instrument :
BNA_M
ClientSampleId :
DFTPP45

254.20	941.0
255.00	150400.0
256.00	23216.0
257.00	2381.0
257.90	11853.0
259.10	1956.0
260.00	439.0
264.90	5002.0
265.90	372.0
267.10	351.0
271.10	450.0
271.90	795.0
273.00	3813.0
274.00	12296.0
275.00	61032.0
276.00	8684.0
276.90	6534.0
278.00	1680.0
281.00	786.0
282.90	545.0
284.10	342.0
284.90	525.0
293.00	1214.0
294.00	693.0
296.00	20888.0
296.90	2944.0
303.10	1945.0
303.90	1042.0
305.20	335.0
307.80	390.0
313.90	711.0
315.10	2672.0
316.00	1334.0
323.10	7193.0
324.00	1097.0
327.00	1165.0
327.90	995.0
332.00	1289.0
333.00	644.0
334.00	3814.0
335.00	1010.0
341.00	481.0
345.90	1247.0
346.90	402.0
351.90	2252.0
353.00	940.0
353.90	1893.0
355.30	385.0
365.00	9802.0
365.80	884.0
366.10	953.0
372.00	3025.0
373.00	749.0
382.80	304.0
385.00	358.0
400.90	381.0
402.10	1162.0
403.10	1982.0
404.00	754.0
404.90	448.0
421.00	1491.0
422.10	603.0
423.00	9498.0
424.00	3090.0
424.90	390.0
441.00	21200.0
442.10	138176.0
443.10	28480.0
444.10	2157.0

Instrument :
BNA_M
ClientSampleId :
DFTPP45

m/z	Abundance
37.00	364.0
38.00	1271.0
39.10	7235.0
40.00	986.0
41.10	341.0
43.90	1266.0
45.00	445.0
49.00	549.0
50.00	24232.0
51.00	93128.0
52.10	4146.0
55.00	756.0
56.10	2479.0
57.00	6300.0
60.90	1266.0
62.00	1544.0
63.00	3594.0
65.10	1807.0
68.10	1239.0
69.00	84976.0
69.90	454.0
73.20	1558.0
74.00	12303.0
75.00	15305.0
76.00	5714.0
77.10	96976.0
78.00	6062.0
79.00	6280.0
80.00	5124.0
81.00	6714.0
82.00	1950.0
83.00	2679.0
85.00	2082.0
86.00	2547.0
86.90	1188.0
88.90	421.0
90.90	1980.0
91.90	2030.0
92.90	12793.0
93.90	974.0
96.10	964.0
96.80	549.0
98.00	10301.0
99.00	7744.0
100.00	969.0
101.00	3189.0
103.10	1985.0
104.00	2685.0
105.00	3270.0
106.00	807.0
107.00	31224.0
108.00	3981.0
110.00	45920.0
111.00	7472.0
112.00	1431.0
112.80	850.0
116.00	1574.0
117.00	33736.0
118.00	2710.0
118.90	482.0
120.10	447.0
122.00	2071.0
122.90	3032.0
123.90	914.0
124.90	1493.0
127.00	76448.0
128.00	5710.0
129.00	37112.0
129.90	3612.0
130.80	476.0
133.80	329.0
134.90	2944.0
136.10	1090.0
136.90	2057.0
140.00	478.0
141.00	5542.0
141.90	1649.0
142.90	977.0
144.90	395.0
146.00	661.0
146.90	2638.0
148.00	9597.0
149.00	1712.0
150.00	596.0
151.30	439.0
152.90	1630.0
153.90	1005.0
155.00	1989.0
156.00	4186.0
157.00	492.0
158.00	1457.0

Instrument :
BNA_M
ClientSampleId :
DFTPP45

159.10	522.0
159.90	1167.0
161.10	2411.0
161.70	654.0
163.20	357.0
165.00	1857.0
166.10	2361.0
167.00	13945.0
168.00	7814.0
168.90	1645.0
170.00	313.0
170.70	308.0
171.90	966.0
172.80	928.0
173.90	2142.0
175.10	3573.0
175.90	2736.0
176.90	1886.0
178.00	811.0
179.00	8452.0
179.90	5472.0
181.00	3174.0
182.10	609.0
184.00	723.0
185.10	5037.0
186.00	30904.0
187.00	9052.0
188.00	956.0
188.90	2003.0
189.90	700.0
191.00	1180.0
191.90	2171.0
193.00	2282.0
193.90	379.0
196.00	5967.0
197.00	1622.0
197.90	186688.0
198.90	12224.0
200.00	1069.0
201.40	537.0
201.90	320.0
202.80	1924.0
203.90	10003.0
205.00	14567.0
206.00	57248.0
207.00	9009.0
208.00	3337.0
209.10	692.0
210.00	1400.0
211.00	2586.0
212.00	984.0
215.00	560.0
216.00	1339.0
216.90	16236.0
218.00	2326.0
218.90	302.0
219.80	510.0
221.00	7898.0
222.00	2516.0
223.10	3069.0
224.00	31392.0
225.00	9237.0
226.00	801.0
226.90	13893.0
228.00	2088.0
229.00	4610.0
229.80	385.0
230.90	1303.0
231.80	433.0
233.90	1465.0
235.00	1092.0
235.80	620.0
237.10	2147.0
238.90	933.0
239.90	633.0
241.10	1677.0
241.90	1868.0
243.00	1989.0
244.00	24296.0
245.00	2838.0
246.00	6573.0
247.00	1877.0
248.80	664.0
252.90	621.0
254.00	1403.0
255.00	135936.0
256.00	19632.0
256.90	2099.0
258.00	10132.0
258.90	984.0
260.10	622.0
260.90	394.0
265.00	3346.0
265.90	330.0
272.10	336.0

Instrument :

BNA_M

ClientSampleId :

DFTPP45

273.00	3946.0
273.90	11718.0
275.00	55744.0
275.90	5132.0
276.90	6393.0
278.10	1418.0
281.10	658.0
282.80	559.0
283.20	612.0
284.80	553.0
293.00	2136.0
293.90	543.0
296.00	20928.0
296.80	2298.0
297.90	406.0
301.90	368.0
303.00	2667.0
304.10	688.0
305.20	361.0
308.80	329.0
314.00	571.0
315.00	3143.0
315.80	1617.0
320.80	403.0
321.10	305.0
323.00	4928.0
324.00	535.0
327.00	1219.0
327.80	508.0
328.90	322.0
331.70	592.0
332.90	538.0
334.10	3669.0
335.00	740.0
341.10	867.0
342.00	381.0
345.90	975.0
352.20	1287.0
353.10	1240.0
354.10	1749.0
355.20	372.0
365.00	9509.0
365.90	1261.0
372.10	4011.0
372.90	411.0
373.20	499.0
377.00	327.0
382.90	1103.0
384.00	694.0
389.90	374.0
401.90	853.0
403.00	1395.0
404.00	663.0
421.00	1422.0
422.00	1203.0
423.00	8733.0
424.00	2718.0
439.00	387.0
441.00	27760.0
442.00	150976.0
443.10	28384.0
444.00	2613.0

Instrument :
BNA_M
ClientSampleId :
DFTPP45

Scan 2476 (17.651 min): BM009105.D
DFTPP45

m/z	Abundance
39.90	812.0
44.00	880.0
73.00	1148.0
206.80	1911.0
208.10	385.0
281.00	376.0

Instrument :
BNA_M
ClientSampleId :
DFTPP45