

DDT Breakdown

Date	Instrument Name	DFTPP Data File
3/8/2021	BNA_M	<u>BM029017.D</u>
Compound Name	Response	Retention Time
DDT	36145	20.56
DDD	230	20.12
DDE	0	19.95
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
230	36375	0.63