

Data Path : Z:\HPCHEM1\BNA_M\Data\BM030917\
 Data File : BM009141.D
 Acq On : 09 Mar 2017 12:33
 Operator : SJ/MA
 Sample : I2062-12
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C0AD4

Quant Time: Mar 09 15:25:15 2017
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM030317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 09 15:10:54 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.88	152	156	0.40	ng/ul	0.00
2) Naphthalene-d8	10.68	136	507	0.40	ng/ul	-0.01
6) Acenaphthene-d10	14.52	164	296	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.36	188	47063	0.40	ng/ul	0.10
16) Chrysene-d12	21.44	240	742	0.40	ng/ul	-0.01
20) Perylene-d12	23.63	264	51359	0.40	ng/ul	-0.16
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.27	152	228	0.27	ng/ul	0.00
14) Fluoranthene-d10	19.29	212	592	0.00	ng/ul	-0.01
Target Compounds						
5) 2-Methylnaphthalene	12.34	142	24	0.02	ng/ul#	53
17) Pyrene	19.69	202	299	0.12	ng/ul#	60
18) Benzo(a)anthracene	21.42	228	164	0.07	ng/ul#	53
19) Chrysene	21.48	228	124	0.05	ng/ul#	45

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\Data\BM030917\
Data File : BM009141.D
Acq On : 09 Mar 2017 12:33
Operator : SJ/MA
Sample : I2062-12
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
C0AD4

Quant Time: Mar 09 15:25:15 2017
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM030317.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 09 15:10:54 2017
Response via : Initial Calibration

