

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM030917\
 Data File : BM009149.D
 Acq On : 09 Mar 2017 17:21
 Operator : SJ/MA
 Sample : I2069-16
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C0AJ1

Manual Integrations
 APPROVED

mohammad
 3/10/2017 6:01:02 PM

Quant Time: Mar 09 18:21:44 2017
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM030317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 09 02:44:03 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.88	152	218	0.40	ng/ul	0.00
2) Naphthalene-d8	10.68	136	693	0.40	ng/ul	-0.01
6) Acenaphthene-d10	14.52	164	426	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.26	188	945m	0.40	ng/ul	0.00
16) Chrysene-d12	21.44	240	1014	0.40	ng/ul	0.00
20) Perylene-d12	23.78	264	858m	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.27	152	250	0.22	ng/ul	0.00
14) Fluoranthene-d10	19.29	212	791m	0.27	ng/ul	0.00
Target Compounds						
15) Fluoranthene	19.32	202	273	0.07	ng/ul	Qvalue 62
18) Benzo(a)anthracene	21.42	228	141	0.04	ng/ul#	42
19) Chrysene	21.47	228	168	0.05	ng/ul#	50

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM030917\
Data File : BM009149.D
Acq On : 09 Mar 2017 17:21
Operator : SJ/MA
Sample : I2069-16
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
C0AJ1

Manual Integrations
APPROVED

mohammad
3/10/2017 6:01:02 PM

Quant Time: Mar 09 18:21:44 2017
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM030317.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 09 02:44:03 2017
Response via : Initial Calibration

