

Data Path : Z:\HPCHEM1\BNA_M\Data\BM030917\
 Data File : BM009138.D
 Acq On : 09 Mar 2017 10:38
 Operator : SJ/MA
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.451

Quant Time: Mar 09 15:09:30 2017
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM030317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 09 02:44:03 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	179#	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	175#	0.00
3	Naphthalene	1.153	1.177	-2.1	182#	0.00
4 SURR	2-Methylnaphthalene-d10	0.664	0.640	3.6	173#	0.00
5	2-Methylnaphthalene	0.829	0.804	3.0	175#	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	160#	0.00
7	Acenaphthylene	2.082	2.113	-1.5	169#	0.00
8 C	Acenaphthene	1.598	1.557	2.6	163#	0.00
9	Fluorene	1.869	1.787	4.4	151#	0.02
10 I	Phenanthrene-d10	1.000	1.000	0.0	149	0.00
11	Pentachlorophenol	0.194	0.178	8.2	130	0.00
12	Phenanthrene	1.256	1.237	1.5	150#	0.02
13	Anthracene	1.241	1.180	4.9	142	0.00
14 SURR	Fluoranthene-d10	1.251	1.198	4.2	143	0.00
15 C	Fluoranthene	1.638	1.508	7.9	137	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	119	0.00
17	Pyrene	1.376	1.505	-9.4	136	0.00
18	Benzo(a)anthracene	1.329	1.398	-5.2	123	0.00
19	Chrysene	1.261	1.328	-5.3	127	0.00
20 I	Perylene-d12	1.000	1.000	0.0	108	0.00
21	Benzo(b)fluoranthene	1.571	1.568	0.2	113	0.00
22	Benzo(k)fluoranthene	1.389	1.475	-6.2	121	0.00
23 C	Benzo(a)pyrene	1.414	1.442	-2.0	112	0.00
24	Indeno(1,2,3-cd)pyrene	1.798	1.778	1.1	104	0.00
25	Dibenzo(a,h)anthracene	1.510	1.481	1.9	102	0.00
26	Benzo(g,h,i)perylene	1.534	1.530	0.3	104	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0