

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM030917\
 Data File : BM009152.D
 Acq On : 09 Mar 2017 19:09
 Operator : SJ/MA
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.452

Quant Time: Mar 10 09:56:24 2017
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM030317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 10 03:16:46 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	162#	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	180#	0.00
3	Naphthalene	1.153	1.143	0.9	181#	0.00
4 SURR	2-Methylnaphthalene-d10	0.664	0.627	5.6	173#	0.00
5	2-Methylnaphthalene	0.829	0.788	4.9	176#	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	169#	0.00
7	Acenaphthylene	2.082	2.119	-1.8	179#	0.00
8 C	Acenaphthene	1.598	1.547	3.2	171#	0.00
9	Fluorene	1.869	1.824	2.4	163#	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	165#	0.00
11	Pentachlorophenol	0.194	0.187	3.6	151#	0.00
12	Phenanthrene	1.256	1.243	1.0	167#	0.00
13	Anthracene	1.241	1.193	3.9	159#	0.00
14 SURR	Fluoranthene-d10	1.251	1.190	4.9	157#	0.00
15 C	Fluoranthene	1.638	1.536	6.2	154#	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	133	0.00
17	Pyrene	1.376	1.535	-11.6	156#	0.00
18	Benzo(a)anthracene	1.329	1.435	-8.0	141	0.00
19	Chrysene	1.261	1.335	-5.9	143	0.00
20 I	Perylene-d12	1.000	1.000	0.0	125	0.00
21	Benzo(b)fluoranthene	1.571	1.612	-2.6	134	0.00
22	Benzo(k)fluoranthene	1.389	1.415	-1.9	135	0.00
23 C	Benzo(a)pyrene	1.414	1.435	-1.5	130	0.00
24	Indeno(1,2,3-cd)pyrene	1.798	1.751	2.6	119	0.00
25	Dibenzo(a,h)anthracene	1.510	1.426	5.6	113	0.00
26	Benzo(g,h,i)perylene	1.534	1.477	3.7	116	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0