

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM030917\
 Data File : BM009165.D
 Acq On : 10 Mar 2017 02:57
 Operator : SJ/MA
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.453

Quant Time: Mar 10 04:22:14 2017
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM030317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 10 03:16:46 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	197#	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	221#	0.00
3	Naphthalene	1.153	1.157	-0.3	225#	0.00
4 SURR	2-Methylnaphthalene-d10	0.664	0.626	5.7	213#	0.00
5	2-Methylnaphthalene	0.829	0.783	5.5	215#	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	206#	0.00
7	Acenaphthylene	2.082	2.183	-4.9	225#	0.00
8 C	Acenaphthene	1.598	1.541	3.6	208#	0.00
9	Fluorene	1.869	1.799	3.7	197#	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	190#	0.00
11	Pentachlorophenol	0.194	0.180	7.2	167#	0.00
12	Phenanthrene	1.256	1.263	-0.6	196#	0.00
13	Anthracene	1.241	1.261	-1.6	193#	0.00
14 SURR	Fluoranthene-d10	1.251	1.229	1.8	187#	0.00
15 C	Fluoranthene	1.638	1.559	4.8	180#	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	163#	0.00
17	Pyrene	1.376	1.474	-7.1	183#	0.00
18	Benzo(a)anthracene	1.329	1.408	-5.9	170#	0.00
19	Chrysene	1.261	1.318	-4.5	173#	0.00
20 I	Perylene-d12	1.000	1.000	0.0	159#	0.00
21	Benzo(b)fluoranthene	1.571	1.508	4.0	159#	0.00
22	Benzo(k)fluoranthene	1.389	1.415	-1.9	171#	0.00
23 C	Benzo(a)pyrene	1.414	1.433	-1.3	164#	0.00
24	Indeno(1,2,3-cd)pyrene	1.798	1.751	2.6	150#	0.00
25	Dibenzo(a,h)anthracene	1.510	1.438	4.8	145	0.00
26	Benzo(g,h,i)perylene	1.534	1.479	3.6	147	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0