


Data Path : Z:\HPCHEM1\BNA_M\DATA\BM030917\
 Data File : BM009137.D
 Acq On : 09 Mar 2017 10:01
 Operator : SJ/MA
 Sample : DFTPP46
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP46

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM121416.M
 Title : SVOA CALIBRATION
 Last Update : Mon Dec 19 15:00:56 2016



AutoFind: Scans 2482, 2483, 2484; Background Corrected with Scan 2476

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	48.6	93588	PASS
68	69	0.00	2	1.9	1725	PASS
69	198	0.00	100	48.2	92696	PASS
70	69	0.00	2	0.3	260	PASS
127	198	10	80	44.0	84664	PASS
197	198	0.00	2	0.5	926	PASS
198	198	100	100	100.0	192405	PASS
199	198	5	9	6.5	12571	PASS
275	198	10	60	27.7	53346	PASS
365	198	1	100	4.4	8511	PASS
441	443	0.01	100	86.0	21933	PASS
442	198	50	100	70.6	135746	PASS
443	442	15	24	18.8	25517	PASS

m/z	Abundance
38.00	1291.0
39.00	7094.0
39.90	1030.0
40.90	477.0
44.00	1523.0
44.90	641.0
48.90	783.0
50.10	28944.0
51.00	113712.0
52.10	5426.0
53.10	535.0
54.80	943.0
56.00	3254.0
57.00	6771.0
58.10	410.0
61.10	1636.0
62.00	1473.0
63.00	5331.0
64.10	335.0
65.00	1946.0
68.00	1868.0
69.00	107904.0
70.00	782.0
73.10	2020.0
74.00	12584.0
75.00	16688.0
76.10	5954.0
77.10	114456.0
78.10	9189.0
79.00	8373.0
80.00	6547.0
80.90	8650.0
82.00	2042.0
83.00	1578.0
83.80	314.0
85.00	2081.0
86.00	2501.0
87.00	1146.0
91.00	2288.0
91.90	2763.0
93.00	15304.0
93.80	955.0
95.00	466.0
96.10	923.0
96.70	456.0
97.20	390.0
98.00	9382.0
99.00	9730.0
99.90	1397.0
101.00	3501.0
102.80	1384.0
103.80	3157.0
105.00	2138.0
105.90	1209.0
107.00	37944.0
108.00	5043.0
109.10	969.0
110.00	55176.0
111.00	7727.0
112.00	950.0
113.10	512.0
114.90	882.0
116.00	2239.0
117.00	35136.0
118.00	2446.0
119.90	701.0
122.00	2332.0
123.00	2961.0
123.80	1293.0
125.00	1915.0
126.10	602.0
127.00	87248.0
128.00	7625.0
129.00	39176.0
130.00	3216.0
130.80	589.0
132.00	625.0
133.00	677.0
134.00	1632.0
135.10	2978.0
136.10	1544.0
137.00	1671.0
139.90	474.0
141.00	7582.0
142.00	1412.0
142.80	2099.0
144.70	303.0
145.90	1703.0
147.00	3589.0
147.90	8418.0
148.90	2014.0

Instrument :
BNA_M
ClientSampleId :
DFTPP46

150.00	344.0
150.80	760.0
151.20	763.0
151.80	667.0
152.90	1295.0
154.10	1330.0
154.90	2986.0
156.00	4934.0
157.00	1048.0
157.90	1940.0
159.00	937.0
159.90	1906.0
161.00	2805.0
161.90	1002.0
164.90	2160.0
165.90	1677.0
167.00	12816.0
168.00	9166.0
168.80	1677.0
171.00	671.0
171.70	1276.0
172.90	778.0
174.00	2636.0
175.00	4315.0
175.90	1604.0
176.90	2297.0
178.10	1259.0
178.90	6168.0
179.90	5975.0
181.00	3031.0
182.10	496.0
184.10	317.0
184.90	3267.0
186.00	30792.0
187.00	8834.0
187.80	1225.0
188.90	3476.0
190.00	494.0
191.00	1272.0
191.90	2722.0
193.10	3397.0
193.80	724.0
195.00	414.0
195.90	6945.0
197.00	1498.0
198.00	197824.0
199.00	12774.0
199.90	2198.0
201.10	457.0
201.50	872.0
203.00	2115.0
204.10	8975.0
205.00	14145.0
206.00	56792.0
207.00	10678.0
208.00	2198.0
209.20	1435.0
209.90	963.0
210.20	905.0
210.90	1839.0
212.10	532.0
214.90	816.0
216.00	1425.0
217.00	17016.0
217.90	2341.0
219.20	327.0
220.90	7253.0
221.90	1925.0
223.00	4569.0
224.00	34496.0
225.00	8474.0
226.00	1294.0
227.00	16113.0
227.90	2413.0
228.90	2503.0
231.00	904.0
233.80	1665.0
235.10	919.0
235.90	1303.0
237.10	2678.0
239.00	900.0
240.10	850.0
241.00	1206.0
242.10	1544.0
242.90	2071.0
244.00	22176.0
245.10	2713.0
246.00	5340.0
246.90	1032.0
248.00	387.0
249.00	532.0
253.10	716.0
254.10	802.0
255.00	122728.0
256.00	21320.0

Instrument :

BNA_M

ClientSampleId :

DFTPP46

257.10	1738.0
258.00	8452.0
259.00	1020.0
259.80	336.0
264.90	3920.0
265.70	460.0
272.10	1145.0
273.00	3240.0
274.10	8941.0
275.00	49920.0
276.00	6181.0
276.90	4170.0
278.00	1071.0
281.10	467.0
281.90	392.0
283.00	1358.0
284.80	890.0
292.00	399.0
293.10	1849.0
296.00	18984.0
297.00	2607.0
301.90	367.0
302.90	2097.0
303.80	302.0
307.90	468.0
313.90	818.0
315.00	1843.0
316.10	1207.0
317.10	468.0
321.10	470.0
322.00	470.0
323.00	5135.0
324.00	1130.0
327.10	1139.0
327.90	747.0
333.00	519.0
334.00	3617.0
335.00	648.0
341.10	859.0
342.10	496.0
345.90	1052.0
351.90	1340.0
352.90	715.0
354.00	1225.0
355.10	476.0
365.00	7960.0
366.00	1439.0
371.10	325.0
372.00	2213.0
372.80	533.0
373.10	548.0
382.90	423.0
402.10	777.0
403.10	1286.0
421.00	830.0
422.10	1316.0
423.10	5754.0
424.10	3388.0
424.90	338.0
441.10	17184.0
442.10	106056.0
443.10	21848.0
444.10	1611.0

Instrument :
BNA_M
ClientSampleId :
DFTPP46

m/z	Abundance
37.10	830.0
37.90	1470.0
39.10	6615.0
40.00	1189.0
41.00	619.0
43.00	332.0
44.10	792.0
44.90	381.0
49.10	748.0
50.00	27288.0
51.10	110472.0
52.00	4886.0
56.10	3254.0
57.00	7248.0
58.00	686.0
60.90	1433.0
62.00	2181.0
63.00	3571.0
64.20	803.0
65.00	2035.0
68.00	1896.0
69.00	108728.0
73.00	2466.0
74.00	13561.0
75.00	17056.0
76.10	6354.0
77.10	123328.0
78.00	8401.0
79.00	7318.0
80.00	5996.0
81.00	8049.0
82.10	1777.0
83.00	2171.0
84.00	375.0
85.20	877.0
86.00	3615.0
87.00	1301.0
88.90	806.0
91.00	3207.0
91.90	1837.0
93.10	14618.0
93.90	834.0
97.00	358.0
98.00	11004.0
99.00	10816.0
99.80	590.0
100.30	405.0
101.00	4114.0
103.00	1648.0
104.00	3493.0
105.00	2613.0
105.90	774.0
107.00	36312.0
108.00	6848.0
108.90	1157.0
110.00	58080.0
111.00	8354.0
111.90	799.0
115.90	1360.0
117.00	38912.0
118.00	2207.0
119.80	419.0
122.00	2403.0
123.00	3180.0
124.00	1629.0
125.00	1245.0
127.00	102704.0
128.00	5798.0
129.00	43976.0
130.00	3487.0
130.90	1428.0
132.00	387.0
133.90	958.0
135.00	2759.0
136.00	2069.0
137.00	2097.0
137.80	753.0
139.80	351.0
140.90	6251.0
141.90	2369.0
142.90	1258.0
143.80	317.0
145.90	1810.0
146.90	3176.0
148.00	10527.0
149.00	1520.0
149.90	444.0
150.90	392.0
151.50	476.0
152.00	518.0
152.80	2703.0

Instrument :
BNA_M
ClientSampleId :
DFTPP46

153.90	1154.0
155.00	2911.0
156.00	5716.0
156.80	978.0
157.90	2464.0
159.00	765.0
159.90	1822.0
161.10	2831.0
162.00	992.0
163.10	342.0
165.00	2796.0
165.90	1897.0
167.00	15024.0
168.00	9191.0
168.80	1195.0
170.10	538.0
171.00	756.0
172.00	1194.0
173.10	1062.0
173.90	2959.0
175.10	5208.0
176.30	1060.0
177.00	2378.0
178.00	797.0
178.90	9222.0
180.00	8669.0
180.90	2870.0
181.80	496.0
182.70	368.0
183.80	914.0
185.00	4081.0
186.00	34760.0
187.00	9938.0
187.90	1026.0
189.00	2703.0
190.00	739.0
191.10	1207.0
192.00	3045.0
193.00	4479.0
194.00	1399.0
194.90	779.0
196.10	7673.0
196.80	1280.0
198.00	236544.0
198.90	14763.0
200.00	1894.0
200.80	478.0
201.40	731.0
203.00	2073.0
204.00	10129.0
205.00	16456.0
206.00	63008.0
207.10	10423.0
208.10	2754.0
208.90	1124.0
209.80	1359.0
210.80	2497.0
212.00	875.0
214.90	906.0
215.90	1590.0
216.90	21160.0
218.00	2491.0
220.10	418.0
221.10	9317.0
221.80	2338.0
223.00	4816.0
224.00	37216.0
225.00	9950.0
226.00	1873.0
227.00	22304.0
228.00	2252.0
229.00	3268.0
230.00	404.0
230.90	1269.0
233.00	485.0
233.80	822.0
234.90	1549.0
236.10	1265.0
237.00	2468.0
239.00	438.0
240.00	385.0
241.10	586.0
242.10	2038.0
243.10	2572.0
244.00	30872.0
245.00	2947.0
246.00	6358.0
246.80	1176.0
248.00	394.0
249.00	2181.0
249.90	425.0
251.20	570.0
253.20	1037.0
254.10	1479.0
255.00	160064.0

Instrument :
BNA_M
ClientSampleId :
DFTPP46

256.00	20096.0
256.90	1942.0
257.90	9593.0
259.00	1312.0
260.00	333.0
264.10	343.0
265.00	4181.0
266.10	651.0
266.90	474.0
272.00	818.0
273.00	3753.0
274.00	11985.0
275.00	63104.0
276.00	9108.0
277.00	8152.0
277.80	865.0
281.00	1122.0
281.80	455.0
282.80	504.0
284.10	468.0
284.90	1744.0
286.10	497.0
289.00	319.0
292.90	1753.0
295.00	703.0
296.00	22696.0
297.00	4191.0
298.20	317.0
303.10	2239.0
314.00	960.0
314.90	2077.0
316.00	962.0
317.10	461.0
321.00	914.0
322.10	361.0
323.10	5686.0
324.10	1673.0
327.10	1266.0
328.00	772.0
332.90	399.0
334.00	5554.0
335.00	1481.0
340.90	1168.0
342.10	337.0
345.90	1388.0
346.80	354.0
352.00	2156.0
353.00	941.0
354.00	1605.0
355.10	415.0
365.00	9214.0
366.00	888.0
367.00	323.0
371.00	817.0
372.10	3478.0
372.80	554.0
373.10	555.0
383.00	937.0
390.00	399.0
402.00	1256.0
403.10	2073.0
403.80	478.0
421.00	1087.0
422.00	1025.0
423.00	9261.0
424.00	3407.0
425.10	645.0
438.30	421.0
441.10	29784.0
442.10	169280.0
443.10	31984.0
444.00	2192.0
444.90	358.0

Instrument :
BNA_M
ClientSampleId :
DFTPP46

m/z	Abundance
38.00	1192.0
39.00	3557.0
40.00	1152.0
42.90	326.0
44.00	737.0
48.80	566.0
50.00	16267.0
51.00	57952.0
52.10	2685.0
55.00	690.0
55.90	1696.0
57.00	4161.0
60.90	429.0
62.00	903.0
63.00	3258.0
65.00	912.0
67.90	1411.0
69.00	61456.0
73.00	1783.0
74.00	7428.0
75.00	9342.0
76.10	4370.0
77.10	67888.0
78.10	4461.0
79.00	5438.0
80.00	3923.0
81.00	5911.0
82.00	1401.0
83.00	1165.0
83.90	468.0
84.80	928.0
86.00	1755.0
86.70	754.0
87.10	495.0
90.90	1117.0
91.90	1826.0
93.00	9886.0
94.10	573.0
95.00	322.0
96.20	428.0
98.00	7012.0
99.00	6153.0
100.90	2744.0
102.80	978.0
104.00	1827.0
105.10	1581.0
106.10	529.0
107.00	21992.0
108.00	3192.0
109.00	717.0
110.00	29384.0
111.00	4872.0
111.80	667.0
116.00	1490.0
117.00	23960.0
117.90	1651.0
118.60	304.0
122.10	1097.0
122.90	3071.0
124.00	938.0
125.10	908.0
127.00	64040.0
128.00	5400.0
129.00	26616.0
130.00	1976.0
130.90	493.0
132.80	622.0
133.90	1116.0
135.10	2220.0
135.90	311.0
137.10	1610.0
137.70	391.0
139.80	357.0
141.00	3314.0
142.10	628.0
142.90	1374.0
145.00	488.0
145.80	626.0
146.90	2512.0
148.00	6466.0
148.80	1341.0
150.00	747.0
151.00	561.0
152.90	2053.0
153.90	1471.0
155.00	2759.0
156.00	2755.0
157.10	827.0
158.20	928.0
159.10	503.0
159.90	767.0

Instrument :
BNA_M
ClientSampleId :
DFTPP46

160.90	2341.0
162.00	351.0
165.00	2041.0
166.00	1184.0
167.00	8050.0
168.00	5243.0
169.20	836.0
169.80	501.0
171.90	850.0
172.90	1016.0
174.10	1654.0
175.00	3663.0
175.90	1013.0
177.10	1902.0
177.80	527.0
178.90	5689.0
180.00	4412.0
180.90	1545.0
182.10	396.0
184.00	791.0
185.00	3006.0
186.00	24176.0
187.00	6958.0
188.30	608.0
188.90	1484.0
189.70	664.0
191.10	899.0
192.00	2657.0
193.00	2109.0
193.90	505.0
195.00	876.0
195.90	4291.0
197.90	142848.0
199.00	10178.0
200.00	1451.0
201.20	784.0
201.60	650.0
203.00	1632.0
204.00	6716.0
205.00	11626.0
206.00	47016.0
207.10	7868.0
208.10	1177.0
208.90	1578.0
209.90	781.0
210.30	824.0
211.10	1484.0
214.80	708.0
215.90	1089.0
216.90	14487.0
218.00	1975.0
220.10	604.0
221.20	6091.0
222.00	1842.0
223.00	3499.0
224.00	27864.0
225.00	7380.0
226.00	506.0
226.90	13256.0
227.90	1856.0
228.90	2950.0
230.70	781.0
233.70	1044.0
234.80	763.0
235.80	527.0
236.10	572.0
236.90	1240.0
238.80	538.0
240.10	357.0
240.90	636.0
242.00	1450.0
243.10	1568.0
244.00	20336.0
245.00	2799.0
245.90	4214.0
247.00	920.0
249.00	525.0
252.90	1025.0
253.80	1054.0
255.00	116312.0
256.00	14143.0
257.00	2181.0
257.90	6634.0
259.00	923.0
259.90	308.0
265.00	3354.0
265.70	417.0
272.00	317.0
273.00	3080.0
273.90	8039.0
275.00	47016.0
276.00	4355.0
276.90	4748.0
277.90	441.0
281.00	355.0

Instrument :
BNA_M
ClientSampleId :
DFTPP46

281.80	376.0
283.00	1058.0
284.10	898.0
285.10	722.0
292.90	1007.0
293.90	544.0
296.00	13918.0
296.90	2363.0
301.20	327.0
303.10	1350.0
304.00	384.0
312.80	542.0
314.00	670.0
315.00	1487.0
316.00	682.0
321.00	581.0
323.00	4581.0
324.00	876.0
327.10	946.0
327.70	418.0
331.90	434.0
333.00	594.0
334.00	2519.0
334.90	613.0
340.90	658.0
346.00	945.0
352.00	1794.0
353.10	1320.0
354.10	1431.0
354.90	539.0
365.00	8360.0
365.80	1335.0
372.00	3263.0
383.00	531.0
402.00	993.0
403.00	1585.0
403.90	404.0
421.00	1304.0
422.00	1214.0
423.00	5967.0
424.10	1804.0
425.00	690.0
441.00	18832.0
442.10	131904.0
443.10	22720.0
444.00	2990.0

Instrument :
BNA_M
ClientSampleId :
DFTPP46

m/z	Abundance
40.00	1343.0
44.00	939.0
45.30	326.0
51.00	457.0
72.90	1270.0
77.10	316.0
84.10	364.0
133.00	514.0
147.10	578.0
192.90	445.0
206.90	1732.0
220.90	308.0
281.00	693.0
355.00	340.0

Instrument :

BNA_M

ClientSampleId :

DFTPP46