

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM030920\
 Data File : BM025506.D
 Acq On : 12 Mar 2020 00:10
 Operator : CG/JU
 Sample : PB127316BL
 Misc :
 ALS Vial : 99 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK16

Manual Integrations
 APPROVED

mohammad
 3/13/2020 8:50:03 AM

Quant Time: Mar 12 02:40:47 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM030920.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 12 02:03:23 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.66	152	3598	0.40	ng/ul	0.00
2) Naphthalene-d8	10.42	136	15401	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.28	164	11218	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.03	188	24195m	0.40	ng/ul	0.00
16) Chrysene-d12	21.22	240	15833	0.40	ng/ul	0.00
20) Perylene-d12	23.40	264	16919	0.40	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Methylnaphthalene-d10	12.02	152	7792	0.28	ng/ul	0.00
14) Fluoranthene-d10	19.06	212	18386	0.28	ng/ul	0.00

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM030920\
Data File : BM025506.D
Acq On : 12 Mar 2020 00:10
Operator : CG/JU
Sample : PB127316BL
Misc :
ALS Vial : 99 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK16

Manual Integrations
APPROVED

mohammad
3/13/2020 8:50:03 AM

Quant Time: Mar 12 02:40:47 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM030920.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 12 02:03:23 2020
Response via : Initial Calibration

