

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM030921\
 Data File : BM029039.D
 Acq On : 09 Mar 2021 10:04
 Operator : CG/JU
 Sample : PB135024BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 PB135024BL

Quant Time: Mar 09 10:35:06 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM022321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 09 10:08:02 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.669	152	9604	20.00	ng	0.00
21) Naphthalene-d8	10.463	136	34871	20.00	ng	# 0.00
39) Acenaphthene-d10	14.339	164	18904	20.00	ng	0.00
64) Phenanthrene-d10	17.086	188	38721	20.00	ng	0.00
76) Chrysene-d12	21.262	240	39565	20.00	ng	# 0.00
86) Perylene-d12	23.415	264	43651	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.246	112	83332	143.82	ng	0.00
7) Phenol-d6	6.869	99	95164	124.35	ng	0.00
23) Nitrobenzene-d5	8.845	82	62309	96.40	ng	0.00
42) 2,4,6-Tribromophenol	15.839	330	30838	137.64	ng	0.00
45) 2-Fluorobiphenyl	12.951	172	129741	91.37	ng	0.00
79) Terphenyl-d14	19.709	244	195964	91.46	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM030921\
Data File : BM029039.D
Acq On : 09 Mar 2021 10:04
Operator : CG/JU
Sample : PB135024BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
PB135024BL

Quant Time: Mar 09 10:35:06 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM022321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 09 10:08:02 2021
Response via : Initial Calibration

