

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM030921\
 Data File : BM029048.D
 Acq On : 09 Mar 2021 16:24
 Operator : CG/JU
 Sample : M1552-01
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 24368

Quant Time: Mar 09 17:01:18 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM022321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 09 10:08:02 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.669	152	9719	20.00	ng	0.00
21) Naphthalene-d8	10.463	136	36545	20.00	ng	# 0.00
39) Acenaphthene-d10	14.333	164	21726	20.00	ng	0.00
64) Phenanthrene-d10	17.080	188	47977	20.00	ng	0.00
76) Chrysene-d12	21.256	240	47742	20.00	ng	# 0.00
86) Perylene-d12	23.415	264	49314	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.245	112	76877	131.11	ng	0.00
7) Phenol-d6	6.869	99	84572	109.20	ng	0.00
23) Nitrobenzene-d5	8.845	82	60650	89.54	ng	0.00
42) 2,4,6-Tribromophenol	15.839	330	39943	155.12	ng	0.00
45) 2-Fluorobiphenyl	12.951	172	136259	83.49	ng	0.00
79) Terphenyl-d14	19.709	244	251130	97.13	ng	0.00
Target Compounds						
10) Phenol	6.898	94	1767	2.30	ng	Qvalue 92
50) Dimethylphthalate	13.804	163	7551	4.59	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM030921\
Data File : BM029048.D
Acq On : 09 Mar 2021 16:24
Operator : CG/JU
Sample : M1552-01
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
24368

Quant Time: Mar 09 17:01:18 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM022321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 09 10:08:02 2021
Response via : Initial Calibration

