

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM030921\
 Data File : BM029051.D
 Acq On : 09 Mar 2021 18:15
 Operator : CG/JU
 Sample : M1576-11
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 IDW-GW806-SO-030421

Quant Time: Mar 09 23:39:15 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM022321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 09 10:08:02 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.669	152	9348	20.00	ng	0.00
21) Naphthalene-d8	10.457	136	33912	20.00	ng	# 0.00
39) Acenaphthene-d10	14.333	164	19147	20.00	ng	0.00
64) Phenanthrene-d10	17.080	188	39418	20.00	ng	0.00
76) Chrysene-d12	21.256	240	40438	20.00	ng	# 0.00
86) Perylene-d12	23.415	264	42414	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.246	112	66739	118.34	ng	0.00
7) Phenol-d6	6.863	99	70061	94.05	ng	0.00
23) Nitrobenzene-d5	8.840	82	53494	85.11	ng	0.00
42) 2,4,6-Tribromophenol	15.833	330	32589	143.61	ng	0.00
45) 2-Fluorobiphenyl	12.945	172	115175	80.08	ng	0.00
79) Terphenyl-d14	19.709	244	212904	97.22	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM030921\
Data File : BM029051.D
Acq On : 09 Mar 2021 18:15
Operator : CG/JU
Sample : M1576-11
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
IDW-GW806-SO-030421

Quant Time: Mar 09 23:39:15 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM022321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 09 10:08:02 2021
Response via : Initial Calibration

