

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM030923\
 Data File : BM038943.D
 Acq On : 09 Mar 2023 13:31
 Operator : CG/JU
 Sample : 01784-01
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DCAD0

Quant Time: Mar 09 14:44:11 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM030323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 09 13:36:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.995	152	11142	0.400	ng/ul	0.00
4) Naphthalene-d8	10.811	136	38072	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.633	164	21750	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.375	188	45121	0.400	ng/ul	0.00
17) Chrysene-d12	21.553	240	32926	0.400	ng/ul	0.00
23) Perylene-d12	24.000	264	25894	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.392	96	141199	11.200	ng/ul	0.01
6) 2-Methylnaphthalene-d10	12.389	152	12940	0.281	ng/ul	0.00
18) Fluoranthene-d10	19.398	212	30439	0.388	ng/ul	0.00
Target Compounds						
					Qvalue	
5) Naphthalene	10.861	128	5468601	60.079	ng/ul	99
7) 2-Methylnaphthalene	12.466	142	501294	9.076	ng/ul	100
8) 1-Methylnaphthalene	12.686	142	418043	7.215	ng/ul	98
10) Acenaphthylene	14.355	152	7464	0.094	ng/ul#	85
11) Acenaphthene	14.693	153	476187	7.321	ng/ul	97
12) Fluorene	15.678	166	156362	2.129	ng/ul	99
15) Phenanthrene	17.417	178	5350	0.045	ng/ul	94
16) Anthracene	17.505	178	2115	0.022	ng/ul#	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM030923\
Data File : BM038943.D
Acq On : 09 Mar 2023 13:31
Operator : CG/JU
Sample : 01784-01
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DCAD0

Quant Time: Mar 09 14:44:11 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM030323.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 09 13:36:03 2023
Response via : Initial Calibration

