

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM030923\
 Data File : BM038963.D
 Acq On : 10 Mar 2023 02:49
 Operator : CG/JU
 Sample : 01729-09
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C0BY1

Quant Time: Mar 10 03:55:20 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM030323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 09 13:36:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.995	152	10292	0.400	ng/ul	0.00
4) Naphthalene-d8	10.806	136	36645	0.400	ng/ul	# 0.00
9) Acenaphthene-d10	14.628	164	21786	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.374	188	45162	0.400	ng/ul	0.00
17) Chrysene-d12	21.550	240	33866	0.400	ng/ul	0.00
23) Perylene-d12	23.994	264	26474	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.375	96	47421	4.072	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.389	152	12763	0.288	ng/ul	0.00
18) Fluoranthene-d10	19.394	212	29474	0.366	ng/ul	0.00
Target Compounds						Qvalue
5) Naphthalene	10.855	128	11959	0.137	ng/ul#	92
7) 2-Methylnaphthalene	12.461	142	26486	0.498	ng/ul	99
8) 1-Methylnaphthalene	12.681	142	14652	0.263	ng/ul	92
12) Fluorene	15.674	166	2412	0.033	ng/ul#	72
15) Phenanthrene	17.412	178	5399	0.045	ng/ul#	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM030923\
Data File : BM038963.D
Acq On : 10 Mar 2023 02:49
Operator : CG/JU
Sample : 01729-09
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
C0BY1

Quant Time: Mar 10 03:55:20 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM030323.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 09 13:36:03 2023
Response via : Initial Calibration

