


Data Path : Z:\HPCHEM1\BNA\_M\DATA\BM031017\  
 Data File : BM009166.D  
 Acq On : 10 Mar 2017 08:37  
 Operator : SJ/MA  
 Sample : DFTPP47  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_M  
 ClientSampleId :  
 DFTPP47

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA\_M\METHODS\SOM02.2-TEST.M  
 Title : SVOA CALIBRATION  
 Last Update : Wed Nov 30 00:32:18 2016



AutoFind: Scans 2482, 2483, 2484; Background Corrected with Scan 2476

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 46.5      | 112264  | PASS             |
| 68          | 69           | 0.00         | 2            | 1.9       | 2283    | PASS             |
| 69          | 198          | 0.00         | 100          | 48.8      | 117936  | PASS             |
| 70          | 69           | 0.00         | 2            | 0.6       | 671     | PASS             |
| 127         | 198          | 10           | 80           | 45.7      | 110416  | PASS             |
| 197         | 198          | 0.00         | 2            | 0.4       | 1040    | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 241493  | PASS             |
| 199         | 198          | 5            | 9            | 6.8       | 16507   | PASS             |
| 275         | 198          | 10           | 60           | 27.2      | 65800   | PASS             |
| 365         | 198          | 1            | 100          | 3.8       | 9205    | PASS             |
| 441         | 443          | 0.01         | 100          | 81.1      | 26026   | PASS             |
| 442         | 198          | 50           | 100          | 67.2      | 162328  | PASS             |
| 443         | 442          | 15           | 24           | 19.8      | 32077   | PASS             |

| m/z    | Abundance |
|--------|-----------|
| 36.80  | 617.0     |
| 38.10  | 1399.0    |
| 39.00  | 8637.0    |
| 40.00  | 2134.0    |
| 41.10  | 694.0     |
| 43.00  | 313.0     |
| 44.10  | 933.0     |
| 45.20  | 777.0     |
| 47.10  | 605.0     |
| 48.90  | 560.0     |
| 50.10  | 31408.0   |
| 51.10  | 116936.0  |
| 52.10  | 6162.0    |
| 55.00  | 1029.0    |
| 56.00  | 3724.0    |
| 57.00  | 8584.0    |
| 57.90  | 309.0     |
| 61.00  | 1540.0    |
| 62.00  | 1431.0    |
| 63.00  | 3874.0    |
| 64.10  | 568.0     |
| 64.90  | 2463.0    |
| 67.00  | 322.0     |
| 67.90  | 2053.0    |
| 69.00  | 122984.0  |
| 70.00  | 429.0     |
| 73.10  | 1923.0    |
| 74.00  | 14123.0   |
| 75.00  | 20144.0   |
| 76.10  | 8627.0    |
| 77.10  | 135808.0  |
| 78.00  | 10282.0   |
| 79.00  | 10154.0   |
| 80.00  | 8271.0    |
| 81.00  | 8485.0    |
| 82.00  | 2482.0    |
| 83.10  | 2597.0    |
| 83.90  | 446.0     |
| 84.90  | 966.0     |
| 85.90  | 3713.0    |
| 87.10  | 1227.0    |
| 88.10  | 575.0     |
| 88.80  | 328.0     |
| 89.60  | 334.0     |
| 90.90  | 2356.0    |
| 92.00  | 3626.0    |
| 93.00  | 17528.0   |
| 94.10  | 1188.0    |
| 94.90  | 445.0     |
| 95.90  | 670.0     |
| 97.00  | 724.0     |
| 98.00  | 12288.0   |
| 99.00  | 10075.0   |
| 99.90  | 973.0     |
| 101.10 | 4134.0    |
| 102.00 | 388.0     |
| 103.10 | 1585.0    |
| 104.00 | 3164.0    |
| 105.10 | 3261.0    |
| 106.00 | 950.0     |
| 107.00 | 36056.0   |
| 108.00 | 4208.0    |
| 109.10 | 1601.0    |
| 110.00 | 57040.0   |
| 111.00 | 10681.0   |
| 112.00 | 1949.0    |
| 114.90 | 394.0     |
| 116.00 | 2672.0    |
| 117.00 | 41000.0   |
| 117.90 | 3069.0    |
| 118.90 | 408.0     |
| 119.80 | 378.0     |
| 121.90 | 3020.0    |
| 122.90 | 3624.0    |
| 123.90 | 1465.0    |
| 125.00 | 1681.0    |
| 126.00 | 307.0     |
| 127.00 | 110040.0  |
| 128.00 | 8011.0    |
| 129.00 | 47584.0   |
| 129.90 | 4302.0    |
| 130.80 | 582.0     |
| 131.90 | 662.0     |
| 132.90 | 590.0     |
| 133.90 | 1049.0    |
| 135.00 | 2922.0    |
| 136.00 | 1696.0    |
| 136.90 | 3227.0    |
| 138.20 | 455.0     |
| 139.90 | 515.0     |
| 141.00 | 5310.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP47

|        |          |
|--------|----------|
| 142.00 | 1035.0   |
| 143.00 | 1645.0   |
| 144.10 | 474.0    |
| 145.00 | 626.0    |
| 146.00 | 1887.0   |
| 146.90 | 3620.0   |
| 147.90 | 6396.0   |
| 149.00 | 1219.0   |
| 149.90 | 452.0    |
| 151.00 | 1368.0   |
| 151.70 | 716.0    |
| 153.00 | 1656.0   |
| 154.10 | 713.0    |
| 155.00 | 3449.0   |
| 156.00 | 4080.0   |
| 157.10 | 1450.0   |
| 158.10 | 1504.0   |
| 159.10 | 1555.0   |
| 160.00 | 1413.0   |
| 160.90 | 2469.0   |
| 161.90 | 1710.0   |
| 163.00 | 370.0    |
| 164.20 | 451.0    |
| 165.10 | 1983.0   |
| 165.90 | 2801.0   |
| 167.00 | 14471.0  |
| 168.00 | 10159.0  |
| 169.00 | 1462.0   |
| 169.80 | 708.0    |
| 171.10 | 925.0    |
| 172.00 | 1847.0   |
| 173.00 | 1321.0   |
| 173.90 | 2769.0   |
| 175.00 | 5045.0   |
| 176.10 | 1409.0   |
| 176.90 | 1559.0   |
| 177.90 | 662.0    |
| 179.00 | 10708.0  |
| 180.00 | 7905.0   |
| 181.00 | 3434.0   |
| 182.90 | 310.0    |
| 183.90 | 855.0    |
| 184.90 | 5324.0   |
| 186.00 | 34248.0  |
| 187.10 | 10067.0  |
| 188.00 | 1073.0   |
| 188.90 | 1912.0   |
| 189.80 | 599.0    |
| 190.20 | 356.0    |
| 191.20 | 1485.0   |
| 192.00 | 3482.0   |
| 193.10 | 4006.0   |
| 194.00 | 687.0    |
| 195.10 | 436.0    |
| 195.90 | 6411.0   |
| 197.10 | 1852.0   |
| 198.00 | 216832.0 |
| 199.00 | 14701.0  |
| 199.90 | 852.0    |
| 201.20 | 880.0    |
| 201.90 | 445.0    |
| 203.00 | 1782.0   |
| 204.00 | 10881.0  |
| 205.00 | 17560.0  |
| 206.00 | 61768.0  |
| 207.10 | 10211.0  |
| 208.00 | 2311.0   |
| 209.00 | 1197.0   |
| 209.80 | 1052.0   |
| 210.30 | 840.0    |
| 211.00 | 1995.0   |
| 211.70 | 395.0    |
| 215.00 | 854.0    |
| 216.00 | 2036.0   |
| 217.00 | 18416.0  |
| 217.90 | 2902.0   |
| 219.00 | 384.0    |
| 219.80 | 489.0    |
| 220.90 | 7558.0   |
| 222.00 | 2231.0   |
| 222.90 | 4171.0   |
| 224.00 | 33856.0  |
| 225.00 | 8259.0   |
| 225.80 | 972.0    |
| 227.00 | 13638.0  |
| 227.90 | 2681.0   |
| 229.00 | 2080.0   |
| 230.00 | 497.0    |
| 230.90 | 2238.0   |
| 234.00 | 2418.0   |
| 235.00 | 1644.0   |
| 235.90 | 1036.0   |
| 236.90 | 2182.0   |
| 238.10 | 486.0    |
| 239.00 | 1161.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP47

|        |          |
|--------|----------|
| 240.00 | 707.0    |
| 240.80 | 954.0    |
| 242.20 | 1611.0   |
| 243.00 | 2058.0   |
| 244.00 | 25144.0  |
| 245.10 | 4717.0   |
| 245.90 | 5753.0   |
| 246.90 | 1597.0   |
| 248.00 | 361.0    |
| 249.00 | 486.0    |
| 251.10 | 501.0    |
| 252.90 | 421.0    |
| 254.00 | 888.0    |
| 255.00 | 132416.0 |
| 256.00 | 20280.0  |
| 257.10 | 1780.0   |
| 258.00 | 8050.0   |
| 258.90 | 765.0    |
| 264.90 | 4432.0   |
| 266.00 | 978.0    |
| 270.70 | 671.0    |
| 273.00 | 3880.0   |
| 274.00 | 11601.0  |
| 275.00 | 53264.0  |
| 276.00 | 7541.0   |
| 277.00 | 6501.0   |
| 277.90 | 1265.0   |
| 281.00 | 952.0    |
| 281.80 | 309.0    |
| 282.90 | 1035.0   |
| 283.90 | 319.0    |
| 285.10 | 1265.0   |
| 290.00 | 303.0    |
| 293.10 | 1347.0   |
| 294.00 | 1041.0   |
| 296.00 | 16424.0  |
| 296.90 | 3299.0   |
| 300.80 | 430.0    |
| 303.00 | 2239.0   |
| 304.30 | 490.0    |
| 313.90 | 1198.0   |
| 315.00 | 1978.0   |
| 316.00 | 762.0    |
| 321.00 | 616.0    |
| 323.00 | 4069.0   |
| 324.00 | 1096.0   |
| 327.10 | 1325.0   |
| 327.90 | 790.0    |
| 331.90 | 372.0    |
| 333.00 | 380.0    |
| 334.00 | 3634.0   |
| 334.70 | 504.0    |
| 335.10 | 721.0    |
| 341.00 | 369.0    |
| 346.10 | 682.0    |
| 352.00 | 2147.0   |
| 353.20 | 372.0    |
| 354.10 | 1892.0   |
| 355.20 | 519.0    |
| 365.00 | 6864.0   |
| 365.90 | 908.0    |
| 370.90 | 455.0    |
| 372.00 | 1551.0   |
| 373.00 | 442.0    |
| 382.60 | 399.0    |
| 383.00 | 390.0    |
| 390.90 | 436.0    |
| 402.10 | 731.0    |
| 402.90 | 1400.0   |
| 404.00 | 475.0    |
| 421.10 | 1069.0   |
| 422.00 | 735.0    |
| 423.10 | 7167.0   |
| 424.10 | 3452.0   |
| 425.00 | 780.0    |
| 441.00 | 18096.0  |
| 442.10 | 110920.0 |
| 443.10 | 19392.0  |
| 444.10 | 2130.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP47

| m/z    | Abundance |
|--------|-----------|
| 36.80  | 594.0     |
| 38.10  | 1139.0    |
| 39.10  | 9869.0    |
| 40.00  | 1148.0    |
| 40.90  | 794.0     |
| 42.90  | 479.0     |
| 44.10  | 1309.0    |
| 49.00  | 702.0     |
| 50.00  | 35224.0   |
| 51.10  | 134144.0  |
| 52.00  | 7242.0    |
| 55.00  | 1630.0    |
| 56.00  | 4272.0    |
| 57.00  | 7942.0    |
| 58.00  | 430.0     |
| 61.10  | 1320.0    |
| 62.10  | 1516.0    |
| 63.00  | 6151.0    |
| 64.10  | 778.0     |
| 65.00  | 3019.0    |
| 68.00  | 2601.0    |
| 69.00  | 140160.0  |
| 70.00  | 950.0     |
| 73.10  | 2858.0    |
| 74.10  | 16832.0   |
| 75.00  | 29152.0   |
| 76.00  | 7294.0    |
| 77.10  | 153856.0  |
| 78.10  | 9531.0    |
| 79.00  | 10267.0   |
| 80.00  | 8936.0    |
| 81.00  | 14046.0   |
| 82.00  | 3713.0    |
| 83.00  | 1665.0    |
| 83.90  | 383.0     |
| 85.00  | 1786.0    |
| 86.00  | 2526.0    |
| 86.80  | 737.0     |
| 88.00  | 920.0     |
| 88.80  | 401.0     |
| 90.20  | 356.0     |
| 91.00  | 3379.0    |
| 92.00  | 3751.0    |
| 93.00  | 20744.0   |
| 93.90  | 1461.0    |
| 94.90  | 384.0     |
| 95.90  | 1112.0    |
| 96.90  | 439.0     |
| 98.00  | 16672.0   |
| 99.00  | 11817.0   |
| 99.90  | 1162.0    |
| 101.00 | 5793.0    |
| 101.80 | 451.0     |
| 103.00 | 2415.0    |
| 104.00 | 4475.0    |
| 105.10 | 3349.0    |
| 106.00 | 1328.0    |
| 107.00 | 45688.0   |
| 108.00 | 7265.0    |
| 109.00 | 1478.0    |
| 110.00 | 78848.0   |
| 111.00 | 12828.0   |
| 112.10 | 1462.0    |
| 114.90 | 362.0     |
| 116.00 | 1944.0    |
| 117.00 | 51496.0   |
| 118.00 | 3270.0    |
| 119.00 | 734.0     |
| 120.10 | 528.0     |
| 121.90 | 2738.0    |
| 122.90 | 4204.0    |
| 124.00 | 2902.0    |
| 125.00 | 1425.0    |
| 126.20 | 1581.0    |
| 127.00 | 135552.0  |
| 128.00 | 10215.0   |
| 129.00 | 59504.0   |
| 130.00 | 7200.0    |
| 131.10 | 1136.0    |
| 132.10 | 796.0     |
| 132.90 | 932.0     |
| 134.00 | 1265.0    |
| 134.90 | 5001.0    |
| 136.00 | 2278.0    |
| 136.90 | 1692.0    |
| 137.10 | 1751.0    |
| 138.10 | 568.0     |
| 139.10 | 450.0     |
| 140.10 | 1012.0    |
| 140.90 | 7746.0    |
| 141.90 | 1903.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP47

|        |          |
|--------|----------|
| 143.00 | 1741.0   |
| 143.90 | 813.0    |
| 145.90 | 1636.0   |
| 147.00 | 4623.0   |
| 148.00 | 12216.0  |
| 149.00 | 2863.0   |
| 149.90 | 983.0    |
| 150.80 | 818.0    |
| 151.70 | 566.0    |
| 152.90 | 1455.0   |
| 154.00 | 1880.0   |
| 155.10 | 5260.0   |
| 156.00 | 4976.0   |
| 157.20 | 1179.0   |
| 157.90 | 1896.0   |
| 159.00 | 1194.0   |
| 160.00 | 3000.0   |
| 161.00 | 2779.0   |
| 161.90 | 864.0    |
| 163.20 | 603.0    |
| 165.00 | 3634.0   |
| 165.90 | 1832.0   |
| 167.00 | 17896.0  |
| 168.00 | 11960.0  |
| 168.90 | 1348.0   |
| 170.20 | 425.0    |
| 170.90 | 835.0    |
| 171.90 | 1148.0   |
| 172.90 | 2091.0   |
| 174.00 | 4568.0   |
| 175.10 | 6234.0   |
| 175.90 | 1935.0   |
| 176.90 | 3091.0   |
| 177.90 | 824.0    |
| 178.90 | 11859.0  |
| 180.00 | 8822.0   |
| 181.10 | 4563.0   |
| 181.90 | 1054.0   |
| 183.90 | 1221.0   |
| 185.00 | 5980.0   |
| 186.00 | 42920.0  |
| 187.00 | 12752.0  |
| 188.00 | 1583.0   |
| 189.00 | 3768.0   |
| 190.00 | 379.0    |
| 190.50 | 321.0    |
| 191.10 | 1927.0   |
| 192.00 | 5212.0   |
| 193.00 | 4869.0   |
| 194.00 | 1403.0   |
| 195.00 | 782.0    |
| 196.00 | 6579.0   |
| 197.00 | 1269.0   |
| 198.00 | 285376.0 |
| 198.90 | 20256.0  |
| 200.00 | 2129.0   |
| 201.50 | 538.0    |
| 203.00 | 3012.0   |
| 204.00 | 14896.0  |
| 205.00 | 20080.0  |
| 206.00 | 82360.0  |
| 207.00 | 13726.0  |
| 208.00 | 2661.0   |
| 209.00 | 2394.0   |
| 210.10 | 2033.0   |
| 210.80 | 3695.0   |
| 214.80 | 653.0    |
| 215.90 | 2830.0   |
| 217.00 | 23160.0  |
| 218.00 | 2513.0   |
| 219.90 | 512.0    |
| 221.10 | 11362.0  |
| 221.90 | 3690.0   |
| 223.00 | 4131.0   |
| 224.00 | 50256.0  |
| 225.10 | 11123.0  |
| 225.90 | 1700.0   |
| 227.00 | 24656.0  |
| 227.90 | 2848.0   |
| 229.10 | 4491.0   |
| 230.00 | 423.0    |
| 231.10 | 1405.0   |
| 234.00 | 1679.0   |
| 235.00 | 1913.0   |
| 236.00 | 850.0    |
| 237.00 | 2053.0   |
| 238.10 | 417.0    |
| 239.00 | 1137.0   |
| 240.00 | 391.0    |
| 240.90 | 1059.0   |
| 242.00 | 2322.0   |
| 243.10 | 1905.0   |
| 244.00 | 35944.0  |
| 245.00 | 4892.0   |
| 245.90 | 9808.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP47

|        |          |
|--------|----------|
| 247.00 | 1589.0   |
| 247.90 | 356.0    |
| 249.00 | 1379.0   |
| 250.90 | 437.0    |
| 252.10 | 390.0    |
| 253.00 | 1143.0   |
| 254.00 | 1406.0   |
| 255.00 | 192704.0 |
| 256.00 | 28472.0  |
| 257.00 | 2561.0   |
| 258.00 | 13491.0  |
| 259.00 | 2859.0   |
| 261.00 | 309.0    |
| 263.90 | 689.0    |
| 265.00 | 4831.0   |
| 265.90 | 602.0    |
| 270.90 | 767.0    |
| 272.10 | 1213.0   |
| 273.00 | 5868.0   |
| 274.00 | 16378.0  |
| 275.00 | 80840.0  |
| 276.00 | 10356.0  |
| 276.90 | 8495.0   |
| 278.00 | 1701.0   |
| 281.00 | 1020.0   |
| 282.00 | 450.0    |
| 283.10 | 996.0    |
| 284.00 | 806.0    |
| 285.20 | 1465.0   |
| 285.80 | 408.0    |
| 293.00 | 1475.0   |
| 293.90 | 830.0    |
| 295.00 | 695.0    |
| 296.00 | 26184.0  |
| 297.00 | 3276.0   |
| 300.80 | 475.0    |
| 303.10 | 2791.0   |
| 304.20 | 652.0    |
| 307.90 | 364.0    |
| 309.00 | 705.0    |
| 313.90 | 721.0    |
| 315.00 | 2698.0   |
| 315.90 | 1556.0   |
| 320.90 | 1146.0   |
| 323.00 | 7003.0   |
| 324.00 | 1382.0   |
| 326.90 | 1258.0   |
| 327.90 | 726.0    |
| 331.90 | 493.0    |
| 333.20 | 487.0    |
| 334.00 | 4575.0   |
| 335.00 | 1100.0   |
| 341.10 | 439.0    |
| 345.90 | 1882.0   |
| 347.10 | 452.0    |
| 352.00 | 1873.0   |
| 353.00 | 1300.0   |
| 354.10 | 1999.0   |
| 355.00 | 806.0    |
| 365.00 | 9975.0   |
| 365.90 | 1663.0   |
| 371.00 | 944.0    |
| 372.00 | 3780.0   |
| 373.00 | 731.0    |
| 382.90 | 739.0    |
| 384.20 | 371.0    |
| 390.10 | 393.0    |
| 402.10 | 1475.0   |
| 403.00 | 1058.0   |
| 403.90 | 633.0    |
| 420.70 | 1045.0   |
| 421.10 | 1214.0   |
| 422.10 | 1483.0   |
| 423.10 | 11694.0  |
| 424.00 | 4298.0   |
| 425.10 | 487.0    |
| 441.10 | 31616.0  |
| 442.10 | 195648.0 |
| 443.10 | 41256.0  |
| 444.00 | 4330.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP47

| m/z    | Abundance |
|--------|-----------|
| 36.90  | 315.0     |
| 38.10  | 936.0     |
| 39.10  | 6623.0    |
| 40.10  | 952.0     |
| 41.00  | 394.0     |
| 43.00  | 431.0     |
| 44.00  | 729.0     |
| 49.10  | 618.0     |
| 50.10  | 22640.0   |
| 51.00  | 85712.0   |
| 52.00  | 4697.0    |
| 55.00  | 627.0     |
| 56.00  | 2602.0    |
| 57.00  | 7041.0    |
| 57.90  | 521.0     |
| 61.00  | 1710.0    |
| 62.10  | 1425.0    |
| 63.00  | 4028.0    |
| 64.10  | 463.0     |
| 65.00  | 2007.0    |
| 67.90  | 2196.0    |
| 69.00  | 90664.0   |
| 70.00  | 635.0     |
| 73.10  | 2317.0    |
| 74.00  | 9195.0    |
| 75.00  | 15512.0   |
| 76.10  | 4820.0    |
| 77.10  | 93928.0   |
| 78.00  | 7308.0    |
| 79.00  | 7418.0    |
| 80.00  | 5774.0    |
| 81.00  | 7655.0    |
| 82.00  | 2518.0    |
| 82.90  | 1739.0    |
| 85.00  | 1214.0    |
| 86.00  | 1850.0    |
| 87.10  | 749.0     |
| 87.90  | 553.0     |
| 91.00  | 1777.0    |
| 92.00  | 2110.0    |
| 93.00  | 13303.0   |
| 94.20  | 874.0     |
| 95.30  | 599.0     |
| 95.90  | 1045.0    |
| 98.00  | 9954.0    |
| 99.00  | 8965.0    |
| 99.90  | 619.0     |
| 101.00 | 3782.0    |
| 103.00 | 2059.0    |
| 104.00 | 2254.0    |
| 105.00 | 3005.0    |
| 106.00 | 761.0     |
| 107.00 | 33560.0   |
| 108.00 | 4819.0    |
| 109.00 | 1233.0    |
| 110.00 | 48328.0   |
| 111.00 | 8330.0    |
| 111.80 | 780.0     |
| 115.90 | 2374.0    |
| 117.00 | 33088.0   |
| 118.00 | 2991.0    |
| 119.00 | 517.0     |
| 120.10 | 666.0     |
| 121.90 | 2702.0    |
| 122.90 | 2872.0    |
| 123.90 | 1366.0    |
| 125.00 | 1555.0    |
| 127.00 | 85656.0   |
| 128.00 | 6619.0    |
| 129.00 | 36584.0   |
| 130.00 | 3544.0    |
| 131.00 | 1981.0    |
| 131.90 | 317.0     |
| 134.10 | 1550.0    |
| 135.00 | 3083.0    |
| 136.10 | 1369.0    |
| 137.00 | 1917.0    |
| 137.90 | 685.0     |
| 139.00 | 543.0     |
| 141.00 | 7238.0    |
| 142.00 | 2254.0    |
| 142.90 | 1917.0    |
| 143.90 | 678.0     |
| 146.10 | 1087.0    |
| 146.90 | 3444.0    |
| 148.00 | 9322.0    |
| 149.00 | 1217.0    |
| 151.30 | 496.0     |
| 153.00 | 2199.0    |
| 154.00 | 961.0     |
| 154.90 | 3663.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP47

|        |          |
|--------|----------|
| 156.00 | 4501.0   |
| 157.20 | 811.0    |
| 158.00 | 1337.0   |
| 158.80 | 1196.0   |
| 159.90 | 1588.0   |
| 161.00 | 2045.0   |
| 162.00 | 937.0    |
| 162.90 | 436.0    |
| 164.90 | 2031.0   |
| 165.90 | 2011.0   |
| 167.00 | 13480.0  |
| 168.00 | 6772.0   |
| 168.80 | 477.0    |
| 169.30 | 677.0    |
| 169.90 | 644.0    |
| 171.30 | 498.0    |
| 171.90 | 1395.0   |
| 173.20 | 961.0    |
| 174.10 | 1946.0   |
| 175.00 | 3709.0   |
| 175.90 | 2157.0   |
| 176.90 | 2722.0   |
| 177.90 | 1246.0   |
| 178.90 | 7719.0   |
| 180.00 | 7300.0   |
| 181.00 | 3152.0   |
| 182.50 | 352.0    |
| 183.80 | 827.0    |
| 185.00 | 4402.0   |
| 186.00 | 26624.0  |
| 187.00 | 9168.0   |
| 187.90 | 1129.0   |
| 188.90 | 2474.0   |
| 190.00 | 551.0    |
| 191.00 | 2943.0   |
| 192.00 | 3194.0   |
| 193.00 | 3607.0   |
| 194.00 | 385.0    |
| 194.90 | 541.0    |
| 196.00 | 6885.0   |
| 196.70 | 1381.0   |
| 198.00 | 222272.0 |
| 198.90 | 14566.0  |
| 200.10 | 1345.0   |
| 201.30 | 1021.0   |
| 202.90 | 2330.0   |
| 204.00 | 9478.0   |
| 205.00 | 13827.0  |
| 206.00 | 63104.0  |
| 207.00 | 10337.0  |
| 207.80 | 2525.0   |
| 208.80 | 780.0    |
| 209.20 | 1178.0   |
| 210.00 | 1720.0   |
| 211.00 | 3435.0   |
| 214.80 | 1038.0   |
| 216.00 | 1217.0   |
| 217.00 | 22048.0  |
| 218.10 | 2033.0   |
| 221.10 | 7809.0   |
| 222.00 | 1680.0   |
| 223.00 | 4862.0   |
| 224.00 | 36064.0  |
| 225.00 | 9375.0   |
| 226.10 | 1180.0   |
| 227.00 | 19800.0  |
| 228.00 | 2349.0   |
| 229.00 | 3432.0   |
| 230.00 | 587.0    |
| 230.90 | 1775.0   |
| 231.90 | 310.0    |
| 232.80 | 472.0    |
| 234.00 | 1024.0   |
| 235.00 | 1134.0   |
| 236.00 | 1351.0   |
| 236.90 | 1397.0   |
| 238.00 | 356.0    |
| 239.00 | 566.0    |
| 240.00 | 643.0    |
| 241.10 | 1396.0   |
| 242.10 | 2033.0   |
| 243.00 | 2002.0   |
| 244.00 | 28688.0  |
| 245.00 | 4127.0   |
| 246.00 | 7151.0   |
| 246.90 | 987.0    |
| 248.90 | 581.0    |
| 249.90 | 359.0    |
| 251.20 | 388.0    |
| 252.00 | 516.0    |
| 252.70 | 666.0    |
| 253.80 | 1252.0   |
| 255.00 | 150080.0 |
| 256.00 | 20976.0  |
| 257.10 | 1893.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP47

|        |          |
|--------|----------|
| 257.90 | 8846.0   |
| 258.90 | 1063.0   |
| 265.00 | 4243.0   |
| 266.00 | 546.0    |
| 270.80 | 406.0    |
| 273.00 | 4536.0   |
| 273.90 | 10907.0  |
| 275.00 | 63296.0  |
| 276.00 | 8494.0   |
| 277.10 | 5360.0   |
| 277.90 | 827.0    |
| 281.00 | 438.0    |
| 281.90 | 989.0    |
| 283.10 | 1075.0   |
| 284.00 | 746.0    |
| 285.10 | 1110.0   |
| 288.90 | 487.0    |
| 293.00 | 1546.0   |
| 294.10 | 670.0    |
| 295.00 | 305.0    |
| 296.00 | 19560.0  |
| 296.90 | 3363.0   |
| 300.80 | 455.0    |
| 302.00 | 622.0    |
| 303.10 | 1426.0   |
| 304.10 | 475.0    |
| 307.90 | 421.0    |
| 310.30 | 374.0    |
| 313.80 | 320.0    |
| 315.00 | 2498.0   |
| 315.90 | 1063.0   |
| 320.80 | 1007.0   |
| 323.10 | 5262.0   |
| 324.00 | 1439.0   |
| 326.80 | 1072.0   |
| 328.00 | 315.0    |
| 333.10 | 715.0    |
| 334.00 | 4091.0   |
| 334.80 | 1003.0   |
| 340.90 | 479.0    |
| 345.90 | 2282.0   |
| 352.00 | 1872.0   |
| 353.00 | 1434.0   |
| 354.00 | 2352.0   |
| 355.00 | 483.0    |
| 365.00 | 10776.0  |
| 365.90 | 1306.0   |
| 371.00 | 445.0    |
| 372.00 | 3087.0   |
| 373.00 | 354.0    |
| 382.80 | 1179.0   |
| 390.00 | 579.0    |
| 401.90 | 1371.0   |
| 403.00 | 1502.0   |
| 403.90 | 597.0    |
| 421.00 | 1356.0   |
| 421.90 | 986.0    |
| 423.10 | 11437.0  |
| 424.10 | 3973.0   |
| 441.10 | 28368.0  |
| 442.10 | 180416.0 |
| 443.10 | 35584.0  |
| 444.10 | 3450.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP47

| m/z    | Abundance |
|--------|-----------|
| 40.00  | 1329.0    |
| 41.10  | 487.0     |
| 44.00  | 1031.0    |
| 47.10  | 365.0     |
| 56.00  | 331.0     |
| 66.80  | 353.0     |
| 73.00  | 1640.0    |
| 74.10  | 349.0     |
| 75.10  | 400.0     |
| 76.90  | 480.0     |
| 91.20  | 405.0     |
| 132.90 | 311.0     |
| 148.90 | 516.0     |
| 162.70 | 303.0     |
| 193.00 | 493.0     |
| 207.10 | 3454.0    |
| 207.80 | 571.0     |
| 209.10 | 635.0     |
| 249.10 | 378.0     |
| 255.00 | 525.0     |
| 280.90 | 920.0     |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP47