

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM031022\
 Data File : BM034207.D
 Acq On : 10 Mar 2022 19:21
 Operator : CG/JU
 Sample : SSTD02048
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020048

Quant Time: Mar 11 00:26:54 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM031022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Mar 11 00:23:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.954	152	179899	20.000	ng/u1	0.00
20) Naphthalene-d8	10.766	136	707088	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.595	164	457812	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.336	188	950325	20.000	ng/u1	0.00
79) Chrysene-d12	21.500	240	1002620	20.000	ng/u1	0.00
88) Perylene-d12	23.924	264	1074527	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.325	96	32126	8.213	ng/uL	0.00
4) Pyridine-d5	3.743	84	186761	16.403	ng/u1	0.00
7) Phenol-d5	7.107	99	225942	16.029	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.278	67	137790	18.161	ng/u1	0.00
11) 2-Chlorophenol-d4	7.484	132	215679	18.458	ng/u1	0.00
15) 4-Methylphenol-d8	8.660	113	198592	16.814	ng/u1	0.00
21) Nitrobenzene-d5	9.113	128	101386	18.506	ng/u1	0.00
24) 2-Nitrophenol-d4	9.842	143	112163	18.850	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.383	165	226595	18.466	ng/u1	0.00
31) 4-Chloroaniline-d4	10.895	131	266250	15.881	ng/u1	0.00
46) Dimethylphthalate-d6	14.001	166	666532	19.186	ng/u1	0.00
49) Acenaphthylene-d8	14.289	160	846070	19.476	ng/u1	0.00
54) 4-Nitrophenol-d4	14.771	143	96877	14.987	ng/u1	0.00
60) Fluorene-d10	15.583	176	576384	18.642	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.695	200	110764	18.061	ng/u1	0.00
73) Anthracene-d10	17.430	188	886950	18.947	ng/u1	0.00
81) Pyrene-d10	19.718	212	1048735	19.029	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.765	264	1065188	18.687	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.360	88	31482	7.435	ng/uL	100
5) Pyridine	3.766	79	196953	17.109	ng/u1	100
6) Benzaldehyde	7.084	77	153290	21.189	ng/u1	100
8) Phenol	7.131	94	223916	15.704	ng/u1	100
10) Bis(2-Chloroethyl)ether	7.366	93	191822	17.336	ng/u1	100
12) 2-Chlorophenol	7.513	128	214095	17.738	ng/u1	100
13) 2-Methylphenol	8.395	108	182079	16.378	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.489	45	180244	18.352	ng/u1	100
16) Acetophenone	8.778	105	314373	16.894	ng/u1	100
17) N-Nitroso-di-n-propyla...	8.766	70	152403	17.928	ng/u1	100
18) 4-Methylphenol	8.725	108	196370	16.172	ng/u1	100
19) Hexachloroethane	9.042	117	94028	18.944	ng/u1	100
22) Nitrobenzene	9.154	77	221778	17.901	ng/u1	100
23) Isophorone	9.689	82	413217	18.499	ng/u1	100
25) 2-Nitrophenol	9.872	139	116942	18.035	ng/u1	100
26) 2,4-Dimethylphenol	9.936	107	236657	18.045	ng/u1	100
27) Bis(2-Chloroethoxy)met...	10.172	93	254125	18.147	ng/u1	100
29) 2,4-Dichlorophenol	10.407	162	217465	17.931	ng/u1	100
30) Naphthalene	10.819	128	689090	18.202	ng/u1	100
32) 4-Chloroaniline	10.919	127	261987	15.345	ng/u1	100
33) Hexachlorobutadiene	11.113	225	176438	19.103	ng/u1	100
34) Caprolactam	11.677	113	57266	15.692	ng/u1	100
35) 4-Chloro-3-methylphenol	12.042	107	205971	17.381	ng/u1	100
36) 2-Methylnaphthalene	12.424	142	478908	17.638	ng/u1	100


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM031022\
 Data File : BM034207.D
 Acq On : 10 Mar 2022 19:21
 Operator : CG/JU
 Sample : SSTD02048
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020048

Quant Time: Mar 11 00:26:54 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM031022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Mar 11 00:23:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.648	142	486708	17.495	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.795	216	309241	19.337	ng/ul	100
40) Hexachlorocyclopentadiene	12.783	237	210192	19.468	ng/ul	100
41) 2,4,6-Trichlorophenol	13.030	196	186745	19.130	ng/ul	100
42) 2,4,5-Trichlorophenol	13.095	196	200280	18.602	ng/ul	100
43) 1,1'-Biphenyl	13.430	154	664945	18.655	ng/ul	100
44) 2-Chloronaphthalene	13.471	162	522433	18.648	ng/ul	100
45) 2-Nitroaniline	13.665	65	110556	17.149	ng/ul	100
47) Dimethylphthalate	14.048	163	647710	18.433	ng/ul	100
48) 2,6-Dinitrotoluene	14.160	165	130757	18.722	ng/ul	100
50) Acenaphthylene	14.318	152	819922	18.719	ng/ul	100
51) 3-Nitroaniline	14.489	138	113089	17.704	ng/ul	100
52) Acenaphthene	14.660	153	534722	18.438	ng/ul	100
53) 2,4-Dinitrophenol	14.689	184	91354	20.673	ng/ul	100
55) 4-Nitrophenol	14.789	109	84933	15.150	ng/ul	100
56) Dibenzofuran	14.989	168	780564	18.274	ng/ul	100
57) 2,4-Dinitrotoluene	14.948	165	182409	17.861	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	15.212	232	174725	18.723	ng/ul	100
59) Diethylphthalate	15.407	149	628473	18.221	ng/ul	100
61) Fluorene	15.636	166	639896	18.211	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.630	204	343850	18.176	ng/ul	100
63) 4-Nitroaniline	15.648	138	97248	16.457	ng/ul	100
66) 4,6-Dinitro-2-methylph...	15.707	198	109302	17.649	ng/ul	100
67) N-Nitrosodiphenylamine	15.842	169	534118	18.849	ng/ul	100
68) 4-Bromophenyl-phenylether	16.524	248	209735	20.186	ng/ul	100
69) Hexachlorobenzene	16.642	284	261552	19.349	ng/ul	100
70) Atrazine	16.789	200	210031	18.307	ng/ul	100
71) Pentachlorophenol	16.983	266	151597	17.672	ng/ul	100
72) Phenanthrene	17.377	178	994546	18.498	ng/ul	100
74) Anthracene	17.465	178	1000121	18.408	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.401	216	316929	19.462	ng/ul	100
76) Pentachlorobenzene	14.912	250	302011	19.806	ng/ul	100
77) Carbazole	17.730	167	778504	16.069	ng/ul	100
78) Di-n-butylphthalate	18.300	149	1003190	18.997	ng/ul	100
80) Fluoranthene	19.383	202	1192183	18.429	ng/ul	100
82) Pyrene	19.747	202	1223090	18.278	ng/ul	100
83) Butylbenzylphthalate	20.641	149	426954	18.645	ng/ul	100
84) 3,3'-Dichlorobenzidine	21.418	252	387748	19.884	ng/ul	100
85) Benzo(a)anthracene	21.488	228	1161486	17.572	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.418	149	646827	19.298	ng/ul	100
87) Chrysene	21.541	228	1221217	18.914	ng/ul	100
89) Di-n-octyl phthalate	22.347	149	1103596	16.530	ng/ul	100
90) Benzo(b)fluoranthene	23.183	252	1250895	16.818	ng/ul	100
91) Benzo(k)fluoranthene	23.235	252	1300275	18.603	ng/ul	100
93) Benzo(a)pyrene	23.818	252	1261267	18.060	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.447	276	1408909	19.325	ng/ul	100
95) Dibenzo(a,h)anthracene	26.465	278	1198528	19.053	ng/ul	100
96) Benzo(g,h,i)perylene	27.223	276	1283643	21.069	ng/ul	100





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 7.435 ng/uL
RT: 3.360 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM034207.D
Acq: 10 Mar 2022 19:21




Instrument :
BNA_M
ClientSampleId :
SSTD020048


Tgt Ion: 88 Resp: 31482
Ion Ratio Lower Upper
88 100
43 26.3 21.0 31.6
58 59.8 47.8 71.8



#5
Pyridine
Concen: 17.109 ng/uL
RT: 3.766 min Scan# 90
Delta R.T. 0.000 min
Lab File: BM034207.D
Acq: 10 Mar 2022 19:21

Tgt Ion: 79 Resp: 196953
Ion Ratio Lower Upper
79 100
52 59.6 47.7 71.5
51 32.0 25.6 38.4









#6
Benzaldehyde
Concen: 21.189 ng/ul
RT: 7.084 min Scan# 61
Delta R.T. 0.000 min
Lab File: BM034207.D
Acq: 10 Mar 2022 19:21




Instrument :
BNA_M
ClientSampleId :
SSTD020048


Tgt Ion: 77 Resp: 153290
Ion Ratio Lower Upper
77 100
105 101.0 80.8 121.2
106 95.8 76.6 115.0



#8
Phenol
Concen: 15.704 ng/ul
RT: 7.131 min Scan# 662
Delta R.T. 0.000 min
Lab File: BM034207.D
Acq: 10 Mar 2022 19:21

Tgt Ion: 94 Resp: 223916
Ion Ratio Lower Upper
94 100
65 28.6 22.9 34.3
66 42.1 33.7 50.5









#10
Bis(2-Chloroethyl)ether
Concen: 17.336 ng/ul
RT: 7.366 min Scan# 702
Delta R.T. 0.000 min
Lab File: BM034207.D
Acq: 10 Mar 2022 19:21




Instrument :
BNA_M
ClientSampleId :
SSTD020048


Tgt Ion: 93 Resp: 191822
Ion Ratio Lower Upper
93 100
63 68.4 54.7 82.1
95 33.2 26.6 39.8



#12
2-Chlorophenol
Concen: 17.738 ng/ul
RT: 7.513 min Scan# 727
Delta R.T. 0.000 min
Lab File: BM034207.D
Acq: 10 Mar 2022 19:21

Tgt Ion:128 Resp: 214095
Ion Ratio Lower Upper
128 100
64 38.3 30.6 46.0
130 32.2 25.8 38.6





#13

2-Methylphenol

Concen: 16.378 ng/ul

RT: 8.395 min Scan# 877

Delta R.T. 0.000 min

Lab File: BM034207.D

Acq: 10 Mar 2022 19:21

Instrument :

BNA_M

ClientSampleId :





SSTD020048

Tgt Ion:108 Resp: 182079

Ion Ratio Lower Upper

108 100

107 89.9 71.9 107.9



#14

2,2'-oxybis(1-Chloropropane)

Concen: 18.352 ng/ul

RT: 8.489 min Scan# 893

Delta R.T. 0.000 min

Lab File: BM034207.D

Acq: 10 Mar 2022 19:21




Tgt Ion: 45 Resp: 180244


Ion Ratio Lower Upper

45 100

77 21.4 17.1 25.7

79 16.0 12.8 19.2









#16
Acetophenone
Concen: 16.894 ng/ul
RT: 8.778 min Scan# 942
Delta R.T. 0.000 min
Lab File: BM034207.D
Acq: 10 Mar 2022 19:21




Instrument :
BNA_M
ClientSampleId :
SSTD020048


Tgt Ion:105 Resp: 314373
Ion Ratio Lower Upper
105 100
77 74.3 59.4 89.2
51 24.1 19.3 28.9



#17
N-Nitroso-di-n-propylamine
Concen: 17.928 ng/ul
RT: 8.766 min Scan# 940
Delta R.T. 0.000 min
Lab File: BM034207.D
Acq: 10 Mar 2022 19:21

Tgt Ion: 70 Resp: 152403
Ion Ratio Lower Upper
70 100
42 50.7 40.6 60.8
101 10.2 8.2 12.2
130 24.9 19.9 29.9









#18
4-Methylphenol
Concen: 16.172 ng/ul
RT: 8.725 min Scan# 911
Delta R.T. 0.000 min
Lab File: BM034207.D
Acq: 10 Mar 2022 19:21




Instrument :
BNA_M
ClientSampleId :
SSTD020048


Tgt Ion:108 Resp: 196370
Ion Ratio Lower Upper
108 100
107 111.2 89.0 133.4



#19
Hexachloroethane
Concen: 18.944 ng/ul
RT: 9.042 min Scan# 987
Delta R.T. 0.000 min
Lab File: BM034207.D
Acq: 10 Mar 2022 19:21

Tgt Ion:117 Resp: 94028
Ion Ratio Lower Upper
117 100
201 112.4 89.9 134.9
199 71.5 57.2 85.8









#22
Nitrobenzene
Concen: 17.901 ng/ul
RT: 9.154 min Scan# 1006
Delta R.T. 0.000 min
Lab File: BM034207.D
Acq: 10 Mar 2022 19:21




Instrument :
BNA_M
ClientSampleId :
SSTD020048


Tgt Ion: 77 Resp: 221778
Ion Ratio Lower Upper
77 100
123 47.7 38.2 57.2
65 13.2 10.6 15.8



#23
Isophorone
Concen: 18.499 ng/ul
RT: 9.689 min Scan# 1097
Delta R.T. 0.000 min
Lab File: BM034207.D
Acq: 10 Mar 2022 19:21

Tgt Ion: 82 Resp: 413217
Ion Ratio Lower Upper
82 100
95 8.5 6.8 10.2
138 20.6 16.5 24.7









#25
2-Nitrophenol
Concen: 18.035 ng/ul
RT: 9.872 min Scan# 1128
Delta R.T. 0.000 min
Lab File: BM034207.D
Acq: 10 Mar 2022 19:21




Instrument :
BNA_M
ClientSampleId :
SSTD020048


Tgt Ion:139 Resp: 116942
Ion Ratio Lower Upper
139 100
65 47.2 37.8 56.6
109 33.1 26.5 39.7



#26
2,4-Dimethylphenol
Concen: 18.045 ng/ul
RT: 9.936 min Scan# 1139
Delta R.T. 0.000 min
Lab File: BM034207.D
Acq: 10 Mar 2022 19:21

Tgt Ion:107 Resp: 236657
Ion Ratio Lower Upper
107 100
121 51.0 40.8 61.2
122 86.3 69.0 103.6





#27

Bis(2-Chloroethoxy)methane

Concen: 18.147 ng/ul

RT: 10.172 min Scan# 1179

Delta R.T. 0.000 min

Lab File: BM034207.D

Acq: 10 Mar 2022 19:21

Instrument :

BNA_M

ClientSampleId :

SSTD020048





Tgt Ion: 93 Resp: 254125

Ion Ratio Lower Upper

93 100

95 31.8 25.4 38.2

123 11.9 9.5 14.3



#29

2,4-Dichlorophenol

Concen: 17.931 ng/ul

RT: 10.407 min Scan# 1219

Delta R.T. 0.000 min

Lab File: BM034207.D

Acq: 10 Mar 2022 19:21




Tgt Ion: 162 Resp: 217465


Ion Ratio Lower Upper

162 100

164 66.0 52.8 79.2

98 32.5 26.0 39.0









#30
Naphthalene
Concen: 18.202 ng/ul
RT: 10.819 min Scan# 1289
Delta R.T. 0.000 min
Lab File: BM034207.D
Acq: 10 Mar 2022 19:21




Instrument :
BNA_M
ClientSampleId :
SSTD020048


Tgt Ion:128 Resp: 689090
Ion Ratio Lower Upper
128 100
129 10.9 8.7 13.1
127 12.4 9.9 14.9



#32
4-Chloroaniline
Concen: 15.345 ng/ul
RT: 10.919 min Scan# 1306
Delta R.T. 0.000 min
Lab File: BM034207.D
Acq: 10 Mar 2022 19:21

Tgt Ion:127 Resp: 261987
Ion Ratio Lower Upper
127 100
129 31.4 25.1 37.7









#33
Hexachlorobutadiene
Concen: 19.103 ng/ul
RT: 11.113 min Scan# 11
Delta R.T. 0.000 min
Lab File: BM034207.D
Acq: 10 Mar 2022 19:21




Instrument :
BNA_M
ClientSampleId :
SSTD020048


Tgt Ion:225 Resp: 176438
Ion Ratio Lower Upper
225 100
223 63.2 50.6 75.8
227 62.9 50.3 75.5



#34
Caprolactam
Concen: 15.692 ng/ul
RT: 11.677 min Scan# 1435
Delta R.T. 0.000 min
Lab File: BM034207.D
Acq: 10 Mar 2022 19:21

Tgt Ion:113 Resp: 57266
Ion Ratio Lower Upper
113 100
55 126.1 100.9 151.3
56 101.2 81.0 121.4









#35
 4-Chloro-3-methylphenol
 Concen: 17.381 ng/ul
 RT: 12.042 min Scan# 1497
 Delta R.T. 0.000 min
 Lab File: BM034207.D
 Acq: 10 Mar 2022 19:21




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020048


Tgt Ion:107 Resp: 205971
 Ion Ratio Lower Upper
 107 100
 144 28.7 23.0 34.4
 142 87.3 69.8 104.8



#36
 2-Methylnaphthalene
 Concen: 17.638 ng/ul
 RT: 12.424 min Scan# 1562
 Delta R.T. 0.000 min
 Lab File: BM034207.D
 Acq: 10 Mar 2022 19:21

Tgt Ion:142 Resp: 478908
 Ion Ratio Lower Upper
 142 100
 141 89.2 71.4 107.0









#37
1-Methylnaphthalene
Concen: 17.495 ng/ul
RT: 12.648 min Scan# 1600
Delta R.T. 0.000 min
Lab File: BM034207.D
Acq: 10 Mar 2022 19:21




Instrument :
BNA_M
ClientSampleId :
SSTD020048


Tgt Ion:142 Resp: 486708
Ion Ratio Lower Upper
142 100
141 93.1 74.5 111.7
116 3.8 3.0 4.6



#39
1,2,4,5-Tetrachlorobenzene
Concen: 19.337 ng/ul
RT: 12.795 min Scan# 1625
Delta R.T. 0.000 min
Lab File: BM034207.D
Acq: 10 Mar 2022 19:21

Tgt Ion:216 Resp: 309241
Ion Ratio Lower Upper
216 100
214 77.7 62.2 93.2
179 19.3 15.4 23.2
108 14.7 11.8 17.6









#40
Hexachlorocyclopentadiene
Concen: 19.468 ng/ul
RT: 12.783 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM034207.D
Acq: 10 Mar 2022 19:21




Instrument :
BNA_M
ClientSampleId :
SSTD020048


Tgt Ion:237 Resp: 210192
Ion Ratio Lower Upper
237 100
235 66.1 52.9 79.3
272 14.8 11.8 17.8



#41
2,4,6-Trichlorophenol
Concen: 19.130 ng/ul
RT: 13.030 min Scan# 1665
Delta R.T. 0.000 min
Lab File: BM034207.D
Acq: 10 Mar 2022 19:21

Tgt Ion:196 Resp: 186745
Ion Ratio Lower Upper
196 100
198 101.0 80.8 121.2
200 33.1 26.5 39.7









#42
2,4,5-Trichlorophenol
Concen: 18.602 ng/u1
RT: 13.095 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM034207.D
Acq: 10 Mar 2022 19:21




Instrument :
BNA_M
ClientSampleId :
SSTD020048


Tgt Ion:196 Resp: 200280
Ion Ratio Lower Upper
196 100
198 97.5 78.0 117.0
200 30.6 24.5 36.7



#43
1,1'-Biphenyl
Concen: 18.655 ng/u1
RT: 13.430 min Scan# 1733
Delta R.T. 0.000 min
Lab File: BM034207.D
Acq: 10 Mar 2022 19:21

Tgt Ion:154 Resp: 664945
Ion Ratio Lower Upper
154 100
153 40.8 32.6 49.0
76 9.6 7.7 11.5









#44
2-Chloronaphthalene
Concen: 18.648 ng/ul
RT: 13.471 min Scan# 1740
Delta R.T. 0.000 min
Lab File: BM034207.D
Acq: 10 Mar 2022 19:21




Instrument :
BNA_M
ClientSampleId :
SSTD020048


Tgt Ion:162 Resp: 522433
Ion Ratio Lower Upper
162 100
164 33.3 26.6 40.0
127 37.0 29.6 44.4



#45
2-Nitroaniline
Concen: 17.149 ng/ul
RT: 13.665 min Scan# 1773
Delta R.T. 0.000 min
Lab File: BM034207.D
Acq: 10 Mar 2022 19:21

Tgt Ion: 65 Resp: 110556
Ion Ratio Lower Upper
65 100
92 71.8 57.4 86.2
138 118.5 94.8 142.2









#47
Dimethylphthalate
Concen: 18.433 ng/ul
RT: 14.048 min Scan# 1838
Delta R.T. 0.000 min
Lab File: BM034207.D
Acq: 10 Mar 2022 19:21




Instrument :
BNA_M
ClientSampleId :
SSTD020048


Tgt Ion:163 Resp: 647710
Ion Ratio Lower Upper
163 100
194 4.3 3.4 5.2
164 9.8 7.8 11.8



#48
2,6-Dinitrotoluene
Concen: 18.722 ng/ul
RT: 14.160 min Scan# 1857
Delta R.T. 0.000 min
Lab File: BM034207.D
Acq: 10 Mar 2022 19:21

Tgt Ion:165 Resp: 130757
Ion Ratio Lower Upper
165 100
89 55.3 44.2 66.4
121 20.9 16.7 25.1









#50
Acenaphthylene
Concen: 18.719 ng/ul
RT: 14.318 min Scan# 1884
Delta R.T. 0.000 min
Lab File: BM034207.D
Acq: 10 Mar 2022 19:21




Instrument :
BNA_M
ClientSampleId :
SSTD020048


Tgt Ion:152 Resp: 819922
Ion Ratio Lower Upper
152 100
151 19.8 15.8 23.8
153 12.7 10.2 15.2



#51
3-Nitroaniline
Concen: 17.704 ng/ul
RT: 14.489 min Scan# 1913
Delta R.T. 0.000 min
Lab File: BM034207.D
Acq: 10 Mar 2022 19:21

Tgt Ion:138 Resp: 113089
Ion Ratio Lower Upper
138 100
108 10.4 8.3 12.5
92 108.7 87.0 130.4









#52
Acenaphthene
Concen: 18.438 ng/ul
RT: 14.660 min Scan# 1942
Delta R.T. 0.000 min
Lab File: BM034207.D
Acq: 10 Mar 2022 19:21




Instrument :
BNA_M
ClientSampleId :
SSTD020048


Tgt Ion:153 Resp: 534722
Ion Ratio Lower Upper
153 100
152 46.7 37.4 56.0
154 88.7 71.0 106.4



#53
2,4-Dinitrophenol
Concen: 20.673 ng/ul
RT: 14.689 min Scan# 1947
Delta R.T. 0.000 min
Lab File: BM034207.D
Acq: 10 Mar 2022 19:21

Tgt Ion:184 Resp: 91354
Ion Ratio Lower Upper
184 100
63 54.7 43.8 65.6
154 62.4 49.9 74.9









#55
4-Nitrophenol
Concen: 15.150 ng/ul
RT: 14.789 min Scan# 1964
Delta R.T. 0.000 min
Lab File: BM034207.D
Acq: 10 Mar 2022 19:21




Instrument :
BNA_M
ClientSampleId :
SSTD020048


Tgt Ion:109 Resp: 84933
Ion Ratio Lower Upper
109 100
139 118.3 94.6 142.0
65 104.4 83.5 125.3



#56
Dibenzofuran
Concen: 18.274 ng/ul
RT: 14.989 min Scan# 1998
Delta R.T. 0.000 min
Lab File: BM034207.D
Acq: 10 Mar 2022 19:21

Tgt Ion:168 Resp: 780564
Ion Ratio Lower Upper
168 100
139 37.2 29.8 44.6









#57
2,4-Dinitrotoluene
Concen: 17.861 ng/ul
RT: 14.948 min Scan# 1991
Delta R.T. 0.000 min
Lab File: BM034207.D
Acq: 10 Mar 2022 19:21




Instrument :
BNA_M
ClientSampleId :
SSTD020048


Tgt Ion:165 Resp: 182409
Ion Ratio Lower Upper
165 100
63 37.8 30.2 45.4
182 2.7 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 18.723 ng/ul
RT: 15.212 min Scan# 2036
Delta R.T. 0.000 min
Lab File: BM034207.D
Acq: 10 Mar 2022 19:21

Tgt Ion:232 Resp: 174725
Ion Ratio Lower Upper
232 100
131 40.3 32.2 48.4
130 2.4 1.9 2.9
166 28.7 23.0 34.4









#59
Diethylphthalate
Concen: 18.221 ng/ul
RT: 15.407 min Scan# 2069
Delta R.T. 0.000 min
Lab File: BM034207.D
Acq: 10 Mar 2022 19:21




Instrument :
BNA_M
ClientSampleId :
SSTD020048


Tgt Ion:149 Resp: 628473
Ion Ratio Lower Upper
149 100
177 22.4 17.9 26.9
150 12.7 10.2 15.2



#61
Fluorene
Concen: 18.211 ng/ul
RT: 15.636 min Scan# 2108
Delta R.T. 0.000 min
Lab File: BM034207.D
Acq: 10 Mar 2022 19:21

Tgt Ion:166 Resp: 639896
Ion Ratio Lower Upper
166 100
165 96.1 76.9 115.3
167 13.5 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 18.176 ng/ul
RT: 15.630 min Scan# 2107
Delta R.T. 0.000 min
Lab File: BM034207.D
Acq: 10 Mar 2022 19:21




Instrument :
BNA_M
ClientSampleId :
SSTD020048


Tgt Ion:204 Resp: 343850
Ion Ratio Lower Upper
204 100
206 31.9 25.5 38.3
141 55.2 44.2 66.2



#63
4-Nitroaniline
Concen: 16.457 ng/ul
RT: 15.648 min Scan# 2110
Delta R.T. 0.000 min
Lab File: BM034207.D
Acq: 10 Mar 2022 19:21

Tgt Ion:138 Resp: 97248
Ion Ratio Lower Upper
138 100
92 45.4 36.3 54.5
108 86.1 68.9 103.3









#66
 4,6-Dinitro-2-methylphenol
 Concen: 17.649 ng/ul
 RT: 15.707 min Scan# 2120
 Delta R.T. 0.000 min
 Lab File: BM034207.D
 Acq: 10 Mar 2022 19:21




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020048


Tgt Ion:198 Resp: 109302
 Ion Ratio Lower Upper
 198 100
 182 4.7 3.8 5.6
 77 23.1 18.5 27.7



#67
 N-Nitrosodiphenylamine
 Concen: 18.849 ng/ul
 RT: 15.842 min Scan# 2143
 Delta R.T. 0.000 min
 Lab File: BM034207.D
 Acq: 10 Mar 2022 19:21

Tgt Ion:169 Resp: 534118
 Ion Ratio Lower Upper
 169 100
 168 66.0 52.8 79.2
 167 35.1 28.1 42.1









#68
4-Bromophenyl-phenylether
Concen: 20.186 ng/ul
RT: 16.524 min Scan# 2259
Delta R.T. 0.000 min
Lab File: BM034207.D
Acq: 10 Mar 2022 19:21




Instrument :
BNA_M
ClientSampleId :
SSTD020048


Tgt Ion:248 Resp: 209735
Ion Ratio Lower Upper
248 100
250 98.0 78.4 117.6
141 64.3 51.4 77.2



#69
Hexachlorobenzene
Concen: 19.349 ng/ul
RT: 16.642 min Scan# 2279
Delta R.T. 0.000 min
Lab File: BM034207.D
Acq: 10 Mar 2022 19:21

Tgt Ion:284 Resp: 261552
Ion Ratio Lower Upper
284 100
142 30.7 24.6 36.8
249 30.1 24.1 36.1









#70
Atrazine
Concen: 18.307 ng/ul
RT: 16.789 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM034207.D
Acq: 10 Mar 2022 19:21




Instrument :
BNA_M
ClientSampleId :
SSTD020048


Tgt Ion:200 Resp: 210031
Ion Ratio Lower Upper
200 100
173 25.3 20.2 30.4
215 48.0 38.4 57.6



#71
Pentachlorophenol
Concen: 17.672 ng/ul
RT: 16.983 min Scan# 2337
Delta R.T. 0.000 min
Lab File: BM034207.D
Acq: 10 Mar 2022 19:21

Tgt Ion:266 Resp: 151597
Ion Ratio Lower Upper
266 100
264 61.6 49.3 73.9
268 62.9 50.3 75.5









#72
 Phenanthrene
 Concen: 18.498 ng/u1
 RT: 17.377 min Scan# 2404
 Delta R.T. 0.000 min
 Lab File: BM034207.D
 Acq: 10 Mar 2022 19:21




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020048


Tgt Ion:178 Resp: 994546
 Ion Ratio Lower Upper
 178 100
 179 14.6 11.7 17.5
 176 18.7 15.0 22.4



#74
 Anthracene
 Concen: 18.408 ng/u1
 RT: 17.465 min Scan# 2419
 Delta R.T. 0.000 min
 Lab File: BM034207.D
 Acq: 10 Mar 2022 19:21

Tgt Ion:178 Resp: 1000121
 Ion Ratio Lower Upper
 178 100
 179 14.9 11.9 17.9
 176 18.0 14.4 21.6





#75

1,2,3,4-Tetrachlorobenzene

Concen: 19.462 ng/uL

RT: 13.401 min Scan# 11

Delta R.T. 0.000 min

Lab File: BM034207.D

Acq: 10 Mar 2022 19:21

Instrument :

BNA_M

ClientSampleId :

SSTD020048





Tgt Ion:216 Resp: 316929

Ion Ratio Lower Upper

216 100

214 77.5 62.0 93.0

218 47.3 37.8 56.8



#76

Pentachlorobenzene

Concen: 19.806 ng/uL

RT: 14.912 min Scan# 1985

Delta R.T. 0.000 min

Lab File: BM034207.D

Acq: 10 Mar 2022 19:21




Tgt Ion:250 Resp: 302011


Ion Ratio Lower Upper

250 100

248 63.2 50.6 75.8

252 62.8 50.2 75.4









#77
 Carbazole
 Concen: 16.069 ng/ul
 RT: 17.730 min Scan# 2464
 Delta R.T. 0.000 min
 Lab File: BM034207.D
 Acq: 10 Mar 2022 19:21




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020048


Tgt Ion:167 Resp: 778504
 Ion Ratio Lower Upper
 167 100
 166 20.8 16.6 25.0
 139 12.4 9.9 14.9



#78
 Di-n-butylphthalate
 Concen: 18.997 ng/ul
 RT: 18.300 min Scan# 2561
 Delta R.T. 0.000 min
 Lab File: BM034207.D
 Acq: 10 Mar 2022 19:21

Tgt Ion:149 Resp: 1003190
 Ion Ratio Lower Upper
 149 100
 150 9.2 7.4 11.0
 104 5.2 4.2 6.2









#80
 Fluoranthene
 Concen: 18.429 ng/ul
 RT: 19.383 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: BM034207.D
 Acq: 10 Mar 2022 19:21




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020048


Tgt Ion:202 Resp: 1192183
 Ion Ratio Lower Upper
 202 100
 101 7.7 6.2 9.2
 100 5.9 4.7 7.1



#82
 Pyrene
 Concen: 18.278 ng/ul
 RT: 19.747 min Scan# 2807
 Delta R.T. 0.000 min
 Lab File: BM034207.D
 Acq: 10 Mar 2022 19:21

Tgt Ion:202 Resp: 1223090
 Ion Ratio Lower Upper
 202 100
 101 8.7 7.0 10.4
 100 7.2 5.8 8.6









#83
Butylbenzylphthalate
Concen: 18.645 ng/ul
RT: 20.641 min Scan# 2959
Delta R.T. 0.000 min
Lab File: BM034207.D
Acq: 10 Mar 2022 19:21




Instrument :
BNA_M
ClientSampleId :
SSTD020048


Tgt Ion:149 Resp: 426954
Ion Ratio Lower Upper
149 100
91 63.6 50.9 76.3
206 24.1 19.3 28.9



#84
3,3'-Dichlorobenzidine
Concen: 19.884 ng/ul
RT: 21.418 min Scan# 3091
Delta R.T. 0.000 min
Lab File: BM034207.D
Acq: 10 Mar 2022 19:21

Tgt Ion:252 Resp: 387748
Ion Ratio Lower Upper
252 100
254 63.4 50.7 76.1
126 8.2 6.6 9.8





#85

Benzo(a)anthracene

Concen: 17.572 ng/ul

RT: 21.488 min Scan# 3103

Delta R.T. 0.000 min

Lab File: BM034207.D

Acq: 10 Mar 2022 19:21

Instrument :

BNA_M

ClientSampleId :

SSTD020048





Tgt Ion:228 Resp: 1161486

Ion Ratio Lower Upper

228 100

229 19.3 15.4 23.2

226 26.2 21.0 31.4



#86

Bis(2-ethylhexyl)phthalate

Concen: 19.298 ng/ul

RT: 21.418 min Scan# 3091

Delta R.T. 0.000 min

Lab File: BM034207.D

Acq: 10 Mar 2022 19:21




Tgt Ion:149 Resp: 646827


Ion Ratio Lower Upper

149 100

167 27.6 22.1 33.1

279 5.2 4.2 6.2









#87
Chrysene
Concen: 18.914 ng/ul
RT: 21.541 min Scan# 3112
Delta R.T. 0.000 min
Lab File: BM034207.D
Acq: 10 Mar 2022 19:21




Instrument :
BNA_M
ClientSampleId :
SSTD020048


Tgt Ion:228 Resp: 1221217
Ion Ratio Lower Upper
228 100
226 29.3 23.4 35.2
229 19.2 15.4 23.0



#89
Di-n-octyl phthalate
Concen: 16.530 ng/ul
RT: 22.347 min Scan# 3249
Delta R.T. 0.000 min
Lab File: BM034207.D
Acq: 10 Mar 2022 19:21

Tgt Ion:149 Resp: 1103596









#90
Benzo(b)fluoranthene
Concen: 16.818 ng/ul
RT: 23.183 min Scan# 3391
Delta R.T. 0.000 min
Lab File: BM034207.D
Acq: 10 Mar 2022 19:21




Instrument :
BNA_M
ClientSampleId :
SSTD020048


Tgt Ion:252 Resp: 1250895
Ion Ratio Lower Upper
252 100
253 21.7 17.4 26.0
125 7.2 5.8 8.6



#91
Benzo(k)fluoranthene
Concen: 18.603 ng/ul
RT: 23.235 min Scan# 3400
Delta R.T. 0.000 min
Lab File: BM034207.D
Acq: 10 Mar 2022 19:21

Tgt Ion:252 Resp: 1300275
Ion Ratio Lower Upper
252 100
253 21.9 17.5 26.3
125 6.4 5.1 7.7









#93
Benzo(a)pyrene
Concen: 18.060 ng/ul
RT: 23.818 min Scan# 3499
Delta R.T. 0.000 min
Lab File: BM034207.D
Acq: 10 Mar 2022 19:21




Instrument :
BNA_M
ClientSampleId :
SSTD020048


Tgt Ion:252 Resp: 1261267
Ion Ratio Lower Upper
252 100
253 21.9 17.5 26.3
125 7.8 6.2 9.4



#94
Indeno(1,2,3-cd)pyrene
Concen: 19.325 ng/ul
RT: 26.447 min Scan# 3946
Delta R.T. 0.000 min
Lab File: BM034207.D
Acq: 10 Mar 2022 19:21

Tgt Ion:276 Resp: 1408909
Ion Ratio Lower Upper
276 100
138 17.0 13.6 20.4
277 24.9 19.9 29.9









#95
Dibenzo(a,h)anthracene
Concen: 19.053 ng/ul
RT: 26.465 min Scan# 3949
Delta R.T. 0.000 min
Lab File: BM034207.D
Acq: 10 Mar 2022 19:21

Instrument :
BNA_M
ClientSampleId :
SSTD020048

Tgt Ion:278 Resp: 1198528
Ion Ratio Lower Upper
278 100
139 12.0 9.6 14.4
279 24.1 19.3 28.9



#96
Benzo(g,h,i)perylene
Concen: 21.069 ng/ul
RT: 27.223 min Scan# 4078
Delta R.T. 0.000 min
Lab File: BM034207.D
Acq: 10 Mar 2022 19:21

Tgt Ion:276 Resp: 1283643
Ion Ratio Lower Upper
276 100
138 14.7 11.8 17.6
277 23.5 18.8 28.2

