

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM031022\
 Data File : BM034211.D
 Acq On : 11 Mar 2022 01:23
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SICV003

Quant Time: Mar 11 11:47:59 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM031022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Mar 11 02:00:05 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.954	152	170391	20.000	ng/ul	0.00
20) Naphthalene-d8	10.766	136	677830	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.595	164	435457	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.330	188	851619	20.000	ng/ul	0.00
79) Chrysene-d12	21.500	240	786986	20.000	ng/ul	0.00
88) Perylene-d12	23.918	264	803589	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.319	96	29533	7.566	ng/uL	0.00
4) Pyridine-d5	3.743	84	175117	18.843	ng/ul	0.00
7) Phenol-d5	7.101	99	213753	18.746	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.272	67	129893	19.418	ng/ul	0.00
11) 2-Chlorophenol-d4	7.478	132	199701	20.265	ng/ul	0.00
15) 4-Methylphenol-d8	8.660	113	183457	18.812	ng/ul	0.00
21) Nitrobenzene-d5	9.113	128	97456	20.044	ng/ul	0.00
24) 2-Nitrophenol-d4	9.842	143	107129	19.870	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.384	165	213093	19.716	ng/ul	0.00
31) 4-Chloroaniline-d4	10.895	131	246005	18.302	ng/ul	0.00
46) Dimethylphthalate-d6	14.001	166	593891	18.744	ng/ul	0.00
49) Acenaphthylene-d8	14.283	160	761198	18.915	ng/ul	0.00
54) 4-Nitrophenol-d4	14.771	143	84344	17.090	ng/ul	0.00
60) Fluorene-d10	15.583	176	513478	18.513	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.689	200	96020	17.821	ng/ul	0.00
73) Anthracene-d10	17.430	188	757198	18.825	ng/ul	0.00
81) Pyrene-d10	19.712	212	846964	19.768	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.765	264	778545	19.638	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.355	88	28968	7.490	ng/uL	99
5) Pyridine	3.766	79	183395	19.041	ng/ul	99
6) Benzaldehyde	7.078	77	145623	24.538	ng/ul	98
8) Phenol	7.131	94	210241	18.687	ng/ul	97
10) Bis(2-Chloroethyl)ether	7.366	93	183985	19.616	ng/ul	98
12) 2-Chlorophenol	7.513	128	199676	20.083	ng/ul	98
13) 2-Methylphenol	8.395	108	169181	18.791	ng/ul	99
14) 2,2'-oxybis(1-Chloropr...	8.484	45	166934	19.301	ng/ul	97
16) Acetophenone	8.778	105	291715	19.050	ng/ul	99
17) N-Nitroso-di-n-propyla...	8.760	70	143749	19.951	ng/ul	99
18) 4-Methylphenol	8.719	108	186895	19.255	ng/ul	97
19) Hexachloroethane	9.042	117	87674	19.552	ng/ul	97
22) Nitrobenzene	9.154	77	208118	19.721	ng/ul	98
23) Isophorone	9.684	82	388157	19.461	ng/ul	99
25) 2-Nitrophenol	9.872	139	113614	20.579	ng/ul	100
26) 2,4-Dimethylphenol	9.931	107	222091	19.470	ng/ul	99
27) Bis(2-Chloroethoxy)met...	10.172	93	238852	19.467	ng/ul	99
29) 2,4-Dichlorophenol	10.407	162	200592	19.231	ng/ul	99
30) Naphthalene	10.819	128	641434	18.949	ng/ul	98
32) 4-Chloroaniline	10.919	127	243675	18.227	ng/ul	99
33) Hexachlorobutadiene	11.113	225	161196	18.704	ng/ul	99
34) Caprolactam	11.672	113	52573	18.017	ng/ul	97
35) 4-Chloro-3-methylphenol	12.042	107	187579	19.386	ng/ul	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM031022\
 Data File : BM034211.D
 Acq On : 11 Mar 2022 01:23
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 14 Sample Multiplier: 1


Instrument :
 BNA_M
 ClientSampleId :
 SICV003

Quant Time: Mar 11 11:47:59 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM031022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Mar 11 02:00:05 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
36) 2-Methylnaphthalene	12.425	142	443613	19.082	ng/ul	100
37) 1-Methylnaphthalene	12.642	142	452332	19.050	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.795	216	281654	18.747	ng/ul	99
40) Hexachlorocyclopentadiene	12.783	237	191996	17.913	ng/ul	94
41) 2,4,6-Trichlorophenol	13.030	196	170499	19.368	ng/ul	98
42) 2,4,5-Trichlorophenol	13.095	196	184336	19.321	ng/ul	97
43) 1,1'-Biphenyl	13.430	154	609200	18.834	ng/ul	99
44) 2-Chloronaphthalene	13.472	162	480299	18.786	ng/ul	99
45) 2-Nitroaniline	13.666	65	100421	19.803	ng/ul	98
47) Dimethylphthalate	14.048	163	571265	18.468	ng/ul	100
48) 2,6-Dinitrotoluene	14.160	165	114085	18.915	ng/ul	99
50) Acenaphthylene	14.313	152	742409	18.832	ng/ul	99
51) 3-Nitroaniline	14.483	138	98362	18.730	ng/ul	99
52) Acenaphthene	14.654	153	484639	18.743	ng/ul	99
53) 2,4-Dinitrophenol	14.689	184	80172	20.444	ng/ul	98
55) 4-Nitrophenol	14.783	109	71320	17.364	ng/ul	94
56) Dibenzofuran	14.989	168	696225	18.591	ng/ul	100
57) 2,4-Dinitrotoluene	14.942	165	162661	19.599	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	15.213	232	154324	18.919	ng/ul	100
59) Diethylphthalate	15.407	149	553490	18.728	ng/ul	100
61) Fluorene	15.636	166	564719	18.543	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.630	204	304442	18.438	ng/ul	99
63) 4-Nitroaniline	15.642	138	83797	18.849	ng/ul	96
66) 4,6-Dinitro-2-methylph...	15.707	198	94033	17.854	ng/ul#	96
67) N-Nitrosodiphenylamine	15.842	169	473760	19.235	ng/ul	100
68) 4-Bromophenyl-phenylether	16.524	248	182847	19.057	ng/ul	100
69) Hexachlorobenzene	16.642	284	224522	18.760	ng/ul	98
70) Atrazine	16.789	200	178877	18.432	ng/ul	99
71) Pentachlorophenol	16.983	266	126516	17.574	ng/ul	98
72) Phenanthrene	17.371	178	851555	18.781	ng/ul	99
74) Anthracene	17.465	178	859359	19.097	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.395	216	286011	18.994	ng/uL	98
76) Pentachlorobenzene	14.913	250	269501	19.026	ng/uL	99
77) Carbazole	17.730	167	650311	17.783	ng/ul	99
78) Di-n-butylphthalate	18.301	149	865432	19.793	ng/ul	99
80) Fluoranthene	19.383	202	961285	19.719	ng/ul	99
82) Pyrene	19.742	202	996933	19.706	ng/ul	99
83) Butylbenzylphthalate	20.636	149	337610	19.936	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.418	252	313408	20.088	ng/ul	99
85) Benzo(a)anthracene	21.483	228	898583	19.537	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.418	149	499264	19.823	ng/ul	99
87) Chrysene	21.536	228	931477	18.812	ng/ul	99
89) Di-n-octyl phthalate	22.347	149	806423	18.728	ng/ul	100
90) Benzo(b)fluoranthene	23.183	252	929905	19.738	ng/ul	100
91) Benzo(k)fluoranthene	23.230	252	945032	19.680	ng/ul	99
93) Benzo(a)pyrene	23.812	252	905326	19.266	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.441	276	1017298	19.800	ng/ul	98
95) Dibenzo(a,h)anthracene	26.459	278	857393	19.918	ng/ul	99
96) Benzo(g,h,i)perylene	27.218	276	939747	19.669	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed





Instrument :
BNA_M
ClientSampleId :
SICV003



1,4-Dioxane
 Concen: 7.490 ng/uL
 RT: 3.355 min Scan# 1
 Delta R.T. -0.005 min
 Lab File: BM034211.D
 Acq: 11 Mar 2022 01:23




Instrument :
 BNA_M
 ClientSampleId :
 SICV003


Tgt Ion: 88 Resp: 28968
 Ion Ratio Lower Upper
 88 100
 43 26.1 21.0 31.6
 58 60.6 47.8 71.8



Pyridine
 Concen: 19.041 ng/uL
 RT: 3.766 min Scan# 90
 Delta R.T. 0.000 min
 Lab File: BM034211.D
 Acq: 11 Mar 2022 01:23

Tgt Ion: 79 Resp: 183395
 Ion Ratio Lower Upper
 79 100
 52 58.7 47.7 71.5
 51 31.5 25.6 38.4









Benzaldehyde
 Concen: 24.538 ng/ul
 RT: 7.078 min Scan# 653
 Delta R.T. -0.006 min
 Lab File: BM034211.D
 Acq: 11 Mar 2022 01:23




Instrument :
 BNA_M
 ClientSampleId :
 SICV003


Tgt Ion: 77 Resp: 145623
 Ion Ratio Lower Upper
 77 100
 105 98.2 80.8 121.2
 106 94.2 76.6 115.0



Phenol
 Concen: 18.687 ng/ul
 RT: 7.131 min Scan# 662
 Delta R.T. 0.000 min
 Lab File: BM034211.D
 Acq: 11 Mar 2022 01:23

Tgt Ion: 94 Resp: 210241
 Ion Ratio Lower Upper
 94 100
 65 29.8 22.9 34.3
 66 39.9 33.7 50.5









Bis(2-Chloroethyl)ether
 Concen: 19.616 ng/ul
 RT: 7.366 min Scan#
 Delta R.T. 0.000 min
 Lab File: BM034211.D
 Acq: 11 Mar 2022 01:23




Instrument :
 BNA_M
 ClientSampleId :
 SICV003


Tgt Ion:	93	Resp:	183985
Ion	Ratio	Lower	Upper
93	100		
63	67.5	54.7	82.1
95	30.5	26.6	39.8



2-Chlorophenol
 Concen: 20.083 ng/ul
 RT: 7.513 min Scan# 727
 Delta R.T. 0.000 min
 Lab File: BM034211.D
 Acq: 11 Mar 2022 01:23

Tgt Ion:	128	Resp:	199676
Ion	Ratio	Lower	Upper
128	100		
64	36.9	30.6	46.0
130	32.9	25.8	38.6









#13
2-Methylphenol
Concen: 18.791 ng/ul
RT: 8.395 min Scan# 877
Delta R.T. 0.000 min
Lab File: BM034211.D
Acq: 11 Mar 2022 01:23




Instrument :
BNA_M
ClientSampleId :
SICV003


Tgt Ion:108 Resp: 169181
Ion Ratio Lower Upper
108 100
107 89.3 71.9 107.9



#14
2,2'-oxybis(1-Chloropropane)
Concen: 19.301 ng/ul
RT: 8.484 min Scan# 892
Delta R.T. -0.006 min
Lab File: BM034211.D
Acq: 11 Mar 2022 01:23

Tgt Ion: 45 Resp: 166934
Ion Ratio Lower Upper
45 100
77 23.1 17.1 25.7
79 15.5 12.8 19.2





Acetophenone

Concen: 19.050 ng/ul

RT: 8.778 min Scan# 939

Delta R.T. 0.000 min

Lab File: BM034211.D

Acq: 11 Mar 2022 01:23

Instrument :

BNA_M

ClientSampleId :

SICV003





Tgt Ion:105 Resp: 291715

Ion Ratio Lower Upper

105 100

77 73.8 59.4 89.2

51 22.7 19.3 28.9



N-Nitroso-di-n-propylamine

Concen: 19.951 ng/ul

RT: 8.760 min Scan# 939

Delta R.T. -0.006 min

Lab File: BM034211.D

Acq: 11 Mar 2022 01:23

Tgt Ion: 70 Resp: 143749




Ion Ratio Lower Upper


70 100

42 52.0 40.6 60.8

101 10.5 8.2 12.2

130 24.7 19.9 29.9









#18
4-Methylphenol
Concen: 19.255 ng/ul
RT: 8.719 min Scan# 933
Delta R.T. -0.006 min
Lab File: BM034211.D
Acq: 11 Mar 2022 01:23




Instrument :
BNA_M
ClientSampleId :
SICV003


Tgt Ion:108 Resp: 186895
Ion Ratio Lower Upper
108 100
107 114.5 89.0 133.4



#19
Hexachloroethane
Concen: 19.552 ng/ul
RT: 9.042 min Scan# 987
Delta R.T. 0.000 min
Lab File: BM034211.D
Acq: 11 Mar 2022 01:23

Tgt Ion:117 Resp: 87674
Ion Ratio Lower Upper
117 100
201 117.4 89.9 134.9
199 70.8 57.2 85.8





Nitrobenzene

Concen: 19.721 ng/ul

RT: 9.154 min Scan# 1006

Delta R.T. 0.000 min

Lab File: BM034211.D

Acq: 11 Mar 2022 01:23

Instrument :

BNA_M

ClientSampleId :

SICV003





Tgt Ion: 77 Resp: 208118

Ion Ratio Lower Upper

77 100

123 46.0 38.2 57.2

65 13.2 10.6 15.8



Isophorone

Concen: 19.461 ng/ul

RT: 9.684 min Scan# 1096

Delta R.T. -0.006 min

Lab File: BM034211.D

Acq: 11 Mar 2022 01:23




Tgt Ion: 82 Resp: 388157

Ion Ratio Lower Upper


82 100

95 8.1 6.8 10.2

138 21.3 16.5 24.7






Abundance Scan 1128 (9.872 min): BM034207.D\data.ms (-1125)




2-Nitrophenol
Concen: 20.579 ng/ul
RT: 9.872 min Scan# 1125
Delta R.T. 0.000 min
Lab File: BM034211.D
Acq: 11 Mar 2022 01:23

Instrument :
BNA_M
ClientSampleId :
SICV003

Tgt Ion:	139	Resp:	113614
Ion	Ratio	Lower	Upper
139	100		
65	46.7	37.8	56.6
109	33.2	26.5	39.7







Abundance Scan 1139 (9.936 min): BM034207.D\data.ms (-1126)



2,4-Dimethylphenol
Concen: 19.470 ng/ul
RT: 9.931 min Scan# 1138
Delta R.T. -0.006 min
Lab File: BM034211.D
Acq: 11 Mar 2022 01:23

Tgt Ion:	107	Resp:	222091
Ion	Ratio	Lower	Upper
107	100		
121	51.3	40.8	61.2
122	85.5	69.0	103.6









Bis(2-Chloroethoxy)methane
 Concen: 19.467 ng/ul
 RT: 10.172 min Scan# 1179
 Delta R.T. 0.000 min
 Lab File: BM034211.D
 Acq: 11 Mar 2022 01:23




Instrument :
 BNA_M
 ClientSampleId :
 SICV003


Tgt Ion:	93	Resp:	238852
Ion	Ratio	Lower	Upper
93	100		
95	32.8	25.4	38.2
123	11.7	9.5	14.3



2,4-Dichlorophenol
 Concen: 19.231 ng/ul
 RT: 10.407 min Scan# 1219
 Delta R.T. 0.000 min
 Lab File: BM034211.D
 Acq: 11 Mar 2022 01:23

Tgt Ion:	162	Resp:	200592
Ion	Ratio	Lower	Upper
162	100		
164	65.2	52.8	79.2
98	31.6	26.0	39.0









Scan 1289 (10.819 min): BM034207.D\data.ms (-1289) (-)
 Naphthalene
 Concen: 18.949 ng/ul
 RT: 10.819 min Scan# 1289
 Delta R.T. 0.000 min
 Lab File: BM034211.D
 Acq: 11 Mar 2022 01:23




Instrument :
 BNA_M
 ClientSampleId :
 SICV003


Tgt Ion:	128	Resp:	641434
Ion	Ratio	Lower	Upper
128	100		
129	11.0	8.7	13.1
127	13.5	9.9	14.9



Scan 1306 (10.919 min): BM034207.D\data.ms (-1306) (-)
 4-Chloroaniline
 Concen: 18.227 ng/ul
 RT: 10.919 min Scan# 1306
 Delta R.T. 0.000 min
 Lab File: BM034211.D
 Acq: 11 Mar 2022 01:23

Tgt Ion:	127	Resp:	243675
Ion	Ratio	Lower	Upper
127	100		
129	32.0	25.1	37.7









Hexachlorobutadiene
 Concen: 18.704 ng/ul
 RT: 11.113 min Scan# 1339
 Delta R.T. 0.000 min
 Lab File: BM034211.D
 Acq: 11 Mar 2022 01:23




Instrument :
 BNA_M
 ClientSampleId :
 SICV003


Tgt Ion:	225	Resp:	161196
Ion	Ratio	Lower	Upper
225	100		
223	63.7	50.6	75.8
227	64.3	50.3	75.5



Caprolactam
 Concen: 18.017 ng/ul
 RT: 11.672 min Scan# 1434
 Delta R.T. -0.006 min
 Lab File: BM034211.D
 Acq: 11 Mar 2022 01:23

Tgt Ion:	113	Resp:	52573
Ion	Ratio	Lower	Upper
113	100		
55	127.2	100.9	151.3
56	96.2	81.0	121.4









4-Chloro-3-methylphenol
Concen: 19.386 ng/ul
RT: 12.042 min Scan# 1497
Delta R.T. 0.000 min
Lab File: BM034211.D
Acq: 11 Mar 2022 01:23




Instrument :
BNA_M
ClientSampleId :
SICV003

Tgt Ion:	107	Resp:	187579
Ion	Ratio	Lower	Upper
107	100		
144	27.8	23.0	34.4
142	84.1	69.8	104.8




2-Methylnaphthalene
Concen: 19.082 ng/ul
RT: 12.425 min Scan# 1562
Delta R.T. 0.000 min
Lab File: BM034211.D
Acq: 11 Mar 2022 01:23

Tgt Ion:	142	Resp:	443613
Ion	Ratio	Lower	Upper
142	100		
141	89.6	71.4	107.0






Abundance Scan 1600 (12.648 min): BM034207.D\data.ms (-1533) (-)




1-Methylnaphthalene
Concen: 19.050 ng/ul
RT: 12.642 min Scan# 1599
Delta R.T. -0.006 min
Lab File: BM034211.D
Acq: 11 Mar 2022 01:23

Instrument :
BNA_M
ClientSampleId :
SICV003

Tgt Ion:	142	Resp:	452332
Ion	Ratio	Lower	Upper
142	100		
141	90.7	74.5	111.7
116	4.0	3.0	4.6






Abundance Scan 1625 (12.795 min): BM034207.D\data.ms (-1533) (-)



1,2,4,5-Tetrachlorobenzene
Concen: 18.747 ng/ul
RT: 12.795 min Scan# 1625
Delta R.T. 0.000 min
Lab File: BM034211.D
Acq: 11 Mar 2022 01:23


Tgt Ion:	216	Resp:	281654
Ion	Ratio	Lower	Upper
216	100		
214	77.2	62.2	93.2
179	20.5	15.4	23.2
108	14.6	11.8	17.6





Abundance Scan 1623 (12.783 min): BM034207.D\data.ms (-1540) (-)

Hexachlorocyclopentadiene
Concen: 17.913 ng/ul
RT: 12.783 min Scan# 1623
Delta R.T. 0.000 min
Lab File: BM034211.D
Acq: 11 Mar 2022 01:23


Instrument :
BNA_M
ClientSampleId :
SICV003



Tgt Ion	Ratio	Lower	Upper
237	100		
235	61.0	52.9	79.3
272	13.7	11.8	17.8





Abundance Scan 1623 (12.783 min): BM034211.D\data.ms (-1589) (-)




Abundance Scan 1665 (13.030 min): BM034207.D\data.ms (-1644) (-)

2,4,6-Trichlorophenol
Concen: 19.368 ng/ul
RT: 13.030 min Scan# 1665
Delta R.T. 0.000 min
Lab File: BM034211.D
Acq: 11 Mar 2022 01:23


Tgt Ion	Ratio	Lower	Upper
196	100		
198	97.9	80.8	121.2
200	32.7	26.5	39.7



Abundance Scan 1665 (13.030 min): BM034211.D\data.ms (-1631) (-)





Abundance Scan 1676 (13.095 min): BM034207.D\data.ms (-1642) (-)




2,4,5-Trichlorophenol
Concen: 19.321 ng/ul
RT: 13.095 min Scan# 1676
Delta R.T. 0.000 min
Lab File: BM034211.D
Acq: 11 Mar 2022 01:23

Instrument :
BNA_M
ClientSampleId :
SICV003


Tgt Ion:	196	Resp:	184336
Ion	Ratio	Lower	Upper
196	100		
198	94.6	78.0	117.0
200	30.0	24.5	36.7



Abundance Scan 1676 (13.095 min): BM034211.D\data.ms (-1659) (-)





Abundance Scan 1733 (13.430 min): BM034207.D\data.ms (-1743) (-)





1,1'-Biphenyl
Concen: 18.834 ng/ul
RT: 13.430 min Scan# 1733
Delta R.T. 0.000 min
Lab File: BM034211.D
Acq: 11 Mar 2022 01:23

Tgt Ion:	154	Resp:	609200
Ion	Ratio	Lower	Upper
154	100		
153	40.2	32.6	49.0
76	9.9	7.7	11.5



Abundance Scan 1733 (13.430 min): BM034211.D\data.ms (-1699) (-)









2-Chloronaphthalene
Concen: 18.786 ng/ul
RT: 13.472 min Scan# 1740
Delta R.T. 0.000 min
Lab File: BM034211.D
Acq: 11 Mar 2022 01:23




Instrument :
BNA_M
ClientSampleId :
SICV003


Tgt Ion:	162	Resp:	480299
Ion	Ratio	Lower	Upper
162	100		
164	33.1	26.6	40.0
127	35.6	29.6	44.4



2-Nitroaniline
Concen: 19.803 ng/ul
RT: 13.666 min Scan# 1773
Delta R.T. 0.000 min
Lab File: BM034211.D
Acq: 11 Mar 2022 01:23

Tgt Ion:	65	Resp:	100421
Ion	Ratio	Lower	Upper
65	100		
92	73.0	57.4	86.2
138	120.6	94.8	142.2









Dimethylphthalate
Concen: 18.468 ng/ul
RT: 14.048 min Scan# 1838
Delta R.T. 0.000 min
Lab File: BM034211.D
Acq: 11 Mar 2022 01:23




Instrument :
BNA_M
ClientSampleId :
SICV003


Tgt Ion:	163	Resp:	571265
Ion	Ratio	Lower	Upper
163	100		
194	4.5	3.4	5.2
164	9.8	7.8	11.8



2,6-Dinitrotoluene
Concen: 18.915 ng/ul
RT: 14.160 min Scan# 1857
Delta R.T. 0.000 min
Lab File: BM034211.D
Acq: 11 Mar 2022 01:23

Tgt Ion:	165	Resp:	114085
Ion	Ratio	Lower	Upper
165	100		
89	55.9	44.2	66.4
121	21.1	16.7	25.1









Acenaphthylene
 Concen: 18.832 ng/ul
 RT: 14.313 min Scan# 1883
 Delta R.T. -0.006 min
 Lab File: BM034211.D
 Acq: 11 Mar 2022 01:23




Instrument :
 BNA_M
 ClientSampleId :
 SICV003

Tgt Ion:	152	Resp:	742409
Ion	Ratio	Lower	Upper
152	100		
151	19.6	15.8	23.8
153	13.0	10.2	15.2



3-Nitroaniline
 Concen: 18.730 ng/ul
 RT: 14.483 min Scan# 1912
 Delta R.T. -0.006 min
 Lab File: BM034211.D
 Acq: 11 Mar 2022 01:23

Tgt Ion:	138	Resp:	98362
Ion	Ratio	Lower	Upper
138	100		
108	11.2	8.3	12.5
92	107.9	87.0	130.4



Abundance Scan 1942 (14.660 min): BM034207.D\data.ms (-1552) (-)

Acenaphthene

Concen: 18.743 ng/ul

RT: 14.654 min Scan#

Delta R.T. -0.006 min

Lab File: BM034211.D

Acq: 11 Mar 2022 01:23

Instrument :

BNA_M

ClientSampleId :

SICV003

Tgt Ion:153 Resp: 484639

Ion Ratio Lower Upper

153 100

152 47.3 37.4 56.0

154 90.1 71.0 106.4

Abundance Scan 1941 (14.654 min): BM034211.D\data.ms

Ref 50

76.0 111.0 153.1 192.9

m/z-->

Abundance Scan 1941 (14.654 min): BM034211.D\data.ms

Raw 50

76.1 111.0 153.1 208.0 280.9

m/z-->

Abundance Scan 1941 (14.654 min): BM034211.D\data.ms (-1908) (-)

Sub 50

76.1 111.0 153.1 208.1

m/z-->

Abundance

14.654

300000

200000

100000

0

Time-->

Abundance Scan 1947 (14.689 min): BM034207.D\data.ms (-1552) (-)

2,4-Dinitrophenol

Concen: 20.444 ng/ul

RT: 14.689 min Scan# 1947

Delta R.T. 0.000 min

Lab File: BM034211.D

Acq: 11 Mar 2022 01:23

Tgt Ion:184 Resp: 80172

Ion Ratio Lower Upper

184 100

63 54.2 43.8 65.6

154 60.1 49.9 74.9

Abundance Scan 1947 (14.689 min): BM034211.D\data.ms

Ref 50

63.0 119.9 184.0 281.0

m/z-->

Abundance Scan 1947 (14.689 min): BM034211.D\data.ms

Raw 50

63.0 119.9 184.0 281.0

m/z-->

Abundance Scan 1947 (14.689 min): BM034211.D\data.ms (-1913) (-)

Sub 50

53.0 107.0 184.0 281.0

m/z-->

Abundance


14.689

200000

100000

0





Time-->



4-Nitrophenol
 Concen: 17.364 ng/ul
 RT: 14.783 min Scan# 1964
 Delta R.T. -0.006 min
 Lab File: BM034211.D
 Acq: 11 Mar 2022 01:23




Instrument :
 BNA_M
 ClientSampleId :
 SICV003


Tgt Ion:109 Resp: 71320
 Ion Ratio Lower Upper
 109 100
 139 120.4 94.6 142.0
 65 114.7 83.5 125.3



Dibenzofuran
 Concen: 18.591 ng/ul
 RT: 14.989 min Scan# 1998
 Delta R.T. 0.000 min
 Lab File: BM034211.D
 Acq: 11 Mar 2022 01:23

Tgt Ion:168 Resp: 696225
 Ion Ratio Lower Upper
 168 100
 139 37.4 29.8 44.6









Scan 1991 (14.948 min): BM034207.D\data.ms (-1957) (-)
 2,4-Dinitrotoluene
 Concen: 19.599 ng/ul
 RT: 14.942 min Scan# 1991
 Delta R.T. -0.006 min
 Lab File: BM034211.D
 Acq: 11 Mar 2022 01:23




Instrument :
 BNA_M
 ClientSampleId :
 SICV003


Tgt Ion:	165	Resp:	162661
Ion	Ratio	Lower	Upper
165	100		
63	37.8	30.2	45.4
182	3.1	2.2	3.2



Scan 2036 (15.212 min): BM034207.D\data.ms (-2056) (-)
 2,3,4,6-Tetrachlorophenol
 Concen: 18.919 ng/ul
 RT: 15.213 min Scan# 2036
 Delta R.T. 0.000 min
 Lab File: BM034211.D
 Acq: 11 Mar 2022 01:23

Tgt Ion:	232	Resp:	154324
Ion	Ratio	Lower	Upper
232	100		
131	40.5	32.2	48.4
130	2.2	1.9	2.9
166	28.3	23.0	34.4









Scan 2069 (15.407 min): BM034207.D\data.ms (-2059) (-)
Diethylphthalate
 Concen: 18.728 ng/ul
 RT: 15.407 min Scan# 2069
 Delta R.T. 0.000 min
 Lab File: BM034211.D
 Acq: 11 Mar 2022 01:23




Instrument :
 BNA_M
ClientSampleId :
 SICV003

Tgt Ion:	149	Resp:	553490
Ion	Ratio	Lower	Upper
149	100		
177	22.4	17.9	26.9
150	12.6	10.2	15.2




Scan 2108 (15.636 min): BM034207.D\data.ms (-2061) (-)
Fluorene
 Concen: 18.543 ng/ul
 RT: 15.636 min Scan# 2108
 Delta R.T. 0.000 min
 Lab File: BM034211.D
 Acq: 11 Mar 2022 01:23

Tgt Ion:	166	Resp:	564719
Ion	Ratio	Lower	Upper
166	100		
165	96.6	76.9	115.3
167	13.6	10.8	16.2



Abundance Scan 2107 (15.630 min): BM034207.D\data.ms (-2107) (-)





4-Chlorophenyl-phenylether
Concen: 18.438 ng/ul
RT: 15.630 min Scan# 2107
Delta R.T. 0.000 min
Lab File: BM034211.D
Acq: 11 Mar 2022 01:23


Instrument :
BNA_M
ClientSampleId :
SICV003

Tgt Ion:	204	Resp:	304442
Ion	Ratio	Lower	Upper
204	100		
206	33.3	25.5	38.3
141	55.8	44.2	66.2


Abundance Scan 2107 (15.630 min): BM034211.D\data.ms



Abundance Scan 2107 (15.630 min): BM034211.D\data.ms (-2073) (-)





Abundance Scan 2110 (15.648 min): BM034207.D\data.ms (-2109) (-)




4-Nitroaniline
Concen: 18.849 ng/ul
RT: 15.642 min Scan# 2109
Delta R.T. -0.006 min
Lab File: BM034211.D
Acq: 11 Mar 2022 01:23

Tgt Ion:	138	Resp:	83797
Ion	Ratio	Lower	Upper
138	100		
92	46.8	36.3	54.5
108	82.2	68.9	103.3

Abundance Scan 2109 (15.642 min): BM034211.D\data.ms



Abundance Scan 2109 (15.642 min): BM034211.D\data.ms (-2076) (-)



Abundance Scan 2120 (15.707 min): BM034207.D\data.ms (-2150) (-)

4,6-Dinitro-2-methylphenol

Concen: 17.854 ng/ul

RT: 15.707 min Scan# 2120

Delta R.T. 0.000 min

Lab File: BM034211.D


Acq: 11 Mar 2022 01:23

Instrument :

BNA_M

ClientSampleId :

SICV003






Tgt Ion:198 Resp: 94033

Ion Ratio Lower Upper

198 100

182 3.5 3.8 5.6#

77 20.9 18.5 27.7



Abundance Scan 2143 (15.842 min): BM034207.D\data.ms (-2157) (-)

N-Nitrosodiphenylamine

Concen: 19.235 ng/ul

RT: 15.842 min Scan# 2143

Delta R.T. 0.000 min

Lab File: BM034211.D

Acq: 11 Mar 2022 01:23




Tgt Ion:169 Resp: 473760

Ion Ratio Lower Upper

169 100

168 66.1 52.8 79.2


167 35.6 28.1 42.1





Abundance Scan 2259 (16.524 min): BM034207.D\data.ms (-2259) (-)

4-Bromophenyl-phenylether
Concen: 19.057 ng/ul
RT: 16.524 min Scan# 2259
Delta R.T. 0.000 min
Lab File: BM034211.D
Acq: 11 Mar 2022 01:23


Instrument :
BNA_M
ClientSampleId :
SICV003



Tgt Ion:	248	Resp:	182847
Ion	Ratio	Lower	Upper
248	100		
250	98.3	78.4	117.6
141	63.9	51.4	77.2







Abundance Scan 2259 (16.524 min): BM034211.D\data.ms (-2225) (-)




Abundance Scan 2279 (16.642 min): BM034207.D\data.ms (-2279) (-)

Hexachlorobenzene
Concen: 18.760 ng/ul
RT: 16.642 min Scan# 2279
Delta R.T. 0.000 min
Lab File: BM034211.D
Acq: 11 Mar 2022 01:23

Tgt Ion:	284	Resp:	224522
Ion	Ratio	Lower	Upper
284	100		
142	29.2	24.6	36.8
249	30.9	24.1	36.1









Atrazine
Concen: 18.432 ng/ul
RT: 16.789 min Scan# 2304
Delta R.T. 0.000 min
Lab File: BM034211.D
Acq: 11 Mar 2022 01:23




Instrument :
BNA_M
ClientSampleId :
SICV003

Tgt Ion:	200	Resp:	178877
Ion	Ratio	Lower	Upper
200	100		
173	24.8	20.2	30.4
215	49.0	38.4	57.6




Pentachlorophenol
Concen: 17.574 ng/ul
RT: 16.983 min Scan# 2337
Delta R.T. 0.000 min
Lab File: BM034211.D
Acq: 11 Mar 2022 01:23

Tgt Ion:	266	Resp:	126516
Ion	Ratio	Lower	Upper
266	100		
264	63.1	49.3	73.9
268	64.8	50.3	75.5



Abundance Scan 2404 (17.377 min): BM034207.D\data.ms (-2372) (-)





Phenanthrene
Concen: 18.781 ng/ul
RT: 17.371 min Scan# 2404
Delta R.T. -0.006 min
Lab File: BM034211.D
Acq: 11 Mar 2022 01:23


Instrument :
BNA_M
ClientSampleId :
SICV003

Tgt Ion:	178	Resp:	851555
Ion	Ratio	Lower	Upper
178	100		
179	15.4	11.7	17.5
176	18.6	15.0	22.4


Abundance Scan 2403 (17.371 min): BM034211.D\data.ms



Abundance Scan 2403 (17.371 min): BM034211.D\data.ms (-2376) (-)





Abundance Scan 2419 (17.465 min): BM034207.D\data.ms (-2374) (-)




Anthracene
Concen: 19.097 ng/ul
RT: 17.465 min Scan# 2419
Delta R.T. 0.000 min
Lab File: BM034211.D
Acq: 11 Mar 2022 01:23


Tgt Ion:	178	Resp:	859359
Ion	Ratio	Lower	Upper
178	100		
179	15.2	11.9	17.9
176	17.8	14.4	21.6

Abundance Scan 2419 (17.465 min): BM034211.D\data.ms



Abundance Scan 2419 (17.465 min): BM034211.D\data.ms (-2385) (-)









1,2,3,4-Tetrachlorobenzene
Concen: 18.994 ng/uL
RT: 13.395 min Scan# 1725
Delta R.T. -0.006 min
Lab File: BM034211.D
Acq: 11 Mar 2022 01:23




Instrument :
BNA_M
ClientSampleId :
SICV003

Tgt Ion:	216	Resp:	286011
Ion	Ratio	Lower	Upper
216	100		
214	79.4	62.0	93.0
218	49.3	37.8	56.8




Pentachlorobenzene
Concen: 19.026 ng/uL
RT: 14.913 min Scan# 1985
Delta R.T. 0.000 min
Lab File: BM034211.D
Acq: 11 Mar 2022 01:23

Tgt Ion:	250	Resp:	269501
Ion	Ratio	Lower	Upper
250	100		
248	63.6	50.6	75.8
252	63.3	50.2	75.4






Abundance Scan 2464 (17.730 min): BM034207.D\data.ms (-2438) (-)




Carbazole
Concen: 17.783 ng/ul
RT: 17.730 min Scan# 2464
Delta R.T. 0.000 min
Lab File: BM034211.D
Acq: 11 Mar 2022 01:23

Instrument :
BNA_M
ClientSampleId :
SICV003

Tgt Ion:	167	Resp:	650311
Ion	Ratio	Lower	Upper
167	100		
166	21.2	16.6	25.0
139	11.9	9.9	14.9






Abundance Scan 2561 (18.301 min): BM034207.D\data.ms (-2556) (-)




Di-n-butylphthalate
Concen: 19.793 ng/ul
RT: 18.301 min Scan# 2561
Delta R.T. 0.000 min
Lab File: BM034211.D
Acq: 11 Mar 2022 01:23

Tgt Ion:	149	Resp:	865432
Ion	Ratio	Lower	Upper
149	100		
150	8.9	7.4	11.0
104	5.4	4.2	6.2



Abundance Scan 2745 (19.383 min): BM034207.D\data.ms (-2739) (-)



Fluoranthene

Concen: 19.719 ng/ul

RT: 19.383 min Scan# 2745

Delta R.T. 0.000 min

Lab File: BM034211.D

Acq: 11 Mar 2022 01:23

Instrument :

BNA_M

ClientSampleId :

SICV003




Tgt Ion:202 Resp: 961285

Ion Ratio Lower Upper


202 100

101 7.3 6.2 9.2

100 6.1 4.7 7.1



Abundance Scan 2807 (19.747 min): BM034207.D\data.ms (-2780) (-)



Pyrene

Concen: 19.706 ng/ul

RT: 19.742 min Scan# 2806

Delta R.T. -0.006 min

Lab File: BM034211.D

Acq: 11 Mar 2022 01:23




Tgt Ion:202 Resp: 996933


Ion Ratio Lower Upper

202 100

101 9.3 7.0 10.4

100 7.6 5.8 8.6









Butylbenzylphthalate
 Concen: 19.936 ng/ul
 RT: 20.636 min Scan# 2959
 Delta R.T. -0.006 min
 Lab File: BM034211.D
 Acq: 11 Mar 2022 01:23




Instrument :
 BNA_M
 ClientSampleId :
 SICV003


Tgt Ion:	149	Resp:	337610
Ion	Ratio	Lower	Upper
149	100		
91	64.8	50.9	76.3
206	22.3	19.3	28.9



3,3'-Dichlorobenzidine
 Concen: 20.088 ng/ul
 RT: 21.418 min Scan# 3091
 Delta R.T. 0.000 min
 Lab File: BM034211.D
 Acq: 11 Mar 2022 01:23

Tgt Ion:	252	Resp:	313408
Ion	Ratio	Lower	Upper
252	100		
254	64.4	50.7	76.1
126	7.9	6.6	9.8









Benzo(a)anthracene
 Concen: 19.537 ng/ul
 RT: 21.483 min Scan# 3103
 Delta R.T. -0.006 min
 Lab File: BM034211.D
 Acq: 11 Mar 2022 01:23




Instrument :
 BNA_M
 ClientSampleId :
 SICV003


Tgt Ion:	228	Resp:	898583
Ion	Ratio	Lower	Upper
228	100		
229	19.4	15.4	23.2
226	26.3	21.0	31.4



Bis(2-ethylhexyl)phthalate
 Concen: 19.823 ng/ul
 RT: 21.418 min Scan# 3091
 Delta R.T. 0.000 min
 Lab File: BM034211.D
 Acq: 11 Mar 2022 01:23

Tgt Ion:	149	Resp:	499264
Ion	Ratio	Lower	Upper
149	100		
167	27.9	22.1	33.1
279	5.6	4.2	6.2









Chrysene
Concen: 18.812 ng/ul
RT: 21.536 min Scan# 3111
Delta R.T. -0.006 min
Lab File: BM034211.D
Acq: 11 Mar 2022 01:23




Instrument :
BNA_M
ClientSampleId :
SICV003


Tgt Ion:	228	Resp:	931477
Ion	Ratio	Lower	Upper
228	100		
226	28.6	23.4	35.2
229	19.7	15.4	23.0



Di-n-octyl phthalate
Concen: 18.728 ng/ul
RT: 22.347 min Scan# 3249
Delta R.T. 0.000 min
Lab File: BM034211.D
Acq: 11 Mar 2022 01:23

Tgt Ion:149 Resp: 806423









Benzo(b)fluoranthene
Concen: 19.738 ng/ul
RT: 23.183 min Scan# 3391
Delta R.T. 0.000 min
Lab File: BM034211.D
Acq: 11 Mar 2022 01:23




Instrument :
BNA_M
ClientSampleId :
SICV003


Tgt Ion:	252	Resp:	929905
Ion Ratio	Lower	Upper	
252	100		
253	21.6	17.4	26.0
125	6.9	5.8	8.6



Benzo(k)fluoranthene
Concen: 19.680 ng/ul
RT: 23.230 min Scan# 3399
Delta R.T. -0.006 min
Lab File: BM034211.D
Acq: 11 Mar 2022 01:23

Tgt Ion:	252	Resp:	945032
Ion Ratio	Lower	Upper	
252	100		
253	22.3	17.5	26.3
125	7.1	5.1	7.7









Benzo(a)pyrene
Concen: 19.266 ng/ul
RT: 23.812 min Scan# 3499
Delta R.T. -0.006 min
Lab File: BM034211.D
Acq: 11 Mar 2022 01:23




Instrument :
BNA_M
ClientSampleId :
SICV003


Tgt Ion:	252	Resp:	905326
Ion	Ratio	Lower	Upper
252	100		
253	21.9	17.5	26.3
125	7.9	6.2	9.4



Indeno(1,2,3-cd)pyrene
Concen: 19.800 ng/ul
RT: 26.441 min Scan# 3945
Delta R.T. -0.006 min
Lab File: BM034211.D
Acq: 11 Mar 2022 01:23

Tgt Ion:	276	Resp:	1017298
Ion	Ratio	Lower	Upper
276	100		
138	16.3	13.6	20.4
277	24.1	19.9	29.9









Dibenzo(a,h)anthracene
Concen: 19.918 ng/ul
RT: 26.459 min Scan# 4077
Delta R.T. -0.006 min
Lab File: BM034211.D
Acq: 11 Mar 2022 01:23

Instrument :
BNA_M
ClientSampleId :
SICV003

Tgt Ion:	278	Resp:	857393
Ion	Ratio	Lower	Upper
278	100		
139	11.6	9.6	14.4
279	24.3	19.3	28.9



Benzo(g,h,i)perylene
Concen: 19.669 ng/ul
RT: 27.218 min Scan# 4077
Delta R.T. -0.006 min
Lab File: BM034211.D
Acq: 11 Mar 2022 01:23

Tgt Ion:	276	Resp:	939747
Ion	Ratio	Lower	Upper
276	100		
138	15.4	11.8	17.6
277	23.8	18.8	28.2

