


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM031022\
 Data File : BM034203.D
 Acq On : 10 Mar 2022 16:05
 Operator : CG/JU
 Sample : DFTPP041
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP041

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM031022.M
 Title : SVOA CALIBRATION
 Last Update : Fri Mar 11 02:00:05 2022



AutoFind: Scans 2472, 2473, 2474; Background Corrected with Scan 2466

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	28.1	55840	PASS
68	69	0.00	2	1.9	1288	PASS
69	198	0.00	100	33.8	67133	PASS
70	69	0.00	2	0.6	389	PASS
127	198	10	80	43.3	86080	PASS
197	198	0.00	2	0.6	1108	PASS
198	198	100	100	100.0	198656	PASS
199	198	5	9	6.8	13420	PASS
275	198	10	60	31.5	62539	PASS
365	198	1	100	5.2	10274	PASS
441	198	0.01	100	14.4	28581	PASS
442	198	50	100	93.9	186603	PASS
443	442	15	24	19.2	35787	PASS

m/z	Abundance
39.90	587.0
43.90	544.0
48.90	261.0
73.00	649.0
76.90	183.0
133.10	188.0
146.90	232.0
191.00	225.0
192.90	226.0
207.00	2175.0
208.00	345.0
209.10	274.0
281.00	571.0

Instrument :
BNA_M
ClientSampleId :
DFTPP041

m/z	Abundance
37.00	279.0
38.00	845.0
39.00	4329.0
40.00	589.0
41.00	215.0
44.00	837.0
44.90	216.0
48.90	696.0
50.00	17152.0
51.00	60744.0
52.00	3170.0
52.90	201.0
54.90	309.0
56.00	1966.0
57.00	4298.0
57.90	185.0
61.00	994.0
62.00	1222.0
63.00	2642.0
63.90	411.0
65.00	1163.0
68.00	1425.0
69.00	70040.0
70.00	532.0
73.00	930.0
74.00	8087.0
75.00	12325.0
76.00	3774.0
77.10	75096.0
78.10	5124.0
79.00	5883.0
80.00	5210.0
81.00	6424.0
82.00	1573.0
83.00	1231.0
84.00	316.0
85.00	853.0
85.90	2097.0
86.90	1050.0
88.00	275.0
91.00	1337.0
91.90	1650.0
93.00	11465.0
93.90	735.0
94.90	344.0
95.90	575.0
97.00	410.0
98.00	8854.0
99.00	7071.0
99.90	773.0
101.00	3416.0
101.90	244.0
103.00	1183.0
103.90	2435.0
104.90	2211.0
106.00	735.0
107.00	26488.0
107.90	3941.0
108.90	954.0
110.00	42216.0
111.00	7614.0
112.00	912.0
112.90	299.0
115.10	201.0
115.90	1448.0
117.00	27176.0
118.00	2080.0
118.90	247.0
119.90	376.0
120.90	162.0
121.90	2086.0
123.00	2913.0
124.00	1594.0
125.00	1084.0
127.00	87872.0
128.00	7760.0
129.00	36512.0
130.00	3154.0
131.00	502.0
132.00	449.0
132.90	419.0
134.10	831.0
135.00	3326.0
136.10	1177.0
137.00	2016.0
138.00	294.0
139.00	272.0
139.90	756.0
141.00	4993.0
142.00	1529.0
143.00	1178.0
143.90	375.0

Instrument :
BNA_M
ClientSampleId :
DFTPP041

145.00	352.0
145.90	849.0
146.90	2594.0
147.90	6505.0
149.00	1434.0
150.00	367.0
151.10	717.0
151.90	404.0
152.90	1499.0
154.00	1480.0
155.00	2791.0
156.00	5042.0
157.00	748.0
158.00	872.0
158.80	777.0
160.00	1863.0
161.00	2364.0
162.10	722.0
163.00	427.0
163.80	429.0
164.90	1985.0
166.00	2021.0
167.00	10211.0
168.00	6641.0
169.00	1175.0
170.00	400.0
170.50	291.0
171.00	422.0
171.90	953.0
173.00	1118.0
174.00	1908.0
175.00	4174.0
175.90	1232.0
176.90	1960.0
177.70	471.0
179.00	7840.0
180.00	5558.0
181.00	2438.0
181.80	509.0
183.00	307.0
184.00	661.0
185.00	4098.0
186.00	27624.0
187.00	9188.0
188.00	1161.0
189.00	1649.0
189.90	333.0
190.90	1211.0
192.00	2068.0
193.00	3110.0
194.10	756.0
195.00	335.0
196.00	5022.0
197.00	1296.0
197.90	185216.0
198.90	13093.0
199.90	1104.0
201.40	916.0
203.00	1438.0
204.00	7939.0
205.00	12471.0
206.00	52392.0
207.00	9179.0
207.90	2055.0
209.00	1038.0
210.00	927.0
211.00	2310.0
214.80	502.0
216.00	1236.0
217.00	14612.0
218.00	1908.0
219.00	305.0
221.00	7148.0
222.00	1762.0
223.00	3072.0
224.00	29888.0
225.00	7655.0
225.90	1107.0
227.00	14225.0
227.90	2026.0
228.90	2668.0
229.90	421.0
231.00	924.0
232.90	272.0
233.80	822.0
235.00	1020.0
236.00	619.0
237.10	1444.0
238.00	180.0
239.00	610.0
239.90	618.0
240.90	788.0
242.00	1890.0
243.00	1145.0
244.00	22000.0
245.00	3679.0

Instrument :
BNA_M
ClientSampleId :
DFTPP041

246.00	5059.0
247.00	885.0
247.90	270.0
248.90	677.0
249.90	189.0
251.00	287.0
252.00	252.0
252.90	383.0
254.10	867.0
255.00	121552.0
256.00	17248.0
257.00	1335.0
258.00	7336.0
258.90	1563.0
259.90	201.0
260.80	349.0
264.00	158.0
264.90	2767.0
265.90	581.0
267.00	210.0
271.00	381.0
271.80	269.0
273.00	3845.0
274.00	8789.0
275.00	53448.0
276.00	6910.0
277.00	4865.0
278.00	1092.0
279.00	247.0
281.00	1008.0
282.00	357.0
283.00	622.0
284.00	615.0
285.00	823.0
285.90	222.0
288.90	320.0
289.90	263.0
293.00	1213.0
294.00	395.0
295.00	368.0
296.00	16067.0
297.00	2412.0
297.90	195.0
300.90	203.0
301.90	219.0
303.00	1622.0
303.90	594.0
307.90	187.0
308.80	196.0
309.90	291.0
312.90	178.0
314.10	914.0
314.90	2045.0
315.90	793.0
321.00	621.0
321.90	334.0
323.10	4752.0
324.10	961.0
325.90	193.0
327.00	902.0
327.90	392.0
331.90	472.0
332.90	606.0
334.00	3183.0
335.00	914.0
340.90	732.0
342.10	210.0
346.00	1240.0
347.00	211.0
352.00	1543.0
353.00	946.0
354.10	2118.0
355.00	549.0
365.00	8046.0
366.00	1345.0
369.90	175.0
371.00	479.0
372.00	2706.0
373.00	873.0
383.00	767.0
383.90	301.0
390.00	360.0
391.10	233.0
392.20	244.0
401.00	204.0
402.00	1196.0
403.00	1640.0
404.00	760.0
404.90	177.0
421.00	1248.0
422.00	1158.0
423.00	8280.0
424.00	3036.0
425.00	556.0
440.10	217.0
441.10	21584.0

Instrument :
BNA_M
ClientSampleId :
DFTPP041

442.00	135104.0
443.10	27440.0
444.00	2763.0
444.80	155.0

Instrument :
BNA_M
ClientSampleId :
DFTPP041

m/z	Abundance
37.00	391.0
38.00	751.0
39.00	4706.0
39.90	893.0
41.00	174.0
43.00	226.0
44.00	646.0
45.00	309.0
48.90	546.0
50.00	18288.0
51.10	65720.0
52.00	3469.0
52.90	238.0
54.90	584.0
56.00	2283.0
57.00	4372.0
58.00	276.0
61.00	896.0
62.00	1402.0
63.00	3113.0
63.90	358.0
65.00	1292.0
68.00	1462.0
69.00	80496.0
70.00	332.0
73.00	1266.0
74.00	8806.0
75.00	14012.0
76.10	4059.0
77.10	87968.0
78.00	6073.0
79.00	6456.0
80.00	5760.0
81.00	6533.0
82.00	1962.0
83.00	1386.0
84.00	357.0
84.90	966.0
86.00	2408.0
87.00	1088.0
88.00	391.0
88.90	250.0
90.90	1616.0
92.00	1692.0
93.00	12840.0
93.90	752.0
95.00	289.0
95.90	437.0
97.10	327.0
98.00	10840.0
99.00	8019.0
100.00	653.0
101.00	4096.0
102.00	270.0
102.90	1279.0
104.00	2566.0
105.00	2563.0
105.90	859.0
107.00	32624.0
108.00	4925.0
108.90	830.0
110.00	48752.0
111.00	8291.0
112.00	1439.0
112.80	257.0
114.80	203.0
116.10	1960.0
117.00	32840.0
117.90	2119.0
119.00	266.0
119.90	372.0
121.90	2147.0
123.00	3486.0
124.00	1763.0
125.00	1329.0
127.00	102840.0
128.00	8792.0
129.00	43888.0
130.00	4163.0
130.90	818.0
131.90	340.0
133.00	331.0
134.00	1761.0
135.00	4126.0
135.90	1686.0
137.00	2547.0
137.90	471.0
139.10	243.0
140.00	538.0
141.00	5619.0
142.00	2023.0
143.00	1428.0

Instrument :
BNA_M
ClientSampleId :
DFTPP041

144.10	391.0
144.90	215.0
146.00	914.0
147.00	3043.0
148.00	7170.0
149.00	1321.0
149.90	356.0
151.00	903.0
152.00	700.0
153.00	1579.0
154.00	1414.0
155.00	3367.0
156.00	5167.0
157.00	612.0
157.90	1306.0
158.90	811.0
160.00	2308.0
161.00	2917.0
162.00	788.0
163.00	265.0
164.00	263.0
165.00	2655.0
165.90	1907.0
167.00	12139.0
168.00	7347.0
169.00	981.0
170.00	339.0
170.90	581.0
171.90	943.0
173.00	1836.0
174.00	2809.0
175.00	4857.0
176.00	1195.0
177.00	2262.0
177.90	818.0
178.90	9407.0
180.00	7289.0
180.90	2850.0
181.90	478.0
183.00	197.0
183.90	747.0
185.00	5338.0
186.00	34376.0
187.00	10088.0
188.00	1384.0
189.00	2919.0
190.00	440.0
191.00	1540.0
192.00	3137.0
193.00	3458.0
193.90	983.0
194.90	270.0
196.00	6552.0
196.90	1207.0
197.90	240640.0
198.90	16664.0
200.00	1231.0
201.60	1142.0
203.00	2016.0
204.00	9790.0
205.00	16184.0
206.00	68808.0
207.00	10346.0
208.00	3028.0
209.00	1172.0
210.00	1120.0
211.00	2763.0
214.90	939.0
215.90	1610.0
216.90	18936.0
218.00	2278.0
218.80	224.0
220.10	344.0
221.00	8783.0
221.90	2269.0
222.90	4241.0
224.00	38408.0
225.00	10174.0
226.10	1026.0
227.00	17608.0
227.90	2724.0
229.00	3805.0
230.00	686.0
231.00	1397.0
232.00	207.0
232.90	289.0
234.00	866.0
235.00	1437.0
236.00	878.0
237.00	2084.0
238.00	295.0
239.10	795.0
240.00	523.0
241.00	1212.0
241.90	2460.0
243.00	2358.0

Instrument :
BNA_M
ClientSampleId :
DFTPP041

244.00	30272.0
245.00	4597.0
246.00	6751.0
247.00	1272.0
247.80	385.0
249.00	1019.0
250.00	273.0
250.90	266.0
251.90	454.0
252.90	802.0
254.00	1194.0
255.00	165760.0
256.00	24912.0
257.00	2170.0
258.00	10532.0
259.00	1769.0
260.00	163.0
260.90	342.0
263.90	255.0
265.00	4484.0
265.90	765.0
266.90	177.0
270.90	516.0
271.90	360.0
272.90	5076.0
274.00	11837.0
275.00	75864.0
276.00	9891.0
276.90	6883.0
278.00	1238.0
278.90	286.0
280.90	618.0
282.00	419.0
283.00	813.0
283.90	593.0
285.10	948.0
285.80	290.0
289.10	376.0
289.90	336.0
291.80	355.0
293.00	1452.0
294.00	593.0
294.90	640.0
296.00	24544.0
297.00	3605.0
297.90	195.0
300.90	206.0
301.80	291.0
303.00	3008.0
304.00	520.0
307.90	336.0
308.80	173.0
310.00	438.0
313.00	202.0
314.00	1124.0
314.90	2977.0
316.00	1497.0
317.00	291.0
321.10	934.0
321.90	276.0
323.00	6895.0
324.00	1383.0
325.00	162.0
326.90	1392.0
327.90	853.0
332.10	670.0
333.00	1021.0
334.00	4948.0
335.00	1041.0
336.10	166.0
338.90	196.0
341.00	1022.0
342.10	274.0
346.00	1761.0
347.00	355.0
352.00	2421.0
353.00	1834.0
354.00	2237.0
355.10	608.0
359.00	158.0
365.00	12042.0
366.00	1661.0
366.90	178.0
370.00	334.0
370.90	480.0
372.00	3468.0
373.10	1010.0
373.90	190.0
383.00	1137.0
383.90	453.0
390.00	429.0
391.00	399.0
392.00	360.0
400.80	229.0
402.00	1710.0
403.10	2483.0

Instrument :
BNA_M
ClientSampleId :
DFTPP041

404.00	873.0
405.00	263.0
421.00	2220.0
422.00	2028.0
423.10	14128.0
424.00	3723.0
425.00	558.0
440.10	266.0
441.10	33688.0
442.10	225024.0
443.10	41960.0
444.00	3708.0
445.00	254.0

Instrument :
BNA_M
ClientSampleId :
DFTPP041

m/z	Abundance
36.90	249.0
38.00	565.0
39.10	3158.0
40.00	679.0
43.00	188.0
44.00	570.0
44.80	154.0
49.00	546.0
50.10	11838.0
51.00	41056.0
52.00	2531.0
55.00	329.0
56.00	1570.0
57.00	2964.0
61.00	710.0
62.00	805.0
63.00	1850.0
64.00	218.0
65.00	952.0
68.00	976.0
69.00	50864.0
70.00	302.0
73.00	1044.0
74.00	5386.0
75.00	8593.0
76.00	2939.0
77.00	54424.0
78.00	3759.0
79.00	4612.0
80.00	3201.0
81.00	4611.0
81.80	818.0
83.00	898.0
84.00	223.0
84.90	830.0
86.00	1393.0
86.90	622.0
88.10	311.0
88.90	155.0
91.00	1052.0
92.00	1422.0
93.00	8972.0
93.90	384.0
94.90	154.0
95.80	365.0
98.00	6478.0
99.00	4827.0
99.90	532.0
101.00	2755.0
102.00	170.0
103.00	857.0
104.00	1861.0
105.00	1751.0
105.90	633.0
107.00	21848.0
108.00	3170.0
109.10	410.0
110.00	33080.0
111.00	5415.0
112.00	710.0
112.90	274.0
116.00	1490.0
117.00	20744.0
118.00	1687.0
119.00	240.0
119.90	333.0
121.00	193.0
122.00	1221.0
123.00	1987.0
124.00	884.0
125.00	904.0
126.00	256.0
127.00	67528.0
128.00	5333.0
129.00	30176.0
130.00	2306.0
131.00	381.0
131.90	279.0
133.00	273.0
134.00	927.0
135.00	2297.0
136.00	1031.0
137.00	1541.0
138.10	389.0
140.10	275.0
141.00	4342.0
142.00	1189.0
142.90	927.0
144.00	310.0
145.00	161.0
145.90	869.0
147.00	2392.0

Instrument :
BNA_M
ClientSampleId :
DFTPP041

148.00	5651.0
148.90	907.0
149.90	338.0
151.10	477.0
151.60	305.0
152.10	160.0
152.90	1162.0
154.00	1104.0
155.00	2380.0
156.00	3290.0
157.00	832.0
157.90	788.0
159.00	735.0
160.00	1413.0
161.00	1943.0
162.00	477.0
163.00	274.0
163.90	391.0
165.00	1730.0
166.00	904.0
167.00	8336.0
168.00	4328.0
169.10	939.0
169.90	263.0
171.00	355.0
171.90	1026.0
173.00	1096.0
174.00	1835.0
175.00	3539.0
176.10	1084.0
176.90	1697.0
177.70	314.0
178.10	495.0
178.90	7287.0
180.00	4447.0
181.00	2243.0
182.10	394.0
184.10	395.0
185.00	2828.0
186.00	24960.0
187.00	6720.0
188.10	830.0
188.90	1770.0
190.00	414.0
190.90	1095.0
191.90	2069.0
193.00	2466.0
194.00	590.0
194.80	370.0
196.00	4644.0
197.00	822.0
197.90	170112.0
199.00	10503.0
199.90	1130.0
201.40	1069.0
202.90	1333.0
204.00	7086.0
205.00	12450.0
206.00	48656.0
207.00	8143.0
208.00	1943.0
208.90	828.0
210.10	1157.0
211.00	1719.0
211.90	346.0
213.00	221.0
214.90	684.0
216.00	1023.0
217.00	15273.0
218.00	2201.0
219.00	206.0
219.80	181.0
221.00	6468.0
221.80	1608.0
223.00	3263.0
224.00	29344.0
225.00	6974.0
226.00	935.0
227.00	12867.0
227.90	2078.0
229.00	2843.0
230.00	259.0
231.00	952.0
231.90	186.0
233.00	334.0
233.90	871.0
235.00	1006.0
236.00	734.0
237.00	1551.0
237.90	214.0
239.00	682.0
239.90	356.0
240.90	1076.0
241.90	1616.0
243.00	1736.0
244.00	21920.0

Instrument :
BNA_M
ClientSampleId :
DFTPP041

245.00	2686.0
246.00	5069.0
246.90	1252.0
247.90	262.0
249.00	723.0
250.10	164.0
250.80	231.0
251.90	218.0
253.00	678.0
254.00	907.0
255.00	124192.0
256.00	18296.0
256.90	1534.0
258.00	7837.0
259.00	1648.0
259.90	251.0
260.90	198.0
263.90	330.0
265.00	3373.0
265.90	500.0
266.80	180.0
270.00	208.0
271.00	237.0
271.80	249.0
273.00	4076.0
274.00	10015.0
275.00	58304.0
276.00	7684.0
277.00	5124.0
278.00	1008.0
278.80	232.0
281.00	801.0
282.00	229.0
282.90	681.0
284.00	440.0
285.00	728.0
286.20	160.0
289.10	234.0
290.10	159.0
291.80	183.0
292.90	1104.0
294.00	386.0
295.00	550.0
296.00	19784.0
297.00	2911.0
297.90	274.0
301.00	153.0
301.80	266.0
303.00	2133.0
304.00	668.0
307.90	400.0
308.90	206.0
309.90	245.0
313.10	251.0
314.00	998.0
314.90	2062.0
316.10	1390.0
321.00	550.0
322.10	467.0
323.00	5653.0
324.00	1122.0
325.90	208.0
327.00	1165.0
327.90	755.0
331.90	509.0
333.00	610.0
334.00	4022.0
335.00	894.0
340.90	812.0
342.00	353.0
345.90	1367.0
346.90	332.0
352.00	1939.0
352.90	1457.0
354.00	2032.0
354.90	400.0
365.00	10733.0
365.90	1280.0
371.00	695.0
372.00	3381.0
373.00	817.0
383.00	1227.0
383.80	223.0
390.00	448.0
391.00	275.0
391.90	247.0
401.00	246.0
402.00	1589.0
403.00	1919.0
404.00	624.0
405.00	180.0
414.90	157.0
421.00	1797.0
422.00	1734.0
423.00	11529.0
424.10	3269.0

Instrument :
BNA_M
ClientSampleId :
DFTPP041

425.10	456.0
441.10	30472.0
442.10	199680.0
443.10	37960.0
444.00	3303.0
444.90	156.0

Instrument :
BNA_M
ClientSampleId :
DFTPP041