


Data Path : Z:\HPCHEM1\BNA_M\Data\BM031115\
 Data File : BM000689.D
 Acq On : 11 Mar 2015 09:57
 Operator : TP/IZ
 Sample : DFTPP35
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP35

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\Methods\SOM02.2-EPA-BM031115.M
 Title : SVOA CALIBRATION
 Last Update : Thu Mar 12 07:14:19 2015



AutoFind: Scans 2518, 2519, 2520; Background Corrected with Scan 2513

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	39.9	87226	PASS
68	69	0.00	2	1.6	1291	PASS
69	198	0.00	100	36.7	80210	PASS
70	69	0.00	2	0.5	424	PASS
127	198	10	80	47.4	103570	PASS
197	198	0.00	2	0.4	858	PASS
198	198	100	100	100.0	218730	PASS
199	198	5	9	6.4	14106	PASS
275	198	10	60	22.8	49826	PASS
365	198	1	100	2.4	5293	PASS
441	443	0.01	100	78.9	28162	PASS
442	198	50	100	85.8	187682	PASS
443	442	15	24	19.0	35685	PASS

m/z	Abundance
38.00	1019.0
39.10	6374.0
40.00	3438.0
44.00	4699.0
49.00	714.0
50.10	21176.0
51.10	88656.0
52.10	4923.0
55.10	333.0
56.00	2664.0
57.00	6204.0
61.00	1096.0
62.00	1126.0
63.10	3583.0
64.10	670.0
65.10	1767.0
68.00	1410.0
69.00	79520.0
70.00	486.0
73.00	1037.0
74.10	8214.0
75.00	11793.0
76.10	4104.0
77.10	91912.0
78.10	6021.0
79.00	6120.0
80.00	5232.0
81.00	7246.0
82.00	1841.0
83.00	1504.0
84.10	396.0
85.00	1101.0
86.00	1917.0
87.00	996.0
91.00	1825.0
92.00	2200.0
93.00	9604.0
94.00	625.0
96.00	556.0
98.00	7728.0
99.00	6613.0
100.20	399.0
101.00	3750.0
102.90	1111.0
104.00	2074.0
105.00	2589.0
106.10	662.0
107.00	27200.0
108.00	4427.0
109.00	626.0
110.00	52080.0
111.00	7630.0
111.90	673.0
113.00	572.0
116.00	1526.0
117.00	19872.0
118.10	1247.0
120.10	484.0
122.00	1991.0
123.00	2847.0
124.00	1142.0
124.90	1029.0
126.10	530.0
127.00	98416.0
128.00	7210.0
129.00	35520.0
130.00	2964.0
130.90	669.0
132.10	407.0
134.10	878.0
135.10	3502.0
136.00	1081.0
137.10	1945.0
139.90	535.0
141.00	5030.0
142.10	1440.0
143.00	978.0
146.00	723.0
147.10	2164.0
148.00	5225.0
149.00	1230.0
151.00	451.0
151.80	316.0
152.90	1329.0
154.10	1168.0
155.00	2929.0
156.00	3547.0
157.00	714.0
157.90	1013.0
159.00	859.0
160.00	1414.0

Instrument :
BNA_M
ClientSampleId :
DFTPP35

161.00	1908.0
161.90	792.0
163.10	342.0
164.10	303.0
165.00	1588.0
166.00	1649.0
167.00	8195.0
168.00	6142.0
169.10	1055.0
170.80	373.0
172.00	925.0
173.00	905.0
174.10	2273.0
175.10	3582.0
176.00	1212.0
177.00	1831.0
177.90	676.0
179.00	6392.0
180.00	6825.0
181.10	1923.0
182.10	481.0
184.10	423.0
185.00	3001.0
186.10	24776.0
187.00	7267.0
188.10	1227.0
189.10	1558.0
190.90	883.0
192.00	1716.0
193.00	2288.0
193.90	497.0
196.00	5507.0
196.70	1119.0
198.00	191552.0
199.00	12556.0
199.90	731.0
201.70	998.0
203.10	1408.0
204.00	6031.0
205.10	10530.0
206.10	47824.0
207.10	7722.0
208.10	1531.0
208.90	684.0
209.90	668.0
211.20	1580.0
211.90	490.0
215.00	434.0
216.10	759.0
217.00	12322.0
218.00	1522.0
221.10	7294.0
221.70	1920.0
223.00	2474.0
224.10	22888.0
225.10	5879.0
226.10	946.0
227.00	8072.0
228.10	1305.0
229.00	2231.0
231.00	680.0
234.00	668.0
234.90	1118.0
235.90	519.0
237.00	1476.0
238.90	379.0
239.90	353.0
241.00	416.0
242.10	1521.0
243.00	1014.0
244.10	18992.0
245.00	2562.0
246.00	3281.0
247.00	708.0
249.10	482.0
252.90	362.0
253.30	331.0
254.10	811.0
255.00	92520.0
256.00	13474.0
256.90	1287.0
258.00	4919.0
258.90	689.0
265.10	1763.0
272.10	366.0
273.00	2662.0
274.00	6569.0
275.00	39264.0
276.00	5319.0
277.00	2607.0
278.00	643.0
283.00	324.0
285.10	507.0
293.00	575.0
294.10	387.0

Instrument :
BNA_M
ClientSampleId :
DFTPP35

296.00	9082.0
297.00	1438.0
303.00	961.0
304.10	310.0
314.10	454.0
315.10	740.0
316.00	593.0
320.90	409.0
323.10	3069.0
324.00	770.0
327.00	664.0
328.00	366.0
334.00	1739.0
335.00	357.0
345.90	550.0
352.00	800.0
353.10	1162.0
354.10	1083.0
365.00	3717.0
366.00	504.0
372.00	1573.0
372.80	387.0
382.90	409.0
402.00	1027.0
403.00	927.0
403.90	307.0
421.00	805.0
422.00	826.0
423.00	6585.0
424.10	3293.0
425.00	387.0
441.00	18128.0
442.00	124712.0
443.00	24928.0
443.90	2128.0

Instrument :
BNA_M
ClientSampleId :
DFTPP35

m/z	Abundance
37.00	402.0
38.00	1073.0
39.10	8190.0
40.00	3319.0
44.00	3979.0
49.00	786.0
50.10	25464.0
51.10	104248.0
52.00	5654.0
54.90	340.0
56.00	3126.0
57.00	7048.0
58.00	430.0
61.10	1091.0
62.00	1528.0
63.00	3377.0
64.00	654.0
65.10	1871.0
68.00	1418.0
69.00	97096.0
70.00	394.0
73.00	1024.0
74.10	9811.0
75.00	15360.0
76.10	5035.0
77.10	111632.0
78.10	7615.0
79.00	7192.0
80.00	4823.0
81.00	7229.0
82.00	2340.0
83.00	1955.0
85.00	1521.0
86.00	2409.0
87.00	912.0
88.00	350.0
91.00	1767.0
92.00	1997.0
93.00	11951.0
94.10	674.0
96.00	675.0
98.00	8333.0
99.00	7822.0
99.90	611.0
101.00	4359.0
103.00	1425.0
104.00	2453.0
105.10	2855.0
106.00	822.0
107.00	34880.0
108.00	4607.0
109.00	861.0
110.00	64488.0
111.00	10118.0
112.00	958.0
112.90	376.0
116.10	1757.0
117.00	26208.0
118.00	1860.0
119.00	344.0
120.00	307.0
122.00	2149.0
123.00	3696.0
123.90	1450.0
124.90	1148.0
126.20	527.0
127.00	126040.0
128.00	9098.0
129.00	46864.0
130.00	3493.0
131.00	930.0
132.00	540.0
134.00	1267.0
135.00	3234.0
136.00	1302.0
137.10	2012.0
137.90	460.0
139.00	301.0
140.00	558.0
141.00	5270.0
142.00	1654.0
143.00	956.0
144.00	498.0
145.10	355.0
145.90	1052.0
147.00	3397.0
148.00	6268.0
149.00	1594.0
150.00	436.0
151.20	768.0
151.70	541.0

Instrument :
BNA_M
ClientSampleId :
DFTPP35

152.90	2026.0
154.00	1220.0
155.00	3590.0
156.00	5217.0
157.00	994.0
158.00	909.0
159.00	933.0
160.00	1920.0
161.00	2728.0
162.00	785.0
164.00	319.0
165.00	2167.0
166.10	1692.0
167.00	11409.0
168.00	6260.0
169.00	1099.0
170.10	414.0
170.90	362.0
172.10	934.0
173.00	1245.0
174.00	2546.0
175.10	4412.0
176.00	1682.0
176.90	1961.0
178.00	857.0
179.00	9319.0
180.00	7636.0
181.00	3383.0
183.10	409.0
183.90	648.0
185.00	4278.0
186.00	34424.0
187.10	9994.0
187.90	1283.0
189.00	1748.0
190.10	384.0
190.90	1154.0
192.00	2749.0
193.10	3392.0
194.00	638.0
195.00	524.0
196.00	7105.0
197.00	1497.0
198.00	264256.0
199.00	17176.0
200.00	1291.0
201.40	1228.0
203.00	2033.0
204.00	8285.0
205.10	14585.0
206.10	65600.0
207.10	10649.0
208.10	2230.0
209.10	1075.0
210.10	797.0
211.00	2634.0
211.80	485.0
214.80	651.0
216.00	1306.0
217.00	15628.0
218.00	2021.0
219.00	317.0
221.10	11276.0
221.80	1929.0
223.00	3662.0
224.10	34488.0
225.10	9006.0
226.00	1342.0
227.00	13391.0
228.00	1636.0
229.00	2830.0
229.90	300.0
231.00	1654.0
234.00	895.0
234.90	994.0
236.00	871.0
237.00	2114.0
239.00	627.0
240.00	495.0
241.00	712.0
242.00	1842.0
243.10	1722.0
244.10	26920.0
245.10	4004.0
246.00	5097.0
247.00	966.0
249.00	1142.0
253.00	527.0
254.10	933.0
255.00	134720.0
256.10	20640.0
257.10	1670.0
258.00	6906.0
258.90	1273.0
260.00	327.0

Instrument :
BNA_M
ClientSampleId :
DFTPP35

265.00	2427.0
265.90	561.0
272.20	393.0
273.00	3900.0
274.00	11574.0
275.00	62504.0
276.10	7733.0
277.00	5127.0
278.00	927.0
283.00	744.0
284.10	468.0
285.00	1208.0
293.00	929.0
294.00	479.0
295.00	439.0
296.00	16006.0
297.00	2008.0
302.00	390.0
303.00	2180.0
304.10	494.0
314.10	658.0
315.00	1362.0
315.90	1116.0
321.00	529.0
323.10	5172.0
324.10	911.0
327.00	1140.0
328.00	673.0
332.00	364.0
333.00	592.0
334.00	3000.0
335.00	630.0
341.00	675.0
345.90	1051.0
352.00	1337.0
353.00	1264.0
354.10	1899.0
355.00	457.0
365.00	6293.0
365.90	885.0
370.80	315.0
372.00	2868.0
373.00	519.0
382.90	875.0
402.00	1042.0
403.00	2154.0
404.00	1063.0
421.00	1460.0
422.00	1707.0
423.10	10648.0
424.00	4740.0
425.10	594.0
441.00	33752.0
442.00	218048.0
443.10	42720.0
444.00	3703.0

Instrument :
BNA_M
ClientSampleId :
DFTPP35

m/z	Abundance
38.00	724.0
39.10	4630.0
40.00	3463.0
44.00	4029.0
49.00	626.0
50.10	15756.0
51.10	68776.0
52.10	3205.0
55.00	436.0
56.00	2096.0
57.00	5044.0
61.10	763.0
62.00	809.0
63.10	2811.0
64.10	421.0
65.00	1186.0
68.10	1046.0
69.00	64016.0
70.00	394.0
73.10	711.0
74.00	6841.0
75.00	10599.0
76.10	3510.0
77.10	72128.0
78.10	5091.0
79.00	4531.0
79.90	3590.0
81.00	5046.0
82.00	1474.0
82.90	1397.0
84.90	1005.0
86.00	1731.0
87.00	489.0
91.00	1082.0
92.00	1251.0
93.00	8398.0
94.10	728.0
96.00	469.0
98.00	5751.0
99.00	5441.0
99.90	395.0
101.00	2676.0
103.00	1038.0
104.00	2252.0
105.00	1669.0
105.90	661.0
107.00	24016.0
108.00	3570.0
109.10	528.0
110.00	45544.0
111.00	6525.0
112.00	1073.0
116.00	954.0
117.00	16680.0
118.10	1545.0
122.00	1150.0
123.00	2884.0
123.90	1103.0
125.00	1332.0
126.20	485.0
127.00	86256.0
128.00	6645.0
129.00	32752.0
129.90	2238.0
131.00	446.0
134.00	1196.0
135.00	3025.0
136.00	1080.0
136.90	1559.0
137.80	455.0
139.90	328.0
141.00	3297.0
142.00	1254.0
143.00	963.0
146.00	881.0
146.90	1967.0
148.00	4827.0
149.00	1262.0
151.00	464.0
152.00	409.0
153.00	1472.0
154.00	865.0
154.90	2175.0
156.00	3628.0
156.90	620.0
157.80	1013.0
159.00	729.0
160.00	1174.0
161.10	2254.0
162.10	456.0
162.90	305.0

Instrument :
BNA_M
ClientSampleId :
DFTPP35

165.00	1878.0
166.00	1187.0
167.00	7849.0
168.00	5606.0
169.00	847.0
170.90	543.0
172.00	690.0
173.00	949.0
173.90	1862.0
175.10	3843.0
175.90	869.0
176.90	1454.0
178.10	615.0
179.00	7292.0
180.00	5129.0
180.90	2206.0
183.00	356.0
183.80	483.0
185.10	3006.0
186.00	25192.0
187.00	6902.0
188.00	1009.0
188.90	1855.0
190.90	824.0
192.00	1589.0
193.00	3138.0
194.10	402.0
196.10	5652.0
196.70	1259.0
197.10	1078.0
198.00	200384.0
199.00	12586.0
200.10	981.0
201.30	958.0
201.80	448.0
203.00	1170.0
204.00	6918.0
205.00	11518.0
206.10	48728.0
207.00	7330.0
208.00	1522.0
209.00	882.0
210.10	722.0
211.10	2024.0
211.80	560.0
215.00	409.0
216.20	720.0
217.00	12424.0
218.10	1394.0
221.00	8310.0
221.80	2184.0
223.10	3372.0
224.10	27328.0
225.10	6543.0
225.90	956.0
227.00	10776.0
227.90	1623.0
229.00	1901.0
230.00	324.0
231.00	899.0
233.90	574.0
234.90	726.0
235.90	536.0
237.00	1108.0
239.00	470.0
240.00	399.0
241.00	502.0
242.00	1196.0
243.00	1569.0
244.10	20880.0
245.20	2681.0
246.00	3766.0
247.00	902.0
249.10	829.0
252.90	751.0
254.10	781.0
255.00	108880.0
256.00	14642.0
257.00	1537.0
258.00	6126.0
259.00	1477.0
265.00	2056.0
266.20	325.0
271.00	305.0
271.90	327.0
273.00	3725.0
274.00	9478.0
275.00	47712.0
276.10	6668.0
277.00	3436.0
277.90	801.0
282.90	419.0
284.10	318.0
285.00	664.0
292.90	1223.0

Instrument :
BNA_M
ClientSampleId :
DFTPP35

294.00	507.0
294.90	359.0
296.00	12500.0
297.00	1706.0
303.00	1285.0
304.20	479.0
314.00	609.0
315.10	1354.0
316.00	577.0
321.00	374.0
323.10	4567.0
324.00	805.0
327.10	826.0
333.00	395.0
334.00	2575.0
335.00	682.0
340.90	379.0
346.00	840.0
352.00	1636.0
353.10	954.0
354.10	1432.0
355.00	476.0
365.00	5870.0
366.00	769.0
370.90	402.0
372.00	2909.0
373.00	735.0
382.90	808.0
384.00	443.0
389.80	365.0
401.90	1085.0
403.00	1817.0
404.00	687.0
421.00	1575.0
422.00	1214.0
423.10	10845.0
424.10	3129.0
424.90	478.0
441.00	32608.0
442.00	220288.0
443.00	39408.0
444.10	4640.0

Instrument :
BNA_M
ClientSampleId :
DFTPP35

Scan 2513 (17.468 min): BM000689.D
DFTPP35

m/z	Abundance
40.00	3394.0
44.00	4992.0
73.10	475.0
207.00	1622.0

Instrument :
BNA_M
ClientSampleId :
DFTPP35