

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM031117\
 Data File : BM009215.D
 Acq On : 11 Mar 2017 19:29
 Operator : SJ/MA
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.459

Quant Time: Mar 13 12:22:31 2017
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM030317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 13 11:21:07 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	197#	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	211#	0.00
3	Naphthalene	1.153	1.119	2.9	207#	0.00
4 SURR	2-Methylnaphthalene-d10	0.664	0.619	6.8	201#	0.00
5	2-Methylnaphthalene	0.829	0.786	5.2	206#	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	204#	0.00
7	Acenaphthylene	2.082	2.112	-1.4	216#	0.00
8 C	Acenaphthene	1.598	1.508	5.6	202#	0.00
9	Fluorene	1.869	1.781	4.7	193#	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	195#	0.00
11	Pentachlorophenol	0.194	0.167	13.9	158#	0.00
12	Phenanthrene	1.256	1.235	1.7	196#	0.00
13	Anthracene	1.241	1.169	5.8	184#	0.00
14 SURR	Fluoranthene-d10	1.251	1.168	6.6	182#	0.00
15 C	Fluoranthene	1.638	1.478	9.8	175#	-0.01
16 I	Chrysene-d12	1.000	1.000	0.0	163#	0.00
17	Pyrene	1.376	1.470	-6.8	183#	-0.01
18	Benzo(a)anthracene	1.329	1.386	-4.3	167#	0.00
19	Chrysene	1.261	1.295	-2.7	170#	0.00
20 I	Perylene-d12	1.000	1.000	0.0	154#	0.00
21	Benzo(b)fluoranthene	1.571	1.597	-1.7	164#	-0.01
22	Benzo(k)fluoranthene	1.389	1.336	3.8	157#	0.00
23 C	Benzo(a)pyrene	1.414	1.398	1.1	156#	0.00
24	Indeno(1,2,3-cd)pyrene	1.798	1.709	4.9	143	0.00
25	Dibenzo(a,h)anthracene	1.510	1.372	9.1	134	-0.01
26	Benzo(g,h,i)perylene	1.534	1.444	5.9	140	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0