

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM031121\
 Data File : BM029087.D
 Acq On : 11 Mar 2021 10:05
 Operator : CG/JU
 Sample : PB135092BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 PB135092BL

Quant Time: Mar 11 10:37:42 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM022321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 09 10:08:02 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.663	152	9598	20.00	ng	0.00
21) Naphthalene-d8	10.457	136	35487	20.00	ng	# 0.00
39) Acenaphthene-d10	14.333	164	20274	20.00	ng	0.00
64) Phenanthrene-d10	17.080	188	41141	20.00	ng	0.00
76) Chrysene-d12	21.250	240	40753	20.00	ng	#-0.01
86) Perylene-d12	23.409	264	42505	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.240	112	80930	139.76	ng	0.00
7) Phenol-d6	6.863	99	91799	120.03	ng	0.00
23) Nitrobenzene-d5	8.839	82	60162	91.47	ng	0.00
42) 2,4,6-Tribromophenol	15.833	330	31654	131.73	ng	0.00
45) 2-Fluorobiphenyl	12.945	172	130867	85.93	ng	0.00
79) Terphenyl-d14	19.703	244	194219	88.00	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM031121\
Data File : BM029087.D
Acq On : 11 Mar 2021 10:05
Operator : CG/JU
Sample : PB135092BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
PB135092BL

Quant Time: Mar 11 10:37:42 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM022321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 09 10:08:02 2021
Response via : Initial Calibration

