

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM031122\
 Data File : BM034231.D
 Acq On : 11 Mar 2022 21:01
 Operator : CG/JU
 Sample : N1816-03
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DBSD0

Quant Time: Mar 11 23:42:32 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM031022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Mar 11 02:00:05 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.948	152	158830	20.000	ng/u1	0.00
20) Naphthalene-d8	10.766	136	615515	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.589	164	394734	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.330	188	773490	20.000	ng/u1	0.00
79) Chrysene-d12	21.500	240	720068	20.000	ng/u1	0.00
88) Perylene-d12	23.918	264	820296	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.319	96	22814	6.270	ng/uL	0.00
4) Pyridine-d5	3.743	84	276833	31.956	ng/u1	0.00
7) Phenol-d5	7.101	99	348473	32.785	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.272	67	213417	34.226	ng/u1	0.00
11) 2-Chlorophenol-d4	7.478	132	322435	35.101	ng/u1	0.00
15) 4-Methylphenol-d8	8.654	113	300085	33.011	ng/u1	0.00
21) Nitrobenzene-d5	9.113	128	150204	34.021	ng/u1	0.00
24) 2-Nitrophenol-d4	9.837	143	168444	34.406	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.378	165	323541	32.965	ng/u1	0.00
31) 4-Chloroaniline-d4	10.895	131	387413	31.740	ng/u1	0.00
46) Dimethylphthalate-d6	13.995	166	962589	33.515	ng/u1	0.00
49) Acenaphthylene-d8	14.283	160	1225690	33.600	ng/u1	0.00
54) 4-Nitrophenol-d4	14.766	143	100138	22.383	ng/u1	0.00
60) Fluorene-d10	15.577	176	843165	33.536	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.689	200	124504	25.441	ng/u1	0.00
73) Anthracene-d10	17.430	188	1234637	33.795	ng/u1	0.00
81) Pyrene-d10	19.712	212	1449986	36.988	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.765	264	1365313	33.737	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM031122\
Data File : BM034231.D
Acq On : 11 Mar 2022 21:01
Operator : CG/JU
Sample : N1816-03
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DBSD0

Quant Time: Mar 11 23:42:32 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM031022.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Mar 11 02:00:05 2022
Response via : Initial Calibration

