

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM031122\
 Data File : BM034217.D
 Acq On : 11 Mar 2022 11:58
 Operator : CG/JU
 Sample : N1816-09MSD
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DBSD4MSD

Quant Time: Mar 11 23:39:42 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM031022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Mar 11 02:00:05 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.954	152	147220	20.000	ng/u1	0.00
20) Naphthalene-d8	10.766	136	582553	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.595	164	371044	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.330	188	720771	20.000	ng/u1	0.00
79) Chrysene-d12	21.501	240	698468	20.000	ng/u1	0.00
88) Perylene-d12	23.918	264	751731	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.325	96	21667	6.424	ng/uL	0.00
4) Pyridine-d5	3.743	84	264651	32.959	ng/u1	0.00
7) Phenol-d5	7.107	99	394447	40.037	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.272	67	228775	39.582	ng/u1	0.00
11) 2-Chlorophenol-d4	7.478	132	352846	41.441	ng/u1	0.00
15) 4-Methylphenol-d8	8.660	113	341045	40.475	ng/u1	0.00
21) Nitrobenzene-d5	9.113	128	175340	41.961	ng/u1	0.00
24) 2-Nitrophenol-d4	9.842	143	199004	42.948	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.384	165	385157	41.464	ng/u1	0.00
31) 4-Chloroaniline-d4	10.895	131	404927	35.052	ng/u1	0.00
46) Dimethylphthalate-d6	14.001	166	1063551	39.394	ng/u1	0.00
49) Acenaphthylene-d8	14.289	160	1370128	39.957	ng/u1	0.00
54) 4-Nitrophenol-d4	14.772	143	156647	37.250	ng/u1	0.00
60) Fluorene-d10	15.583	176	920570	38.952	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.689	200	177805	38.990	ng/u1	0.00
73) Anthracene-d10	17.430	188	1367286	40.163	ng/u1	0.00
81) Pyrene-d10	19.718	212	1573815	41.389	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.765	264	1525495	41.133	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.355	88	46729	13.984	ng/uL	94
5) Pyridine	3.761	79	308278	37.046	ng/u1	98
6) Benzaldehyde	7.078	77	233310	45.502	ng/u1	98
8) Phenol	7.131	94	373338	38.405	ng/u1	97
10) Bis(2-Chloroethyl)ether	7.366	93	303490	37.449	ng/u1	100
12) 2-Chlorophenol	7.513	128	334435	38.931	ng/u1	100
13) 2-Methylphenol	8.390	108	291569	37.481	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.490	45	282775	37.839	ng/u1	99
16) Acetophenone	8.778	105	482669	36.481	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.766	70	240020	38.555	ng/u1	97
18) 4-Methylphenol	8.725	108	318212	37.944	ng/u1	100
19) Hexachloroethane	9.043	117	142403	36.756	ng/u1	95
22) Nitrobenzene	9.154	77	354240	39.058	ng/u1	100
23) Isophorone	9.690	82	645695	37.668	ng/u1	100
25) 2-Nitrophenol	9.872	139	191025	40.260	ng/u1	100
26) 2,4-Dimethylphenol	9.931	107	367223	37.458	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.172	93	397034	37.652	ng/u1	99
29) 2,4-Dichlorophenol	10.407	162	346468	38.650	ng/u1	99
30) Naphthalene	10.813	128	1064555	36.591	ng/u1	99
32) 4-Chloroaniline	10.919	127	370197	32.220	ng/u1	99
33) Hexachlorobutadiene	11.113	225	264471	35.706	ng/u1	99
34) Caprolactam	11.678	113	91782	36.597	ng/u1	98
35) 4-Chloro-3-methylphenol	12.042	107	319005	38.361	ng/u1	97
36) 2-Methylnaphthalene	12.425	142	736682	36.871	ng/u1	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM031122\
 Data File : BM034217.D
 Acq On : 11 Mar 2022 11:58
 Operator : CG/JU
 Sample : N1816-09MSD
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DBSD4MSD

Quant Time: Mar 11 23:39:42 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM031022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Mar 11 02:00:05 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.642	142	752967	36.897	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.795	216	460836	35.998	ng/ul	99
40) Hexachlorocyclopentadiene	12.778	237	305662	33.468	ng/ul	95
41) 2,4,6-Trichlorophenol	13.031	196	286316	38.170	ng/ul	94
42) 2,4,5-Trichlorophenol	13.095	196	310056	38.139	ng/ul	98
43) 1,1'-Biphenyl	13.430	154	1015476	36.844	ng/ul	99
44) 2-Chloronaphthalene	13.472	162	792843	36.394	ng/ul	100
45) 2-Nitroaniline	13.666	65	174960	40.492	ng/ul	97
47) Dimethylphthalate	14.048	163	952178	36.126	ng/ul	99
48) 2,6-Dinitrotoluene	14.160	165	200789	39.070	ng/ul	98
50) Acenaphthylene	14.313	152	1231925	36.673	ng/ul	98
51) 3-Nitroaniline	14.489	138	154608	34.551	ng/ul	96
52) Acenaphthene	14.654	153	801197	36.364	ng/ul	99
53) 2,4-Dinitrophenol	14.689	184	109257	32.698	ng/ul	97
55) 4-Nitrophenol	14.789	109	125381	35.826	ng/ul	98
56) Dibenzofuran	14.989	168	1152761	36.126	ng/ul	100
57) 2,4-Dinitrotoluene	14.942	165	276077	39.040	ng/ul	98
58) 2,3,4,6-Tetrachlorophenol	15.213	232	257127	36.994	ng/ul	99
59) Diethylphthalate	15.407	149	914199	36.303	ng/ul	100
61) Fluorene	15.636	166	938481	36.165	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.630	204	508143	36.117	ng/ul	100
63) 4-Nitroaniline	15.642	138	149790	39.543	ng/ul	96
66) 4,6-Dinitro-2-methylph...	15.707	198	162244	36.398	ng/ul#	97
67) N-Nitrosodiphenylamine	15.842	169	787500	37.777	ng/ul	99
68) 4-Bromophenyl-phenylether	16.524	248	306100	37.694	ng/ul	99
69) Hexachlorobenzene	16.642	284	371838	36.709	ng/ul	99
70) Atrazine	16.789	200	297526	36.223	ng/ul	99
71) Pentachlorophenol	16.983	266	222108	36.454	ng/ul	98
72) Phenanthrene	17.371	178	1410724	36.761	ng/ul	99
74) Anthracene	17.466	178	1406566	36.932	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.395	216	470084	36.886	ng/ul	99
76) Pentachlorobenzene	14.913	250	462356	38.566	ng/ul	100
77) Carbazole	17.730	167	1115894	36.053	ng/ul	99
78) Di-n-butylphthalate	18.301	149	1425020	38.507	ng/ul	100
80) Fluoranthene	19.383	202	1636898	37.833	ng/ul	100
82) Pyrene	19.748	202	1664622	37.075	ng/ul	99
83) Butylbenzylphthalate	20.636	149	601028	39.989	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.418	252	434411	31.373	ng/ul	98
85) Benzo(a)anthracene	21.483	228	1539091	37.704	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.418	149	872421	39.028	ng/ul	99
87) Chrysene	21.542	228	1584655	36.059	ng/ul	99
89) Di-n-octyl phthalate	22.348	149	1469039	36.470	ng/ul	100
90) Benzo(b)fluoranthene	23.183	252	1654449	37.540	ng/ul	100
91) Benzo(k)fluoranthene	23.230	252	1675088	37.289	ng/ul	99
93) Benzo(a)pyrene	23.818	252	1659574	37.754	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.447	276	1863882	38.780	ng/ul	99
95) Dibenzo(a,h)anthracene	26.465	278	1546078	38.395	ng/ul	98
96) Benzo(g,h,i)perylene	27.218	276	1655338	37.036	ng/ul	99





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 13.984 ng/uL
RT: 3.355 min Scan# 20
Delta R.T. -0.005 min
Lab File: BM034217.D
Acq: 11 Mar 2022 11:58




Instrument :
BNA_M
ClientSampleId :
DBSD4MSD


Tgt Ion: 88 Resp: 46729
Ion Ratio Lower Upper
88 100
43 28.7 21.0 31.6
58 64.3 47.8 71.8



#5
Pyridine
Concen: 37.046 ng/uL
RT: 3.761 min Scan# 89
Delta R.T. -0.006 min
Lab File: BM034217.D
Acq: 11 Mar 2022 11:58

Tgt Ion: 79 Resp: 308278
Ion Ratio Lower Upper
79 100
52 61.4 47.7 71.5
51 32.5 25.6 38.4









#6
Benzaldehyde
Concen: 45.502 ng/ul
RT: 7.078 min Scan# 61
Delta R.T. -0.006 min
Lab File: BM034217.D
Acq: 11 Mar 2022 11:58




Instrument :
BNA_M
ClientSampleId :
DBSD4MSD


Tgt Ion: 77 Resp: 233310
Ion Ratio Lower Upper
77 100
105 104.0 80.8 121.2
106 96.7 76.6 115.0



#8
Phenol
Concen: 38.405 ng/ul
RT: 7.131 min Scan# 662
Delta R.T. 0.000 min
Lab File: BM034217.D
Acq: 11 Mar 2022 11:58

Tgt Ion: 94 Resp: 373338
Ion Ratio Lower Upper
94 100
65 28.4 22.9 34.3
66 39.1 33.7 50.5









#10
Bis(2-Chloroethyl)ether
Concen: 37.449 ng/ul
RT: 7.366 min Scan# 702
Delta R.T. 0.000 min
Lab File: BM034217.D
Acq: 11 Mar 2022 11:58




Instrument :
BNA_M
ClientSampleId :
DBSD4MSD


Tgt Ion: 93 Resp: 303490
Ion Ratio Lower Upper
93 100
63 68.4 54.7 82.1
95 32.8 26.6 39.8



#12
2-Chlorophenol
Concen: 38.931 ng/ul
RT: 7.513 min Scan# 727
Delta R.T. 0.000 min
Lab File: BM034217.D
Acq: 11 Mar 2022 11:58

Tgt Ion:128 Resp: 334435
Ion Ratio Lower Upper
128 100
64 38.5 30.6 46.0
130 32.3 25.8 38.6









#13
2-Methylphenol
Concen: 37.481 ng/ul
RT: 8.390 min Scan# 81
Delta R.T. -0.006 min
Lab File: BM034217.D
Acq: 11 Mar 2022 11:58




Instrument :
BNA_M
ClientSampleId :
DBSD4MSD


Tgt Ion:108 Resp: 291569
Ion Ratio Lower Upper
108 100
107 91.3 71.9 107.9



#14
2,2'-oxybis(1-Chloropropane)
Concen: 37.839 ng/ul
RT: 8.490 min Scan# 893
Delta R.T. 0.000 min
Lab File: BM034217.D
Acq: 11 Mar 2022 11:58

Tgt Ion: 45 Resp: 282775
Ion Ratio Lower Upper
45 100
77 21.1 17.1 25.7
79 16.3 12.8 19.2









#16
 Acetophenone
 Concen: 36.481 ng/ul
 RT: 8.778 min Scan# 941
 Delta R.T. 0.000 min
 Lab File: BM034217.D
 Acq: 11 Mar 2022 11:58




Instrument :
 BNA_M
 ClientSampleId :
 DBSD4MSD


Tgt Ion:105 Resp: 482669
 Ion Ratio Lower Upper
 105 100
 77 74.8 59.4 89.2
 51 23.1 19.3 28.9



#17
 N-Nitroso-di-n-propylamine
 Concen: 38.555 ng/ul
 RT: 8.766 min Scan# 940
 Delta R.T. 0.000 min
 Lab File: BM034217.D
 Acq: 11 Mar 2022 11:58

Tgt Ion: 70 Resp: 240020
 Ion Ratio Lower Upper
 70 100
 42 48.9 40.6 60.8
 101 10.6 8.2 12.2
 130 22.6 19.9 29.9





#18

4-Methylphenol

Concen: 37.944 ng/ul

RT: 8.725 min Scan# 91

Delta R.T. 0.000 min

Lab File: BM034217.D

Acq: 11 Mar 2022 11:58

Instrument :

BNA_M

ClientSampleId :





DBSD4MSD

Tgt Ion:108 Resp: 318212

Ion Ratio Lower Upper

108 100

107 111.2 89.0 133.4



#19

Hexachloroethane

Concen: 36.756 ng/ul

RT: 9.043 min Scan# 987

Delta R.T. 0.000 min

Lab File: BM034217.D

Acq: 11 Mar 2022 11:58




Tgt Ion:117 Resp: 142403


Ion Ratio Lower Upper

117 100

201 120.0 89.9 134.9

199 73.6 57.2 85.8









#22
Nitrobenzene
Concen: 39.058 ng/ul
RT: 9.154 min Scan# 1006
Delta R.T. 0.000 min
Lab File: BM034217.D
Acq: 11 Mar 2022 11:58




Instrument :
BNA_M
ClientSampleId :
DBSD4MSD


Tgt Ion: 77 Resp: 354240
Ion Ratio Lower Upper
77 100
123 47.6 38.2 57.2
65 13.3 10.6 15.8



#23
Isophorone
Concen: 37.668 ng/ul
RT: 9.690 min Scan# 1097
Delta R.T. 0.000 min
Lab File: BM034217.D
Acq: 11 Mar 2022 11:58

Tgt Ion: 82 Resp: 645695
Ion Ratio Lower Upper
82 100
95 8.3 6.8 10.2
138 20.5 16.5 24.7









#25
2-Nitrophenol
Concen: 40.260 ng/ul
RT: 9.872 min Scan# 1128
Delta R.T. 0.000 min
Lab File: BM034217.D
Acq: 11 Mar 2022 11:58




Instrument :
BNA_M
ClientSampleId :
DBSD4MSD


Tgt Ion:139 Resp: 191025
Ion Ratio Lower Upper
139 100
65 46.7 37.8 56.6
109 33.2 26.5 39.7



#26
2,4-Dimethylphenol
Concen: 37.458 ng/ul
RT: 9.931 min Scan# 1138
Delta R.T. -0.006 min
Lab File: BM034217.D
Acq: 11 Mar 2022 11:58

Tgt Ion:107 Resp: 367223
Ion Ratio Lower Upper
107 100
121 50.2 40.8 61.2
122 84.9 69.0 103.6









#27
 Bis(2-Chloroethoxy)methane
 Concen: 37.652 ng/ul
 RT: 10.172 min Scan# 1179
 Delta R.T. 0.000 min
 Lab File: BM034217.D
 Acq: 11 Mar 2022 11:58




Instrument :
 BNA_M
 ClientSampleId :
 DBSD4MSD


Tgt Ion: 93 Resp: 397034
 Ion Ratio Lower Upper
 93 100
 95 32.1 25.4 38.2
 123 12.0 9.5 14.3



#29
 2,4-Dichlorophenol
 Concen: 38.650 ng/ul
 RT: 10.407 min Scan# 1219
 Delta R.T. 0.000 min
 Lab File: BM034217.D
 Acq: 11 Mar 2022 11:58

Tgt Ion: 162 Resp: 346468
 Ion Ratio Lower Upper
 162 100
 164 64.9 52.8 79.2
 98 32.0 26.0 39.0









#30
Naphthalene
Concen: 36.591 ng/u1
RT: 10.813 min Scan# 1288
Delta R.T. -0.006 min
Lab File: BM034217.D
Acq: 11 Mar 2022 11:58




Instrument :
BNA_M
ClientSampleId :
DBSD4MSD


Tgt Ion:128 Resp: 1064555
Ion Ratio Lower Upper
128 100
129 11.0 8.7 13.1
127 12.9 9.9 14.9



#32
4-Chloroaniline
Concen: 32.220 ng/u1
RT: 10.919 min Scan# 1306
Delta R.T. 0.000 min
Lab File: BM034217.D
Acq: 11 Mar 2022 11:58

Tgt Ion:127 Resp: 370197
Ion Ratio Lower Upper
127 100
129 31.9 25.1 37.7









#33
Hexachlorobutadiene
Concen: 35.706 ng/ul
RT: 11.113 min Scan# 11
Delta R.T. 0.000 min
Lab File: BM034217.D
Acq: 11 Mar 2022 11:58




Instrument :
BNA_M
ClientSampleId :
DBSD4MSD


Tgt Ion:225 Resp: 264471
Ion Ratio Lower Upper
225 100
223 62.5 50.6 75.8
227 63.7 50.3 75.5



#34
Caprolactam
Concen: 36.597 ng/ul
RT: 11.678 min Scan# 1435
Delta R.T. 0.000 min
Lab File: BM034217.D
Acq: 11 Mar 2022 11:58

Tgt Ion:113 Resp: 91782
Ion Ratio Lower Upper
113 100
55 122.4 100.9 151.3
56 102.2 81.0 121.4









#35
4-Chloro-3-methylphenol
Concen: 38.361 ng/ul
RT: 12.042 min Scan# 1496
Delta R.T. 0.000 min
Lab File: BM034217.D
Acq: 11 Mar 2022 11:58




Instrument :
BNA_M
ClientSampleId :
DBSD4MSD


Tgt Ion:107 Resp: 319005
Ion Ratio Lower Upper
107 100
144 28.2 23.0 34.4
142 83.5 69.8 104.8



#36
2-Methylnaphthalene
Concen: 36.871 ng/ul
RT: 12.425 min Scan# 1562
Delta R.T. 0.000 min
Lab File: BM034217.D
Acq: 11 Mar 2022 11:58

Tgt Ion:142 Resp: 736682
Ion Ratio Lower Upper
142 100
141 90.5 71.4 107.0









#37
1-Methylnaphthalene
Concen: 36.897 ng/ul
RT: 12.642 min Scan# 1599
Delta R.T. -0.006 min
Lab File: BM034217.D
Acq: 11 Mar 2022 11:58




Instrument :
BNA_M
ClientSampleId :
DBSD4MSD


Tgt Ion:142 Resp: 752967
Ion Ratio Lower Upper
142 100
141 92.6 74.5 111.7
116 3.7 3.0 4.6



#39
1,2,4,5-Tetrachlorobenzene
Concen: 35.998 ng/ul
RT: 12.795 min Scan# 1625
Delta R.T. 0.000 min
Lab File: BM034217.D
Acq: 11 Mar 2022 11:58

Tgt Ion:216 Resp: 460836
Ion Ratio Lower Upper
216 100
214 76.7 62.2 93.2
179 20.3 15.4 23.2
108 14.5 11.8 17.6









#40
Hexachlorocyclopentadiene
Concen: 33.468 ng/ul
RT: 12.778 min Scan# 1622
Delta R.T. -0.006 min
Lab File: BM034217.D
Acq: 11 Mar 2022 11:58




Instrument :
BNA_M
ClientSampleId :
DBSD4MSD


Tgt Ion:237 Resp: 305662
Ion Ratio Lower Upper
237 100
235 61.9 52.9 79.3
272 13.0 11.8 17.8



#41
2,4,6-Trichlorophenol
Concen: 38.170 ng/ul
RT: 13.031 min Scan# 1665
Delta R.T. 0.000 min
Lab File: BM034217.D
Acq: 11 Mar 2022 11:58

Tgt Ion:196 Resp: 286316
Ion Ratio Lower Upper
196 100
198 94.0 80.8 121.2
200 31.2 26.5 39.7









#42
2,4,5-Trichlorophenol
Concen: 38.139 ng/ul
RT: 13.095 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM034217.D
Acq: 11 Mar 2022 11:58




Instrument :
BNA_M
ClientSampleId :
DBSD4MSD


Tgt Ion:196 Resp: 310056
Ion Ratio Lower Upper
196 100
198 94.5 78.0 117.0
200 30.2 24.5 36.7



#43
1,1'-Biphenyl
Concen: 36.844 ng/ul
RT: 13.430 min Scan# 1733
Delta R.T. 0.000 min
Lab File: BM034217.D
Acq: 11 Mar 2022 11:58

Tgt Ion:154 Resp: 1015476
Ion Ratio Lower Upper
154 100
153 39.9 32.6 49.0
76 9.6 7.7 11.5





#44

2-Chloronaphthalene

Concen: 36.394 ng/ul

RT: 13.472 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM034217.D

Acq: 11 Mar 2022 11:58

Instrument :

BNA_M

ClientSampleId :

DBSD4MSD





Tgt Ion:162 Resp: 792843

Ion Ratio Lower Upper

162 100

164 33.2 26.6 40.0

127 37.1 29.6 44.4



#45

2-Nitroaniline

Concen: 40.492 ng/ul

RT: 13.666 min Scan# 1773

Delta R.T. 0.000 min

Lab File: BM034217.D

Acq: 11 Mar 2022 11:58




Tgt Ion: 65 Resp: 174960


Ion Ratio Lower Upper

65 100

92 71.7 57.4 86.2

138 123.6 94.8 142.2









#47
Dimethylphthalate
Concen: 36.126 ng/ul
RT: 14.048 min Scan# 1838
Delta R.T. 0.000 min
Lab File: BM034217.D
Acq: 11 Mar 2022 11:58




Instrument :
BNA_M
ClientSampleId :
DBSD4MSD


Tgt Ion:163 Resp: 952178
Ion Ratio Lower Upper
163 100
194 4.3 3.4 5.2
164 10.3 7.8 11.8



#48
2,6-Dinitrotoluene
Concen: 39.070 ng/ul
RT: 14.160 min Scan# 1857
Delta R.T. 0.000 min
Lab File: BM034217.D
Acq: 11 Mar 2022 11:58

Tgt Ion:165 Resp: 200789
Ion Ratio Lower Upper
165 100
89 53.2 44.2 66.4
121 20.3 16.7 25.1









#50
Acenaphthylene
Concen: 36.673 ng/ul
RT: 14.313 min Scan# 1883
Delta R.T. -0.006 min
Lab File: BM034217.D
Acq: 11 Mar 2022 11:58




Instrument :
BNA_M
ClientSampleId :
DBSD4MSD


Tgt Ion:152 Resp: 1231925
Ion Ratio Lower Upper
152 100
151 18.9 15.8 23.8
153 13.3 10.2 15.2



#51
3-Nitroaniline
Concen: 34.551 ng/ul
RT: 14.489 min Scan# 1913
Delta R.T. 0.000 min
Lab File: BM034217.D
Acq: 11 Mar 2022 11:58

Tgt Ion:138 Resp: 154608
Ion Ratio Lower Upper
138 100
108 10.4 8.3 12.5
92 103.9 87.0 130.4









#52
Acenaphthene
Concen: 36.364 ng/ul
RT: 14.654 min Scan# 1941
Delta R.T. -0.006 min
Lab File: BM034217.D
Acq: 11 Mar 2022 11:58




Instrument :
BNA_M
ClientSampleId :
DBSD4MSD


Tgt Ion:153 Resp: 801197
Ion Ratio Lower Upper
153 100
152 47.1 37.4 56.0
154 89.3 71.0 106.4



#53
2,4-Dinitrophenol
Concen: 32.698 ng/ul
RT: 14.689 min Scan# 1947
Delta R.T. 0.000 min
Lab File: BM034217.D
Acq: 11 Mar 2022 11:58

Tgt Ion:184 Resp: 109257
Ion Ratio Lower Upper
184 100
63 58.5 43.8 65.6
154 63.0 49.9 74.9









#55
4-Nitrophenol
Concen: 35.826 ng/ul
RT: 14.789 min Scan# 1963
Delta R.T. 0.000 min
Lab File: BM034217.D
Acq: 11 Mar 2022 11:58




Instrument :
BNA_M
ClientSampleId :
DBSD4MSD


Tgt Ion:109 Resp: 125381
Ion Ratio Lower Upper
109 100
139 121.1 94.6 142.0
65 106.9 83.5 125.3



#56
Dibenzofuran
Concen: 36.126 ng/ul
RT: 14.989 min Scan# 1998
Delta R.T. 0.000 min
Lab File: BM034217.D
Acq: 11 Mar 2022 11:58

Tgt Ion:168 Resp: 1152761
Ion Ratio Lower Upper
168 100
139 37.2 29.8 44.6









#57
2,4-Dinitrotoluene
Concen: 39.040 ng/ul
RT: 14.942 min Scan# 1990
Delta R.T. -0.006 min
Lab File: BM034217.D
Acq: 11 Mar 2022 11:58




Instrument :
BNA_M
ClientSampleId :
DBSD4MSD


Tgt Ion:165 Resp: 276077
Ion Ratio Lower Upper
165 100
63 39.3 30.2 45.4
182 3.0 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 36.994 ng/ul
RT: 15.213 min Scan# 2036
Delta R.T. 0.000 min
Lab File: BM034217.D
Acq: 11 Mar 2022 11:58

Tgt Ion:232 Resp: 257127
Ion Ratio Lower Upper
232 100
131 41.0 32.2 48.4
130 2.2 1.9 2.9
166 28.7 23.0 34.4









#59
Diethylphthalate
Concen: 36.303 ng/ul
RT: 15.407 min Scan# 2069
Delta R.T. 0.000 min
Lab File: BM034217.D
Acq: 11 Mar 2022 11:58




Instrument :
BNA_M
ClientSampleId :
DBSD4MSD


Tgt Ion:149 Resp: 914199
Ion Ratio Lower Upper
149 100
177 22.7 17.9 26.9
150 12.6 10.2 15.2



#61
Fluorene
Concen: 36.165 ng/ul
RT: 15.636 min Scan# 2108
Delta R.T. 0.000 min
Lab File: BM034217.D
Acq: 11 Mar 2022 11:58

Tgt Ion:166 Resp: 938481
Ion Ratio Lower Upper
166 100
165 96.0 76.9 115.3
167 13.4 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 36.117 ng/ul
RT: 15.630 min Scan# 2107
Delta R.T. 0.000 min
Lab File: BM034217.D
Acq: 11 Mar 2022 11:58




Instrument :
BNA_M
ClientSampleId :
DBSD4MSD


Tgt Ion:204 Resp: 508143
Ion Ratio Lower Upper
204 100
206 32.5 25.5 38.3
141 55.3 44.2 66.2



#63
4-Nitroaniline
Concen: 39.543 ng/ul
RT: 15.642 min Scan# 2109
Delta R.T. -0.006 min
Lab File: BM034217.D
Acq: 11 Mar 2022 11:58

Tgt Ion:138 Resp: 149790
Ion Ratio Lower Upper
138 100
92 46.3 36.3 54.5
108 81.5 68.9 103.3









#66
4,6-Dinitro-2-methylphenol
Concen: 36.398 ng/ul
RT: 15.707 min Scan# 2119
Delta R.T. 0.000 min
Lab File: BM034217.D
Acq: 11 Mar 2022 11:58




Instrument :
BNA_M
ClientSampleId :
DBSD4MSD


Tgt Ion:198 Resp: 162244
Ion Ratio Lower Upper
198 100
182 3.5 3.8 5.6#
77 21.8 18.5 27.7



#67
N-Nitrosodiphenylamine
Concen: 37.777 ng/ul
RT: 15.842 min Scan# 2143
Delta R.T. 0.000 min
Lab File: BM034217.D
Acq: 11 Mar 2022 11:58

Tgt Ion:169 Resp: 787500
Ion Ratio Lower Upper
169 100
168 65.4 52.8 79.2
167 34.5 28.1 42.1









#68
4-Bromophenyl-phenylether
Concen: 37.694 ng/ul
RT: 16.524 min Scan# 2259
Delta R.T. 0.000 min
Lab File: BM034217.D
Acq: 11 Mar 2022 11:58




Instrument :
BNA_M
ClientSampleId :
DBSD4MSD


Tgt Ion:248 Resp: 306100
Ion Ratio Lower Upper
248 100
250 97.4 78.4 117.6
141 62.2 51.4 77.2



#69
Hexachlorobenzene
Concen: 36.709 ng/ul
RT: 16.642 min Scan# 2279
Delta R.T. 0.000 min
Lab File: BM034217.D
Acq: 11 Mar 2022 11:58

Tgt Ion:284 Resp: 371838
Ion Ratio Lower Upper
284 100
142 30.2 24.6 36.8
249 30.6 24.1 36.1





#70

Atrazine

Concen: 36.223 ng/ul

RT: 16.789 min Scan# 21

Delta R.T. 0.000 min

Lab File: BM034217.D

Acq: 11 Mar 2022 11:58

Instrument :

BNA_M

ClientSampleId :

DBSD4MSD





Tgt Ion:200 Resp: 297526

Ion Ratio Lower Upper

200 100

173 24.7 20.2 30.4

215 48.8 38.4 57.6



#71

Pentachlorophenol

Concen: 36.454 ng/ul

RT: 16.983 min Scan# 2337

Delta R.T. 0.000 min

Lab File: BM034217.D

Acq: 11 Mar 2022 11:58




Tgt Ion:266 Resp: 222108


Ion Ratio Lower Upper

266 100

264 62.8 49.3 73.9

268 65.4 50.3 75.5









#72
Phenanthrene
Concen: 36.761 ng/u1
RT: 17.371 min Scan# 2403
Delta R.T. -0.006 min
Lab File: BM034217.D
Acq: 11 Mar 2022 11:58




Instrument :
BNA_M
ClientSampleId :
DBSD4MSD


Tgt Ion:178 Resp: 1410724
Ion Ratio Lower Upper
178 100
179 15.3 11.7 17.5
176 18.8 15.0 22.4



#74
Anthracene
Concen: 36.932 ng/u1
RT: 17.466 min Scan# 2419
Delta R.T. 0.000 min
Lab File: BM034217.D
Acq: 11 Mar 2022 11:58

Tgt Ion:178 Resp: 1406566
Ion Ratio Lower Upper
178 100
179 15.5 11.9 17.9
176 18.2 14.4 21.6









#75
1,2,3,4-Tetrachlorobenzene
Concen: 36.886 ng/uL
RT: 13.395 min Scan# 1727
Delta R.T. -0.006 min
Lab File: BM034217.D
Acq: 11 Mar 2022 11:58




Instrument :
BNA_M
ClientSampleId :
DBSD4MSD


Tgt Ion:216 Resp: 470084
Ion Ratio Lower Upper
216 100
214 78.4 62.0 93.0
218 47.6 37.8 56.8



#76
Pentachlorobenzene
Concen: 38.566 ng/uL
RT: 14.913 min Scan# 1985
Delta R.T. 0.000 min
Lab File: BM034217.D
Acq: 11 Mar 2022 11:58

Tgt Ion:250 Resp: 462356
Ion Ratio Lower Upper
250 100
248 63.2 50.6 75.8
252 63.5 50.2 75.4









#77
 Carbazole
 Concen: 36.053 ng/ul
 RT: 17.730 min Scan# 2464
 Delta R.T. 0.000 min
 Lab File: BM034217.D
 Acq: 11 Mar 2022 11:58




Instrument :
 BNA_M
 ClientSampleId :
 DBSD4MSD


Tgt Ion:167 Resp: 1115894
 Ion Ratio Lower Upper
 167 100
 166 21.5 16.6 25.0
 139 12.4 9.9 14.9



#78
 Di-n-butylphthalate
 Concen: 38.507 ng/ul
 RT: 18.301 min Scan# 2561
 Delta R.T. 0.000 min
 Lab File: BM034217.D
 Acq: 11 Mar 2022 11:58

Tgt Ion:149 Resp: 1425020
 Ion Ratio Lower Upper
 149 100
 150 9.3 7.4 11.0
 104 5.4 4.2 6.2





#80

Fluoranthene

Concen: 37.833 ng/ul

RT: 19.383 min Scan# 21

Delta R.T. 0.000 min

Lab File: BM034217.D

Acq: 11 Mar 2022 11:58

Instrument :

BNA_M

ClientSampleId :

DBSD4MSD





Tgt Ion:202 Resp: 1636898

Ion Ratio Lower Upper

202 100

101 7.7 6.2 9.2

100 6.1 4.7 7.1



#82

Pyrene

Concen: 37.075 ng/ul

RT: 19.748 min Scan# 2807

Delta R.T. 0.000 min

Lab File: BM034217.D

Acq: 11 Mar 2022 11:58




Tgt Ion:202 Resp: 1664622


Ion Ratio Lower Upper

202 100

101 8.8 7.0 10.4

100 7.5 5.8 8.6





#83

Butylbenzylphthalate

Concen: 39.989 ng/ul

RT: 20.636 min Scan# 2958

Delta R.T. -0.006 min

Lab File: BM034217.D

Acq: 11 Mar 2022 11:58

Instrument :

BNA_M

ClientSampleId :

DBSD4MSD





Tgt Ion:149 Resp: 601028

Ion Ratio Lower Upper

149 100

91 65.6 50.9 76.3

206 22.9 19.3 28.9



#84

3,3'-Dichlorobenzidine

Concen: 31.373 ng/ul

RT: 21.418 min Scan# 3091

Delta R.T. 0.000 min

Lab File: BM034217.D

Acq: 11 Mar 2022 11:58




Tgt Ion:252 Resp: 434411


Ion Ratio Lower Upper

252 100

254 64.8 50.7 76.1

126 8.2 6.6 9.8









#85
Benzo(a)anthracene
Concen: 37.704 ng/ul
RT: 21.483 min Scan# 3102
Delta R.T. -0.006 min
Lab File: BM034217.D
Acq: 11 Mar 2022 11:58




Instrument :
BNA_M
ClientSampleId :
DBSD4MSD


Tgt Ion:228 Resp: 1539091
Ion Ratio Lower Upper
228 100
229 19.6 15.4 23.2
226 26.5 21.0 31.4



#86
Bis(2-ethylhexyl)phthalate
Concen: 39.028 ng/ul
RT: 21.418 min Scan# 3091
Delta R.T. 0.000 min
Lab File: BM034217.D
Acq: 11 Mar 2022 11:58

Tgt Ion:149 Resp: 872421
Ion Ratio Lower Upper
149 100
167 28.1 22.1 33.1
279 5.3 4.2 6.2









#87
Chrysene
Concen: 36.059 ng/ul
RT: 21.542 min Scan# 3112
Delta R.T. 0.000 min
Lab File: BM034217.D
Acq: 11 Mar 2022 11:58




Instrument :
BNA_M
ClientSampleId :
DBSD4MSD


Tgt Ion:228 Resp: 1584655
Ion Ratio Lower Upper
228 100
226 28.9 23.4 35.2
229 19.5 15.4 23.0



#89
Di-n-octyl phthalate
Concen: 36.470 ng/ul
RT: 22.348 min Scan# 3249
Delta R.T. 0.000 min
Lab File: BM034217.D
Acq: 11 Mar 2022 11:58

Tgt Ion:149 Resp: 1469039









#90
Benzo(b)fluoranthene
Concen: 37.540 ng/ul
RT: 23.183 min Scan# 3391
Delta R.T. 0.000 min
Lab File: BM034217.D
Acq: 11 Mar 2022 11:58




Instrument :
BNA_M
ClientSampleId :
DBSD4MSD


Tgt Ion:252 Resp: 1654449
Ion Ratio Lower Upper
252 100
253 21.6 17.4 26.0
125 6.9 5.8 8.6



#91
Benzo(k)fluoranthene
Concen: 37.289 ng/ul
RT: 23.230 min Scan# 3399
Delta R.T. -0.006 min
Lab File: BM034217.D
Acq: 11 Mar 2022 11:58

Tgt Ion:252 Resp: 1675088
Ion Ratio Lower Upper
252 100
253 22.2 17.5 26.3
125 7.3 5.1 7.7









#93
Benzo(a)pyrene
Concen: 37.754 ng/ul
RT: 23.818 min Scan# 3498
Delta R.T. 0.000 min
Lab File: BM034217.D
Acq: 11 Mar 2022 11:58




Instrument :
BNA_M
ClientSampleId :
DBSD4MSD


Tgt Ion:252 Resp: 1659574
Ion Ratio Lower Upper
252 100
253 21.5 17.5 26.3
125 7.8 6.2 9.4



#94
Indeno(1,2,3-cd)pyrene
Concen: 38.780 ng/ul
RT: 26.447 min Scan# 3946
Delta R.T. 0.000 min
Lab File: BM034217.D
Acq: 11 Mar 2022 11:58

Tgt Ion:276 Resp: 1863882
Ion Ratio Lower Upper
276 100
138 17.7 13.6 20.4
277 25.2 19.9 29.9









#95
 Dibenzo(a,h)anthracene
 Concen: 38.395 ng/ul
 RT: 26.465 min Scan# 3949
 Delta R.T. 0.000 min
 Lab File: BM034217.D
 Acq: 11 Mar 2022 11:58

Instrument :
 BNA_M
 ClientSampleId :
 DBSD4MSD

Tgt Ion:278 Resp: 1546078
 Ion Ratio Lower Upper
 278 100
 139 11.4 9.6 14.4
 279 23.4 19.3 28.9



#96
 Benzo(g,h,i)perylene
 Concen: 37.036 ng/ul
 RT: 27.218 min Scan# 4077
 Delta R.T. -0.006 min
 Lab File: BM034217.D
 Acq: 11 Mar 2022 11:58

Tgt Ion:276 Resp: 1655338
 Ion Ratio Lower Upper
 276 100
 138 15.4 11.8 17.6
 277 23.6 18.8 28.2

