

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM031215\
 Data File : BM000699.D
 Acq On : 12 Mar 2015 15:06
 Operator : TP/IZ
 Sample : SSTD0.473
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD0.473

Quant Time: Mar 13 09:02:04 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-SIM-BM021815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 13 08:51:18 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.65	152	4500	0.40	ng/ul	0.00
2) Naphthalene-d8	10.43	136	15944	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.31	164	2495	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.05	188	20037	0.40	ng/ul	0.00
16) Chrysene-d12	21.25	240	18247	0.40	ng/ul	0.00
20) Perylene-d12	23.44	264	16290	0.40	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Methylnaphthalene-d10	12.03	152	8600	0.40	ng/ul	0.00
14) Fluoranthene-d10	19.08	212	18319	0.42	ng/ul	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) Naphthalene	10.48	128	17263	0.42	ng/ul	100
5) 2-Methylnaphthalene	12.11	142	11664	0.40	ng/ul	100
7) Acenaphthylene	14.02	152	16114	0.95	ng/ul	100
8) Acenaphthene	14.36	153	12516	0.96	ng/ul	100
9) Fluorene	15.35	166	15238	0.95	ng/ul	100
11) Pentachlorophenol	16.71	266	945	0.18	ng/ul	100
12) Phenanthrene	17.08	178	24369	0.42	ng/ul	100
13) Anthracene	17.17	178	22065	0.40	ng/ul	100
15) Fluoranthene	19.11	202	27861	0.42	ng/ul	100
17) Pyrene	19.48	202	28387	0.34	ng/ul	100
18) Benzo(a)anthracene	21.23	228	22798	0.39	ng/ul	100
19) Chrysene	21.28	228	27016	0.42	ng/ul	100
21) Benzo(b)fluoranthene	22.78	252	24929	0.40	ng/ul	100
22) Benzo(k)fluoranthene	22.82	252	24918	0.41	ng/ul	100
23) Benzo(a)pyrene	23.34	252	22915	0.41	ng/ul	100
24) Indeno(1,2,3-cd)pyrene	25.65	276	24265	0.45	ng/ul	100
25) Dibenzo(a,h)anthracene	25.66	278	19219	0.45	ng/ul	100
26) Benzo(g,h,i)perylene	26.31	276	21597	0.46	ng/ul	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM031215\
Data File : BM000699.D
Acq On : 12 Mar 2015 15:06
Operator : TP/IZ
Sample : SSTD0.473
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD0.473

Quant Time: Mar 13 09:02:04 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-SIM-BM021815.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 13 08:51:18 2015
Response via : Initial Calibration

