

Data Path : Z:\HPCHEM1\BNA M\DATA\BM031218\
 Data File : BM014528.D
 Acq On : 12 Mar 2018 21:34
 Operator : SJ/JU
 Sample : J1813-03
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 F6F18

Manual Integrations
 APPROVED

Sohil
 3/13/2018 4:24:09 PM

Quant Time: Mar 13 04:35:31 2018
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-BM021418.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Mar 13 02:49:59 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.47	152	66946	20.00	ng/ul	0.00
18) Naphthalene-d8	10.25	136	295406	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.14	164	181959	20.00	ng/ul	0.00
61) Phenanthrene-d10	16.90	188	448425	20.00	ng/ul	0.00
75) Chrysene-d12	21.13	240	516105	20.00	ng/ul	0.00
83) Perylene-d12	23.29	264	515621	20.00	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.05	96	4425	2.83	ng/uL	0.00
5) Phenol-d5	6.69	99	139491	24.65	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.83	67	90109	26.31	ng/ul	0.00
9) 2-Chlorophenol-d4	7.02	132	119941	27.64	ng/ul	0.00
13) 4-Methylphenol-d8	8.21	113	106743	21.56	ng/ul	0.00
19) Nitrobenzene-d5	8.65	128	62890	28.81	ng/ul	0.00
22) 2-Nitrophenol-d4	9.36	143	72965	32.29	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	9.90	165	112570	24.05	ng/ul	0.00
29) 4-Chloroaniline-d4	10.42	131	156595	33.96	ng/ul	0.00
43) Dimethylphthalate-d6	13.57	166	525215	34.95	ng/ul	0.00
46) Acenaphthylene-d8	13.83	160	600019	32.38	ng/ul	0.00
51) 4-Nitrophenol-d4	14.44	143	64860m	31.26	ng/ul	-0.02
57) Fluorene-d10	15.14	176	422637	31.39	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.33	200	62886	23.38	ng/ul	0.00
70) Anthracene-d10	17.00	188	703511m	32.52	ng/ul	0.00
76) Pyrene-d10	19.32	212	814073	30.58	ng/ul	0.00
87) Benzo(a)pyrene-d12	23.16	264	819345	32.65	ng/ul	0.00
Target Compounds						
44) Dimethylphthalate	13.62	163	70703	4.772	ng/ul#	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA M\DATA\BM031218\
Data File : BM014528.D
Acq On : 12 Mar 2018 21:34
Operator : SJ/JU
Sample : J1813-03
Misc :
ALS Vial : 12 Sample Multiplier: 1




Instrument :
BNA_M
ClientSampleId :
F6F18

Manual Integrations
APPROVED

Sohil
3/13/2018 4:24:09 PM

Quant Time: Mar 13 04:35:31 2018
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-BM021418.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Mar 13 02:49:59 2018
Response via : Initial Calibration





#44
 Dimethylphthalate
 Concen: 4.772 ng/ul
 RT: 13.62 min Scan# 1824
 Delta R.T. -0.00 min
 Lab File: BM014528.D
 Acq: 12 Mar 2018 21:34

Instrument :
 BNA_M
 ClientSampleId :
 F6F18

Tot Ion	Ratio	Resp Lower	Upper
163	100		
194	5.9	3.5	5.3#
164	11.2	8.2	12.4

Manual Integrations
 APPROVED

Sohil
 3/13/2018 4:24:09 PM

