

Data Path : Z:\HPCHEM1\BNA M\DATA\BM031218\
 Data File : BM014532.D
 Acq On : 12 Mar 2018 23:59
 Operator : SJ/JU
 Sample : J1813-14
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 F6F10

Manual Integrations
 APPROVED

Sohil
 3/13/2018 4:24:19 PM

Quant Time: Mar 13 05:03:27 2018
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-BM021418.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Mar 13 02:49:59 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.47	152	91507	20.00	ng/ul	0.00
18) Naphthalene-d8	10.25	136	389498	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.14	164	230891	20.00	ng/ul	0.00
61) Phenanthrene-d10	16.90	188	559683	20.00	ng/ul	0.00
75) Chrysene-d12	21.13	240	589668	20.00	ng/ul	0.00
83) Perylene-d12	23.29	264	496169	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.06	96	6987m	3.27	ng/uL	0.00
5) Phenol-d5	6.68	99	211033	27.29	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.83	67	130879	27.96	ng/ul	0.00
9) 2-Chlorophenol-d4	7.02	132	171488	28.92	ng/ul	0.00
13) 4-Methylphenol-d8	8.21	113	167793	24.79	ng/ul	0.00
19) Nitrobenzene-d5	8.64	128	90877	31.57	ng/ul	0.00
22) 2-Nitrophenol-d4	9.36	143	104109	34.95	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	9.90	165	176332	28.58	ng/ul	0.00
29) 4-Chloroaniline-d4	10.42	131	169539	27.89	ng/ul	0.00
43) Dimethylphthalate-d6	13.56	166	666402	34.95	ng/ul	0.00
46) Acenaphthylene-d8	13.83	160	820801	34.91	ng/ul	0.00
51) 4-Nitrophenol-d4	14.43	143	78321	29.75	ng/ul	-0.03
57) Fluorene-d10	15.14	176	563650	33.00	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.32	200	87371	26.02	ng/ul	0.00
70) Anthracene-d10	17.00	188	863653	31.98	ng/ul	0.00
76) Pyrene-d10	19.32	212	997408	32.79	ng/ul	0.00
87) Benzo(a)pyrene-d12	23.16	264	815000	33.75	ng/ul	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
44) Dimethylphthalate	13.62	163	142440	7.576	ng/ul#	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA M\DATA\BM031218\
Data File : BM014532.D
Acq On : 12 Mar 2018 23:59
Operator : SJ/JU
Sample : J1813-14
Misc :
ALS Vial : 16 Sample Multiplier: 1


Instrument :
BNA_M
ClientSampleId :
F6F10

Quant Time: Mar 13 05:03:27 2018
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-BM021418.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Mar 13 02:49:59 2018
Response via : Initial Calibration

Manual Integrations
APPROVED

Sohil
3/13/2018 4:24:19 PM





#44
 Dimethylphthalate
 Concen: 7.576 ng/ul
 RT: 13.62 min Scan# 1824
 Delta R.T. -0.00 min
 Lab File: BM014532.D
 Acq: 12 Mar 2018 23:59

Instrument :
 BNA_M
 ClientSampleId :
 F6F10

Tot Ion	Ratio	Resp Lower	Upper
163	100		
194	5.4	3.5	5.3#
164	11.3	8.2	12.4

Manual Integrations
 APPROVED

Sohil
 3/13/2018 4:24:19 PM

