


Data Path : Z:\HPCHEM1\BNA_M\Data\BM031218\
Data File : BM014516.D
Acq On : 12 Mar 2018 12:59
Operator : SJ/JU
Sample : DFTPP10
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP10

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM021418.M
Title : SVOA CALIBRATION
Last Update : Tue Mar 13 02:49:59 2018



AutoFind: Scans 2458, 2459, 2460; Background Corrected with Scan 2453

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	52.9	73984	PASS
68	69	0.00	2	1.7	1237	PASS
69	198	0.00	100	51.9	72666	PASS
70	69	0.00	2	0.4	299	PASS
127	198	10	80	53.5	74893	PASS
197	198	0.00	2	0.8	1142	PASS
198	198	100	100	100.0	139922	PASS
199	198	5	9	6.3	8829	PASS
275	198	10	60	28.9	40477	PASS
365	198	1	100	5.5	7722	PASS
441	443	0.01	100	74.8	16048	PASS
442	198	50	100	75.0	104893	PASS
443	442	15	24	20.4	21448	PASS

m/z	Abundance
37.10	565.0
38.00	891.0
39.10	5357.0
40.20	379.0
41.00	306.0
44.00	869.0
49.00	468.0
50.10	23920.0
51.10	82512.0
52.00	5230.0
54.90	342.0
56.00	3318.0
57.00	5409.0
58.00	333.0
61.10	1055.0
62.10	1082.0
63.10	2783.0
64.20	356.0
65.00	1155.0
68.10	1257.0
69.00	74376.0
69.90	599.0
73.00	1534.0
74.00	10196.0
75.00	12988.0
76.00	5769.0
77.10	90656.0
78.00	5695.0
79.00	7342.0
80.00	4691.0
81.00	8092.0
81.80	809.0
82.10	855.0
83.00	1641.0
83.90	384.0
84.90	932.0
86.00	2391.0
86.90	1651.0
91.00	2213.0
92.00	1937.0
93.00	15486.0
93.90	1106.0
95.80	631.0
98.00	12736.0
99.00	7821.0
99.90	792.0
101.00	3570.0
103.00	1298.0
104.10	2254.0
105.00	3197.0
106.00	848.0
107.00	29432.0
108.00	4542.0
108.90	700.0
110.00	37040.0
111.00	7905.0
111.90	1131.0
112.90	469.0
116.00	1743.0
117.00	40536.0
118.00	2701.0
119.90	381.0
122.00	1926.0
123.00	4276.0
124.00	1223.0
124.90	1361.0
126.00	328.0
127.00	78384.0
128.00	6397.0
129.00	35824.0
130.00	2214.0
131.10	583.0
131.90	439.0
132.90	468.0
133.80	981.0
135.00	3122.0
136.00	1170.0
137.00	2045.0
138.00	392.0
139.00	468.0
140.20	410.0
140.90	5585.0
141.90	1848.0
143.00	1321.0
143.80	588.0
145.10	343.0
146.10	868.0
147.00	2651.0
148.00	7928.0
149.00	1333.0
150.10	595.0

Instrument :

BNA_M

ClientSampleId :

DFTPP10

150.80	434.0
151.30	560.0
152.10	307.0
152.90	1497.0
154.00	1155.0
155.10	1945.0
156.00	4659.0
158.00	1898.0
159.10	836.0
160.00	2118.0
161.10	2143.0
161.90	876.0
165.00	2535.0
166.00	1542.0
167.00	11452.0
168.00	7708.0
169.00	713.0
170.00	352.0
170.90	375.0
171.90	1061.0
173.10	1263.0
174.00	3077.0
175.00	4004.0
176.00	935.0
177.10	1313.0
177.90	335.0
179.00	6606.0
179.90	4259.0
181.00	2293.0
183.80	1252.0
185.00	3045.0
186.00	23896.0
187.10	6934.0
187.80	609.0
189.00	2809.0
189.90	390.0
190.90	1075.0
192.00	2723.0
193.00	2244.0
194.10	669.0
196.00	4423.0
197.10	1283.0
198.00	135296.0
199.00	8464.0
200.00	1012.0
203.00	1374.0
204.00	7911.0
205.00	9144.0
206.00	40448.0
207.00	9190.0
208.00	2313.0
209.00	800.0
210.00	1370.0
211.00	1965.0
211.90	312.0
213.00	307.0
214.90	830.0
216.00	1258.0
217.00	15645.0
218.00	1446.0
220.90	3684.0
221.80	2483.0
223.00	2689.0
224.00	22536.0
225.00	5563.0
226.10	455.0
227.00	10914.0
227.90	1378.0
228.90	1707.0
229.90	434.0
230.90	1077.0
233.90	1219.0
235.00	1288.0
235.90	923.0
236.80	474.0
238.90	665.0
239.80	886.0
240.90	674.0
242.00	1163.0
243.00	1875.0
244.00	18848.0
245.00	2216.0
246.10	5286.0
246.90	1388.0
249.00	1157.0
253.30	333.0
254.10	747.0
255.00	89536.0
256.00	13862.0
256.90	1402.0
258.00	7124.0
259.00	1767.0
265.00	2978.0
265.80	756.0
272.90	2084.0

Instrument :
BNA_M
ClientSampleId :
DFTPP10

274.00	7242.0
275.00	36208.0
276.10	3751.0
277.00	4910.0
277.90	799.0
279.20	442.0
281.00	315.0
282.90	326.0
284.00	456.0
284.90	891.0
290.80	333.0
292.80	1029.0
294.90	492.0
296.00	14719.0
297.00	2426.0
300.80	301.0
303.10	1842.0
303.90	371.0
314.00	781.0
314.80	1662.0
315.90	730.0
320.80	899.0
322.40	617.0
323.00	3913.0
324.10	704.0
327.90	455.0
332.10	395.0
332.80	392.0
334.00	2275.0
334.90	344.0
335.10	323.0
335.90	359.0
341.00	593.0
346.00	1008.0
351.90	1130.0
353.10	603.0
354.10	337.0
365.00	6099.0
366.00	641.0
370.90	450.0
371.90	1044.0
402.00	878.0
402.90	1072.0
404.00	445.0
421.00	763.0
422.10	675.0
423.10	3757.0
424.20	1108.0
424.90	337.0
441.00	11184.0
442.10	67312.0
443.00	16752.0
444.10	1717.0

Instrument :
BNA_M
ClientSampleId :
DFTPP10

m/z	Abundance
37.10	303.0
38.00	2008.0
39.00	6356.0
39.90	712.0
44.10	726.0
49.10	904.0
50.10	27648.0
51.00	88632.0
52.00	4436.0
55.20	678.0
56.00	1973.0
57.00	5187.0
61.10	965.0
62.10	1228.0
63.00	3983.0
63.90	599.0
65.10	1281.0
67.90	1275.0
69.00	87440.0
70.10	454.0
73.10	1676.0
74.00	12764.0
75.00	16007.0
76.10	7066.0
77.00	93960.0
78.10	5839.0
79.00	10663.0
80.00	5034.0
81.00	6940.0
81.80	2286.0
82.90	1387.0
84.90	1976.0
86.10	2220.0
87.00	795.0
88.10	498.0
91.00	1941.0
92.10	2774.0
93.00	17048.0
94.10	1441.0
96.00	531.0
98.00	12028.0
99.00	9045.0
100.10	1022.0
101.00	4294.0
101.80	393.0
103.10	2155.0
103.90	2330.0
104.90	1900.0
106.10	894.0
107.00	33408.0
108.00	5284.0
109.10	951.0
110.00	47928.0
111.10	6823.0
112.10	1900.0
113.00	537.0
114.10	407.0
114.80	320.0
116.10	2056.0
117.00	51536.0
118.00	4172.0
119.80	622.0
121.90	1930.0
122.90	3739.0
123.90	1397.0
125.00	1809.0
125.90	391.0
126.20	480.0
127.00	92360.0
128.00	7670.0
129.00	42360.0
130.00	4136.0
131.10	961.0
132.20	618.0
133.00	484.0
134.00	1047.0
135.00	3776.0
135.90	1783.0
137.00	1420.0
138.00	586.0
138.80	372.0
140.10	747.0
141.00	7374.0
142.10	2163.0
142.90	1577.0
144.00	367.0
145.10	392.0
145.90	1795.0
147.00	3636.0
148.00	9960.0
149.00	2045.0

Instrument :
BNA_M
ClientSampleId :
DFTPP10

150.00	366.0
150.90	742.0
151.50	624.0
153.10	2402.0
154.10	1685.0
155.00	2537.0
156.10	3540.0
157.00	2069.0
157.90	1194.0
158.80	624.0
160.00	1974.0
160.90	2739.0
162.00	712.0
163.00	358.0
164.90	1560.0
166.00	1255.0
167.00	13521.0
168.10	7340.0
168.90	1144.0
170.00	713.0
171.00	863.0
172.00	1387.0
173.00	1153.0
173.90	2145.0
175.10	4521.0
176.10	658.0
176.90	2178.0
178.20	811.0
179.00	10538.0
180.00	5675.0
180.90	1559.0
181.90	653.0
184.00	998.0
184.90	4101.0
186.00	29512.0
187.00	9159.0
188.00	1282.0
188.90	3914.0
189.90	873.0
190.70	932.0
191.10	829.0
192.10	2707.0
193.10	3691.0
193.90	952.0
196.00	5628.0
197.10	1445.0
198.00	168896.0
199.00	10912.0
200.00	713.0
201.50	992.0
203.00	2207.0
204.00	8968.0
205.10	12926.0
206.00	59112.0
207.00	8373.0
208.10	2539.0
209.10	1092.0
209.90	815.0
210.90	2725.0
211.80	747.0
215.10	773.0
216.10	1681.0
217.00	20608.0
217.90	2138.0
221.10	4918.0
222.10	1761.0
223.00	3795.0
224.00	32728.0
225.00	8523.0
226.00	1423.0
227.00	18768.0
228.00	2438.0
228.90	2763.0
230.10	418.0
231.00	1184.0
233.00	359.0
234.10	1222.0
235.00	1549.0
236.00	847.0
237.10	1001.0
239.00	893.0
240.00	347.0
241.10	743.0
242.10	1476.0
243.00	1927.0
244.00	22464.0
245.00	3188.0
245.90	6634.0
247.00	833.0
248.00	372.0
248.70	327.0
252.00	647.0
253.00	1130.0
254.10	1367.0
255.00	118016.0

Instrument :
BNA_M
ClientSampleId :
DFTPP10

256.00	19088.0
256.90	1951.0
258.00	11082.0
258.90	1648.0
264.90	3656.0
266.00	598.0
273.10	4192.0
274.00	10818.0
275.00	48960.0
276.00	7033.0
276.90	6632.0
277.80	780.0
281.30	498.0
282.90	719.0
284.10	371.0
285.00	1069.0
286.20	300.0
289.10	443.0
293.10	1105.0
296.00	22488.0
297.00	3313.0
302.20	414.0
303.10	2619.0
303.90	394.0
307.80	313.0
310.00	359.0
314.00	906.0
315.00	2382.0
315.90	996.0
320.90	592.0
321.90	367.0
322.40	451.0
323.00	4697.0
324.00	1014.0
327.00	1396.0
327.90	538.0
332.10	356.0
333.00	605.0
334.10	3002.0
335.00	932.0
340.90	691.0
345.90	1363.0
352.10	1486.0
353.20	866.0
354.00	2196.0
358.90	304.0
365.00	8679.0
366.10	1383.0
370.90	656.0
372.10	2999.0
372.80	314.0
373.20	587.0
383.20	358.0
389.90	631.0
391.00	398.0
402.00	1019.0
403.10	1705.0
404.00	506.0
421.10	754.0
422.00	827.0
423.10	6889.0
424.10	1753.0
424.90	306.0
441.10	17640.0
442.10	133056.0
443.10	25048.0
444.00	2274.0

Instrument :
BNA_M
ClientSampleId :
DFTPP10

m/z	Abundance
38.00	756.0
39.10	3508.0
40.00	500.0
44.00	891.0
49.00	620.0
50.10	15424.0
51.00	50808.0
52.00	2216.0
56.00	1942.0
57.00	3477.0
61.00	792.0
62.00	1097.0
63.10	2440.0
65.10	1010.0
67.90	1181.0
69.00	56184.0
69.90	300.0
72.90	1285.0
74.00	8616.0
75.10	9980.0
76.00	3096.0
77.10	54592.0
78.10	3637.0
79.00	5025.0
79.90	2866.0
81.10	5201.0
81.90	1792.0
82.90	1000.0
85.00	803.0
86.00	2337.0
86.90	1054.0
91.00	940.0
92.00	2014.0
93.00	10324.0
93.90	1296.0
96.10	521.0
98.00	8204.0
99.10	5914.0
99.90	331.0
101.00	2624.0
102.80	657.0
104.00	1685.0
105.00	2640.0
105.90	749.0
107.00	20168.0
108.00	2656.0
109.20	757.0
110.00	31496.0
111.00	4281.0
112.00	391.0
116.00	1234.0
117.00	33680.0
118.10	2152.0
119.90	459.0
122.10	2264.0
123.00	2197.0
124.00	1495.0
125.00	1518.0
127.00	53936.0
127.90	4929.0
129.00	27592.0
129.90	2301.0
131.10	835.0
133.90	749.0
135.00	2757.0
136.00	840.0
136.90	1257.0
137.80	334.0
138.90	364.0
141.00	5306.0
141.60	858.0
142.10	754.0
142.90	720.0
144.10	354.0
146.00	956.0
146.90	2697.0
148.00	7136.0
149.10	957.0
149.90	352.0
151.10	827.0
152.20	317.0
153.00	1066.0
153.90	1071.0
155.10	2029.0
156.10	3397.0
157.10	997.0
158.10	486.0
158.90	593.0
160.00	1151.0
161.10	2079.0
161.90	515.0

Instrument :
BNA_M
ClientSampleId :
DFTPP10

164.90	1527.0
166.10	581.0
167.00	9996.0
168.00	5804.0
168.80	598.0
172.10	546.0
172.90	620.0
174.00	1692.0
175.00	3259.0
175.80	592.0
176.20	705.0
176.90	1970.0
177.90	733.0
178.90	6887.0
180.00	4910.0
180.90	1956.0
181.90	533.0
184.10	741.0
185.00	3012.0
186.00	21208.0
187.00	4811.0
187.90	540.0
188.20	613.0
189.00	1868.0
189.80	549.0
191.00	1004.0
192.20	1818.0
193.00	1534.0
193.20	1550.0
193.90	579.0
196.00	3466.0
197.00	699.0
198.00	115576.0
199.00	7111.0
199.80	600.0
201.70	596.0
202.90	1831.0
204.00	7405.0
205.00	12360.0
206.10	37184.0
207.10	4641.0
208.00	2237.0
209.00	876.0
210.00	667.0
211.10	1523.0
215.00	903.0
215.90	1274.0
217.00	13751.0
217.90	2055.0
220.20	616.0
220.90	3128.0
221.80	1572.0
222.20	1223.0
223.10	2679.0
224.00	25064.0
225.00	5564.0
226.00	716.0
227.00	13142.0
228.10	2064.0
229.00	2088.0
230.90	1224.0
234.10	842.0
234.90	1382.0
236.00	428.0
236.90	802.0
239.10	513.0
241.00	797.0
242.10	854.0
243.10	2065.0
244.00	15566.0
245.00	1494.0
246.00	6061.0
247.00	1131.0
249.10	322.0
252.00	734.0
253.00	443.0
253.80	605.0
255.00	88648.0
256.00	12671.0
257.00	1160.0
258.00	8659.0
259.00	770.0
265.00	3010.0
265.90	303.0
267.00	466.0
272.90	2412.0
274.00	7045.0
275.00	36264.0
276.00	5351.0
277.00	5439.0
278.00	667.0
281.10	698.0
282.80	625.0
283.80	323.0
284.90	576.0

Instrument :
BNA_M
ClientSampleId :
DFTPP10

291.80	374.0
292.90	751.0
294.00	497.0
296.00	16488.0
297.00	2091.0
301.70	338.0
303.10	1462.0
303.90	536.0
314.10	877.0
314.90	1833.0
315.90	689.0
321.10	804.0
323.10	4378.0
324.00	756.0
327.00	1009.0
328.10	387.0
333.00	666.0
334.00	3544.0
334.90	705.0
341.10	651.0
346.10	1519.0
351.90	578.0
353.20	676.0
354.00	1190.0
355.00	387.0
365.00	8388.0
366.00	1861.0
371.00	304.0
372.10	1929.0
372.80	688.0
383.00	956.0
402.10	621.0
403.00	1087.0
404.00	469.0
421.10	1131.0
421.90	764.0
423.10	6003.0
424.10	1475.0
438.80	419.0
441.10	19320.0
442.10	114312.0
443.10	22544.0
444.10	1803.0

Instrument :
BNA_M
ClientSampleId :
DFTPP10

Scan 2453 (17.315 min): BM014516.D
DFTPP10

m/z	Abundance
44.10	620.0
73.20	665.0
190.90	391.0
207.00	1315.0
281.20	432.0

Instrument :
BNA_M
ClientSampleId :
DFTPP10