

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM031219\
 Data File : BM019124.D
 Acq On : 12 Mar 2019 22:34
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 13 05:56:06 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM031119.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 11 17:30:46 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	104	0.00
2	1,4-Dioxane	0.549	0.521	5.1	104	0.00
3	Pyridine	1.495	1.617	-8.2	107	0.00
4	n-Nitrosodimethylamine	0.465	0.461	0.9	103	0.00
5 S	2-Fluorophenol	1.235	1.229	0.5	104	0.00
6	Aniline	2.332	2.388	-2.4	107	0.00
7 S	Phenol-d6	1.806	1.871	-3.6	107	0.00
8	2-Chlorophenol	1.484	1.524	-2.7	108	0.00
9	Benzaldehyde	1.136	1.185	-4.3	108	0.00
10 C	Phenol	1.947	2.032	-4.4	108	0.00
11	bis(2-Chloroethyl)ether	1.486	1.512	-1.7	105	0.00
12	1,3-Dichlorobenzene	1.571	1.568	0.2	106	0.00
13 C	1,4-Dichlorobenzene	1.601	1.585	1.0	105	0.00
14	1,2-Dichlorobenzene	1.558	1.544	0.9	105	0.00
15	Benzyl Alcohol	1.495	1.604	-7.3	110	0.00
16	2,2'-oxybis(1-Chloropropane	1.517	1.536	-1.3	108	-0.02
17	2-Methylphenol	1.269	1.329	-4.7	109	0.00
18	Hexachloroethane	0.595	0.605	-1.7	106	0.00
19 P	n-Nitroso-di-n-propylamine	1.482	1.584	-6.9	110	0.00
20	3+4-Methylphenols	1.722	1.850	-7.4	110	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	109	0.00
22	Acetophenone	0.553	0.551	0.4	108	0.00
23 S	Nitrobenzene-d5	0.423	0.424	-0.2	110	0.00
24	Nitrobenzene	0.440	0.442	-0.5	109	0.00
25	Isophorone	0.807	0.821	-1.7	111	0.00
26 C	2-Nitrophenol	0.182	0.188	-3.3	108	0.00
27	2,4-Dimethylphenol	0.271	0.275	-1.5	110	0.00
28	bis(2-Chloroethoxy)methane	0.478	0.480	-0.4	110	0.00
29 C	2,4-Dichlorophenol	0.300	0.315	-5.0	112	0.00
30	1,2,4-Trichlorobenzene	0.339	0.333	1.8	110	0.00
31	Naphthalene	1.054	1.041	1.2	110	0.00
32	Benzoic acid	0.230	0.234	-1.7	111	0.00
33	4-Chloroaniline	0.432	0.446	-3.2	111	0.00
34 C	Hexachlorobutadiene	0.194	0.190	2.1	109	0.00
35	Caprolactam	0.107	0.114	-6.5	111	0.00
36 C	4-Chloro-3-methylphenol	0.371	0.387	-4.3	113	0.00
37	2-Methylnaphthalene	0.766	0.760	0.8	111	0.00
38	1-Methylnaphthalene	0.755	0.750	0.7	110	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	112	0.00
40	1,2,4,5-Tetrachlorobenzene	0.633	0.615	2.8	110	0.00
41 P	Hexachlorocyclopentadiene	0.287	0.280	2.4	108	0.00
42 S	2,4,6-Tribromophenol	0.295	0.303	-2.7	113	0.00
43 C	2,4,6-Trichlorophenol	0.409	0.418	-2.2	112	0.00
44	2,4,5-Trichlorophenol	0.437	0.446	-2.1	111	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM031219\
 Data File : BM019124.D
 Acq On : 12 Mar 2019 22:34
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 13 05:56:06 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM031119.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 11 17:30:46 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.538	1.506	2.1	112	0.00
46	1,1'-Biphenyl	1.747	1.710	2.1	112	0.00
47	2-Chloronaphthalene	1.304	1.279	1.9	112	0.00
48	2-Nitroaniline	0.437	0.460	-5.3	112	0.00
49	Acenaphthylene	2.204	2.190	0.6	112	0.00
50	Dimethylphthalate	1.711	1.680	1.8	112	0.00
51	2,6-Dinitrotoluene	0.370	0.376	-1.6	111	0.00
52 C	Acenaphthene	1.267	1.235	2.5	111	0.00
53	3-Nitroaniline	0.392	0.384	2.0	111	0.00
54 P	2,4-Dinitrophenol	0.215	0.205	4.7	106	0.00
55	Dibenzofuran	2.043	2.003	2.0	113	0.00
56 P	4-Nitrophenol	0.320	0.316	1.3	110	0.00
57	2,4-Dinitrotoluene	0.537	0.528	1.7	112	0.00
58	Fluorene	1.684	1.674	0.6	113	0.00
59	2,3,4,6-Tetrachlorophenol	0.407	0.407	0.0	111	0.00
60	Diethylphthalate	1.733	1.712	1.2	112	0.00
61	4-Chlorophenyl-phenylether	0.818	0.814	0.5	113	0.00
62	4-Nitroaniline	0.395	0.390	1.3	112	0.00
63	Azobenzene	1.823	1.871	-2.6	114	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	111	0.00
65	4,6-Dinitro-2-methylphenol	0.131	0.128	2.3	108	0.00
66 c	n-Nitrosodiphenylamine	0.637	0.642	-0.8	112	0.00
67	4-Bromophenyl-phenylether	0.221	0.223	-0.9	112	0.00
68	Hexachlorobenzene	0.262	0.266	-1.5	114	0.00
69	Atrazine	0.214	0.217	-1.4	111	0.00
70 C	Pentachlorophenol	0.158	0.160	-1.3	113	0.00
71	Phenanthrene	1.172	1.161	0.9	112	0.00
72	Anthracene	1.148	1.148	0.0	111	0.00
73	Carbazole	1.083	1.087	-0.4	112	0.00
74	Di-n-butylphthalate	1.310	1.350	-3.1	113	0.00
75 C	Fluoranthene	1.345	1.322	1.7	111	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	110	0.00
77	Benzidine	0.567	0.618	-9.0	119	0.00
78	Pyrene	1.265	1.305	-3.2	111	0.00
79 S	Terphenyl-d14	1.007	1.044	-3.7	112	0.00
80	Butylbenzylphthalate	0.551	0.592	-7.4	112	0.00
81	Benzo(a)anthracene	1.255	1.236	1.5	109	0.00
82	3,3'-Dichlorobenzidine	0.483	0.493	-2.1	110	0.00
83	Chrysene	1.214	1.191	1.9	110	0.00
84	Bis(2-ethylhexyl)phthalate	0.802	0.860	-7.2	113	0.00
85 c	Di-n-octyl phthalate	1.350	1.406	-4.1	114	0.00
86	Indeno(1,2,3-cd)pyrene	1.307	1.151	11.9	111	0.00
87 I	Perylene-d12	1.000	1.000	0.0	110	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM031219\
Data File : BM019124.D
Acq On : 12 Mar 2019 22:34
Operator : JU/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC040

Quant Time: Mar 13 05:56:06 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM031119.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 11 17:30:46 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.296	1.305	-0.7	112	0.00
89	Benzo(k)fluoranthene	1.192	1.239	-3.9	111	0.00
90 C	Benzo(a)pyrene	1.165	1.167	-0.2	111	0.00
91	Dibenzo(a,h)anthracene	1.114	1.070	3.9	111	0.00
92	Benzo(q,h,i)perylene	1.023	0.964	5.8	109	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0