

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM031220\
 Data File : BM025533.D
 Acq On : 12 Mar 2020 21:37
 Operator : CG/JU
 Sample : L1782-13
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C0AB6

Manual Integrations
 APPROVED

mohammad
 3/16/2020 1:50:32 PM

Quant Time: Mar 13 05:22:11 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM030920.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 09 13:32:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.66	152	3351	0.40	ng/ul	0.00
2) Naphthalene-d8	10.42	136	14615	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.28	164	10974	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.03	188	23264m	0.40	ng/ul	0.00
16) Chrysene-d12	21.21	240	15709	0.40	ng/ul	0.00
20) Perylene-d12	23.40	264	16292	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.01	152	9096	0.35	ng/ul	-0.01
14) Fluoranthene-d10	19.06	212	25683	0.40	ng/ul	0.00
Target Compounds						
3) Naphthalene	10.47	128	5303	0.123	ng/ul	97
5) 2-Methylnaphthalene	12.09	142	2421	0.078	ng/ul	100
12) Phenanthrene	17.07	178	1681	0.023	ng/ul#	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM031220\
Data File : BM025533.D
Acq On : 12 Mar 2020 21:37
Operator : CG/JU
Sample : L1782-13
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
C0AB6

Manual Integrations
APPROVED

mohammad
3/16/2020 1:50:32 PM

Quant Time: Mar 13 05:22:11 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM030920.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 09 13:32:56 2020
Response via : Initial Calibration

