


Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM031220\  
 Data File : BM025513.D  
 Acq On : 12 Mar 2020 07:17  
 Operator : CG/JU  
 Sample : DFTPP85  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_M  
 ClientSampleId :  
 DFTPP85

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA\_M\METHODS\SOM-EPA-SIM-BM030920.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Fri Mar 13 16:53:43 2020



AutoFind: Scans 2483, 2484, 2485; Background Corrected with Scan 2477

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 37.0      | 234496  | PASS             |
| 68          | 69           | 0.00         | 2            | 1.9       | 4719    | PASS             |
| 69          | 198          | 0.00         | 100          | 39.3      | 249386  | PASS             |
| 70          | 69           | 0.00         | 2            | 0.5       | 1309    | PASS             |
| 127         | 198          | 10           | 80           | 47.8      | 303253  | PASS             |
| 197         | 198          | 0.00         | 2            | 0.9       | 5438    | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 634368  | PASS             |
| 199         | 198          | 5            | 9            | 6.9       | 43845   | PASS             |
| 275         | 198          | 10           | 60           | 31.5      | 200069  | PASS             |
| 365         | 198          | 1            | 100          | 4.1       | 26098   | PASS             |
| 441         | 198          | 0.01         | 100          | 14.2      | 89893   | PASS             |
| 442         | 198          | 50           | 100          | 95.5      | 605674  | PASS             |
| 443         | 442          | 15           | 24           | 19.7      | 119101  | PASS             |

Scan 2477 (17.457 min): BM025513.D  
DFTPP85

| m/z    | Abundance |
|--------|-----------|
| 40.00  | 740.0     |
| 43.90  | 304.0     |
| 73.10  | 655.0     |
| 191.00 | 442.0     |
| 207.00 | 1844.0    |
| 208.00 | 567.0     |
| 281.00 | 397.0     |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP85

| m/z    | Abundance |
|--------|-----------|
| 37.10  | 1100.0    |
| 38.10  | 3105.0    |
| 39.10  | 15435.0   |
| 40.00  | 1765.0    |
| 41.10  | 470.0     |
| 44.00  | 593.0     |
| 49.00  | 1763.0    |
| 50.10  | 51152.0   |
| 51.10  | 192768.0  |
| 52.00  | 9495.0    |
| 53.00  | 476.0     |
| 55.00  | 892.0     |
| 56.00  | 5690.0    |
| 57.00  | 12533.0   |
| 58.00  | 850.0     |
| 58.90  | 368.0     |
| 61.10  | 2281.0    |
| 62.10  | 2765.0    |
| 63.00  | 7583.0    |
| 64.00  | 1335.0    |
| 65.00  | 3464.0    |
| 67.10  | 326.0     |
| 68.10  | 4039.0    |
| 69.00  | 201024.0  |
| 70.10  | 721.0     |
| 73.00  | 1884.0    |
| 74.10  | 19328.0   |
| 75.00  | 30960.0   |
| 76.00  | 10756.0   |
| 77.10  | 208448.0  |
| 78.10  | 16087.0   |
| 79.00  | 14722.0   |
| 80.00  | 11721.0   |
| 81.00  | 16213.0   |
| 82.00  | 3619.0    |
| 83.00  | 3585.0    |
| 84.00  | 559.0     |
| 85.00  | 2603.0    |
| 86.00  | 4712.0    |
| 87.00  | 2396.0    |
| 88.10  | 907.0     |
| 89.00  | 659.0     |
| 91.00  | 3181.0    |
| 92.10  | 3848.0    |
| 93.00  | 25496.0   |
| 94.00  | 1756.0    |
| 96.00  | 1294.0    |
| 97.10  | 405.0     |
| 98.00  | 21592.0   |
| 99.00  | 17128.0   |
| 100.00 | 1508.0    |
| 101.00 | 8645.0    |
| 102.10 | 454.0     |
| 103.00 | 3015.0    |
| 104.00 | 6183.0    |
| 105.00 | 5969.0    |
| 106.20 | 1722.0    |
| 107.00 | 65520.0   |
| 108.00 | 10796.0   |
| 109.00 | 1511.0    |
| 110.00 | 120000.0  |
| 111.00 | 17352.0   |
| 112.00 | 2192.0    |
| 113.00 | 750.0     |
| 115.00 | 306.0     |
| 116.00 | 3803.0    |
| 117.00 | 54728.0   |
| 118.00 | 4069.0    |
| 119.00 | 599.0     |
| 119.90 | 957.0     |
| 121.10 | 325.0     |
| 122.00 | 4678.0    |
| 123.00 | 6428.0    |
| 124.00 | 3370.0    |
| 124.90 | 2787.0    |
| 127.00 | 228032.0  |
| 128.00 | 17872.0   |
| 129.00 | 92528.0   |
| 130.10 | 7010.0    |
| 131.00 | 1641.0    |
| 132.00 | 715.0     |
| 132.90 | 510.0     |
| 133.90 | 2477.0    |
| 135.00 | 7092.0    |
| 136.00 | 3121.0    |
| 137.10 | 4209.0    |
| 137.90 | 999.0     |
| 140.00 | 1082.0    |
| 141.00 | 10600.0   |
| 142.10 | 3813.0    |
| 143.00 | 1927.0    |
| 144.00 | 639.0     |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP85

|        |          |
|--------|----------|
| 145.00 | 324.0    |
| 146.00 | 2364.0   |
| 147.00 | 6226.0   |
| 148.00 | 13723.0  |
| 149.00 | 2646.0   |
| 150.10 | 784.0    |
| 151.20 | 1713.0   |
| 151.90 | 810.0    |
| 152.90 | 2880.0   |
| 154.00 | 3186.0   |
| 155.00 | 6574.0   |
| 156.00 | 9397.0   |
| 157.00 | 1976.0   |
| 158.00 | 2521.0   |
| 159.00 | 1680.0   |
| 160.00 | 4202.0   |
| 161.00 | 5273.0   |
| 162.10 | 1451.0   |
| 163.00 | 506.0    |
| 164.00 | 822.0    |
| 165.00 | 4472.0   |
| 165.90 | 2948.0   |
| 167.00 | 21200.0  |
| 168.00 | 13626.0  |
| 169.00 | 1819.0   |
| 170.00 | 851.0    |
| 171.00 | 926.0    |
| 172.00 | 2323.0   |
| 173.00 | 1954.0   |
| 174.00 | 4675.0   |
| 175.10 | 8269.0   |
| 176.10 | 2694.0   |
| 177.00 | 3963.0   |
| 178.00 | 1497.0   |
| 179.00 | 16392.0  |
| 180.00 | 12769.0  |
| 181.00 | 5432.0   |
| 182.00 | 1016.0   |
| 183.00 | 403.0    |
| 184.00 | 1256.0   |
| 185.00 | 7280.0   |
| 186.10 | 59016.0  |
| 187.00 | 18560.0  |
| 188.00 | 2374.0   |
| 189.00 | 3597.0   |
| 190.00 | 650.0    |
| 191.00 | 1963.0   |
| 192.00 | 4381.0   |
| 193.10 | 5359.0   |
| 194.00 | 1051.0   |
| 194.90 | 714.0    |
| 196.00 | 12422.0  |
| 197.20 | 3003.0   |
| 198.00 | 421504.0 |
| 199.00 | 32376.0  |
| 200.10 | 2266.0   |
| 201.40 | 1825.0   |
| 202.20 | 789.0    |
| 203.00 | 3005.0   |
| 204.00 | 16091.0  |
| 205.10 | 26176.0  |
| 206.10 | 106576.0 |
| 207.10 | 17392.0  |
| 208.00 | 4370.0   |
| 209.00 | 1743.0   |
| 210.30 | 1507.0   |
| 211.00 | 4158.0   |
| 212.10 | 868.0    |
| 212.90 | 432.0    |
| 214.90 | 1300.0   |
| 216.00 | 1982.0   |
| 217.00 | 26624.0  |
| 218.00 | 3040.0   |
| 219.00 | 392.0    |
| 221.00 | 17888.0  |
| 223.10 | 6202.0   |
| 224.00 | 58792.0  |
| 225.00 | 15604.0  |
| 226.10 | 2008.0   |
| 227.00 | 25384.0  |
| 228.00 | 3591.0   |
| 229.00 | 5668.0   |
| 230.00 | 549.0    |
| 231.00 | 1950.0   |
| 231.90 | 395.0    |
| 232.90 | 493.0    |
| 234.00 | 1459.0   |
| 235.10 | 1947.0   |
| 236.10 | 1312.0   |
| 237.00 | 3290.0   |
| 237.80 | 468.0    |
| 239.00 | 1128.0   |
| 240.00 | 970.0    |
| 241.00 | 1594.0   |
| 242.00 | 3033.0   |
| 243.00 | 3642.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP85

|        |          |
|--------|----------|
| 244.10 | 47280.0  |
| 245.00 | 6901.0   |
| 246.00 | 10405.0  |
| 246.90 | 1753.0   |
| 248.00 | 478.0    |
| 249.00 | 2159.0   |
| 251.00 | 413.0    |
| 251.90 | 315.0    |
| 252.30 | 358.0    |
| 253.00 | 1087.0   |
| 254.10 | 2361.0   |
| 255.00 | 260416.0 |
| 256.00 | 39248.0  |
| 257.00 | 4183.0   |
| 258.00 | 15173.0  |
| 259.00 | 2129.0   |
| 260.00 | 386.0    |
| 260.90 | 429.0    |
| 264.00 | 326.0    |
| 265.00 | 5692.0   |
| 265.90 | 924.0    |
| 271.10 | 588.0    |
| 272.20 | 525.0    |
| 273.00 | 7676.0   |
| 274.00 | 20504.0  |
| 275.00 | 122192.0 |
| 276.00 | 18280.0  |
| 277.00 | 10539.0  |
| 278.00 | 1951.0   |
| 278.90 | 516.0    |
| 281.10 | 630.0    |
| 282.10 | 448.0    |
| 283.00 | 880.0    |
| 284.00 | 841.0    |
| 285.00 | 1930.0   |
| 286.00 | 378.0    |
| 289.00 | 513.0    |
| 292.00 | 555.0    |
| 293.10 | 2067.0   |
| 293.90 | 660.0    |
| 295.00 | 766.0    |
| 296.00 | 31648.0  |
| 297.00 | 5217.0   |
| 298.00 | 334.0    |
| 301.00 | 401.0    |
| 302.00 | 680.0    |
| 303.00 | 4353.0   |
| 304.00 | 1161.0   |
| 304.90 | 344.0    |
| 307.90 | 509.0    |
| 310.00 | 452.0    |
| 314.00 | 1528.0   |
| 315.00 | 3393.0   |
| 316.00 | 2429.0   |
| 317.00 | 457.0    |
| 321.00 | 1020.0   |
| 322.10 | 803.0    |
| 323.00 | 10739.0  |
| 324.10 | 1791.0   |
| 327.00 | 2041.0   |
| 328.00 | 1105.0   |
| 332.10 | 518.0    |
| 333.10 | 826.0    |
| 334.10 | 6600.0   |
| 335.00 | 1847.0   |
| 336.10 | 404.0    |
| 341.00 | 1190.0   |
| 342.00 | 466.0    |
| 346.00 | 2271.0   |
| 347.00 | 789.0    |
| 352.00 | 2944.0   |
| 353.10 | 1693.0   |
| 354.10 | 3673.0   |
| 355.00 | 1293.0   |
| 365.00 | 14039.0  |
| 366.00 | 2556.0   |
| 371.00 | 1010.0   |
| 372.10 | 5219.0   |
| 373.00 | 1003.0   |
| 382.90 | 950.0    |
| 383.90 | 585.0    |
| 390.00 | 489.0    |
| 391.00 | 545.0    |
| 402.10 | 2354.0   |
| 403.10 | 2693.0   |
| 404.00 | 1081.0   |
| 421.00 | 2231.0   |
| 422.10 | 2137.0   |
| 423.10 | 16364.0  |
| 424.10 | 6583.0   |
| 425.00 | 843.0    |
| 441.10 | 43088.0  |
| 442.10 | 298624.0 |
| 443.10 | 57336.0  |
| 444.00 | 5274.0   |
| 445.00 | 550.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP85

| m/z    | Abundance |
|--------|-----------|
| 37.00  | 795.0     |
| 38.00  | 3502.0    |
| 39.10  | 20792.0   |
| 40.10  | 1747.0    |
| 41.00  | 748.0     |
| 44.00  | 668.0     |
| 45.00  | 497.0     |
| 49.10  | 2178.0    |
| 50.10  | 74240.0   |
| 51.10  | 277440.0  |
| 52.00  | 14082.0   |
| 53.10  | 711.0     |
| 55.00  | 1523.0    |
| 56.10  | 7963.0    |
| 57.00  | 19168.0   |
| 58.00  | 1001.0    |
| 59.00  | 334.0     |
| 61.00  | 3583.0    |
| 62.10  | 4130.0    |
| 63.10  | 12334.0   |
| 64.00  | 1637.0    |
| 65.10  | 5322.0    |
| 66.10  | 353.0     |
| 67.00  | 577.0     |
| 68.10  | 5705.0    |
| 69.00  | 295552.0  |
| 70.10  | 1602.0    |
| 73.10  | 2722.0    |
| 74.00  | 31064.0   |
| 75.00  | 47112.0   |
| 76.10  | 15586.0   |
| 77.10  | 317824.0  |
| 78.10  | 22104.0   |
| 79.00  | 23016.0   |
| 80.00  | 16832.0   |
| 81.00  | 23512.0   |
| 82.00  | 5833.0    |
| 83.00  | 4760.0    |
| 84.00  | 645.0     |
| 85.00  | 3819.0    |
| 86.00  | 6518.0    |
| 87.00  | 3235.0    |
| 88.00  | 1340.0    |
| 88.80  | 495.0     |
| 91.00  | 5442.0    |
| 92.00  | 6373.0    |
| 93.00  | 38472.0   |
| 94.00  | 2833.0    |
| 95.00  | 569.0     |
| 96.00  | 1720.0    |
| 97.20  | 814.0     |
| 98.00  | 30560.0   |
| 99.00  | 25080.0   |
| 100.00 | 2673.0    |
| 101.00 | 12513.0   |
| 102.00 | 786.0     |
| 103.00 | 4976.0    |
| 104.00 | 8806.0    |
| 105.00 | 8992.0    |
| 106.00 | 2361.0    |
| 107.00 | 98336.0   |
| 108.00 | 15641.0   |
| 109.20 | 3193.0    |
| 110.00 | 184640.0  |
| 111.00 | 29192.0   |
| 112.00 | 3548.0    |
| 113.00 | 1106.0    |
| 116.00 | 6104.0    |
| 117.00 | 83768.0   |
| 118.00 | 7311.0    |
| 119.10 | 883.0     |
| 119.90 | 1584.0    |
| 121.00 | 640.0     |
| 122.00 | 6530.0    |
| 123.00 | 10531.0   |
| 124.00 | 4672.0    |
| 125.00 | 4139.0    |
| 127.00 | 356992.0  |
| 128.00 | 26120.0   |
| 129.00 | 142400.0  |
| 130.00 | 11380.0   |
| 131.00 | 2468.0    |
| 131.90 | 1103.0    |
| 133.10 | 580.0     |
| 134.00 | 3979.0    |
| 135.00 | 11265.0   |
| 136.00 | 4689.0    |
| 137.10 | 6457.0    |
| 137.90 | 1516.0    |
| 139.00 | 630.0     |
| 140.00 | 1928.0    |
| 141.00 | 18440.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP85

|        |          |
|--------|----------|
| 142.00 | 5213.0   |
| 143.00 | 4417.0   |
| 143.90 | 1135.0   |
| 144.90 | 1051.0   |
| 146.00 | 3298.0   |
| 147.00 | 10619.0  |
| 148.00 | 20872.0  |
| 149.00 | 4242.0   |
| 150.00 | 1147.0   |
| 151.00 | 2190.0   |
| 151.70 | 1215.0   |
| 152.10 | 986.0    |
| 153.00 | 5719.0   |
| 154.00 | 4784.0   |
| 155.00 | 10355.0  |
| 156.00 | 14359.0  |
| 157.00 | 2978.0   |
| 158.00 | 3594.0   |
| 159.00 | 2698.0   |
| 160.00 | 5629.0   |
| 161.00 | 9343.0   |
| 162.00 | 2163.0   |
| 162.90 | 766.0    |
| 164.00 | 936.0    |
| 165.00 | 7844.0   |
| 166.00 | 6433.0   |
| 167.00 | 32408.0  |
| 168.00 | 19560.0  |
| 169.10 | 3085.0   |
| 169.90 | 1162.0   |
| 171.00 | 1818.0   |
| 172.00 | 3575.0   |
| 173.00 | 4531.0   |
| 174.00 | 8014.0   |
| 175.00 | 15837.0  |
| 176.10 | 3940.0   |
| 177.00 | 6394.0   |
| 177.90 | 2111.0   |
| 179.00 | 28000.0  |
| 180.00 | 20856.0  |
| 181.00 | 8824.0   |
| 182.10 | 1197.0   |
| 183.00 | 745.0    |
| 184.00 | 2436.0   |
| 185.00 | 14144.0  |
| 186.00 | 100568.0 |
| 187.00 | 30256.0  |
| 188.00 | 4247.0   |
| 189.00 | 6922.0   |
| 190.00 | 1142.0   |
| 191.00 | 3287.0   |
| 192.00 | 8546.0   |
| 193.00 | 9606.0   |
| 194.10 | 1716.0   |
| 195.00 | 1129.0   |
| 196.00 | 22304.0  |
| 197.20 | 6616.0   |
| 198.00 | 749504.0 |
| 199.00 | 52280.0  |
| 200.00 | 3324.0   |
| 201.40 | 3069.0   |
| 203.00 | 5094.0   |
| 204.10 | 24856.0  |
| 205.10 | 45704.0  |
| 206.10 | 187072.0 |
| 207.10 | 27888.0  |
| 208.00 | 7303.0   |
| 209.00 | 2612.0   |
| 211.00 | 6480.0   |
| 213.00 | 578.0    |
| 215.00 | 2375.0   |
| 216.00 | 4456.0   |
| 217.00 | 51680.0  |
| 218.00 | 6068.0   |
| 219.00 | 827.0    |
| 221.00 | 32128.0  |
| 221.80 | 6366.0   |
| 223.00 | 10093.0  |
| 224.10 | 101048.0 |
| 225.00 | 27512.0  |
| 226.10 | 3352.0   |
| 227.00 | 44968.0  |
| 228.00 | 5754.0   |
| 229.00 | 8594.0   |
| 230.10 | 978.0    |
| 231.00 | 4002.0   |
| 231.90 | 632.0    |
| 233.10 | 880.0    |
| 234.00 | 3078.0   |
| 235.00 | 2636.0   |
| 236.10 | 1760.0   |
| 237.00 | 4414.0   |
| 237.90 | 594.0    |
| 239.00 | 1555.0   |
| 240.10 | 1464.0   |
| 241.00 | 2304.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP85

|        |          |
|--------|----------|
| 242.00 | 6304.0   |
| 243.10 | 7096.0   |
| 244.10 | 87400.0  |
| 245.00 | 12193.0  |
| 246.00 | 17376.0  |
| 247.00 | 3970.0   |
| 248.00 | 804.0    |
| 249.00 | 3429.0   |
| 250.10 | 601.0    |
| 251.00 | 639.0    |
| 252.10 | 1037.0   |
| 253.10 | 2035.0   |
| 255.00 | 466752.0 |
| 256.00 | 72088.0  |
| 257.00 | 6680.0   |
| 258.00 | 24456.0  |
| 259.00 | 4064.0   |
| 260.00 | 829.0    |
| 261.00 | 607.0    |
| 263.90 | 824.0    |
| 265.00 | 10664.0  |
| 266.00 | 1483.0   |
| 267.00 | 341.0    |
| 269.90 | 439.0    |
| 271.00 | 1573.0   |
| 272.00 | 1055.0   |
| 273.00 | 14787.0  |
| 274.00 | 39776.0  |
| 275.00 | 226048.0 |
| 276.00 | 31768.0  |
| 277.00 | 18328.0  |
| 278.00 | 3653.0   |
| 279.00 | 579.0    |
| 281.00 | 702.0    |
| 282.00 | 557.0    |
| 283.00 | 2131.0   |
| 284.00 | 1586.0   |
| 285.00 | 3489.0   |
| 286.00 | 711.0    |
| 288.90 | 954.0    |
| 290.00 | 561.0    |
| 292.00 | 720.0    |
| 293.00 | 4378.0   |
| 294.00 | 1529.0   |
| 295.10 | 1364.0   |
| 296.00 | 62136.0  |
| 297.00 | 8563.0   |
| 297.90 | 373.0    |
| 301.10 | 1062.0   |
| 302.10 | 1102.0   |
| 303.00 | 7046.0   |
| 304.00 | 2002.0   |
| 305.10 | 402.0    |
| 307.90 | 835.0    |
| 309.00 | 641.0    |
| 310.00 | 652.0    |
| 312.90 | 675.0    |
| 314.00 | 2633.0   |
| 315.00 | 6576.0   |
| 316.10 | 3824.0   |
| 316.90 | 738.0    |
| 321.00 | 2277.0   |
| 322.00 | 1099.0   |
| 323.10 | 20360.0  |
| 324.10 | 3849.0   |
| 325.10 | 437.0    |
| 326.00 | 370.0    |
| 327.00 | 3992.0   |
| 328.00 | 1786.0   |
| 329.00 | 343.0    |
| 332.00 | 1495.0   |
| 333.00 | 2217.0   |
| 334.00 | 12775.0  |
| 335.10 | 3504.0   |
| 341.00 | 2686.0   |
| 342.00 | 691.0    |
| 346.00 | 4021.0   |
| 347.00 | 1283.0   |
| 351.00 | 372.0    |
| 352.10 | 5306.0   |
| 353.00 | 3975.0   |
| 354.10 | 7408.0   |
| 355.00 | 1044.0   |
| 359.10 | 394.0    |
| 365.00 | 28352.0  |
| 366.00 | 4055.0   |
| 370.00 | 571.0    |
| 371.10 | 1388.0   |
| 372.00 | 10192.0  |
| 373.00 | 2035.0   |
| 383.00 | 2704.0   |
| 384.10 | 863.0    |
| 390.10 | 1256.0   |
| 391.00 | 1013.0   |
| 392.00 | 899.0    |
| 401.00 | 458.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP85



|        |          |
|--------|----------|
| 402.00 | 4092.0   |
| 403.00 | 5722.0   |
| 404.10 | 2078.0   |
| 404.90 | 760.0    |
| 421.10 | 4280.0   |
| 422.10 | 4961.0   |
| 423.10 | 34856.0  |
| 424.10 | 10899.0  |
| 424.90 | 898.0    |
| 440.10 | 629.0    |
| 441.10 | 99168.0  |
| 442.10 | 659392.0 |
| 443.10 | 132160.0 |
| 444.10 | 12918.0  |
| 445.10 | 862.0    |

**Instrument :**  
BNA\_M  
**ClientSampleId :**  
DFTPP85

| m/z    | Abundance |
|--------|-----------|
| 37.00  | 1363.0    |
| 38.10  | 3058.0    |
| 39.10  | 18016.0   |
| 40.00  | 1489.0    |
| 41.00  | 573.0     |
| 44.00  | 855.0     |
| 44.90  | 347.0     |
| 49.10  | 2507.0    |
| 50.00  | 62504.0   |
| 51.10  | 233280.0  |
| 52.00  | 11300.0   |
| 53.00  | 410.0     |
| 55.10  | 1120.0    |
| 56.00  | 6710.0    |
| 57.00  | 16656.0   |
| 58.00  | 828.0     |
| 61.00  | 3448.0    |
| 62.00  | 3423.0    |
| 63.00  | 8409.0    |
| 64.00  | 1256.0    |
| 65.10  | 4645.0    |
| 66.10  | 343.0     |
| 66.90  | 347.0     |
| 68.20  | 4413.0    |
| 69.00  | 251584.0  |
| 70.10  | 1606.0    |
| 73.00  | 2457.0    |
| 74.00  | 26344.0   |
| 75.00  | 43168.0   |
| 76.10  | 13256.0   |
| 77.10  | 272704.0  |
| 78.10  | 19120.0   |
| 79.00  | 19680.0   |
| 80.00  | 14733.0   |
| 81.00  | 20608.0   |
| 82.00  | 4842.0    |
| 83.00  | 5180.0    |
| 83.90  | 643.0     |
| 85.00  | 3101.0    |
| 86.00  | 5560.0    |
| 87.00  | 2568.0    |
| 88.00  | 1062.0    |
| 89.10  | 606.0     |
| 91.00  | 5178.0    |
| 92.00  | 4888.0    |
| 93.00  | 33984.0   |
| 94.00  | 2462.0    |
| 95.00  | 542.0     |
| 96.10  | 1319.0    |
| 97.10  | 626.0     |
| 98.00  | 26240.0   |
| 99.00  | 20000.0   |
| 100.10 | 1698.0    |
| 101.00 | 12294.0   |
| 102.00 | 686.0     |
| 103.00 | 4042.0    |
| 104.00 | 6766.0    |
| 105.00 | 6929.0    |
| 106.20 | 2214.0    |
| 107.00 | 92784.0   |
| 108.00 | 14834.0   |
| 109.10 | 2756.0    |
| 110.00 | 161920.0  |
| 111.00 | 25432.0   |
| 112.00 | 2947.0    |
| 113.10 | 728.0     |
| 115.20 | 554.0     |
| 116.00 | 4951.0    |
| 117.00 | 78152.0   |
| 118.00 | 5487.0    |
| 118.90 | 680.0     |
| 120.00 | 1138.0    |
| 120.90 | 313.0     |
| 122.00 | 6292.0    |
| 123.00 | 9470.0    |
| 124.00 | 4363.0    |
| 125.00 | 4543.0    |
| 127.00 | 324736.0  |
| 128.00 | 24416.0   |
| 129.00 | 128048.0  |
| 130.00 | 11337.0   |
| 131.00 | 2648.0    |
| 132.00 | 1112.0    |
| 133.00 | 528.0     |
| 134.00 | 3867.0    |
| 135.00 | 10635.0   |
| 136.00 | 4266.0    |
| 137.10 | 5528.0    |
| 138.10 | 1316.0    |
| 138.90 | 1002.0    |
| 140.00 | 1799.0    |
| 141.00 | 17216.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP85

|        |          |
|--------|----------|
| 142.00 | 5946.0   |
| 143.00 | 3823.0   |
| 144.00 | 1081.0   |
| 145.00 | 1014.0   |
| 146.00 | 3014.0   |
| 147.00 | 8767.0   |
| 148.00 | 20280.0  |
| 149.00 | 4000.0   |
| 150.00 | 1283.0   |
| 151.10 | 2054.0   |
| 151.90 | 1365.0   |
| 153.00 | 5831.0   |
| 154.00 | 3975.0   |
| 155.00 | 9528.0   |
| 156.10 | 13498.0  |
| 157.10 | 2349.0   |
| 158.00 | 3558.0   |
| 159.00 | 2826.0   |
| 160.00 | 5022.0   |
| 161.00 | 7581.0   |
| 162.00 | 2238.0   |
| 163.00 | 727.0    |
| 164.10 | 799.0    |
| 165.00 | 6494.0   |
| 166.00 | 4947.0   |
| 167.00 | 32104.0  |
| 168.00 | 18368.0  |
| 169.00 | 2946.0   |
| 170.10 | 1439.0   |
| 170.90 | 1509.0   |
| 172.00 | 3161.0   |
| 173.00 | 4154.0   |
| 174.00 | 7611.0   |
| 175.00 | 15045.0  |
| 176.00 | 3937.0   |
| 177.00 | 7561.0   |
| 178.00 | 2246.0   |
| 179.00 | 26168.0  |
| 180.00 | 19016.0  |
| 181.10 | 8743.0   |
| 182.10 | 1147.0   |
| 182.90 | 1007.0   |
| 184.00 | 2280.0   |
| 185.00 | 13880.0  |
| 186.10 | 99496.0  |
| 187.00 | 26680.0  |
| 188.00 | 3159.0   |
| 189.00 | 6700.0   |
| 190.00 | 1469.0   |
| 191.10 | 2338.0   |
| 192.00 | 8162.0   |
| 193.10 | 9497.0   |
| 194.10 | 1772.0   |
| 195.00 | 1077.0   |
| 196.00 | 22040.0  |
| 197.20 | 6695.0   |
| 198.00 | 732096.0 |
| 199.00 | 46880.0  |
| 200.00 | 4081.0   |
| 201.50 | 2860.0   |
| 203.00 | 4771.0   |
| 204.10 | 25848.0  |
| 205.00 | 46552.0  |
| 206.10 | 197312.0 |
| 207.10 | 26440.0  |
| 208.00 | 6554.0   |
| 209.00 | 2422.0   |
| 210.30 | 3258.0   |
| 211.00 | 6399.0   |
| 213.00 | 467.0    |
| 215.00 | 1758.0   |
| 216.10 | 4285.0   |
| 217.00 | 51296.0  |
| 218.00 | 6203.0   |
| 218.90 | 530.0    |
| 221.00 | 33344.0  |
| 221.80 | 7643.0   |
| 223.00 | 10725.0  |
| 224.00 | 107912.0 |
| 225.10 | 27328.0  |
| 226.10 | 3104.0   |
| 227.00 | 44680.0  |
| 228.00 | 5762.0   |
| 229.00 | 9818.0   |
| 230.10 | 1316.0   |
| 231.00 | 4290.0   |
| 232.00 | 781.0    |
| 232.90 | 817.0    |
| 234.00 | 3459.0   |
| 235.00 | 3448.0   |
| 236.00 | 2132.0   |
| 237.00 | 4696.0   |
| 238.00 | 805.0    |
| 239.00 | 1872.0   |
| 240.10 | 1647.0   |
| 241.00 | 2625.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP85

|        |          |
|--------|----------|
| 242.00 | 5682.0   |
| 243.10 | 7273.0   |
| 244.00 | 88648.0  |
| 245.00 | 12615.0  |
| 246.00 | 17624.0  |
| 247.00 | 3157.0   |
| 248.00 | 872.0    |
| 249.00 | 3395.0   |
| 250.00 | 678.0    |
| 251.00 | 612.0    |
| 252.00 | 964.0    |
| 253.00 | 2265.0   |
| 254.20 | 5056.0   |
| 255.00 | 497792.0 |
| 256.00 | 72256.0  |
| 257.00 | 5973.0   |
| 258.00 | 27552.0  |
| 259.10 | 4430.0   |
| 260.00 | 620.0    |
| 261.00 | 1158.0   |
| 264.00 | 1110.0   |
| 265.00 | 12071.0  |
| 266.10 | 1699.0   |
| 269.90 | 759.0    |
| 271.00 | 914.0    |
| 272.10 | 1299.0   |
| 273.00 | 15410.0  |
| 274.00 | 39408.0  |
| 275.00 | 251968.0 |
| 276.00 | 32472.0  |
| 277.00 | 17744.0  |
| 278.00 | 3029.0   |
| 279.00 | 654.0    |
| 281.00 | 516.0    |
| 282.10 | 474.0    |
| 283.00 | 2412.0   |
| 284.00 | 1746.0   |
| 285.00 | 3395.0   |
| 286.00 | 1019.0   |
| 289.00 | 1077.0   |
| 290.10 | 505.0    |
| 291.00 | 546.0    |
| 292.00 | 825.0    |
| 293.00 | 4738.0   |
| 294.00 | 1080.0   |
| 295.10 | 1327.0   |
| 296.00 | 67952.0  |
| 297.00 | 8733.0   |
| 298.00 | 441.0    |
| 299.10 | 315.0    |
| 301.00 | 695.0    |
| 302.00 | 1182.0   |
| 303.10 | 7620.0   |
| 304.00 | 2296.0   |
| 304.90 | 388.0    |
| 308.00 | 802.0    |
| 308.90 | 581.0    |
| 310.00 | 1006.0   |
| 312.00 | 321.0    |
| 312.90 | 606.0    |
| 314.00 | 3291.0   |
| 315.00 | 7319.0   |
| 316.00 | 4506.0   |
| 316.90 | 657.0    |
| 321.10 | 2098.0   |
| 322.00 | 1150.0   |
| 323.10 | 22488.0  |
| 324.00 | 4694.0   |
| 325.00 | 469.0    |
| 326.00 | 456.0    |
| 327.00 | 4169.0   |
| 328.00 | 2426.0   |
| 329.10 | 462.0    |
| 332.10 | 1665.0   |
| 333.10 | 1827.0   |
| 334.00 | 14622.0  |
| 335.00 | 3811.0   |
| 336.00 | 592.0    |
| 338.90 | 433.0    |
| 340.00 | 377.0    |
| 341.00 | 2730.0   |
| 342.00 | 860.0    |
| 346.00 | 5002.0   |
| 347.00 | 811.0    |
| 351.10 | 372.0    |
| 352.10 | 6919.0   |
| 353.10 | 4208.0   |
| 354.00 | 8330.0   |
| 355.00 | 1343.0   |
| 359.10 | 637.0    |
| 364.20 | 390.0    |
| 365.00 | 35904.0  |
| 366.00 | 4769.0   |
| 370.10 | 811.0    |
| 371.10 | 1521.0   |
| 372.10 | 13476.0  |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP85

|        |          |
|--------|----------|
| 373.10 | 3070.0   |
| 374.00 | 384.0    |
| 383.00 | 3176.0   |
| 384.10 | 925.0    |
| 385.10 | 380.0    |
| 390.00 | 1421.0   |
| 391.00 | 1019.0   |
| 392.00 | 769.0    |
| 401.10 | 997.0    |
| 402.10 | 5299.0   |
| 403.10 | 7203.0   |
| 404.00 | 2751.0   |
| 404.90 | 545.0    |
| 410.10 | 354.0    |
| 415.20 | 380.0    |
| 421.10 | 5978.0   |
| 422.00 | 6044.0   |
| 423.10 | 42872.0  |
| 424.10 | 11227.0  |
| 425.10 | 1355.0   |
| 440.10 | 453.0    |
| 441.10 | 127424.0 |
| 442.10 | 859008.0 |
| 443.10 | 167808.0 |
| 444.10 | 15459.0  |
| 445.10 | 1041.0   |

**Instrument :**  
BNA\_M  
**ClientSampleId :**  
DFTPP85