

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM031221\
 Data File : BM029106.D
 Acq On : 12 Mar 2021 11:51
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 12 12:30:31 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM022321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 12 12:30:15 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	57	0.00
2	1,4-Dioxane	0.454	0.460	-1.3	56	0.00
3	Pyridine	1.197	1.205	-0.7	57	0.00
4	n-Nitrosodimethylamine	0.669	0.836	-25.0	69	0.00
5 S	2-Fluorophenol	1.207	1.178	2.4	54	0.00
6	Aniline	1.704	1.668	2.1	55	0.00
7 S	Phenol-d6	1.594	1.522	4.5	54	0.00
8	2-Chlorophenol	1.433	1.371	4.3	54	0.00
9	Benzaldehyde	0.869	0.707	18.6	52	0.00
10 C	Phenol	1.584	1.539	2.8	55	0.00
11	bis(2-Chloroethyl)ether	1.201	1.144	4.7	54	0.00
12	1,3-Dichlorobenzene	1.563	1.506	3.6	55	0.00
13 C	1,4-Dichlorobenzene	1.585	1.557	1.8	55	0.00
14	1,2-Dichlorobenzene	1.525	1.509	1.0	56	0.00
15	Benzyl Alcohol	1.140	1.201	-5.4	59	0.00
16	2,2'-oxybis(1-Chloropropane	1.850	2.033	-9.9	63	0.00
17	2-Methylphenol	1.076	1.037	3.6	54	0.00
18	Hexachloroethane	0.576	0.589	-2.3	58	0.00
19 P	n-Nitroso-di-n-propylamine	0.938	0.944	-0.6	56	0.00
20	3+4-Methylphenols	1.447	1.407	2.8	55	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	55	0.00
22	Acetophenone	0.488	0.498	-2.0	56	0.00
23 S	Nitrobenzene-d5	0.371	0.384	-3.5	56	0.00
24	Nitrobenzene	0.377	0.389	-3.2	56	0.00
25	Isophorone	0.652	0.684	-4.9	57	0.00
26 C	2-Nitrophenol	0.190	0.197	-3.7	55	0.00
27	2,4-Dimethylphenol	0.286	0.282	1.4	54	0.00
28	bis(2-Chloroethoxy)methane	0.371	0.363	2.2	53	0.00
29 C	2,4-Dichlorophenol	0.327	0.338	-3.4	56	0.00
30	1,2,4-Trichlorobenzene	0.358	0.367	-2.5	56	0.00
31	Naphthalene	1.103	1.088	1.4	54	0.00
32	Benzoic acid	0.205	0.221	-7.8	55	0.00
33	4-Chloroaniline	0.443	0.439	0.9	54	0.00
34 C	Hexachlorobutadiene	0.215	0.232	-7.9	59	0.00
35	Caprolactam	0.089	0.091	-2.2	55	0.00
36 C	4-Chloro-3-methylphenol	0.329	0.340	-3.3	57	0.00
37	2-Methylnaphthalene	0.776	0.762	1.8	54	0.00
38	1-Methylnaphthalene	0.735	0.719	2.2	54	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	56	0.00
40	1,2,4,5-Tetrachlorobenzene	0.625	0.624	0.2	56	0.00
41 P	Hexachlorocyclopentadiene	0.238	0.210	11.8	48#	0.00
42 S	2,4,6-Tribromophenol	0.237	0.246	-3.8	56	0.00
43 C	2,4,6-Trichlorophenol	0.438	0.443	-1.1	55	0.00
44	2,4,5-Trichlorophenol	0.470	0.468	0.4	55	0.00
45 S	2-Fluorobiphenyl	1.502	1.512	-0.7	57	0.00
46	1,1'-Biphenyl	1.604	1.568	2.2	55	0.00
47	2-Chloronaphthalene	1.309	1.292	1.3	56	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM031221\
 Data File : BM029106.D
 Acq On : 12 Mar 2021 11:51
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 12 12:30:31 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM022321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 12 12:30:15 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.352	0.381	-8.2	58	0.00
49	Acenaphthylene	2.011	2.014	-0.1	56	0.00
50	Dimethylphthalate	1.516	1.553	-2.4	57	0.00
51	2,6-Dinitrotoluene	0.353	0.354	-0.3	54	0.00
52 C	Acenaphthene	1.233	1.188	3.6	54	0.00
53	3-Nitroaniline	0.344	0.348	-1.2	54	0.00
54 P	2,4-Dinitrophenol	0.181	0.178	1.7	52	0.00
55	Dibenzofuran	1.936	1.939	-0.2	56	0.00
56 P	4-Nitrophenol	0.282	0.261	7.4	51	0.00
57	2,4-Dinitrotoluene	0.462	0.494	-6.9	57	0.00
58	Fluorene	1.552	1.545	0.5	55	0.00
59	2,3,4,6-Tetrachlorophenol	0.378	0.387	-2.4	56	0.00
60	Diethylphthalate	1.525	1.625	-6.6	59	0.00
61	4-Chlorophenyl-phenylether	0.740	0.758	-2.4	58	0.00
62	4-Nitroaniline	0.333	0.336	-0.9	53	0.00
63	Azobenzene	1.407	1.505	-7.0	59	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	56	0.00
65	4,6-Dinitro-2-methylphenol	0.141	0.140	0.7	54	0.00
66 c	n-Nitrosodiphenylamine	0.684	0.674	1.5	55	0.00
67	4-Bromophenyl-phenylether	0.224	0.225	-0.4	56	0.00
68	Hexachlorobenzene	0.250	0.250	0.0	57	0.00
69	Atrazine	0.213	0.227	-6.6	59	0.00
70 C	Pentachlorophenol	0.134	0.140	-4.5	55	0.00
71	Phenanthrene	1.191	1.168	1.9	55	0.00
72	Anthracene	1.180	1.173	0.6	55	0.00
73	Carbazole	1.132	1.129	0.3	55	0.00
74	Di-n-butylphthalate	1.287	1.425	-10.7	60	0.00
75 C	Fluoranthene	1.309	1.330	-1.6	56	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	57	0.00
77	Benzidine	0.659	0.563	14.6	42#	0.00
78	Pyrene	1.469	1.404	4.4	56	0.00
79 S	Terphenyl-d14	1.083	1.099	-1.5	59	0.00
80	Butylbenzylphthalate	0.632	0.669	-5.9	60	0.00
81	Benzo(a)anthracene	1.388	1.367	1.5	57	0.00
82	3,3'-Dichlorobenzidine	0.479	0.486	-1.5	57	0.00
83	Chrysene	1.353	1.347	0.4	57	0.00
84	Bis(2-ethylhexyl)phthalate	0.910	1.024	-12.5	63	0.00
85 c	Di-n-octyl phthalate	1.544	1.744	-13.0	62	0.00
86 I	Perylene-d12	1.000	1.000	0.0	58	0.00
87	Indeno(1,2,3-cd)pyrene	1.509	1.508	0.1	57	0.00
88	Benzo(b)fluoranthene	1.413	1.368	3.2	56	0.00
89	Benzo(k)fluoranthene	1.342	1.326	1.2	58	0.00
90 C	Benzo(a)pyrene	1.289	1.256	2.6	57	0.00
91	Dibenzo(a,h)anthracene	1.312	1.313	-0.1	57	0.00
92	Benzo(g,h,i)perylene	1.266	1.245	1.7	56	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM031221\
Data File : BM029106.D
Acq On : 12 Mar 2021 11:51
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC040

Quant Time: Mar 12 12:30:31 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM022321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 12 12:30:15 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0