

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM031224\
Data File : BM044588.D
Acq On : 12 Mar 2024 13:03
Operator : MA/JU
Sample : PB159473BL
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK473

Quant Time: Mar 12 13:31:06 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM031124.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 11 14:02:17 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.724	152	5773	0.400	ng/u1	0.00
4) Naphthalene-d8	10.512	136	18468	0.400	ng/u1	0.00
9) Acenaphthene-d10	14.372	164	10684	0.400	ng/u1	0.00
13) Phenanthrene-d10	17.136	188	24650	0.400	ng/u1	0.02
17) Chrysene-d12	21.328	240	15882	0.400	ng/u1	0.01
23) Perylene-d12	23.584	264	15992	0.400	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.281	96	4715	0.623	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.123	152	8260	0.311	ng/u1	0.02
18) Fluoranthene-d10	19.164	212	18781	0.337	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM031224\
Data File : BM044588.D
Acq On : 12 Mar 2024 13:03
Operator : MA/JU
Sample : PB159473BL
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK473

Quant Time: Mar 12 13:31:06 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM031124.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 11 14:02:17 2024
Response via : Initial Calibration

