

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM031224\
 Data File : BM044585.D
 Acq On : 12 Mar 2024 10:05
 Operator : MA/JU
 Sample : PB159440BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK440

Quant Time: Mar 12 10:37:03 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM031124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 11 14:02:17 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.724	152	5761	0.400	ng/ul	0.00
4) Naphthalene-d8	10.518	136	18166	0.400	ng/ul	0.01
9) Acenaphthene-d10	14.372	164	10277	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.137	188	23147	0.400	ng/ul	0.02
17) Chrysene-d12	21.327	240	14330	0.400	ng/ul	0.01
23) Perylene-d12	23.586	264	14082	0.400	ng/ul	0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.281	96	5140	0.681	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.134	152	8007	0.307	ng/ul	0.03
18) Fluoranthene-d10	19.165	212	18535	0.368	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM031224\
Data File : BM044585.D
Acq On : 12 Mar 2024 10:05
Operator : MA/JU
Sample : PB159440BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK440

Quant Time: Mar 12 10:37:03 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM031124.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 11 14:02:17 2024
Response via : Initial Calibration

