

Data Path : Z:\HPCHEM1\BNA M\DATA\BM031317\
 Data File : BM009239.D
 Acq On : 14 Mar 2017 03:56
 Operator : SJ/MA
 Sample : I2181-10
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C0BF9

Quant Time: Mar 14 04:59:30 2017
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-2016\SOM-EPA-BM031317MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Mar 14 04:57:13 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.87	152	92150	20.00	ng/ul	0.00
18) Naphthalene-d8	10.68	136	438395	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.52	164	327534	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.26	188	873640	20.00	ng/ul	0.00
77) Chrysene-d12	21.44	240	1302129	20.00	ng/ul	0.00
85) Perylene-d12	23.77	264	1363968	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.35	96	12084	5.78	ng/uL	0.00
5) Phenol-d5	7.03	99	54012	5.42	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.21	67	218424	33.64	ng/ul	0.00
9) 2-Chlorophenol-d4	7.40	132	153092	25.13	ng/ul	0.00
13) 4-Methylphenol-d8	8.57	113	113561	14.28	ng/ul	0.00
19) Nitrobenzene-d5	9.05	128	109889	35.10	ng/ul	0.00
22) 2-Nitrophenol-d4	9.76	143	116824	33.62	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.29	165	209877	28.27	ng/ul	0.00
29) 4-Chloroaniline-d4	10.83	131	14901	1.68	ng/ul	0.00
43) Dimethylphthalate-d6	13.92	166	1009914	40.60	ng/ul	0.00
46) Acenaphthylene-d8	14.22	160	1149599	40.19	ng/ul	0.00
51) 4-Nitrophenol-d4	14.70	143	26885	5.44	ng/ul	0.00
57) Fluorene-d10	15.51	176	913937	41.18	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.63	200	159220	30.46	ng/ul	0.00
70) Anthracene-d10	17.36	188	1602472	38.76	ng/ul	0.00
78) Pyrene-d10	19.66	212	2019524	38.43	ng/ul	0.00
89) Benzo(a)pyrene-d12	23.63	264	2381524	38.16	ng/ul	0.00

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM031317\
Data File : BM009239.D
Acq On : 14 Mar 2017 03:56
Operator : SJ/MA
Sample : I2181-10
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
C0BF9

Quant Time: Mar 14 04:59:30 2017
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-2016\SOM-EPA-BM031317MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Mar 14 04:57:13 2017
Response via : Initial Calibration

