


Data Path : Z:\HPCHEM1\BNA_M\Data\BM031318\
 Data File : BM014540.D
 Acq On : 13 Mar 2018 10:10
 Operator : SJ/JU
 Sample : DFTPP11
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP11

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM021418.M
 Title : SVOA CALIBRATION
 Last Update : Wed Mar 14 02:08:18 2018



AutoFind: Scans 2457, 2458, 2459; Background Corrected with Scan 2452

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	72.8	60253	PASS
68	69	0.00	2	1.6	817	PASS
69	198	0.00	100	63.2	52336	PASS
70	69	0.00	2	0.6	290	PASS
127	198	10	80	57.0	47202	PASS
197	198	0.00	2	1.0	866	PASS
198	198	100	100	100.0	82784	PASS
199	198	5	9	6.6	5448	PASS
275	198	10	60	29.6	24514	PASS
365	198	1	100	6.1	5057	PASS
441	443	0.01	100	91.3	8827	PASS
442	198	50	100	63.7	52706	PASS
443	442	15	24	18.3	9666	PASS

m/z	Abundance
36.90	350.0
38.00	848.0
39.00	3814.0
49.00	664.0
50.10	17432.0
51.10	52888.0
52.00	2032.0
55.00	1192.0
55.90	1745.0
57.00	3196.0
61.00	877.0
62.00	526.0
63.00	2289.0
64.10	343.0
64.90	785.0
68.00	814.0
69.00	49600.0
70.10	446.0
73.00	1149.0
74.00	9289.0
75.00	8497.0
76.00	2865.0
77.10	56904.0
78.10	3063.0
79.00	5157.0
79.90	3331.0
81.00	3947.0
82.00	671.0
83.10	449.0
85.10	322.0
85.90	2396.0
86.90	1063.0
91.10	1234.0
92.20	880.0
93.00	12736.0
93.90	671.0
95.80	388.0
98.00	9632.0
99.00	4936.0
100.10	386.0
101.00	2349.0
103.10	936.0
104.00	2087.0
104.90	1400.0
106.10	456.0
107.00	18952.0
108.10	2246.0
109.20	596.0
110.00	22856.0
111.10	3982.0
111.90	787.0
116.10	1526.0
117.00	32936.0
118.00	1715.0
121.80	1612.0
123.00	1912.0
123.90	1672.0
124.90	768.0
127.00	42240.0
128.00	3148.0
128.90	20360.0
130.00	1876.0
130.90	392.0
132.00	362.0
134.00	597.0
134.90	2271.0
135.90	693.0
136.90	1384.0
139.90	716.0
140.90	3026.0
142.10	914.0
142.80	492.0
145.90	700.0
147.00	1805.0
147.90	5445.0
148.90	1309.0
150.00	476.0
153.00	1011.0
153.80	781.0
155.00	2470.0
156.00	2309.0
156.90	485.0
158.00	613.0
158.80	486.0
160.10	703.0
160.90	2295.0
165.00	1365.0
165.90	1246.0
166.90	8061.0
168.00	4708.0
169.00	615.0

Instrument :
BNA_M
ClientSampleId :
DFTPP11

172.00	917.0
173.10	1514.0
173.90	1478.0
175.00	2526.0
175.90	808.0
176.80	588.0
177.90	593.0
179.00	4330.0
180.00	3340.0
181.10	960.0
182.20	510.0
183.90	746.0
184.90	2141.0
186.00	12837.0
186.90	3908.0
187.90	943.0
189.00	1689.0
191.00	358.0
192.00	1284.0
193.00	1711.0
195.90	2536.0
197.00	814.0
197.90	68720.0
199.00	5434.0
199.90	576.0
201.40	469.0
203.10	1093.0
204.00	4153.0
205.00	7419.0
206.10	21520.0
207.00	3471.0
208.00	905.0
209.10	587.0
209.80	674.0
211.00	1439.0
215.80	864.0
217.00	7412.0
218.00	1228.0
221.20	2515.0
222.00	1462.0
223.00	1318.0
224.10	15970.0
225.00	3488.0
226.00	855.0
226.90	8154.0
228.10	1385.0
228.90	1651.0
231.00	459.0
234.00	712.0
235.90	451.0
236.90	702.0
239.00	696.0
240.00	407.0
240.90	366.0
242.10	700.0
242.90	889.0
244.00	8367.0
245.00	1488.0
245.90	2370.0
247.10	1233.0
252.90	606.0
254.10	495.0
255.00	45336.0
256.10	6829.0
257.00	467.0
258.00	4324.0
258.90	1143.0
265.00	2055.0
273.00	1636.0
273.90	4853.0
275.00	17872.0
276.00	3459.0
277.10	2483.0
278.00	683.0
279.20	381.0
283.00	566.0
285.00	329.0
293.00	539.0
296.10	7506.0
296.80	469.0
297.20	540.0
314.00	442.0
314.80	1125.0
316.10	348.0
323.00	1281.0
333.10	400.0
334.00	1548.0
334.90	446.0
351.90	469.0
353.10	631.0
354.20	441.0
364.90	2343.0
365.80	470.0
371.80	472.0
372.20	323.0

Instrument :
BNA_M
ClientSampleId :
DFTPP11

383.20	330.0
403.10	325.0
423.10	2024.0
424.00	404.0
441.10	5178.0
442.00	29424.0
443.00	5794.0
444.20	301.0

Instrument :
BNA_M
ClientSampleId :
DFTPP11

m/z	Abundance
37.10	550.0
38.00	950.0
39.00	4455.0
40.00	639.0
44.00	388.0
49.10	816.0
50.10	23032.0
51.00	74920.0
52.10	3637.0
55.10	824.0
55.90	1218.0
57.00	2932.0
60.90	802.0
61.80	947.0
62.10	971.0
63.00	3827.0
63.80	309.0
65.00	978.0
68.00	1096.0
69.00	62216.0
70.20	426.0
72.90	1036.0
74.10	8208.0
75.00	13838.0
76.00	4800.0
77.00	80144.0
78.10	3837.0
78.90	6189.0
80.00	4668.0
81.00	7399.0
82.00	1395.0
83.00	1218.0
84.90	1504.0
85.90	2576.0
86.90	776.0
91.00	1156.0
92.00	1250.0
93.00	11648.0
94.10	844.0
97.10	504.0
98.00	11786.0
99.00	8289.0
99.90	606.0
101.00	4557.0
102.90	854.0
104.00	2246.0
105.00	2429.0
105.90	453.0
107.00	24216.0
108.00	4003.0
108.90	773.0
110.00	34976.0
111.00	5958.0
112.00	1469.0
113.10	516.0
116.10	1130.0
117.00	42752.0
118.00	2320.0
121.90	2016.0
123.00	2584.0
123.90	923.0
124.90	923.0
125.70	541.0
127.00	56776.0
128.00	4330.0
129.00	28184.0
130.00	1966.0
131.00	910.0
134.00	1286.0
135.00	2503.0
135.90	1010.0
137.00	1404.0
140.00	672.0
141.00	5412.0
142.00	1137.0
142.90	1195.0
145.70	806.0
147.00	3422.0
148.00	8233.0
149.10	1744.0
150.00	340.0
151.00	559.0
151.90	507.0
153.00	986.0
154.10	1185.0
154.80	1448.0
155.20	1355.0
156.00	3320.0
156.90	356.0
157.80	721.0
158.90	537.0

Instrument :
BNA_M
ClientSampleId :
DFTPP11

159.90	1692.0
161.00	1734.0
161.90	789.0
165.00	1525.0
166.00	1224.0
167.00	13943.0
168.00	6485.0
169.00	1200.0
171.80	868.0
172.80	1348.0
173.90	2029.0
175.10	4246.0
175.90	448.0
176.90	2721.0
178.00	792.0
178.90	6831.0
180.00	4306.0
180.90	1587.0
184.00	905.0
185.00	3604.0
186.00	18792.0
187.00	5978.0
188.00	787.0
189.00	2151.0
190.10	549.0
191.00	567.0
192.00	2147.0
192.80	1453.0
193.10	1402.0
194.00	896.0
195.90	3153.0
197.00	994.0
198.00	97048.0
199.00	5899.0
200.00	1158.0
201.30	333.0
203.10	1600.0
204.00	6756.0
205.00	10646.0
206.00	36040.0
207.00	4281.0
208.00	1781.0
210.00	546.0
210.90	1225.0
215.00	599.0
215.90	1359.0
217.00	13407.0
218.00	1819.0
220.20	308.0
221.10	1951.0
221.80	1044.0
223.10	1930.0
224.00	22992.0
225.00	5712.0
225.90	367.0
227.00	10788.0
227.90	2863.0
229.00	1955.0
231.00	1013.0
234.20	503.0
234.80	512.0
236.00	753.0
237.00	1155.0
239.00	576.0
240.00	413.0
241.10	674.0
242.10	1360.0
242.90	1378.0
244.00	15812.0
245.00	2156.0
246.00	5191.0
247.00	1330.0
249.10	762.0
253.00	596.0
254.10	822.0
255.00	74088.0
256.00	11282.0
256.80	756.0
258.00	7244.0
258.80	1648.0
265.00	4435.0
266.00	796.0
273.10	1925.0
274.10	5409.0
275.00	28664.0
276.00	3355.0
277.00	4621.0
278.00	607.0
278.90	364.0
283.00	463.0
285.00	874.0
293.00	1024.0
293.90	382.0
295.10	500.0
296.00	14645.0

Instrument :
BNA_M
ClientSampleId :
DFTPP11

297.00	2713.0
303.10	1158.0
313.10	304.0
314.00	369.0
315.00	1106.0
316.10	346.0
323.10	3238.0
323.90	639.0
326.90	945.0
328.90	444.0
332.10	403.0
333.20	390.0
334.10	2321.0
334.90	770.0
340.80	414.0
346.10	704.0
351.90	400.0
352.80	799.0
354.10	1291.0
365.00	6853.0
366.10	712.0
371.90	2421.0
401.90	890.0
403.10	878.0
404.00	492.0
421.20	891.0
421.90	387.0
423.00	3796.0
424.10	889.0
439.20	324.0
441.10	11983.0
442.10	64272.0
443.10	11019.0
443.80	544.0
444.20	533.0

Instrument :
BNA_M
ClientSampleId :
DFTPP11

m/z	Abundance
37.10	310.0
38.00	566.0
39.10	3205.0
44.10	386.0
49.10	365.0
50.00	16976.0
51.10	52952.0
52.10	2838.0
55.90	1733.0
57.00	2184.0
60.90	614.0
61.90	1016.0
62.90	1791.0
63.90	724.0
64.90	708.0
68.10	541.0
69.00	45192.0
71.10	357.0
72.90	1509.0
74.00	7926.0
75.00	8712.0
76.10	3147.0
77.00	51824.0
78.10	3457.0
78.90	5271.0
80.00	2906.0
80.90	4334.0
82.00	543.0
83.00	1125.0
84.00	370.0
84.80	540.0
86.00	1718.0
87.00	742.0
90.90	1298.0
92.00	1162.0
93.00	10961.0
96.00	333.0
97.90	9388.0
99.00	6441.0
100.10	536.0
100.90	2187.0
103.00	1185.0
104.00	2145.0
104.90	1431.0
107.00	19552.0
108.00	2994.0
110.00	20160.0
111.00	3801.0
112.00	836.0
115.90	1183.0
117.00	31824.0
118.00	2159.0
121.90	1825.0
123.00	2770.0
123.90	1265.0
125.00	790.0
127.00	42592.0
128.00	3572.0
129.00	22440.0
130.00	2203.0
130.90	549.0
131.90	450.0
133.20	411.0
133.90	787.0
134.90	2151.0
136.10	1186.0
136.90	1180.0
141.00	4377.0
142.00	992.0
143.20	633.0
146.00	1042.0
147.00	1918.0
147.90	5988.0
149.10	1450.0
151.00	784.0
153.00	1057.0
154.00	1105.0
155.00	2584.0
156.00	2378.0
156.80	703.0
158.10	834.0
159.10	1093.0
160.10	1381.0
160.90	975.0
161.80	435.0
165.00	1437.0
165.90	1081.0
167.00	9428.0
168.00	4122.0
169.10	647.0
171.00	549.0

Instrument :
BNA_M
ClientSampleId :
DFTPP11

171.90	917.0
173.00	1047.0
173.90	1751.0
175.10	2068.0
176.10	605.0
176.90	1142.0
178.00	550.0
179.00	6426.0
180.00	4042.0
181.00	1838.0
184.00	826.0
185.00	2603.0
186.00	17152.0
187.10	4642.0
189.00	1736.0
191.00	351.0
191.90	1353.0
192.90	1567.0
194.10	380.0
196.00	2548.0
196.80	435.0
197.10	355.0
197.90	82584.0
198.90	5012.0
199.90	444.0
201.40	369.0
202.90	770.0
204.00	5419.0
205.10	7739.0
206.00	28280.0
207.10	3751.0
207.90	1099.0
209.10	619.0
209.80	931.0
211.00	1274.0
212.20	363.0
215.00	573.0
215.80	968.0
216.90	12554.0
218.10	1018.0
221.10	4978.0
221.90	1234.0
223.00	1771.0
224.00	17296.0
225.00	4356.0
225.90	1187.0
227.00	12289.0
228.00	1468.0
228.90	1559.0
230.10	398.0
231.10	730.0
233.90	835.0
234.90	607.0
235.90	650.0
237.10	354.0
239.20	417.0
241.10	350.0
242.10	951.0
243.00	1230.0
244.00	13645.0
245.00	2366.0
245.90	3799.0
246.90	876.0
249.10	506.0
250.80	308.0
252.90	733.0
253.60	640.0
255.00	63096.0
256.00	9964.0
256.90	1262.0
257.90	5576.0
259.00	1278.0
264.90	2844.0
266.10	355.0
271.00	382.0
273.10	2879.0
274.00	6051.0
275.00	27008.0
276.00	4006.0
277.00	4207.0
278.00	412.0
280.80	380.0
282.80	485.0
283.90	355.0
285.10	1005.0
293.20	306.0
295.10	337.0
296.00	11347.0
297.20	2147.0
301.00	491.0
303.00	1124.0
303.80	359.0
313.80	400.0
315.00	2434.0
316.00	625.0

Instrument :
BNA_M
ClientSampleId :
DFTPP11

323.00	2638.0
324.00	446.0
327.00	956.0
328.00	617.0
332.90	499.0
334.10	2275.0
346.00	827.0
346.90	436.0
351.80	742.0
353.10	924.0
354.10	960.0
365.10	5977.0
366.00	1622.0
371.00	344.0
372.10	1136.0
373.00	528.0
389.90	416.0
402.00	1178.0
402.90	615.0
420.80	1082.0
421.90	313.0
423.10	2466.0
424.00	780.0
439.20	353.0
441.10	9321.0
442.10	64424.0
443.10	12185.0
444.10	1304.0

Instrument :
BNA_M
ClientSampleId :
DFTPP11

m/z	Abundance
72.90	304.0
207.20	577.0

Instrument :
BNA_M
ClientSampleId :
DFTPP11