

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM031319\
 Data File : BM019151.D
 Acq On : 13 Mar 2019 18:16
 Operator : JU/SJ
 Sample : K1861-02
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BF-07-031119

Quant Time: Mar 14 06:56:05 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM031119.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 11 17:30:46 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.66	152	196542	20.00	ng	0.00
21) Naphthalene-d8	10.45	136	818465	20.00	ng	0.00
39) Acenaphthene-d10	14.31	164	554916	20.00	ng	-0.01
64) Phenanthrene-d10	17.07	188	1319749	20.00	ng	0.00
76) Chrysene-d12	21.27	240	1357194	20.00	ng	0.00
87) Perylene-d12	23.50	264	1231091	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.26	112	1776944	146.42	ng	0.00
7) Phenol-d6	6.84	99	2684939	151.25	ng	0.00
23) Nitrobenzene-d5	8.83	82	1618751	93.49	ng	0.00
42) 2,4,6-Tribromophenol	15.82	330	1335894	163.05	ng	0.00
45) 2-Fluorobiphenyl	12.93	172	3484179	81.67	ng	0.00
79) Terphenyl-d14	19.71	244	6069489	88.78	ng	0.00
Target Compounds						
10) Phenol	6.86	94	64036	3.347	ng	Qvalue # 76
50) Dimethylphthalate	13.78	163	297600	6.269	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM031319\
Data File : BM019151.D
Acq On : 13 Mar 2019 18:16
Operator : JU/SJ
Sample : K1861-02
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
BF-07-031119

Quant Time: Mar 14 06:56:05 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM031119.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 11 17:30:46 2019
Response via : Initial Calibration

