

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM031323\
 Data File : BM039004.D
 Acq On : 14 Mar 2023 00:35
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020107

Quant Time: Mar 14 01:44:38 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM031323.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Mar 13 16:25:27 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.987	152	566743	20.000	ng/ul	0.00
20) Naphthalene-d8	10.798	136	2580966	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.628	164	1567905	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.369	188	2988204	20.000	ng/ul	0.00
79) Chrysene-d12	21.551	240	1934481	20.000	ng/ul	0.00
88) Perylene-d12	23.998	264	1802609	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.370	96	131179	8.072	ng/uL	0.00
4) Pyridine-d5	3.793	84	945857	20.575	ng/ul	0.00
7) Phenol-d5	7.140	99	1117191	20.478	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.310	67	730127	20.708	ng/ul	0.00
11) 2-Chlorophenol-d4	7.510	132	875500	21.557	ng/ul	0.00
15) 4-Methylphenol-d8	8.699	113	879280	20.601	ng/ul	0.00
21) Nitrobenzene-d5	9.152	128	455242	21.098	ng/ul	0.00
24) 2-Nitrophenol-d4	9.875	143	481837	22.164	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.416	165	853108	21.437	ng/ul	0.00
31) 4-Chloroaniline-d4	10.934	131	1281435	19.909	ng/ul	0.00
46) Dimethylphthalate-d6	14.034	166	2427269	20.439	ng/ul	0.00
49) Acenaphthylene-d8	14.316	160	2925176	21.209	ng/ul	0.00
54) 4-Nitrophenol-d4	14.816	143	458458	18.529	ng/ul	0.00
60) Fluorene-d10	15.616	176	2071322	19.990	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.733	200	370097	19.554	ng/ul	0.00
73) Anthracene-d10	17.469	188	2993141	20.775	ng/ul	0.00
81) Pyrene-d10	19.757	212	2874722	25.382	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.839	264	1958906	21.168	ng/ul	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.411	88	145055	8.145	ng/uL	94
5) Pyridine	3.817	79	1006630	20.605	ng/ul	98
6) Benzaldehyde	7.122	77	648534	24.259	ng/ul	99
8) Phenol	7.169	94	1185818	20.495	ng/ul	99
10) Bis(2-Chloroethyl)ether	7.405	93	957595	20.778	ng/ul	98
12) 2-Chlorophenol	7.546	128	891495	21.262	ng/ul	98
13) 2-Methylphenol	8.428	108	886481	20.772	ng/ul	98
14) 2,2'-oxybis(1-Chloropr...	8.522	45	1270279	20.447	ng/ul	100
16) Acetophenone	8.816	105	1495432	21.182	ng/ul	99
17) N-Nitroso-di-n-propyla...	8.804	70	787021	21.596	ng/ul	99
18) 4-Methylphenol	8.763	108	985745	20.783	ng/ul	94
19) Hexachloroethane	9.075	117	379190	21.414	ng/ul	96
22) Nitrobenzene	9.199	77	1127296	20.635	ng/ul	97
23) Isophorone	9.728	82	2175134	20.847	ng/ul	100
25) 2-Nitrophenol	9.910	139	516243	21.692	ng/ul	99
26) 2,4-Dimethylphenol	9.969	107	1065771	21.152	ng/ul	99
27) Bis(2-Chloroethoxy)met...	10.210	93	1345208	20.738	ng/ul	100
29) 2,4-Dichlorophenol	10.440	162	850982	21.269	ng/ul	98
30) Naphthalene	10.851	128	3006971	20.593	ng/ul	100
32) 4-Chloroaniline	10.957	127	1286637	19.862	ng/ul	100
33) Hexachlorobutadiene	11.140	225	504662	21.081	ng/ul	99
34) Caprolactam	11.740	113	154423	11.305	ng/ul	98
35) 4-Chloro-3-methylphenol	12.075	107	937820	20.573	ng/ul	98
36) 2-Methylnaphthalene	12.457	142	1937488	20.444	ng/ul	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM031323\
 Data File : BM039004.D
 Acq On : 14 Mar 2023 00:35
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 21 Sample Multiplier: 1


Instrument :
 BNA_M
 ClientSampleId :
 SSTD020107


Quant Time: Mar 14 01:44:38 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM031323.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Mar 13 16:25:27 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.675	142	1950778	20.491	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.822	216	977126	21.245	ng/ul	99
40) Hexachlorocyclopentadiene	12.804	237	643767	21.448	ng/ul	100
41) 2,4,6-Trichlorophenol	13.057	196	652831	21.702	ng/ul	98
42) 2,4,5-Trichlorophenol	13.128	196	714064	21.831	ng/ul	98
43) 1,1'-Biphenyl	13.463	154	2598791	20.857	ng/ul	99
44) 2-Chloronaphthalene	13.504	162	2021597	20.905	ng/ul	100
45) 2-Nitroaniline	13.704	65	613005	21.273	ng/ul	98
47) Dimethylphthalate	14.086	163	2436530	20.258	ng/ul	100
48) 2,6-Dinitrotoluene	14.204	165	485063	21.884	ng/ul	99
50) Acenaphthylene	14.345	152	3173974	20.693	ng/ul	98
51) 3-Nitroaniline	14.528	138	520317	20.255	ng/ul	96
52) Acenaphthene	14.686	153	2166328	20.358	ng/ul	99
53) 2,4-Dinitrophenol	14.728	184	247589	17.811	ng/ul	98
55) 4-Nitrophenol	14.828	109	355824	18.505	ng/ul	97
56) Dibenzofuran	15.022	168	2940749	20.133	ng/ul	99
57) 2,4-Dinitrotoluene	14.986	165	677157	20.639	ng/ul	97
58) 2,3,4,6-Tetrachlorophenol	15.245	232	575265	20.653	ng/ul	98
59) Diethylphthalate	15.445	149	2438724	20.074	ng/ul	99
61) Fluorene	15.669	166	2436489	20.095	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.663	204	1184023	20.418	ng/ul	97
63) 4-Nitroaniline	15.692	138	487769	19.953	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.745	198	370832	19.715	ng/ul	99
67) N-Nitrosodiphenylamine	15.875	169	1988773	21.796	ng/ul	99
68) 4-Bromophenyl-phenylether	16.557	248	658222	21.859	ng/ul	97
69) Hexachlorobenzene	16.675	284	740867	21.110	ng/ul	98
70) Atrazine	16.833	200	638807	20.495	ng/ul	98
71) Pentachlorophenol	17.016	266	445819	19.837	ng/ul	99
72) Phenanthrene	17.410	178	3465727	20.422	ng/ul	99
74) Anthracene	17.504	178	3556718	20.643	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.428	216	984630	23.053	ng/uL	99
76) Pentachlorobenzene	14.939	250	950789	22.431	ng/uL	99
77) Carbazole	17.774	167	3093884	20.004	ng/ul	100
78) Di-n-butylphthalate	18.339	149	3750280	20.761	ng/ul	100
80) Fluoranthene	19.421	202	3437918	25.981	ng/ul	98
82) Pyrene	19.786	202	3573854	25.306	ng/ul	97
83) Butylbenzylphthalate	20.686	149	1297853	23.229	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.468	252	877027	21.139	ng/ul	99
85) Benzo(a)anthracene	21.533	228	2855637	21.235	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.462	149	1788480	21.497	ng/ul	99
87) Chrysene	21.586	228	2679171	20.732	ng/ul	100
89) Di-n-octyl phthalate	22.404	149	2663571	20.106	ng/ul	100
90) Benzo(b)fluoranthene	23.245	252	2537989	21.056	ng/ul	100
91) Benzo(k)fluoranthene	23.298	252	2454418	20.602	ng/ul	99
93) Benzo(a)pyrene	23.886	252	2187733	20.980	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.544	276	2781663	23.133	ng/ul	98
95) Dibenzo(a,h)anthracene	26.562	278	2356926	23.264	ng/ul	100
96) Benzo(g,h,i)perylene	27.327	276	2234940	23.486	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Mar 14 01:44:38 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM031323.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Mar 13 16:25:27 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 8.145 ng/uL
RT: 3.411 min Scan# 21
Delta R.T. 0.006 min
Lab File: BM039004.D
Acq: 14 Mar 2023 00:35




Instrument :
BNA_M
ClientSampleId :
SSTD020107


Tgt Ion: 88 Resp: 145055
Ion Ratio Lower Upper
88 100
43 26.8 24.0 36.0
58 69.7 59.8 89.8



#5
Pyridine
Concen: 20.605 ng/uL
RT: 3.817 min Scan# 90
Delta R.T. 0.000 min
Lab File: BM039004.D
Acq: 14 Mar 2023 00:35

Tgt Ion: 79 Resp: 1006630
Ion Ratio Lower Upper
79 100
52 69.5 56.6 85.0
51 32.9 27.8 41.6









#6
Benzaldehyde
Concen: 24.259 ng/ul
RT: 7.122 min Scan# 652
Delta R.T. 0.000 min
Lab File: BM039004.D
Acq: 14 Mar 2023 00:35




Instrument :
BNA_M
ClientSampleId :
SSTD020107


Tgt Ion: 77 Resp: 648534
Ion Ratio Lower Upper
77 100
105 97.8 77.8 116.8
106 95.2 75.4 113.2



#8
Phenol
Concen: 20.495 ng/ul
RT: 7.169 min Scan# 660
Delta R.T. 0.000 min
Lab File: BM039004.D
Acq: 14 Mar 2023 00:35

Tgt Ion: 94 Resp: 1185818
Ion Ratio Lower Upper
94 100
65 28.0 22.7 34.1
66 35.0 28.2 42.2









#10
Bis(2-Chloroethyl)ether
Concen: 20.778 ng/ul
RT: 7.405 min Scan# 701
Delta R.T. -0.006 min
Lab File: BM039004.D
Acq: 14 Mar 2023 00:35




Instrument :
BNA_M
ClientSampleId :
SSTD020107


Tgt Ion: 93 Resp: 957595
Ion Ratio Lower Upper
93 100
63 78.3 61.1 91.7
95 32.9 25.4 38.2



#12
2-Chlorophenol
Concen: 21.262 ng/ul
RT: 7.546 min Scan# 724
Delta R.T. 0.000 min
Lab File: BM039004.D
Acq: 14 Mar 2023 00:35

Tgt Ion: 128 Resp: 891495
Ion Ratio Lower Upper
128 100
64 47.4 39.7 59.5
130 32.4 26.2 39.2









#13
2-Methylphenol
Concen: 20.772 ng/ul
RT: 8.428 min Scan# 874
Delta R.T. 0.000 min
Lab File: BM039004.D
Acq: 14 Mar 2023 00:35




Instrument :
BNA_M
ClientSampleId :
SSTD020107


Tgt Ion:108 Resp: 886481
Ion Ratio Lower Upper
108 100
107 87.4 71.4 107.0



#14
2,2'-oxybis(1-Chloropropane)
Concen: 20.447 ng/ul
RT: 8.522 min Scan# 890
Delta R.T. 0.000 min
Lab File: BM039004.D
Acq: 14 Mar 2023 00:35

Tgt Ion: 45 Resp: 1270279
Ion Ratio Lower Upper
45 100
77 14.9 11.9 17.9
79 11.4 9.0 13.6









#16
 Acetophenone
 Concen: 21.182 ng/ul
 RT: 8.816 min Scan# 940
 Delta R.T. 0.000 min
 Lab File: BM039004.D
 Acq: 14 Mar 2023 00:35




Instrument :
 BNA_M
 ClientSampleId :
 SST020107


Tgt Ion:105 Resp: 1495432
 Ion Ratio Lower Upper
 105 100
 77 78.4 62.6 93.8
 51 26.6 21.9 32.9



#17
 N-Nitroso-di-n-propylamine
 Concen: 21.596 ng/ul
 RT: 8.804 min Scan# 938
 Delta R.T. 0.000 min
 Lab File: BM039004.D
 Acq: 14 Mar 2023 00:35

Tgt Ion: 70 Resp: 787021
 Ion Ratio Lower Upper
 70 100
 42 55.1 44.7 67.1
 101 9.4 7.3 10.9
 130 19.8 15.5 23.3









#18
4-Methylphenol
Concen: 20.783 ng/ul
RT: 8.763 min Scan# 91
Delta R.T. 0.006 min
Lab File: BM039004.D
Acq: 14 Mar 2023 00:35




Instrument :
BNA_M
ClientSampleId :
SSTD020107


Tgt Ion:108 Resp: 985745
Ion Ratio Lower Upper
108 100
107 106.5 90.0 135.0



#19
Hexachloroethane
Concen: 21.414 ng/ul
RT: 9.075 min Scan# 984
Delta R.T. 0.000 min
Lab File: BM039004.D
Acq: 14 Mar 2023 00:35

Tgt Ion:117 Resp: 379190
Ion Ratio Lower Upper
117 100
201 94.2 71.6 107.4
199 58.6 45.1 67.7









#22
Nitrobenzene
Concen: 20.635 ng/ul
RT: 9.199 min Scan# 1005
Delta R.T. 0.000 min
Lab File: BM039004.D
Acq: 14 Mar 2023 00:35




Instrument :
BNA_M
ClientSampleId :
SSTD020107


Tgt Ion: 77 Resp: 1127296
Ion Ratio Lower Upper
77 100
123 45.1 34.2 51.2
65 13.1 10.6 16.0



#23
Isophorone
Concen: 20.847 ng/ul
RT: 9.728 min Scan# 1095
Delta R.T. 0.000 min
Lab File: BM039004.D
Acq: 14 Mar 2023 00:35

Tgt Ion: 82 Resp: 2175134
Ion Ratio Lower Upper
82 100
95 6.5 5.3 7.9
138 16.6 13.3 19.9









#25
2-Nitrophenol
Concen: 21.692 ng/ul
RT: 9.910 min Scan# 1126
Delta R.T. 0.000 min
Lab File: BM039004.D
Acq: 14 Mar 2023 00:35




Instrument :
BNA_M
ClientSampleId :
SSTD020107


Tgt Ion:139 Resp: 516243
Ion Ratio Lower Upper
139 100
65 52.9 43.4 65.0
109 29.3 23.1 34.7



#26
2,4-Dimethylphenol
Concen: 21.152 ng/ul
RT: 9.969 min Scan# 1136
Delta R.T. 0.000 min
Lab File: BM039004.D
Acq: 14 Mar 2023 00:35

Tgt Ion:107 Resp: 1065771
Ion Ratio Lower Upper
107 100
121 50.1 40.6 60.8
122 87.5 69.7 104.5









#27
 Bis(2-Chloroethoxy)methane
 Concen: 20.738 ng/ul
 RT: 10.210 min Scan# 1177
 Delta R.T. 0.000 min
 Lab File: BM039004.D
 Acq: 14 Mar 2023 00:35

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020107

Tgt Ion: 93 Resp: 1345208




Ion	Ratio	Lower	Upper
93	100		
95	32.3	25.9	38.9
123	11.3	9.1	13.7




#29
 2,4-Dichlorophenol
 Concen: 21.269 ng/ul
 RT: 10.440 min Scan# 1216
 Delta R.T. -0.006 min
 Lab File: BM039004.D
 Acq: 14 Mar 2023 00:35

Tgt Ion: 162 Resp: 850982

Ion	Ratio	Lower	Upper
162	100		
164	64.9	53.1	79.7
98	37.8	29.6	44.4









#30
Naphthalene
Concen: 20.593 ng/ul
RT: 10.851 min Scan# 11
Delta R.T. 0.000 min
Lab File: BM039004.D
Acq: 14 Mar 2023 00:35




Instrument :
BNA_M
ClientSampleId :
SSTD020107


Tgt Ion:128 Resp: 3006971
Ion Ratio Lower Upper
128 100
129 11.0 8.6 13.0
127 13.0 10.3 15.5



#32
4-Chloroaniline
Concen: 19.862 ng/ul
RT: 10.957 min Scan# 1304
Delta R.T. -0.006 min
Lab File: BM039004.D
Acq: 14 Mar 2023 00:35

Tgt Ion:127 Resp: 1286637
Ion Ratio Lower Upper
127 100
129 32.3 25.9 38.9









#33
Hexachlorobutadiene
Concen: 21.081 ng/ul
RT: 11.140 min Scan# 11
Delta R.T. 0.000 min
Lab File: BM039004.D
Acq: 14 Mar 2023 00:35




Instrument :
BNA_M
ClientSampleId :
SSTD020107


Tgt Ion:225 Resp: 504662
Ion Ratio Lower Upper
225 100
223 60.5 49.8 74.6
227 62.6 50.2 75.4



#34
Caprolactam
Concen: 11.305 ng/ul
RT: 11.740 min Scan# 1437
Delta R.T. 0.006 min
Lab File: BM039004.D
Acq: 14 Mar 2023 00:35

Tgt Ion:113 Resp: 154423
Ion Ratio Lower Upper
113 100
55 174.3 141.8 212.8
56 127.0 102.2 153.2









#35
4-Chloro-3-methylphenol
Concen: 20.573 ng/ul
RT: 12.075 min Scan# 1494
Delta R.T. -0.006 min
Lab File: BM039004.D
Acq: 14 Mar 2023 00:35




Instrument :
BNA_M
ClientSampleId :
SSTD020107


Tgt Ion:107 Resp: 937820
Ion Ratio Lower Upper
107 100
144 25.9 19.9 29.9
142 79.5 62.6 93.8



#36
2-Methylnaphthalene
Concen: 20.444 ng/ul
RT: 12.457 min Scan# 1559
Delta R.T. 0.000 min
Lab File: BM039004.D
Acq: 14 Mar 2023 00:35

Tgt Ion:142 Resp: 1937488
Ion Ratio Lower Upper
142 100
141 88.6 70.7 106.1









#37
1-Methylnaphthalene
Concen: 20.491 ng/ul
RT: 12.675 min Scan# 1596
Delta R.T. 0.000 min
Lab File: BM039004.D
Acq: 14 Mar 2023 00:35




Instrument :
BNA_M
ClientSampleId :
SSTD020107


Tgt Ion:142 Resp: 1950778
Ion Ratio Lower Upper
142 100
141 91.5 73.1 109.7
116 3.9 3.0 4.6



#39
1,2,4,5-Tetrachlorobenzene
Concen: 21.245 ng/ul
RT: 12.822 min Scan# 1621
Delta R.T. 0.000 min
Lab File: BM039004.D
Acq: 14 Mar 2023 00:35

Tgt Ion:216 Resp: 977126
Ion Ratio Lower Upper
216 100
214 79.6 62.6 93.8
179 21.0 17.0 25.4
108 19.5 15.9 23.9









#40
Hexachlorocyclopentadiene
Concen: 21.448 ng/ul
RT: 12.804 min Scan# 1618
Delta R.T. 0.000 min
Lab File: BM039004.D
Acq: 14 Mar 2023 00:35




Instrument :
BNA_M
ClientSampleId :
SSTD020107


Tgt Ion:237 Resp: 643767
Ion Ratio Lower Upper
237 100
235 62.8 50.0 75.0
272 13.0 10.7 16.1



#41
2,4,6-Trichlorophenol
Concen: 21.702 ng/ul
RT: 13.057 min Scan# 1661
Delta R.T. -0.006 min
Lab File: BM039004.D
Acq: 14 Mar 2023 00:35

Tgt Ion:196 Resp: 652831
Ion Ratio Lower Upper
196 100
198 96.8 76.1 114.1
200 30.1 25.4 38.0





#42

2,4,5-Trichlorophenol

Concen: 21.831 ng/ul

RT: 13.128 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM039004.D

Acq: 14 Mar 2023 00:35

Instrument :

BNA_M

ClientSampleId :

SSTD020107





Tgt Ion:196 Resp: 714064

Ion Ratio Lower Upper

196 100

198 95.4 78.2 117.2

200 30.6 24.9 37.3



#43

1,1'-Biphenyl

Concen: 20.857 ng/ul

RT: 13.463 min Scan# 1730

Delta R.T. 0.000 min

Lab File: BM039004.D

Acq: 14 Mar 2023 00:35




Tgt Ion:154 Resp: 2598791


Ion Ratio Lower Upper

154 100

153 39.6 32.0 48.0

76 13.9 11.5 17.3









#44
2-Chloronaphthalene
Concen: 20.905 ng/ul
RT: 13.504 min Scan# 1737
Delta R.T. 0.000 min
Lab File: BM039004.D
Acq: 14 Mar 2023 00:35




Instrument :
BNA_M
ClientSampleId :
SSTD020107


Tgt Ion:162 Resp: 2021597
Ion Ratio Lower Upper
162 100
164 33.1 26.6 39.8
127 40.0 32.2 48.2



#45
2-Nitroaniline
Concen: 21.273 ng/ul
RT: 13.704 min Scan# 1771
Delta R.T. -0.006 min
Lab File: BM039004.D
Acq: 14 Mar 2023 00:35

Tgt Ion: 65 Resp: 613005
Ion Ratio Lower Upper
65 100
92 67.3 53.3 79.9
138 100.2 82.9 124.3





#47

Dimethylphthalate

Concen: 20.258 ng/ul

RT: 14.086 min Scan# 1836

Delta R.T. 0.000 min

Lab File: BM039004.D

Acq: 14 Mar 2023 00:35

Instrument :

BNA_M

ClientSampleId :

SSTD020107





Tgt Ion:163 Resp: 2436530

Ion Ratio Lower Upper

163 100

194 4.0 3.1 4.7

164 9.9 8.0 12.0



#48

2,6-Dinitrotoluene

Concen: 21.884 ng/ul

RT: 14.204 min Scan# 1856

Delta R.T. 0.000 min

Lab File: BM039004.D

Acq: 14 Mar 2023 00:35




Tgt Ion:165 Resp: 485063


Ion Ratio Lower Upper

165 100

89 61.4 50.2 75.4

121 22.2 18.0 27.0





#50

Acenaphthylene

Concen: 20.693 ng/ul

RT: 14.345 min Scan# 1

Delta R.T. -0.006 min

Lab File: BM039004.D

Acq: 14 Mar 2023 00:35

Instrument :

BNA_M

ClientSampleId :

SSTD020107





Tgt Ion:152 Resp: 3173974

Ion Ratio Lower Upper

152 100

151 20.1 15.6 23.4

153 13.5 10.2 15.4



#51

3-Nitroaniline

Concen: 20.255 ng/ul

RT: 14.528 min Scan# 1911

Delta R.T. 0.000 min

Lab File: BM039004.D

Acq: 14 Mar 2023 00:35




Tgt Ion:138 Resp: 520317


Ion Ratio Lower Upper

138 100

108 10.2 8.9 13.3

92 118.0 97.9 146.9





#52

Acenaphthene

Concen: 20.358 ng/ul

RT: 14.686 min Scan# 1938

Delta R.T. -0.006 min

Lab File: BM039004.D

Acq: 14 Mar 2023 00:35

Instrument :

BNA_M

ClientSampleId :

SSTD020107





Tgt Ion:153 Resp: 2166328

Ion Ratio Lower Upper

153 100

152 48.2 38.1 57.1

154 88.8 71.4 107.0



#53

2,4-Dinitrophenol

Concen: 17.811 ng/ul

RT: 14.728 min Scan# 1945

Delta R.T. 0.000 min

Lab File: BM039004.D

Acq: 14 Mar 2023 00:35




Tgt Ion:184 Resp: 247589


Ion Ratio Lower Upper

184 100

63 79.0 64.7 97.1

154 64.3 50.8 76.2









#55
4-Nitrophenol
Concen: 18.505 ng/ul
RT: 14.828 min Scan# 1962
Delta R.T. 0.000 min
Lab File: BM039004.D
Acq: 14 Mar 2023 00:35




Instrument :
BNA_M
ClientSampleId :
SSTD020107


Tgt Ion:109 Resp: 355824
Ion Ratio Lower Upper
109 100
139 132.2 103.4 155.0
65 140.7 108.7 163.1



#56
Dibenzofuran
Concen: 20.133 ng/ul
RT: 15.022 min Scan# 1995
Delta R.T. 0.000 min
Lab File: BM039004.D
Acq: 14 Mar 2023 00:35

Tgt Ion:168 Resp: 2940749
Ion Ratio Lower Upper
168 100
139 37.9 30.7 46.1





#57

2,4-Dinitrotoluene

Concen: 20.639 ng/ul

RT: 14.986 min Scan# 1989

Delta R.T. 0.000 min

Lab File: BM039004.D

Acq: 14 Mar 2023 00:35

Instrument :

BNA_M

ClientSampleId :

SSTD020107





Tgt Ion:165 Resp: 677157

Ion Ratio Lower Upper

165 100

63 46.0 38.7 58.1

182 2.5 2.1 3.1



#58

2,3,4,6-Tetrachlorophenol

Concen: 20.653 ng/ul

RT: 15.245 min Scan# 2033

Delta R.T. 0.000 min

Lab File: BM039004.D

Acq: 14 Mar 2023 00:35

Tgt Ion:232 Resp: 575265




Ion Ratio Lower Upper


232 100

131 49.6 41.3 61.9

130 2.6 2.2 3.4

166 32.0 26.5 39.7





#59

Diethylphthalate

Concen: 20.074 ng/ul

RT: 15.445 min Scan# 2067

Delta R.T. 0.000 min

Lab File: BM039004.D

Acq: 14 Mar 2023 00:35

Instrument :

BNA_M

ClientSampleId :

SSTD020107





Tgt Ion:149 Resp: 2438724

Ion Ratio Lower Upper

149 100

177 21.5 16.9 25.3

150 12.6 9.8 14.8



#61

Fluorene

Concen: 20.095 ng/ul

RT: 15.669 min Scan# 2105

Delta R.T. 0.000 min

Lab File: BM039004.D

Acq: 14 Mar 2023 00:35




Tgt Ion:166 Resp: 2436489


Ion Ratio Lower Upper

166 100

165 96.9 77.5 116.3

167 13.5 11.0 16.6









#62
4-Chlorophenyl-phenylether
Concen: 20.418 ng/ul
RT: 15.663 min Scan# 2104
Delta R.T. -0.006 min
Lab File: BM039004.D
Acq: 14 Mar 2023 00:35




Instrument :
BNA_M
ClientSampleId :
SSTD020107


Tgt Ion:204 Resp: 1184023
Ion Ratio Lower Upper
204 100
206 32.3 26.9 40.3
141 60.4 46.6 69.8



#63
4-Nitroaniline
Concen: 19.953 ng/ul
RT: 15.692 min Scan# 2109
Delta R.T. 0.006 min
Lab File: BM039004.D
Acq: 14 Mar 2023 00:35

Tgt Ion:138 Resp: 487769
Ion Ratio Lower Upper
138 100
92 51.7 42.7 64.1
108 78.7 63.5 95.3









#66
4,6-Dinitro-2-methylphenol
Concen: 19.715 ng/ul
RT: 15.745 min Scan# 2118
Delta R.T. 0.000 min
Lab File: BM039004.D
Acq: 14 Mar 2023 00:35




Instrument :
BNA_M
ClientSampleId :
SSTD020107


Tgt Ion:198 Resp: 370832
Ion Ratio Lower Upper
198 100
182 4.3 3.6 5.4
77 28.8 23.7 35.5



#67
N-Nitrosodiphenylamine
Concen: 21.796 ng/ul
RT: 15.875 min Scan# 2140
Delta R.T. -0.006 min
Lab File: BM039004.D
Acq: 14 Mar 2023 00:35

Tgt Ion:169 Resp: 1988773
Ion Ratio Lower Upper
169 100
168 66.1 53.5 80.3
167 36.3 29.0 43.4









#68
4-Bromophenyl-phenylether
Concen: 21.859 ng/ul
RT: 16.557 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM039004.D
Acq: 14 Mar 2023 00:35




Instrument :
BNA_M
ClientSampleId :
SSTD020107


Tgt Ion:248 Resp: 658222
Ion Ratio Lower Upper
248 100
250 98.8 77.3 115.9
141 81.8 62.6 93.8



#69
Hexachlorobenzene
Concen: 21.110 ng/ul
RT: 16.675 min Scan# 2276
Delta R.T. 0.000 min
Lab File: BM039004.D
Acq: 14 Mar 2023 00:35

Tgt Ion:284 Resp: 740867
Ion Ratio Lower Upper
284 100
142 37.3 31.0 46.6
249 30.8 24.9 37.3









#70
Atrazine
Concen: 20.495 ng/ul
RT: 16.833 min Scan# 21
Delta R.T. 0.006 min
Lab File: BM039004.D
Acq: 14 Mar 2023 00:35




Instrument :
BNA_M
ClientSampleId :
SSTD020107


Tgt Ion:200 Resp: 638807
Ion Ratio Lower Upper
200 100
173 26.9 21.5 32.3
215 49.5 38.1 57.1



#71
Pentachlorophenol
Concen: 19.837 ng/ul
RT: 17.016 min Scan# 2334
Delta R.T. 0.000 min
Lab File: BM039004.D
Acq: 14 Mar 2023 00:35

Tgt Ion:266 Resp: 445819
Ion Ratio Lower Upper
266 100
264 61.4 50.0 75.0
268 64.0 50.5 75.7





#72

Phenanthrene

Concen: 20.422 ng/ul

RT: 17.410 min Scan# 2402

Delta R.T. -0.006 min

Lab File: BM039004.D

Acq: 14 Mar 2023 00:35

Instrument :

BNA_M

ClientSampleId :

SSTD020107





Tgt Ion:178 Resp: 3465727

Ion Ratio Lower Upper

178 100

179 15.8 12.1 18.1

176 19.6 15.7 23.5



#74

Anthracene

Concen: 20.643 ng/ul

RT: 17.504 min Scan# 2417

Delta R.T. 0.000 min

Lab File: BM039004.D

Acq: 14 Mar 2023 00:35




Tgt Ion:178 Resp: 3556718


Ion Ratio Lower Upper

178 100

179 15.8 12.4 18.6

176 19.2 15.2 22.8









#75
1,2,3,4-Tetrachlorobenzene
Concen: 23.053 ng/uL
RT: 13.428 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM039004.D
Acq: 14 Mar 2023 00:35




Instrument :
BNA_M
ClientSampleId :
SSTD020107


Tgt Ion:216 Resp: 984630
Ion Ratio Lower Upper
216 100
214 77.7 62.7 94.1
218 48.3 37.8 56.8



#76
Pentachlorobenzene
Concen: 22.431 ng/uL
RT: 14.939 min Scan# 1981
Delta R.T. 0.000 min
Lab File: BM039004.D
Acq: 14 Mar 2023 00:35

Tgt Ion:250 Resp: 950789
Ion Ratio Lower Upper
250 100
248 62.8 50.5 75.7
252 64.0 50.5 75.7





#77

Carbazole

Concen: 20.004 ng/ul

RT: 17.774 min Scan# 2463

Delta R.T. 0.000 min

Lab File: BM039004.D

Acq: 14 Mar 2023 00:35

Instrument :





BNA_M

ClientSampleId :

SSTD020107

Tgt Ion:167 Resp: 3093884

Ion	Ratio	Lower	Upper
167	100		
166	21.2	17.0	25.4
139	13.4	10.4	15.6



#78

Di-n-butylphthalate

Concen: 20.761 ng/ul

RT: 18.339 min Scan# 2559




Delta R.T. 0.000 min


Lab File: BM039004.D

Acq: 14 Mar 2023 00:35

Tgt Ion:149 Resp: 3750280

Ion	Ratio	Lower	Upper
149	100		
150	9.1	7.4	11.0
104	5.2	4.3	6.5









#80
 Fluoranthene
 Concen: 25.981 ng/ul
 RT: 19.421 min Scan# 21
 Delta R.T. -0.006 min
 Lab File: BM039004.D
 Acq: 14 Mar 2023 00:35




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020107


Tgt Ion:202 Resp: 3437918
 Ion Ratio Lower Upper
 202 100
 101 13.2 9.9 14.9
 100 10.0 7.4 11.2



#82
 Pyrene
 Concen: 25.306 ng/ul
 RT: 19.786 min Scan# 2805
 Delta R.T. -0.006 min
 Lab File: BM039004.D
 Acq: 14 Mar 2023 00:35

Tgt Ion:202 Resp: 3573854
 Ion Ratio Lower Upper
 202 100
 101 15.4 11.4 17.2
 100 12.5 9.3 13.9





#83

Butylbenzylphthalate

Concen: 23.229 ng/ul

RT: 20.686 min Scan# 2958

Delta R.T. 0.000 min

Lab File: BM039004.D

Acq: 14 Mar 2023 00:35

Instrument :

BNA_M

ClientSampleId :

SSTD020107





Tgt Ion:149 Resp: 1297853

Ion Ratio Lower Upper

149 100

91 73.7 61.5 92.3

206 20.4 15.8 23.8



#84

3,3'-Dichlorobenzidine

Concen: 21.139 ng/ul

RT: 21.468 min Scan# 3091

Delta R.T. 0.000 min

Lab File: BM039004.D

Acq: 14 Mar 2023 00:35




Tgt Ion:252 Resp: 877027


Ion Ratio Lower Upper

252 100

254 65.0 52.6 79.0

126 11.3 9.8 14.6





#85

Benzo(a)anthracene

Concen: 21.235 ng/ul

RT: 21.533 min Scan# 3102

Delta R.T. 0.000 min

Lab File: BM039004.D

Acq: 14 Mar 2023 00:35

Instrument :

BNA_M

ClientSampleId :

SSTD020107





Tgt Ion:228 Resp: 2855637

Ion Ratio Lower Upper

228 100

229 19.5 15.9 23.9

226 26.9 21.8 32.8



#86

Bis(2-ethylhexyl)phthalate

Concen: 21.497 ng/ul

RT: 21.462 min Scan# 3090

Delta R.T. 0.000 min

Lab File: BM039004.D

Acq: 14 Mar 2023 00:35




Tgt Ion:149 Resp: 1788480


Ion Ratio Lower Upper

149 100

167 27.6 21.8 32.8

279 4.1 2.8 4.2





#87

Chrysene

Concen: 20.732 ng/ul

RT: 21.586 min Scan# 3111

Delta R.T. -0.006 min

Lab File: BM039004.D

Acq: 14 Mar 2023 00:35

Instrument :

BNA_M

ClientSampleId :

SSTD020107





Tgt Ion:228 Resp: 2679171

Ion Ratio Lower Upper

228 100

226 29.9 24.1 36.1

229 19.8 15.8 23.6



#89

Di-n-octyl phthalate

Concen: 20.106 ng/ul




RT: 22.404 min Scan# 3250


Delta R.T. -0.006 min

Lab File: BM039004.D

Acq: 14 Mar 2023 00:35

Tgt Ion:149 Resp: 2663571









#90
Benzo(b)fluoranthene
Concen: 21.056 ng/ul
RT: 23.245 min Scan# 31
Delta R.T. -0.006 min
Lab File: BM039004.D
Acq: 14 Mar 2023 00:35




Instrument :
BNA_M
ClientSampleId :
SSTD020107


Tgt Ion:252 Resp: 2537989
Ion Ratio Lower Upper
252 100
253 22.1 17.4 26.2
125 11.5 9.2 13.8



#91
Benzo(k)fluoranthene
Concen: 20.602 ng/ul
RT: 23.298 min Scan# 3402
Delta R.T. -0.006 min
Lab File: BM039004.D
Acq: 14 Mar 2023 00:35

Tgt Ion:252 Resp: 2454418
Ion Ratio Lower Upper
252 100
253 21.6 17.5 26.3
125 11.3 8.8 13.2





#93

Benzo(a)pyrene

Concen: 20.980 ng/ul

RT: 23.886 min Scan# 3

Delta R.T. -0.006 min

Lab File: BM039004.D

Acq: 14 Mar 2023 00:35

Instrument :

BNA_M

ClientSampleId :

SSTD020107





Tgt Ion:252 Resp: 2187733

Ion Ratio Lower Upper

252 100

253 21.8 17.4 26.2

125 13.0 9.9 14.9



#94

Indeno(1,2,3-cd)pyrene

Concen: 23.133 ng/ul

RT: 26.544 min Scan# 3954

Delta R.T. -0.006 min

Lab File: BM039004.D

Acq: 14 Mar 2023 00:35




Tgt Ion:276 Resp: 2781663


Ion Ratio Lower Upper

276 100

138 26.9 23.1 34.7

277 24.6 19.9 29.9





#95

Dibenzo(a,h)anthracene

Concen: 23.264 ng/ul

RT: 26.562 min Scan# 3958

Delta R.T. -0.006 min

Lab File: BM039004.D

Acq: 14 Mar 2023 00:35

Instrument :

BNA_M

ClientSampleId :

SSTD020107





Tgt Ion:278 Resp: 2356926

Ion Ratio Lower Upper

278 100

139 18.5 14.6 22.0

279 23.7 19.0 28.4



#96

Benzo(g,h,i)perylene

Concen: 23.486 ng/ul

RT: 27.327 min Scan# 4087

Delta R.T. -0.006 min

Lab File: BM039004.D

Acq: 14 Mar 2023 00:35

Tgt Ion:276 Resp: 2234940

Ion Ratio Lower Upper

276 100

138 24.4 20.1 30.1

277 23.4 19.0 28.6

