

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM031323\
 Data File : BM039008.D
 Acq On : 14 Mar 2023 11:28
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020108

Quant Time: Mar 15 01:04:47 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM031323.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Mar 13 16:25:27 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.987	152	482864	20.000	ng/u1	0.00
20) Naphthalene-d8	10.798	136	2178024	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.627	164	1298312	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.368	188	2420351	20.000	ng/u1	0.00
79) Chrysene-d12	21.551	240	1634336	20.000	ng/u1	0.00
88) Perylene-d12	23.997	264	1596852	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.375	96	115015	8.306	ng/uL	0.00
4) Pyridine-d5	3.793	84	802221	20.482	ng/u1	0.00
7) Phenol-d5	7.140	99	940107	20.225	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.310	67	622141	20.711	ng/u1	0.00
11) 2-Chlorophenol-d4	7.510	132	738257	21.335	ng/u1	0.00
15) 4-Methylphenol-d8	8.698	113	740282	20.357	ng/u1	0.00
21) Nitrobenzene-d5	9.151	128	382599	21.012	ng/u1	0.00
24) 2-Nitrophenol-d4	9.875	143	398240	21.707	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.416	165	713340	21.241	ng/u1	0.00
31) 4-Chloroaniline-d4	10.934	131	1067659	19.656	ng/u1	0.00
46) Dimethylphthalate-d6	14.033	166	1987712	20.214	ng/u1	0.00
49) Acenaphthylene-d8	14.316	160	2408300	21.087	ng/u1	0.00
54) 4-Nitrophenol-d4	14.810	143	368573	17.989	ng/u1	0.00
60) Fluorene-d10	15.616	176	1717925	20.022	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.727	200	293419	19.140	ng/u1	0.00
73) Anthracene-d10	17.468	188	2446398	20.964	ng/u1	0.00
81) Pyrene-d10	19.756	212	2324577	24.294	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.839	264	1718550	20.964	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.410	88	125292	8.258	ng/uL	95
5) Pyridine	3.816	79	860958	20.685	ng/u1	98
6) Benzaldehyde	7.122	77	554575	24.348	ng/u1	99
8) Phenol	7.169	94	998720	20.260	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.404	93	817202	20.812	ng/u1	99
12) 2-Chlorophenol	7.546	128	752159	21.055	ng/u1	98
13) 2-Methylphenol	8.428	108	743703	20.454	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.516	45	1089153	20.577	ng/u1	99
16) Acetophenone	8.816	105	1263792	21.010	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.804	70	662569	21.340	ng/u1	99
18) 4-Methylphenol	8.757	108	819485	20.279	ng/u1	98
19) Hexachloroethane	9.075	117	324661	21.519	ng/u1	97
22) Nitrobenzene	9.192	77	962956	20.887	ng/u1	100
23) Isophorone	9.728	82	1831477	20.800	ng/u1	100
25) 2-Nitrophenol	9.910	139	425732	21.199	ng/u1	98
26) 2,4-Dimethylphenol	9.969	107	886439	20.848	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.210	93	1136199	20.757	ng/u1	100
29) 2,4-Dichlorophenol	10.439	162	715355	21.187	ng/u1	99
30) Naphthalene	10.851	128	2554757	20.733	ng/u1	100
32) 4-Chloroaniline	10.957	127	1077878	19.718	ng/u1	100
33) Hexachlorobutadiene	11.139	225	429050	21.238	ng/u1	99
34) Caprolactam	11.733	113	135706	11.773	ng/u1	99
35) 4-Chloro-3-methylphenol	12.075	107	781849	20.325	ng/u1	99
36) 2-Methylnaphthalene	12.457	142	1633495	20.425	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM031323\
 Data File : BM039008.D
 Acq On : 14 Mar 2023 11:28
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020108


Quant Time: Mar 15 01:04:47 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM031323.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Mar 13 16:25:27 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.675	142	1634462	20.345	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.822	216	817776	21.473	ng/ul	99
40) Hexachlorocyclopentadiene	12.804	237	534492	21.505	ng/ul	99
41) 2,4,6-Trichlorophenol	13.057	196	537612	21.583	ng/ul	99
42) 2,4,5-Trichlorophenol	13.128	196	577647	21.327	ng/ul	99
43) 1,1'-Biphenyl	13.463	154	2177984	21.109	ng/ul	100
44) 2-Chloronaphthalene	13.504	162	1680518	20.986	ng/ul	99
45) 2-Nitroaniline	13.704	65	501032	20.998	ng/ul	98
47) Dimethylphthalate	14.080	163	2013476	20.217	ng/ul	99
48) 2,6-Dinitrotoluene	14.204	165	390947	21.300	ng/ul	100
50) Acenaphthylene	14.345	152	2649992	20.864	ng/ul	99
51) 3-Nitroaniline	14.527	138	422072	19.843	ng/ul	96
52) Acenaphthene	14.686	153	1791824	20.336	ng/ul	98
53) 2,4-Dinitrophenol	14.727	184	190041	16.510	ng/ul	98
55) 4-Nitrophenol	14.827	109	291528	18.309	ng/ul	95
56) Dibenzofuran	15.022	168	2445027	20.215	ng/ul	100
57) 2,4-Dinitrotoluene	14.986	165	543205	19.994	ng/ul	97
58) 2,3,4,6-Tetrachlorophenol	15.245	232	464711	20.148	ng/ul	99
59) Diethylphthalate	15.445	149	1990832	19.790	ng/ul	99
61) Fluorene	15.669	166	2008488	20.004	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.663	204	972311	20.249	ng/ul	97
63) 4-Nitroaniline	15.692	138	397260	19.625	ng/ul	97
66) 4,6-Dinitro-2-methylph...	15.745	198	295323	19.384	ng/ul	99
67) N-Nitrosodiphenylamine	15.880	169	1637655	22.159	ng/ul	99
68) 4-Bromophenyl-phenylether	16.563	248	539482	22.119	ng/ul	98
69) Hexachlorobenzene	16.674	284	602474	21.194	ng/ul	98
70) Atrazine	16.827	200	509707	20.190	ng/ul	99
71) Pentachlorophenol	17.015	266	355960	19.555	ng/ul	99
72) Phenanthrene	17.410	178	2836939	20.639	ng/ul	99
74) Anthracene	17.504	178	2897733	20.764	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.428	216	821492	23.746	ng/uL	99
76) Pentachlorobenzene	14.939	250	788101	22.955	ng/uL	100
77) Carbazole	17.774	167	2504249	19.990	ng/ul	100
78) Di-n-butylphthalate	18.339	149	3078341	21.039	ng/ul	100
80) Fluoranthene	19.427	202	2745770	24.561	ng/ul	99
82) Pyrene	19.786	202	2904777	24.346	ng/ul	98
83) Butylbenzylphthalate	20.686	149	1073347	22.739	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.468	252	685377	19.554	ng/ul	98
85) Benzo(a)anthracene	21.539	228	2377425	20.926	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.462	149	1519041	21.611	ng/ul	100
87) Chrysene	21.592	228	2263601	20.733	ng/ul	100
89) Di-n-octyl phthalate	22.409	149	2342438	19.960	ng/ul	100
90) Benzo(b)fluoranthene	23.250	252	2157427	20.205	ng/ul	99
91) Benzo(k)fluoranthene	23.303	252	2201204	20.857	ng/ul	100
93) Benzo(a)pyrene	23.892	252	1923245	20.820	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.550	276	2459762	23.092	ng/ul	98
95) Dibenzo(a,h)anthracene	26.568	278	2080418	23.181	ng/ul	99
96) Benzo(g,h,i)perylene	27.327	276	1983749	23.533	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Mar 15 01:04:47 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM031323.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Mar 13 16:25:27 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 8.258 ng/uL
RT: 3.410 min Scan# 21
Delta R.T. 0.006 min
Lab File: BM039008.D
Acq: 14 Mar 2023 11:28




Instrument :
BNA_M
ClientSampleId :
SSTD020108


Tgt Ion: 88 Resp: 125292
Ion Ratio Lower Upper
88 100
43 27.3 24.0 36.0
58 70.8 59.8 89.8



#5
Pyridine
Concen: 20.685 ng/uL
RT: 3.816 min Scan# 90
Delta R.T. 0.000 min
Lab File: BM039008.D
Acq: 14 Mar 2023 11:28

Tgt Ion: 79 Resp: 860958
Ion Ratio Lower Upper
79 100
52 69.8 56.6 85.0
51 33.1 27.8 41.6









#6
Benzaldehyde
Concen: 24.348 ng/ul
RT: 7.122 min Scan# 652
Delta R.T. 0.000 min
Lab File: BM039008.D
Acq: 14 Mar 2023 11:28




Instrument :
BNA_M
ClientSampleId :
SSTD020108


Tgt Ion: 77 Resp: 554575
Ion Ratio Lower Upper
77 100
105 96.1 77.8 116.8
106 93.1 75.4 113.2



#8
Phenol
Concen: 20.260 ng/ul
RT: 7.169 min Scan# 660
Delta R.T. 0.000 min
Lab File: BM039008.D
Acq: 14 Mar 2023 11:28

Tgt Ion: 94 Resp: 998720
Ion Ratio Lower Upper
94 100
65 27.7 22.7 34.1
66 35.1 28.2 42.2









#10
Bis(2-Chloroethyl)ether
Concen: 20.812 ng/ul
RT: 7.404 min Scan# 700
Delta R.T. -0.006 min
Lab File: BM039008.D
Acq: 14 Mar 2023 11:28




Instrument :
BNA_M
ClientSampleId :
SSTD020108


Tgt Ion: 93 Resp: 817202
Ion Ratio Lower Upper
93 100
63 77.3 61.1 91.7
95 32.5 25.4 38.2



#12
2-Chlorophenol
Concen: 21.055 ng/ul
RT: 7.546 min Scan# 724
Delta R.T. 0.000 min
Lab File: BM039008.D
Acq: 14 Mar 2023 11:28

Tgt Ion:128 Resp: 752159
Ion Ratio Lower Upper
128 100
64 47.3 39.7 59.5
130 32.5 26.2 39.2









#13
2-Methylphenol
Concen: 20.454 ng/ul
RT: 8.428 min Scan# 874
Delta R.T. 0.000 min
Lab File: BM039008.D
Acq: 14 Mar 2023 11:28




Instrument :
BNA_M
ClientSampleId :
SSTD020108


Tgt Ion:108 Resp: 743703
Ion Ratio Lower Upper
108 100
107 88.2 71.4 107.0



#14
2,2'-oxybis(1-Chloropropane)
Concen: 20.577 ng/ul
RT: 8.516 min Scan# 889
Delta R.T. -0.006 min
Lab File: BM039008.D
Acq: 14 Mar 2023 11:28

Tgt Ion: 45 Resp: 1089153
Ion Ratio Lower Upper
45 100
77 14.4 11.9 17.9
79 10.9 9.0 13.6









#16
Acetophenone
Concen: 21.010 ng/ul
RT: 8.816 min Scan# 940
Delta R.T. 0.000 min
Lab File: BM039008.D
Acq: 14 Mar 2023 11:28




Instrument :
BNA_M
ClientSampleId :
SSTD020108


Tgt Ion:105 Resp: 1263792
Ion Ratio Lower Upper
105 100
77 78.5 62.6 93.8
51 26.5 21.9 32.9



#17
N-Nitroso-di-n-propylamine
Concen: 21.340 ng/ul
RT: 8.804 min Scan# 938
Delta R.T. 0.000 min
Lab File: BM039008.D
Acq: 14 Mar 2023 11:28

Tgt Ion: 70 Resp: 662569
Ion Ratio Lower Upper
70 100
42 56.4 44.7 67.1
101 9.3 7.3 10.9
130 19.4 15.5 23.3









#18
4-Methylphenol
Concen: 20.279 ng/ul
RT: 8.757 min Scan# 911
Delta R.T. 0.000 min
Lab File: BM039008.D
Acq: 14 Mar 2023 11:28




Instrument :
BNA_M
ClientSampleId :
SSTD020108


Tgt Ion:108 Resp: 819485
Ion Ratio Lower Upper
108 100
107 110.7 90.0 135.0



#19
Hexachloroethane
Concen: 21.519 ng/ul
RT: 9.075 min Scan# 984
Delta R.T. 0.000 min
Lab File: BM039008.D
Acq: 14 Mar 2023 11:28

Tgt Ion:117 Resp: 324661
Ion Ratio Lower Upper
117 100
201 93.2 71.6 107.4
199 58.0 45.1 67.7









#22
Nitrobenzene
Concen: 20.887 ng/ul
RT: 9.192 min Scan# 1004
Delta R.T. -0.006 min
Lab File: BM039008.D
Acq: 14 Mar 2023 11:28




Instrument :
BNA_M
ClientSampleId :
SSTD020108


Tgt Ion: 77 Resp: 962956
Ion Ratio Lower Upper
77 100
123 42.4 34.2 51.2
65 13.4 10.6 16.0



#23
Isophorone
Concen: 20.800 ng/ul
RT: 9.728 min Scan# 1095
Delta R.T. 0.000 min
Lab File: BM039008.D
Acq: 14 Mar 2023 11:28

Tgt Ion: 82 Resp: 1831477
Ion Ratio Lower Upper
82 100
95 6.6 5.3 7.9
138 16.6 13.3 19.9









#25
2-Nitrophenol
Concen: 21.199 ng/ul
RT: 9.910 min Scan# 1126
Delta R.T. 0.000 min
Lab File: BM039008.D
Acq: 14 Mar 2023 11:28




Instrument :
BNA_M
ClientSampleId :
SSTD020108


Tgt Ion:139 Resp: 425732
Ion Ratio Lower Upper
139 100
65 55.1 43.4 65.0
109 30.1 23.1 34.7



#26
2,4-Dimethylphenol
Concen: 20.848 ng/ul
RT: 9.969 min Scan# 1136
Delta R.T. 0.000 min
Lab File: BM039008.D
Acq: 14 Mar 2023 11:28

Tgt Ion:107 Resp: 886439
Ion Ratio Lower Upper
107 100
121 49.9 40.6 60.8
122 87.8 69.7 104.5









#27
 Bis(2-Chloroethoxy)methane
 Concen: 20.757 ng/ul
 RT: 10.210 min Scan# 1177
 Delta R.T. 0.000 min
 Lab File: BM039008.D
 Acq: 14 Mar 2023 11:28

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020108

Tgt Ion: 93 Resp: 1136199




Ion	Ratio	Lower	Upper
93	100		
95	32.4	25.9	38.9
123	11.1	9.1	13.7




#29
 2,4-Dichlorophenol
 Concen: 21.187 ng/ul
 RT: 10.439 min Scan# 1216
 Delta R.T. -0.006 min
 Lab File: BM039008.D
 Acq: 14 Mar 2023 11:28

Tgt Ion: 162 Resp: 715355

Ion	Ratio	Lower	Upper
162	100		
164	65.4	53.1	79.7
98	37.8	29.6	44.4









#30
Naphthalene
Concen: 20.733 ng/ul
RT: 10.851 min Scan# 1286
Delta R.T. 0.000 min
Lab File: BM039008.D
Acq: 14 Mar 2023 11:28




Instrument :
BNA_M
ClientSampleId :
SSTD020108


Tgt Ion:128 Resp: 2554757
Ion Ratio Lower Upper
128 100
129 11.0 8.6 13.0
127 12.9 10.3 15.5



#32
4-Chloroaniline
Concen: 19.718 ng/ul
RT: 10.957 min Scan# 1304
Delta R.T. -0.006 min
Lab File: BM039008.D
Acq: 14 Mar 2023 11:28

Tgt Ion:127 Resp: 1077878
Ion Ratio Lower Upper
127 100
129 32.2 25.9 38.9





#33

Hexachlorobutadiene

Concen: 21.238 ng/ul

RT: 11.139 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM039008.D

Acq: 14 Mar 2023 11:28

Instrument :

BNA_M

ClientSampleId :

SSTD020108





Tgt Ion:225 Resp: 429050

Ion Ratio Lower Upper

225 100

223 62.6 49.8 74.6

227 64.1 50.2 75.4



#34

Caprolactam

Concen: 11.773 ng/ul

RT: 11.733 min Scan# 1436

Delta R.T. 0.000 min

Lab File: BM039008.D

Acq: 14 Mar 2023 11:28




Tgt Ion:113 Resp: 135706


Ion Ratio Lower Upper

113 100

55 177.0 141.8 212.8

56 131.3 102.2 153.2









#35
4-Chloro-3-methylphenol
Concen: 20.325 ng/ul
RT: 12.075 min Scan# 1494
Delta R.T. -0.006 min
Lab File: BM039008.D
Acq: 14 Mar 2023 11:28




Instrument :
BNA_M
ClientSampleId :
SSTD020108


Tgt Ion:107 Resp: 781849
Ion Ratio Lower Upper
107 100
144 25.8 19.9 29.9
142 77.7 62.6 93.8



#36
2-Methylnaphthalene
Concen: 20.425 ng/ul
RT: 12.457 min Scan# 1559
Delta R.T. 0.000 min
Lab File: BM039008.D
Acq: 14 Mar 2023 11:28

Tgt Ion:142 Resp: 1633495
Ion Ratio Lower Upper
142 100
141 88.8 70.7 106.1









#37
1-Methylnaphthalene
Concen: 20.345 ng/ul
RT: 12.675 min Scan# 1596
Delta R.T. 0.000 min
Lab File: BM039008.D
Acq: 14 Mar 2023 11:28




Instrument :
BNA_M
ClientSampleId :
SSTD020108


Tgt Ion:142 Resp: 1634462
Ion Ratio Lower Upper
142 100
141 91.2 73.1 109.7
116 3.9 3.0 4.6



#39
1,2,4,5-Tetrachlorobenzene
Concen: 21.473 ng/ul
RT: 12.822 min Scan# 1621
Delta R.T. 0.000 min
Lab File: BM039008.D
Acq: 14 Mar 2023 11:28

Tgt Ion:216 Resp: 817776
Ion Ratio Lower Upper
216 100
214 79.2 62.6 93.8
179 21.1 17.0 25.4
108 19.3 15.9 23.9









#40
Hexachlorocyclopentadiene
Concen: 21.505 ng/ul
RT: 12.804 min Scan# 1618
Delta R.T. 0.000 min
Lab File: BM039008.D
Acq: 14 Mar 2023 11:28




Instrument :
BNA_M
ClientSampleId :
SSTD020108


Tgt Ion:237 Resp: 534492
Ion Ratio Lower Upper
237 100
235 62.2 50.0 75.0
272 12.4 10.7 16.1



#41
2,4,6-Trichlorophenol
Concen: 21.583 ng/ul
RT: 13.057 min Scan# 1661
Delta R.T. -0.006 min
Lab File: BM039008.D
Acq: 14 Mar 2023 11:28

Tgt Ion:196 Resp: 537612
Ion Ratio Lower Upper
196 100
198 94.1 76.1 114.1
200 30.4 25.4 38.0





#42

2,4,5-Trichlorophenol

Concen: 21.327 ng/ul

RT: 13.128 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM039008.D


Acq: 14 Mar 2023 11:28

Instrument :

BNA_M

ClientSampleId :

SSTD020108






Tgt Ion:196 Resp: 577647

Ion Ratio Lower Upper

196 100

198 96.5 78.2 117.2

200 30.9 24.9 37.3



#43

1,1'-Biphenyl

Concen: 21.109 ng/ul

RT: 13.463 min Scan# 1730

Delta R.T. 0.000 min

Lab File: BM039008.D

Acq: 14 Mar 2023 11:28




Tgt Ion:154 Resp: 2177984


Ion Ratio Lower Upper

154 100

153 40.2 32.0 48.0

76 14.4 11.5 17.3









#44
2-Chloronaphthalene
Concen: 20.986 ng/ul
RT: 13.504 min Scan# 1737
Delta R.T. 0.000 min
Lab File: BM039008.D
Acq: 14 Mar 2023 11:28




Instrument :
BNA_M
ClientSampleId :
SSTD020108


Tgt Ion:162 Resp: 1680518
Ion Ratio Lower Upper
162 100
164 32.6 26.6 39.8
127 39.5 32.2 48.2



#45
2-Nitroaniline
Concen: 20.998 ng/ul
RT: 13.704 min Scan# 1771
Delta R.T. -0.006 min
Lab File: BM039008.D
Acq: 14 Mar 2023 11:28

Tgt Ion: 65 Resp: 501032
Ion Ratio Lower Upper
65 100
92 66.4 53.3 79.9
138 100.8 82.9 124.3









#47
Dimethylphthalate
Concen: 20.217 ng/ul
RT: 14.080 min Scan# 1835
Delta R.T. -0.006 min
Lab File: BM039008.D
Acq: 14 Mar 2023 11:28




Instrument :
BNA_M
ClientSampleId :
SSTD020108


Tgt Ion:163 Resp: 2013476
Ion Ratio Lower Upper
163 100
194 3.8 3.1 4.7
164 10.4 8.0 12.0



#48
2,6-Dinitrotoluene
Concen: 21.300 ng/ul
RT: 14.204 min Scan# 1856
Delta R.T. 0.000 min
Lab File: BM039008.D
Acq: 14 Mar 2023 11:28

Tgt Ion:165 Resp: 390947
Ion Ratio Lower Upper
165 100
89 63.0 50.2 75.4
121 22.7 18.0 27.0









#50
Acenaphthylene
Concen: 20.864 ng/ul
RT: 14.345 min Scan# 1880
Delta R.T. -0.006 min
Lab File: BM039008.D
Acq: 14 Mar 2023 11:28




Instrument :
BNA_M
ClientSampleId :
SSTD020108


Tgt Ion:152 Resp: 2649992
Ion Ratio Lower Upper
152 100
151 19.9 15.6 23.4
153 13.2 10.2 15.4



#51
3-Nitroaniline
Concen: 19.843 ng/ul
RT: 14.527 min Scan# 1911
Delta R.T. 0.000 min
Lab File: BM039008.D
Acq: 14 Mar 2023 11:28

Tgt Ion:138 Resp: 422072
Ion Ratio Lower Upper
138 100
108 10.5 8.9 13.3
92 117.3 97.9 146.9





#52

Acenaphthene

Concen: 20.336 ng/ul

RT: 14.686 min Scan# 1938

Delta R.T. -0.006 min

Lab File: BM039008.D

Acq: 14 Mar 2023 11:28

Instrument :

BNA_M

ClientSampleId :

SSTD020108





Tgt Ion:153 Resp: 1791824

Ion Ratio Lower Upper

153 100

152 49.0 38.1 57.1

154 90.4 71.4 107.0



#53

2,4-Dinitrophenol

Concen: 16.510 ng/ul

RT: 14.727 min Scan# 1945

Delta R.T. 0.000 min

Lab File: BM039008.D

Acq: 14 Mar 2023 11:28




Tgt Ion:184 Resp: 190041


Ion Ratio Lower Upper

184 100

63 78.4 64.7 97.1

154 63.8 50.8 76.2









#55
4-Nitrophenol
Concen: 18.309 ng/ul
RT: 14.827 min Scan# 1962
Delta R.T. 0.000 min
Lab File: BM039008.D
Acq: 14 Mar 2023 11:28




Instrument :
BNA_M
ClientSampleId :
SSTD020108


Tgt Ion:109 Resp: 291528
Ion Ratio Lower Upper
109 100
139 135.4 103.4 155.0
65 142.4 108.7 163.1



#56
Dibenzofuran
Concen: 20.215 ng/ul
RT: 15.022 min Scan# 1995
Delta R.T. 0.000 min
Lab File: BM039008.D
Acq: 14 Mar 2023 11:28

Tgt Ion:168 Resp: 2445027
Ion Ratio Lower Upper
168 100
139 38.2 30.7 46.1









#57
2,4-Dinitrotoluene
Concen: 19.994 ng/ul
RT: 14.986 min Scan# 1989
Delta R.T. 0.000 min
Lab File: BM039008.D
Acq: 14 Mar 2023 11:28




Instrument :
BNA_M
ClientSampleId :
SSTD020108


Tgt Ion:165 Resp: 543205
Ion Ratio Lower Upper
165 100
63 46.2 38.7 58.1
182 2.5 2.1 3.1



#58
2,3,4,6-Tetrachlorophenol
Concen: 20.148 ng/ul
RT: 15.245 min Scan# 2033
Delta R.T. 0.000 min
Lab File: BM039008.D
Acq: 14 Mar 2023 11:28

Tgt Ion:232 Resp: 464711
Ion Ratio Lower Upper
232 100
131 51.5 41.3 61.9
130 2.4 2.2 3.4
166 32.5 26.5 39.7









#59
Diethylphthalate
Concen: 19.790 ng/ul
RT: 15.445 min Scan# 2067
Delta R.T. 0.000 min
Lab File: BM039008.D
Acq: 14 Mar 2023 11:28




Instrument :
BNA_M
ClientSampleId :
SSTD020108


Tgt Ion:149 Resp: 1990832
Ion Ratio Lower Upper
149 100
177 21.5 16.9 25.3
150 12.2 9.8 14.8



#61
Fluorene
Concen: 20.004 ng/ul
RT: 15.669 min Scan# 2105
Delta R.T. 0.000 min
Lab File: BM039008.D
Acq: 14 Mar 2023 11:28

Tgt Ion:166 Resp: 2008488
Ion Ratio Lower Upper
166 100
165 97.9 77.5 116.3
167 13.8 11.0 16.6









#62
4-Chlorophenyl-phenylether
Concen: 20.249 ng/ul
RT: 15.663 min Scan# 2104
Delta R.T. -0.006 min
Lab File: BM039008.D
Acq: 14 Mar 2023 11:28




Instrument :
BNA_M
ClientSampleId :
SSTD020108


Tgt Ion:204 Resp: 972311
Ion Ratio Lower Upper
204 100
206 32.6 26.9 40.3
141 60.6 46.6 69.8



#63
4-Nitroaniline
Concen: 19.625 ng/ul
RT: 15.692 min Scan# 2109
Delta R.T. 0.006 min
Lab File: BM039008.D
Acq: 14 Mar 2023 11:28

Tgt Ion:138 Resp: 397260
Ion Ratio Lower Upper
138 100
92 53.4 42.7 64.1
108 83.2 63.5 95.3









#66
4,6-Dinitro-2-methylphenol
Concen: 19.384 ng/ul
RT: 15.745 min Scan# 2118
Delta R.T. 0.000 min
Lab File: BM039008.D
Acq: 14 Mar 2023 11:28




Instrument :
BNA_M
ClientSampleId :
SSTD020108


Tgt Ion:198 Resp: 295323
Ion Ratio Lower Upper
198 100
182 4.7 3.6 5.4
77 28.9 23.7 35.5



#67
N-Nitrosodiphenylamine
Concen: 22.159 ng/ul
RT: 15.880 min Scan# 2141
Delta R.T. 0.000 min
Lab File: BM039008.D
Acq: 14 Mar 2023 11:28

Tgt Ion:169 Resp: 1637655
Ion Ratio Lower Upper
169 100
168 65.9 53.5 80.3
167 36.6 29.0 43.4









#68
4-Bromophenyl-phenylether
Concen: 22.119 ng/ul
RT: 16.563 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM039008.D
Acq: 14 Mar 2023 11:28




Instrument :
BNA_M
ClientSampleId :
SSTD020108


Tgt Ion:248 Resp: 539482
Ion Ratio Lower Upper
248 100
250 95.8 77.3 115.9
141 74.4 62.6 93.8



#69
Hexachlorobenzene
Concen: 21.194 ng/ul
RT: 16.674 min Scan# 2276
Delta R.T. 0.000 min
Lab File: BM039008.D
Acq: 14 Mar 2023 11:28

Tgt Ion:284 Resp: 602474
Ion Ratio Lower Upper
284 100
142 37.2 31.0 46.6
249 30.5 24.9 37.3





#70

Atrazine

Concen: 20.190 ng/ul

RT: 16.827 min Scan# 21

Delta R.T. 0.000 min

Lab File: BM039008.D

Acq: 14 Mar 2023 11:28

Instrument :

BNA_M

ClientSampleId :

SSTD020108





Tgt Ion:200 Resp: 509707

Ion Ratio Lower Upper

200 100

173 28.0 21.5 32.3

215 47.5 38.1 57.1



#71

Pentachlorophenol

Concen: 19.555 ng/ul

RT: 17.015 min Scan# 2334

Delta R.T. 0.000 min

Lab File: BM039008.D

Acq: 14 Mar 2023 11:28




Tgt Ion:266 Resp: 355960


Ion Ratio Lower Upper

266 100

264 61.0 50.0 75.0

268 63.2 50.5 75.7









#72
Phenanthrene
Concen: 20.639 ng/ul
RT: 17.410 min Scan# 2401
Delta R.T. -0.006 min
Lab File: BM039008.D
Acq: 14 Mar 2023 11:28




Instrument :
BNA_M
ClientSampleId :
SSTD020108


Tgt Ion:178 Resp: 2836939
Ion Ratio Lower Upper
178 100
179 15.6 12.1 18.1
176 19.6 15.7 23.5



#74
Anthracene
Concen: 20.764 ng/ul
RT: 17.504 min Scan# 2417
Delta R.T. 0.000 min
Lab File: BM039008.D
Acq: 14 Mar 2023 11:28

Tgt Ion:178 Resp: 2897733
Ion Ratio Lower Upper
178 100
179 15.6 12.4 18.6
176 18.9 15.2 22.8





#75

1,2,3,4-Tetrachlorobenzene

Concen: 23.746 ng/uL

RT: 13.428 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM039008.D

Acq: 14 Mar 2023 11:28

Instrument :

BNA_M

ClientSampleId :

SSTD020108





Tgt Ion:216 Resp: 821492

Ion Ratio Lower Upper

216 100

214 79.0 62.7 94.1

218 48.3 37.8 56.8



#76

Pentachlorobenzene

Concen: 22.955 ng/uL

RT: 14.939 min Scan# 1981

Delta R.T. 0.000 min

Lab File: BM039008.D

Acq: 14 Mar 2023 11:28




Tgt Ion:250 Resp: 788101


Ion Ratio Lower Upper

250 100

248 63.2 50.5 75.7

252 63.2 50.5 75.7





#77

Carbazole

Concen: 19.990 ng/ul

RT: 17.774 min Scan# 2463

Delta R.T. 0.000 min

Lab File: BM039008.D

Acq: 14 Mar 2023 11:28

Instrument :

BNA_M

ClientSampleId :

SSTD020108





Tgt Ion:167 Resp: 2504249

Ion Ratio Lower Upper

167 100

166 21.2 17.0 25.4

139 13.4 10.4 15.6



#78

Di-n-butylphthalate

Concen: 21.039 ng/ul

RT: 18.339 min Scan# 2559

Delta R.T. 0.000 min

Lab File: BM039008.D

Acq: 14 Mar 2023 11:28




Tgt Ion:149 Resp: 3078341


Ion Ratio Lower Upper

149 100

150 9.3 7.4 11.0

104 5.2 4.3 6.5









#80
 Fluoranthene
 Concen: 24.561 ng/ul
 RT: 19.427 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: BM039008.D
 Acq: 14 Mar 2023 11:28




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020108


Tgt Ion:202 Resp: 2745770
 Ion Ratio Lower Upper
 202 100
 101 12.1 9.9 14.9
 100 9.4 7.4 11.2



#82
 Pyrene
 Concen: 24.346 ng/ul
 RT: 19.786 min Scan# 2805
 Delta R.T. -0.006 min
 Lab File: BM039008.D
 Acq: 14 Mar 2023 11:28

Tgt Ion:202 Resp: 2904777
 Ion Ratio Lower Upper
 202 100
 101 15.4 11.4 17.2
 100 12.3 9.3 13.9









#83
Butylbenzylphthalate
Concen: 22.739 ng/ul
RT: 20.686 min Scan# 2958
Delta R.T. 0.000 min
Lab File: BM039008.D
Acq: 14 Mar 2023 11:28




Instrument :
BNA_M
ClientSampleId :
SSTD020108


Tgt Ion:149 Resp: 1073347
Ion Ratio Lower Upper
149 100
91 75.7 61.5 92.3
206 19.4 15.8 23.8



#84
3,3'-Dichlorobenzidine
Concen: 19.554 ng/ul
RT: 21.468 min Scan# 3091
Delta R.T. 0.000 min
Lab File: BM039008.D
Acq: 14 Mar 2023 11:28

Tgt Ion:252 Resp: 685377
Ion Ratio Lower Upper
252 100
254 64.5 52.6 79.0
126 12.6 9.8 14.6





#85

Benzo(a)anthracene

Concen: 20.926 ng/ul

RT: 21.539 min Scan# 3102

Delta R.T. 0.006 min

Lab File: BM039008.D

Acq: 14 Mar 2023 11:28

Instrument :

BNA_M

ClientSampleId :

SSTD020108





Tgt Ion:228 Resp: 2377425

Ion Ratio Lower Upper

228 100

229 19.3 15.9 23.9

226 27.2 21.8 32.8



#86

Bis(2-ethylhexyl)phthalate

Concen: 21.611 ng/ul

RT: 21.462 min Scan# 3090

Delta R.T. 0.000 min

Lab File: BM039008.D

Acq: 14 Mar 2023 11:28




Tgt Ion:149 Resp: 1519041


Ion Ratio Lower Upper

149 100

167 27.1 21.8 32.8

279 3.5 2.8 4.2









#87
Chrysene
Concen: 20.733 ng/ul
RT: 21.592 min Scan# 3111
Delta R.T. 0.000 min
Lab File: BM039008.D
Acq: 14 Mar 2023 11:28




Instrument :
BNA_M
ClientSampleId :
SSTD020108


Tgt Ion:228 Resp: 2263601
Ion Ratio Lower Upper
228 100
226 30.0 24.1 36.1
229 19.6 15.8 23.6



#89
Di-n-octyl phthalate
Concen: 19.960 ng/ul
RT: 22.409 min Scan# 3251
Delta R.T. 0.000 min
Lab File: BM039008.D
Acq: 14 Mar 2023 11:28

Tgt Ion:149 Resp: 2342438





#90

Benzo(b)fluoranthene

Concen: 20.205 ng/ul

RT: 23.250 min Scan# 31

Delta R.T. 0.000 min

Lab File: BM039008.D

Acq: 14 Mar 2023 11:28

Instrument :

BNA_M

ClientSampleId :

SSTD020108





Tgt Ion:252 Resp: 2157427

Ion Ratio Lower Upper

252 100

253 22.3 17.4 26.2

125 11.6 9.2 13.8



#91

Benzo(k)fluoranthene

Concen: 20.857 ng/ul

RT: 23.303 min Scan# 3403

Delta R.T. 0.000 min

Lab File: BM039008.D

Acq: 14 Mar 2023 11:28




Tgt Ion:252 Resp: 2201204


Ion Ratio Lower Upper

252 100

253 21.7 17.5 26.3

125 11.2 8.8 13.2





#93

Benzo(a)pyrene

Concen: 20.820 ng/ul

RT: 23.892 min Scan# 3

Delta R.T. 0.000 min

Lab File: BM039008.D

Acq: 14 Mar 2023 11:28

Instrument :

BNA_M

ClientSampleId :

SSTD020108





Tgt Ion:252 Resp: 1923245

Ion Ratio Lower Upper

252 100

253 21.6 17.4 26.2

125 12.8 9.9 14.9



#94

Indeno(1,2,3-cd)pyrene

Concen: 23.092 ng/ul

RT: 26.550 min Scan# 3955

Delta R.T. 0.000 min

Lab File: BM039008.D

Acq: 14 Mar 2023 11:28




Tgt Ion:276 Resp: 2459762


Ion Ratio Lower Upper

276 100

138 27.5 23.1 34.7

277 25.3 19.9 29.9









#95
Dibenzo(a,h)anthracene
Concen: 23.181 ng/ul
RT: 26.568 min Scan# 3957
Delta R.T. 0.000 min
Lab File: BM039008.D
Acq: 14 Mar 2023 11:28

Instrument :
BNA_M
ClientSampleId :
SSTD020108

Tgt Ion:278 Resp: 2080418
Ion Ratio Lower Upper
278 100
139 18.8 14.6 22.0
279 23.8 19.0 28.4



#96
Benzo(g,h,i)perylene
Concen: 23.533 ng/ul
RT: 27.327 min Scan# 4087
Delta R.T. -0.006 min
Lab File: BM039008.D
Acq: 14 Mar 2023 11:28

Tgt Ion:276 Resp: 1983749
Ion Ratio Lower Upper
276 100
138 25.9 20.1 30.1
277 23.8 19.0 28.6

