

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM031325\
 Data File : BM049712.D
 Acq On : 13 Mar 2025 12:58
 Operator : RC/JU
 Sample : SSTDICC060
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC060

Manual Integrations
 APPROVED

Reviewed By :Anahy Claudio 03/14/2025
 Supervised By :Jagrut Upadhyay 03/14/2025

Quant Time: Mar 13 14:43:18 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM031325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 13 14:27:21 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.804	152	315799	20.000	ng	0.00
21) Naphthalene-d8	10.604	136	1024759	20.000	ng	0.00
39) Acenaphthene-d10	14.445	164	618823	20.000	ng	0.00
64) Phenanthrene-d10	17.186	188	1169651	20.000	ng	0.00
76) Chrysene-d12	21.427	240	1173482	20.000	ng	0.00
86) Perylene-d12	24.438	264	1084370	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.387	112	2416717	129.126	ng	0.00
7) Phenol-d6	6.981	99	2994230	127.482	ng	0.00
23) Nitrobenzene-d5	8.963	82	2583706	129.447	ng	0.00
42) 2,4,6-Tribromophenol	15.933	330	1073944	148.622	ng	0.00
45) 2-Fluorobiphenyl	13.069	172	6475865	131.059	ng	0.00
79) Terphenyl-d14	19.809	244	9673323	138.909	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.287	88	500669	61.795	ng	99
3) Pyridine	3.687	79	1281013m	61.832	ng	
4) n-Nitrosodimethylamine	3.593	42	568214	62.064	ng	99
6) Aniline	7.140	93	1180576	59.996	ng	99
8) 2-Chlorophenol	7.375	128	1161936	62.692	ng	99
9) Benzaldehyde	6.945	77	582373	50.156	ng	99
10) Phenol	7.010	94	1429401	62.763	ng	98
11) bis(2-Chloroethyl)ether	7.234	93	1137868	61.698	ng	99
12) 1,3-Dichlorobenzene	7.698	146	1393806	61.881	ng	99
13) 1,4-Dichlorobenzene	7.845	146	1404557	61.813	ng	99
14) 1,2-Dichlorobenzene	8.163	146	1324220	61.320	ng	98
15) Benzyl Alcohol	8.045	79	944038	64.048	ng	98
16) 2,2'-oxybis(1-Chloropr...	8.345	45	1292162	59.613	ng	100
17) 2-Methylphenol	8.257	107	851393	62.967	ng	99
18) Hexachloroethane	8.892	117	493239	61.468	ng	93
19) n-Nitroso-di-n-propyla...	8.616	70	865717	63.565	ng	99
20) 3+4-Methylphenols	8.581	107	1147270	63.315	ng	98
22) Acetophenone	8.628	105	1686019	63.809	ng	# 99
24) Nitrobenzene	9.004	77	1197256	63.615	ng	99
25) Isophorone	9.533	82	1988098	64.875	ng	99
26) 2-Nitrophenol	9.716	139	556011	69.411	ng	100
27) 2,4-Dimethylphenol	9.781	122	669470	65.658	ng	97
28) bis(2-Chloroethoxy)met...	10.016	93	1353615	62.714	ng	100
29) 2,4-Dichlorophenol	10.257	162	1085297	65.322	ng	98
30) 1,2,4-Trichlorobenzene	10.469	180	1314305	63.227	ng	99
31) Naphthalene	10.651	128	3372239	63.997	ng	100
32) Benzoic acid	9.933	122	611385	70.482	ng	98
33) 4-Chloroaniline	10.763	127	1140961	63.446	ng	99
34) Hexachlorobutadiene	10.945	225	820160	63.912	ng	98
35) Caprolactam	11.551	113	267177	69.264	ng	98
36) 4-Chloro-3-methylphenol	11.892	107	952831	66.827	ng	99
37) 2-Methylnaphthalene	12.269	142	2360362	64.061	ng	100
38) 1-Methylnaphthalene	12.486	142	2278919	64.106	ng	99
40) 1,2,4,5-Tetrachloroben...	12.639	216	1483366	65.259	ng	99
41) Hexachlorocyclopentadiene	12.622	237	557389	66.191	ng	97
43) 2,4,6-Trichlorophenol	12.880	196	863951	67.864	ng	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM031325\
 Data File : BM049712.D
 Acq On : 13 Mar 2025 12:58
 Operator : RC/JU
 Sample : SSTDICC060
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC060

Manual Integrations
 APPROVED

Reviewed By :Anahy Claudio 03/14/2025
 Supervised By :Jagrut Upadhyay 03/14/2025

Quant Time: Mar 13 14:43:18 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM031325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 13 14:27:21 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.951	196	936029	67.377	ng	99
46) 1,1'-Biphenyl	13.280	154	3048618	63.921	ng	99
47) 2-Chloronaphthalene	13.321	162	2381008	64.138	ng	99
48) 2-Nitroaniline	13.521	65	578552	69.424	ng	99
49) Acenaphthylene	14.168	152	3384598	65.227	ng	99
50) Dimethylphthalate	13.910	163	2716983	63.771	ng	99
51) 2,6-Dinitrotoluene	14.021	165	563727	66.034	ng	95
52) Acenaphthene	14.510	154	2217160	65.221	ng	99
53) 3-Nitroaniline	14.345	138	556379	68.488	ng	98
54) 2,4-Dinitrophenol	14.551	184	295154	60.822	ng	97
55) Dibenzofuran	14.845	168	3632034	64.040	ng	99
56) 4-Nitrophenol	14.663	139	477486	67.524	ng	99
57) 2,4-Dinitrotoluene	14.804	165	763894	69.539	ng	99
58) Fluorene	15.492	166	3055405	66.168	ng	100
59) 2,3,4,6-Tetrachlorophenol	15.074	232	777060	66.821	ng	100
60) Diethylphthalate	15.274	149	2576429	64.207	ng	99
61) 4-Chlorophenyl-phenyle...	15.492	204	1674356	66.402	ng	98
62) 4-Nitroaniline	15.510	138	598859	60.865	ng	97
63) Azobenzene	15.780	77	2482118	64.760	ng	100
65) 4,6-Dinitro-2-methylph...	15.568	198	417995	61.103	ng	96
66) n-Nitrosodiphenylamine	15.704	169	2302780	64.980	ng	100
67) 4-Bromophenyl-phenylether	16.380	248	881819	65.299	ng	99
68) Hexachlorobenzene	16.498	284	986543	64.964	ng	99
70) Pentachlorophenol	16.839	266	647404	67.839	ng	98
71) Phenanthrene	17.227	178	4268082	65.716	ng	100
72) Anthracene	17.321	178	4187832	66.580	ng	100
73) Carbazole	17.592	167	3826935	67.182	ng	99
74) Di-n-butylphthalate	18.162	149	4214218	67.499	ng	99
75) Fluoranthene	19.245	202	5199093	69.531	ng	100
77) Benzidine	19.427	184	771546	54.727	ng	99
78) Pyrene	19.609	202	5421531	66.127	ng	100
80) Butylbenzylphthalate	20.509	149	1729836	67.988	ng	100
81) Benzo(a)anthracene	21.409	228	5145212	65.644	ng	99
82) 3,3'-Dichlorobenzidine	21.333	252	1687314	66.120	ng	99
83) Chrysene	21.474	228	4892160	65.843	ng	100
84) Bis(2-ethylhexyl)phtha...	21.339	149	2633480	66.414	ng	99
85) Di-n-octyl phthalate	22.480	149	3977588	64.539	ng	99
87) Indeno(1,2,3-cd)pyrene	27.856	276	5301272	67.923	ng	100
88) Benzo(b)fluoranthene	23.491	252	4904977	67.158	ng	100
89) Benzo(k)fluoranthene	23.556	252	4810286	66.934	ng	99
90) Benzo(a)pyrene	24.303	252	4164433	67.883	ng	99
91) Dibenzo(a,h)anthracene	27.909	278	4403728	67.722	ng	99
92) Benzo(g,h,i)perylene	28.915	276	4374098	66.781	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM031325\
Data File : BM049712.D
Acq On : 13 Mar 2025 12:58
Operator : RC/JU
Sample : SSTDICC060
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :

BNA_M

ClientSampleId :

SSTDICC060


Manual Integrations

APPROVED

Reviewed By :Anahy Claudio 03/14/2025

Supervised By :Jagrut Upadhyay 03/14/2025

Quant Time: Mar 13 14:43:18 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM031325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 13 14:27:21 2025
Response via : Initial Calibration




Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM031325\
Data File : BM049712.D
Acq On : 13 Mar 2025 12:58
Operator : RC/JU
Sample : SSTDICC060
Misc :
ALS Vial : 8 Sample Multiplier: 1


Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED

Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025


Quant Time: Mar 13 14:43:18 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM031325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 13 14:27:21 2025
Response via : Initial Calibration





#1
 1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 7.804 min Scan# 819
 Delta R.T. -0.001 min
 Lab File: BM049712.D
 Acq: 13 Mar 2025 12:58




Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC060



Tgt Ion: 152 Resp: 315799
 Ion Ratio Lower Upper
 152 100
 150 152.8 123.8 185.8
 115 52.7 43.0 64.6




Manual Integrations
 APPROVED


Reviewed By :Anahy Claudio 03/14/2025
 Supervised By :Jagrut Upadhyay 03/14/2025



#2
 1,4-Dioxane
 Concen: 61.795 ng
 RT: 3.287 min Scan# 51
 Delta R.T. -0.000 min
 Lab File: BM049712.D
 Acq: 13 Mar 2025 12:58

Tgt Ion: 88 Resp: 500669
 Ion Ratio Lower Upper
 88 100
 58 71.5 56.7 85.1
 43 27.9 22.6 34.0





#3
Pyridine
Concen: 61.832 ng m
RT: 3.687 min Scan# 119
Delta R.T. -0.000 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58

Instrument :

BNA_M

ClientSampleId :





SSTDICC060

Tgt Ion: 79 Resp: 128101
Ion Ratio Lower Upper
79 100
52 63.3 50.1 75.1
51 32.6 25.2 37.8

Manual Integrations




APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025



#4
n-Nitrosodimethylamine
Concen: 62.064 ng
RT: 3.593 min Scan# 103
Delta R.T. -0.000 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58


Tgt Ion: 42 Resp: 568214
Ion Ratio Lower Upper
42 100
74 139.3 112.6 168.8
44 6.1 5.0 7.6





#5
2-Fluorophenol
Concen: 129.126 ng
RT: 5.387 min Scan# 408
Delta R.T. -0.000 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58




Instrument :
BNA_M
ClientSampleId :
SSTDICC060




Tgt Ion:112 Resp: 241671
Ion Ratio Lower Upper
112 100
64 56.1 45.3 67.9
63 29.9 24.4 36.6

Manual Integrations
APPROVED



Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#6
Aniline
Concen: 59.996 ng
RT: 7.140 min Scan# 706
Delta R.T. 0.005 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58




Tgt Ion: 93 Resp: 1180576
Ion Ratio Lower Upper
93 100
66 38.6 30.6 46.0
65 19.8 15.4 23.2





#7
Phenol-d6
Concen: 127.482 ng
RT: 6.981 min Scan# 679
Delta R.T. -0.000 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58




Instrument :
BNA_M
ClientSampleId :
SSTDICC060



Tgt Ion: 99 Resp: 2994230
Ion Ratio Lower Upper
99 100
42 17.0 13.6 20.4
71 34.0 27.0 40.6




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025



#8
2-Chlorophenol
Concen: 62.692 ng
RT: 7.375 min Scan# 746
Delta R.T. -0.000 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58

Tgt Ion:128 Resp: 1161936
Ion Ratio Lower Upper
128 100
130 31.8 12.5 52.5
64 45.5 24.9 64.9









#9
Benzaldehyde
Concen: 50.156 ng
RT: 6.945 min Scan# 61
Delta R.T. -0.001 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58

Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Tgt Ion: 77 Resp: 58237
Ion Ratio Lower Upper
77 100
105 94.5 75.0 115.0
106 91.0 72.3 112.3




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025



#10
Phenol
Concen: 62.763 ng
RT: 7.010 min Scan# 684
Delta R.T. 0.005 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58


Tgt Ion: 94 Resp: 1429401
Ion Ratio Lower Upper
94 100
65 28.8 10.4 50.4
66 36.1 17.1 57.1





#11
bis(2-Chloroethyl)ether
Concen: 61.698 ng
RT: 7.234 min Scan# 721
Delta R.T. 0.005 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58




Instrument :
BNA_M
ClientSampleId :
SSTDICC060



Tgt Ion: 93 Resp: 113786
Ion Ratio Lower Upper
93 100
63 74.2 54.8 94.8
95 32.4 12.8 52.8




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025



#12
1,3-Dichlorobenzene
Concen: 61.881 ng
RT: 7.698 min Scan# 801
Delta R.T. -0.000 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58

Tgt Ion:146 Resp: 1393806
Ion Ratio Lower Upper
146 100
148 65.1 52.3 78.5
75 29.1 24.2 36.2









#13
1,4-Dichlorobenzene
Concen: 61.813 ng
RT: 7.845 min Scan# 81
Delta R.T. 0.005 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58

Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Tgt Ion:146 Resp: 140455
Ion Ratio Lower Upper
146 100
148 64.9 52.6 79.0
111 37.4 30.8 46.2




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025



#14
1,2-Dichlorobenzene
Concen: 61.320 ng
RT: 8.163 min Scan# 880
Delta R.T. 0.005 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58


Tgt Ion:146 Resp: 1324220
Ion Ratio Lower Upper
146 100
148 66.3 50.9 76.3
111 38.8 31.0 46.6





#15
Benzyl Alcohol
Concen: 64.048 ng
RT: 8.045 min Scan# 80
Delta R.T. -0.000 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58




Instrument :
BNA_M
ClientSampleId :
SSTDICC060




Tgt Ion: 79 Resp: 944038
Ion Ratio Lower Upper
79 100
108 75.3 60.7 91.1
77 66.3 55.4 83.2

Manual Integrations
APPROVED



Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#16
2,2'-oxybis(1-Chloropropane)
Concen: 59.613 ng
RT: 8.345 min Scan# 911
Delta R.T. 0.005 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58



Tgt Ion: 45 Resp: 1292162
Ion Ratio Lower Upper
45 100
77 16.2 0.0 36.1
79 12.8 0.0 32.8






#17
2-Methylphenol
Concen: 62.967 ng
RT: 8.257 min Scan# 896
Delta R.T. 0.005 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58

Instrument :

BNA_M

ClientSampleId :

SSTDICC060






Tgt Ion:107 Resp: 85139
Ion Ratio Lower Upper
107 100
108 113.3 89.6 134.4
77 48.3 39.8 59.6
79 46.3 37.4 56.2


Manual Integrations

APPROVED



Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#18
Hexachloroethane
Concen: 61.468 ng
RT: 8.892 min Scan# 1004
Delta R.T. 0.005 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58




Tgt Ion:117 Resp: 493239
Ion Ratio Lower Upper
117 100
119 95.8 73.1 109.7
201 122.6 90.0 135.0





#19
n-Nitroso-di-n-propylamine
Concen: 63.565 ng
RT: 8.616 min Scan# 957
Delta R.T. -0.000 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58




Instrument :
BNA_M
ClientSampleId :
SSTDICC060



Tgt Ion: 70 Resp: 86571
Ion Ratio Lower Upper
70 100
42 51.8 41.8 62.6
101 8.8 7.7 11.5
130 19.7 16.2 24.4




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025



#20
3+4-Methylphenols
Concen: 63.315 ng
RT: 8.581 min Scan# 951
Delta R.T. -0.000 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58


Tgt Ion:107 Resp: 1147270
Ion Ratio Lower Upper
107 100
108 90.2 71.0 111.0
77 34.6 15.1 55.1
79 28.9 11.2 51.2





#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.604 min Scan# 1295
Delta R.T. -0.000 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58




Instrument :
BNA_M
ClientSampleId :
SSTDICC060



Tgt Ion:136 Resp: 1024759
Ion Ratio Lower Upper
136 100
137 10.9 8.6 12.8
54 6.8 5.8 8.8
68 5.3 4.5 6.7




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025



#22
Acetophenone
Concen: 63.809 ng
RT: 8.628 min Scan# 959
Delta R.T. -0.000 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58

Tgt Ion:105 Resp: 1686019
Ion Ratio Lower Upper
105 100
71 5.1 2.5 3.7#
51 24.7 19.9 29.9
120 22.8 18.3 27.5





#23

Nitrobenzene-d5

Concen: 129.447 ng

RT: 8.963 min Scan# 1016

Delta R.T. -0.000 min

Lab File: BM049712.D

Acq: 13 Mar 2025 12:58

Instrument :

BNA_M

ClientSampleId :

SSTDICC060

Tgt Ion: 82 Resp: 2583700

Ion Ratio Lower Upper

82 100

128 42.2 33.4 50.0





54 46.1 36.4 54.6

Manual Integrations

APPROVED

Reviewed By :Anahy Claudio 03/14/2025

Supervised By :Jagrut Upadhyay 03/14/2025



#24

Nitrobenzene

Concen: 63.615 ng

RT: 9.004 min Scan# 1023

Delta R.T. -0.000 min

Lab File: BM049712.D

Acq: 13 Mar 2025 12:58




Tgt Ion: 77 Resp: 1197256


Ion Ratio Lower Upper

77 100

123 44.2 34.9 52.3

65 13.3 10.6 15.8





#25

Isophorone

Concen: 64.875 ng

RT: 9.533 min Scan# 1113

Delta R.T. -0.000 min

Lab File: BM049712.D

Acq: 13 Mar 2025 12:58

Instrument :

BNA_M

ClientSampleId :

SSTDICC060

Tgt Ion: 82 Resp: 1988098

Ion Ratio Lower Upper

82 100

95 6.9 5.3 7.9





138 17.6 14.4 21.6

Manual Integrations

APPROVED

Reviewed By :Anahy Claudio 03/14/2025

Supervised By :Jagrut Upadhyay 03/14/2025



#26

2-Nitrophenol

Concen: 69.411 ng

RT: 9.716 min Scan# 1144

Delta R.T. -0.000 min

Lab File: BM049712.D

Acq: 13 Mar 2025 12:58




Tgt Ion:139 Resp: 556011


Ion Ratio Lower Upper

139 100

109 28.6 23.3 34.9


65 48.3 38.6 58.0





#27
2,4-Dimethylphenol
Concen: 65.658 ng
RT: 9.781 min Scan# 1155
Delta R.T. -0.000 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58




Instrument :
BNA_M
ClientSampleId :
SSTDICC060




Tgt Ion:122 Resp: 669470
Ion Ratio Lower Upper
122 100
107 113.5 94.3 141.5
121 57.4 45.4 68.2

Manual Integrations
APPROVED



Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#28
bis(2-Chloroethoxy)methane
Concen: 62.714 ng
RT: 10.016 min Scan# 1195
Delta R.T. -0.000 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58



Tgt Ion: 93 Resp: 1353615
Ion Ratio Lower Upper
93 100
95 33.2 26.6 40.0
123 11.4 9.0 13.6









#29
2,4-Dichlorophenol
Concen: 65.322 ng
RT: 10.257 min Scan# 11
Delta R.T. 0.005 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58

Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Tgt Ion:162 Resp: 108529
Ion Ratio Lower Upper
162 100
164 65.5 46.3 86.3
98 31.1 12.8 52.8




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025



#30
1,2,4-Trichlorobenzene
Concen: 63.227 ng
RT: 10.469 min Scan# 1272
Delta R.T. 0.005 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58


Tgt Ion:180 Resp: 1314305
Ion Ratio Lower Upper
180 100
182 98.2 77.7 116.5
145 25.8 21.0 31.4





#31
Naphthalene
Concen: 63.997 ng
RT: 10.651 min Scan# 1180
Delta R.T. -0.000 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58




Instrument :
BNA_M
ClientSampleId :
SSTDICC060



Tgt Ion:128 Resp: 3372239
Ion Ratio Lower Upper
128 100
129 11.0 8.6 13.0
127 13.0 10.4 15.6




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025



#32
Benzoic acid
Concen: 70.482 ng
RT: 9.933 min Scan# 1181
Delta R.T. 0.023 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58

Tgt Ion:122 Resp: 611385
Ion Ratio Lower Upper
122 100
105 128.5 105.6 145.6
77 98.8 76.9 116.9






#33
4-Chloroaniline
Concen: 63.446 ng
RT: 10.763 min Scan# 11
Delta R.T. 0.005 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58

Instrument :

BNA_M

ClientSampleId :

SSTDICC060






Tgt Ion:127 Resp: 114096
Ion Ratio Lower Upper
127 100
129 32.8 25.7 38.5
65 29.6 24.1 36.1
92 18.1 14.5 21.7


Manual Integrations

APPROVED



Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#34
Hexachlorobutadiene
Concen: 63.912 ng
RT: 10.945 min Scan# 1353
Delta R.T. -0.000 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58




Tgt Ion:225 Resp: 820160
Ion Ratio Lower Upper
225 100
223 61.7 49.8 74.8
227 62.9 52.1 78.1





#35
 Caprolactam
 Concen: 69.264 ng
 RT: 11.551 min Scan# 1456
 Delta R.T. 0.011 min
 Lab File: BM049712.D
 Acq: 13 Mar 2025 12:58




Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC060




Tgt Ion:113 Resp: 26717
 Ion Ratio Lower Upper
 113 100
 55 161.5 141.8 181.8
 56 122.3 98.0 138.0

Manual Integrations
 APPROVED



Reviewed By :Anahy Claudio 03/14/2025
 Supervised By :Jagrut Upadhyay 03/14/2025




#36
 4-Chloro-3-methylphenol
 Concen: 66.827 ng
 RT: 11.892 min Scan# 1514
 Delta R.T. -0.000 min
 Lab File: BM049712.D
 Acq: 13 Mar 2025 12:58




Tgt Ion:107 Resp: 952831
 Ion Ratio Lower Upper
 107 100
 144 28.5 23.0 34.6
 142 86.7 69.8 104.8





#37
2-Methylnaphthalene
Concen: 64.061 ng
RT: 12.269 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58




Instrument :
BNA_M
ClientSampleId :
SSTDICC060




Tgt Ion:142 Resp: 2360362
Ion Ratio Lower Upper
142 100
141 88.5 70.5 105.7
115 33.4 26.5 39.7

Manual Integrations
APPROVED



Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#38
1-Methylnaphthalene
Concen: 64.106 ng
RT: 12.486 min Scan# 1615
Delta R.T. -0.000 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58




Tgt Ion:142 Resp: 2278919
Ion Ratio Lower Upper
142 100
141 90.5 73.4 110.2
116 3.6 2.8 4.2





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.445 min Scan# 1948
Delta R.T. -0.000 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58




Instrument :
BNA_M
ClientSampleId :
SSTDICC060



Tgt Ion:164 Resp: 618823
Ion Ratio Lower Upper
164 100
162 99.5 80.5 120.7
160 44.9 36.9 55.3




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025



#40
1,2,4,5-Tetrachlorobenzene
Concen: 65.259 ng
RT: 12.639 min Scan# 1641
Delta R.T. 0.005 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58


Tgt Ion:216 Resp: 1483366
Ion Ratio Lower Upper
216 100
214 78.1 63.5 95.3
179 19.1 15.7 23.5
108 14.8 12.2 18.2





#41
Hexachlorocyclopentadiene
Concen: 66.191 ng
RT: 12.622 min Scan# 1638
Delta R.T. -0.000 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58




Instrument :
BNA_M
ClientSampleId :
SSTDICC060




Tgt Ion:237 Resp: 557389
Ion Ratio Lower Upper
237 100
235 63.4 41.1 81.1
272 13.5 0.0 32.5

Manual Integrations
APPROVED



Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#42
2,4,6-Tribromophenol
Concen: 148.622 ng
RT: 15.933 min Scan# 2201
Delta R.T. 0.000 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58




Tgt Ion:330 Resp: 1073944
Ion Ratio Lower Upper
330 100
332 96.7 77.0 115.4
141 32.4 27.8 41.6





#43
2,4,6-Trichlorophenol
Concen: 67.864 ng
RT: 12.880 min Scan# 1682
Delta R.T. 0.005 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58




Instrument :
BNA_M
ClientSampleId :
SSTDICC060




Tgt Ion:196 Resp: 86395
Ion Ratio Lower Upper
196 100
198 95.4 78.3 117.5
200 31.3 25.1 37.7

Manual Integrations
APPROVED



Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#44
2,4,5-Trichlorophenol
Concen: 67.377 ng
RT: 12.951 min Scan# 1694
Delta R.T. -0.000 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58




Tgt Ion:196 Resp: 936029
Ion Ratio Lower Upper
196 100
198 96.7 76.7 115.1
97 41.6 32.9 49.3
132 21.6 16.8 25.2





#45
2-Fluorobiphenyl
Concen: 131.059 ng
RT: 13.069 min Scan# 1714
Delta R.T. -0.000 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58




Instrument :
BNA_M
ClientSampleId :
SSTDICC060




Tgt Ion:172 Resp: 647586
Ion Ratio Lower Upper
172 100
171 34.7 27.9 41.9
170 23.4 18.6 28.0

Manual Integrations
APPROVED



Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#46
1,1'-Biphenyl
Concen: 63.921 ng
RT: 13.280 min Scan# 1750
Delta R.T. -0.000 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58



Tgt Ion:154 Resp: 3048618
Ion Ratio Lower Upper
154 100
153 40.0 19.7 59.7
76 10.9 0.0 31.2






#47
2-Chloronaphthalene
Concen: 64.138 ng
RT: 13.321 min Scan# 1757
Delta R.T. -0.000 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58

Instrument :

BNA_M

ClientSampleId :

SSTDICC060






Tgt Ion:162 Resp: 2381008
Ion Ratio Lower Upper
162 100
127 32.2 25.8 38.6
164 33.2 27.1 40.7


Manual Integrations

APPROVED



Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#48
2-Nitroaniline
Concen: 69.424 ng
RT: 13.521 min Scan# 1791
Delta R.T. -0.000 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58




Tgt Ion: 65 Resp: 578552
Ion Ratio Lower Upper
65 100
92 69.3 53.4 80.2
138 110.0 87.6 131.4





#49
Acenaphthylene
Concen: 65.227 ng
RT: 14.168 min Scan# 1901
Delta R.T. -0.000 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58




Instrument :
BNA_M
ClientSampleId :
SSTDICC060



Tgt Ion:152 Resp: 3384598
Ion Ratio Lower Upper
152 100
151 19.6 15.8 23.8
153 13.1 10.2 15.4




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025



#50
Dimethylphthalate
Concen: 63.771 ng
RT: 13.910 min Scan# 1857
Delta R.T. 0.005 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58

Tgt Ion:163 Resp: 2716983
Ion Ratio Lower Upper
163 100
194 4.4 3.4 5.0
164 10.0 8.2 12.2









#51
2,6-Dinitrotoluene
Concen: 66.034 ng
RT: 14.021 min Scan# 1875
Delta R.T. 0.005 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58

Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Tgt Ion:165 Resp: 56372
Ion Ratio Lower Upper
165 100
63 49.5 42.3 63.5
89 51.1 43.9 65.9




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025



#52
Acenaphthene
Concen: 65.221 ng
RT: 14.510 min Scan# 1959
Delta R.T. -0.000 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58


Tgt Ion:154 Resp: 2217160
Ion Ratio Lower Upper
154 100
153 112.2 90.4 135.6
152 54.7 44.2 66.2





#53
3-Nitroaniline
Concen: 68.488 ng
RT: 14.345 min Scan# 1931
Delta R.T. -0.000 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58




Instrument :
BNA_M
ClientSampleId :
SSTDICC060




Tgt Ion:138 Resp: 556379
Ion Ratio Lower Upper
138 100
108 9.8 7.5 11.3
92 112.1 88.0 132.0

Manual Integrations
APPROVED



Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#54
2,4-Dinitrophenol
Concen: 60.822 ng
RT: 14.551 min Scan# 1966
Delta R.T. 0.006 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58




Tgt Ion:184 Resp: 295154
Ion Ratio Lower Upper
184 100
63 56.9 47.2 70.8
154 60.6 49.9 74.9





#55
Dibenzofuran
Concen: 64.040 ng
RT: 14.845 min Scan# 2016
Delta R.T. -0.000 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58




Instrument :
BNA_M
ClientSampleId :
SSTDICC060




Tgt Ion:168 Resp: 3632034
Ion Ratio Lower Upper
168 100
139 33.5 27.4 41.0
169 13.4 10.7 16.1

Manual Integrations
APPROVED



Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#56
4-Nitrophenol
Concen: 67.524 ng
RT: 14.663 min Scan# 1985
Delta R.T. 0.006 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58



Tgt Ion:139 Resp: 477486
Ion Ratio Lower Upper
139 100
109 76.2 54.7 94.7
65 95.3 76.2 116.2






#57
2,4-Dinitrotoluene
Concen: 69.539 ng
RT: 14.804 min Scan# 2009
Delta R.T. -0.000 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58

Instrument :

BNA_M

ClientSampleId :

SSTDICC060






Tgt Ion:165 Resp: 763894
Ion Ratio Lower Upper
165 100
63 39.0 30.8 46.2
89 65.4 53.1 79.7
182 2.5 2.1 3.1

Manual Integrations




APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025



#58
Fluorene
Concen: 66.168 ng
RT: 15.492 min Scan# 2126
Delta R.T. -0.000 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58


Tgt Ion:166 Resp: 3055405
Ion Ratio Lower Upper
166 100
165 96.6 77.5 116.3
167 13.8 10.9 16.3





#59
2,3,4,6-Tetrachlorophenol
Concen: 66.821 ng
RT: 15.074 min Scan# 2054
Delta R.T. 0.005 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58




Instrument :
BNA_M
ClientSampleId :
SSTDICC060



Tgt Ion:232 Resp: 777060
Ion Ratio Lower Upper
232 100
131 37.2 30.0 45.0
130 2.1 1.6 2.4
166 28.2 22.5 33.7




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025



#60
Diethylphthalate
Concen: 64.207 ng
RT: 15.274 min Scan# 2089
Delta R.T. 0.005 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58


Tgt Ion:149 Resp: 2576429
Ion Ratio Lower Upper
149 100
177 23.5 18.2 27.2
150 12.2 10.0 15.0





#61
4-Chlorophenyl-phenylether
Concen: 66.402 ng
RT: 15.492 min Scan# 2126
Delta R.T. -0.000 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58




Instrument :
BNA_M
ClientSampleId :
SSTDICC060




Tgt Ion:204 Resp: 1674350
Ion Ratio Lower Upper
204 100
206 33.0 25.2 37.8
141 42.1 34.5 51.7

Manual Integrations
APPROVED



Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#62
4-Nitroaniline
Concen: 60.865 ng
RT: 15.510 min Scan# 2129
Delta R.T. 0.006 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58




Tgt Ion:138 Resp: 598859
Ion Ratio Lower Upper
138 100
92 48.6 29.8 69.8
108 78.6 61.3 101.3





#63
Azobenzene
Concen: 64.760 ng
RT: 15.780 min Scan# 2175
Delta R.T. -0.000 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58




Instrument :
BNA_M
ClientSampleId :
SSTDICC060



Tgt Ion: 77 Resp: 2482118
Ion Ratio Lower Upper
77 100
182 27.8 8.0 48.0
105 19.7 0.0 40.0
51 27.4 7.3 47.3




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.186 min Scan# 2414
Delta R.T. -0.000 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58


Tgt Ion:188 Resp: 1169651
Ion Ratio Lower Upper
188 100
94 7.1 5.8 8.6
80 8.3 6.3 9.5





#65
4,6-Dinitro-2-methylphenol
Concen: 61.103 ng
RT: 15.568 min Scan# 2139
Delta R.T. -0.001 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58




Instrument :
BNA_M
ClientSampleId :
SSTDICC060




Tgt Ion:198 Resp: 417991
Ion Ratio Lower Upper
198 100
51 36.1 13.8 53.8
105 36.4 14.0 54.0

Manual Integrations
APPROVED



Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#66
n-Nitrosodiphenylamine
Concen: 64.980 ng
RT: 15.704 min Scan# 2162
Delta R.T. -0.000 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58



Tgt Ion:169 Resp: 2302780
Ion Ratio Lower Upper
169 100
168 68.3 54.8 82.2
167 37.4 29.9 44.9









#67
4-Bromophenyl-phenylether
Concen: 65.299 ng
RT: 16.380 min Scan# 2277
Delta R.T. -0.000 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58

Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Tgt Ion:248 Resp: 881819
Ion Ratio Lower Upper
248 100
250 98.8 77.9 116.9
141 60.7 48.6 73.0




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025



#68
Hexachlorobenzene
Concen: 64.964 ng
RT: 16.498 min Scan# 2297
Delta R.T. -0.000 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58


Tgt Ion:284 Resp: 986543
Ion Ratio Lower Upper
284 100
142 31.0 25.5 38.3
249 30.9 24.2 36.2





#70
Pentachlorophenol
Concen: 67.839 ng
RT: 16.839 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58




Instrument :
BNA_M
ClientSampleId :
SSTDICC060



Tgt Ion:266 Resp: 647404
Ion Ratio Lower Upper
266 100
268 61.5 50.2 75.2
264 62.3 51.3 76.9




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025



#71
Phenanthrene
Concen: 65.716 ng
RT: 17.227 min Scan# 2421
Delta R.T. -0.000 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58


Tgt Ion:178 Resp: 4268082
Ion Ratio Lower Upper
178 100
176 19.6 15.5 23.3
179 15.5 12.5 18.7





#72
 Anthracene
 Concen: 66.580 ng
 RT: 17.321 min Scan# 2437
 Delta R.T. -0.000 min
 Lab File: BM049712.D
 Acq: 13 Mar 2025 12:58




Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC060




Tgt Ion:178 Resp: 418783
 Ion Ratio Lower Upper
 178 100
 176 18.8 15.1 22.7
 179 15.7 12.4 18.6

Manual Integrations
 APPROVED



Reviewed By :Anahy Claudio 03/14/2025
 Supervised By :Jagrut Upadhyay 03/14/2025




#73
 Carbazole
 Concen: 67.182 ng
 RT: 17.592 min Scan# 2483
 Delta R.T. 0.005 min
 Lab File: BM049712.D
 Acq: 13 Mar 2025 12:58




Tgt Ion:167 Resp: 3826935
 Ion Ratio Lower Upper
 167 100
 166 21.6 17.0 25.6
 139 11.4 9.4 14.2





#74
Di-n-butylphthalate
Concen: 67.499 ng
RT: 18.162 min Scan# 2580
Delta R.T. -0.000 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58




Instrument :
BNA_M
ClientSampleId :
SSTDICC060



Tgt Ion:149 Resp: 4214218
Ion Ratio Lower Upper
149 100
150 9.2 7.2 10.8
104 4.8 3.9 5.9




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025



#75
Fluoranthene
Concen: 69.531 ng
RT: 19.245 min Scan# 2764
Delta R.T. -0.000 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58


Tgt Ion:202 Resp: 5199093
Ion Ratio Lower Upper
202 100
101 7.5 0.0 27.8
203 17.7 0.0 37.6





#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.427 min Scan# 3135
Delta R.T. -0.000 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58




Instrument :
BNA_M
ClientSampleId :
SSTDICC060



Tgt Ion:240 Resp: 117348
Ion Ratio Lower Upper
240 100
120 7.9 6.5 9.7
236 26.6 20.7 31.1




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025



#77
Benzidine
Concen: 54.727 ng
RT: 19.427 min Scan# 2795
Delta R.T. -0.000 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58

Tgt Ion:184 Resp: 771546
Ion Ratio Lower Upper
184 100
185 14.3 11.7 17.5
183 11.2 8.8 13.2





#78

Pyrene

Concen: 66.127 ng

RT: 19.609 min Scan# 2825

Delta R.T. 0.005 min

Lab File: BM049712.D

Acq: 13 Mar 2025 12:58

Instrument :

BNA_M

ClientSampleId :

SSTDICC060

Tgt Ion:202 Resp: 542153

Ion Ratio Lower Upper

202 100

200 21.2 17.0 25.4





203 17.6 14.0 21.0

Manual Integrations

APPROVED

Reviewed By :Anahy Claudio 03/14/2025

Supervised By :Jagrut Upadhyay 03/14/2025



#79

Terphenyl-d14

Concen: 138.909 ng

RT: 19.809 min Scan# 2860

Delta R.T. -0.000 min

Lab File: BM049712.D

Acq: 13 Mar 2025 12:58




Tgt Ion:244 Resp: 9673323


Ion Ratio Lower Upper

244 100

212 7.3 5.8 8.6


122 7.2 6.0 9.0





#80
Butylbenzylphthalate
Concen: 67.988 ng
RT: 20.509 min Scan# 2979
Delta R.T. -0.000 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58




Instrument :
BNA_M
ClientSampleId :
SSTDICC060



Tgt Ion:149 Resp: 1729830
Ion Ratio Lower Upper
149 100
91 65.5 52.2 78.4
206 22.4 17.6 26.4




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025



#81
Benzo(a)anthracene
Concen: 65.644 ng
RT: 21.409 min Scan# 3132
Delta R.T. -0.000 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58


Tgt Ion:228 Resp: 5145212
Ion Ratio Lower Upper
228 100
226 28.0 22.0 33.0
229 20.0 15.8 23.8





#82
3,3'-Dichlorobenzidine
Concen: 66.120 ng
RT: 21.333 min Scan# 3119
Delta R.T. -0.000 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58




Instrument :
BNA_M
ClientSampleId :
SSTDICC060




Tgt Ion:252 Resp: 1687314
Ion Ratio Lower Upper
252 100
254 64.9 51.0 76.4
126 7.5 6.3 9.5

Manual Integrations
APPROVED



Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#83
Chrysene
Concen: 65.843 ng
RT: 21.474 min Scan# 3143
Delta R.T. 0.005 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58




Tgt Ion:228 Resp: 4892160
Ion Ratio Lower Upper
228 100
226 30.3 24.5 36.7
229 19.6 15.8 23.6





#84
Bis(2-ethylhexyl)phthalate
Concen: 66.414 ng
RT: 21.339 min Scan# 3120
Delta R.T. -0.000 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58




Instrument :
BNA_M
ClientSampleId :
SSTDICC060




Tgt Ion:149 Resp: 2633480
Ion Ratio Lower Upper
149 100
167 28.9 22.9 34.3
279 4.8 3.5 5.3

Manual Integrations
APPROVED



Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#85
Di-n-octyl phthalate
Concen: 64.539 ng
RT: 22.480 min Scan# 3314
Delta R.T. -0.000 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58



Tgt Ion:149 Resp: 3977588
Ion Ratio Lower Upper
149 100
167 1.6 1.3 1.9
43 8.2 6.8 10.2





#86

Perylene-d12

Concen: 20.000 ng

RT: 24.438 min Scan# 30

Delta R.T. -0.000 min

Lab File: BM049712.D

Acq: 13 Mar 2025 12:58

Instrument :

BNA_M

ClientSampleId :

SSTDICC060

Tgt Ion:264 Resp: 1084370

Ion Ratio Lower Upper

264 100

260 23.7 19.6 29.4





265 22.2 18.2 27.2

Manual Integrations

APPROVED

Reviewed By :Anahy Claudio 03/14/2025

Supervised By :Jagrut Upadhyay 03/14/2025



#87

Indeno(1,2,3-cd)pyrene

Concen: 67.923 ng

RT: 27.856 min Scan# 4228

Delta R.T. 0.011 min

Lab File: BM049712.D

Acq: 13 Mar 2025 12:58




Tgt Ion:276 Resp: 5301272


Ion Ratio Lower Upper

276 100

138 20.0 16.3 24.5

277 24.9 19.9 29.9





#88

Benzo(b)fluoranthene

Concen: 67.158 ng

RT: 23.491 min Scan# 3486

Delta R.T. -0.000 min

Lab File: BM049712.D

Acq: 13 Mar 2025 12:58

Instrument :

BNA_M

ClientSampleId :

SSTDICC060





Tgt Ion:252 Resp: 490497

Ion	Ratio	Lower	Upper
252	100		
253	21.8	17.4	26.0
125	7.3	5.9	8.9

Manual Integrations

APPROVED

Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025



#89

Benzo(k)fluoranthene

Concen: 66.934 ng

RT: 23.556 min Scan# 3497




Delta R.T. 0.005 min


Lab File: BM049712.D

Acq: 13 Mar 2025 12:58

Tgt Ion:252 Resp: 4810286


Ion	Ratio	Lower	Upper
252	100		
253	21.7	17.3	25.9
125	7.0	6.2	9.2





#90
Benzo(a)pyrene
Concen: 67.883 ng
RT: 24.303 min Scan# 30
Delta R.T. 0.006 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58




Instrument :
BNA_M
ClientSampleId :
SSTDICC060




Tgt Ion:252 Resp: 416443
Ion Ratio Lower Upper
252 100
253 21.8 17.2 25.8
125 8.0 7.0 10.6

Manual Integrations
APPROVED



Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#91
Dibenzo(a,h)anthracene
Concen: 67.722 ng
RT: 27.909 min Scan# 4237
Delta R.T. 0.005 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58




Tgt Ion:278 Resp: 4403728
Ion Ratio Lower Upper
278 100
139 12.3 10.1 15.1
279 23.6 19.1 28.7





#92
Benzo(g,h,i)perylene
Concen: 66.781 ng
RT: 28.915 min Scan# 4408
Delta R.T. 0.011 min
Lab File: BM049712.D
Acq: 13 Mar 2025 12:58

Instrument :
BNA_M
ClientSampleId :
SSTDICC060



Tgt Ion:276 Resp: 4374098
Ion Ratio Lower Upper
276 100
277 24.0 18.8 28.2
138 17.2 14.0 21.0

Manual Integrations
APPROVED

Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025

