

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM031424\
 Data File : BM044611.D
 Acq On : 14 Mar 2024 13:33
 Operator : MA/JU
 Sample : PB159579BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK579

Quant Time: Mar 14 15:02:03 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM030524.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Mar 05 23:05:11 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.704	152	169061	20.000	ng/u1	-0.03
20) Naphthalene-d8	10.481	136	694569	20.000	ng/u1	-0.04
38) Acenaphthene-d10	14.345	164	419475	20.000	ng/u1	-0.02
64) Phenanthrene-d10	17.104	188	882941	20.000	ng/u1	-0.02
79) Chrysene-d12	21.298	240	832050	20.000	ng/u1	-0.01
88) Perylene-d12	23.545	264	984561	20.000	ng/u1	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.258	96	27819	5.581	ng/uL	-0.03
4) Pyridine-d5	3.652	84	383253	26.123	ng/u1	-0.03
7) Phenol-d5	6.875	99	461165	26.532	ng/u1	-0.03
9) Bis-(2-Chloroethyl)eth...	7.051	67	302978	28.220	ng/u1	-0.03
11) 2-Chlorophenol-d4	7.234	132	369738	28.947	ng/u1	-0.04
15) 4-Methylphenol-d8	8.404	113	352220	25.803	ng/u1	-0.04
21) Nitrobenzene-d5	8.863	128	183386	31.355	ng/u1	-0.03
24) 2-Nitrophenol-d4	9.575	143	198249	31.443	ng/u1	-0.03
28) 2,4-Dichlorophenol-d3	10.104	165	313455	29.211	ng/u1	-0.04
31) 4-Chloroaniline-d4	10.628	131	510615	28.651	ng/u1	-0.04
46) Dimethylphthalate-d6	13.769	166	1058641	32.172	ng/u1	-0.02
49) Acenaphthylene-d8	14.039	160	1156399	30.729	ng/u1	-0.02
54) 4-Nitrophenol-d4	14.563	143	189683	27.797	ng/u1	-0.02
60) Fluorene-d10	15.345	176	876821	31.709	ng/u1	-0.02
65) 4,6-Dinitro-2-methylph...	15.474	200	165246	28.832	ng/u1	-0.02
73) Anthracene-d10	17.198	188	1400167	33.176	ng/u1	-0.02
81) Pyrene-d10	19.498	212	1618931	31.724	ng/u1	-0.02
92) Benzo(a)pyrene-d12	23.403	264	1711126	33.065	ng/u1	-0.02

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM031424\
Data File : BM044611.D
Acq On : 14 Mar 2024 13:33
Operator : MA/JU
Sample : PB159579BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK579

Quant Time: Mar 14 15:02:03 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM030524.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Mar 05 23:05:11 2024
Response via : Initial Calibration

