

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM031516\
 Data File : BM004627.D
 Acq On : 16 Mar 2016 09:24
 Operator : UM/SJ
 Sample : SSTDCCC020EC
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02050

Quant Time: Mar 16 13:03:43 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM031516.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Mar 16 07:32:20 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
2	1,4-Dioxane	0.676	0.681	-0.7	103	0.00
3 S	1,4-Dioxane-d8	0.580	0.577	0.5	103	0.00
4	Benzaldehyde	0.911	0.973	-6.8	104	0.00
5 S	Phenol-d5	1.643	1.691	-2.9	105	0.00
6	Phenol	1.729	1.783	-3.1	104	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	0.931	0.964	-3.5	103	0.00
8	Bis(2-Chloroethyl)ether	1.346	1.386	-3.0	102	0.00
9 S	2-Chlorophenol-d4	1.323	1.354	-2.3	105	0.00
10	2-Chlorophenol	1.397	1.435	-2.7	105	0.00
11	2-Methylphenol	1.329	1.350	-1.6	102	0.00
12	2,2'-oxybis(1-Chloropropane	1.610	1.638	-1.7	101	0.00
13 S	4-Methylphenol-d8	1.325	1.335	-0.8	100	0.00
14	Acetophenone	2.067	2.131	-3.1	99	0.00
15 P	N-Nitroso-di-n-propylamine	1.050	1.006	4.2	97	0.00
16	4-Methylphenol	1.463	1.484	-1.4	102	0.00
17	Hexachloroethane	0.555	0.546	1.6	100	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	101	0.00
19 S	Nitrobenzene-d5	0.139	0.134	3.6	96	0.00
20	Nitrobenzene	0.343	0.341	0.6	98	0.00
21	Isophorone	0.657	0.641	2.4	97	0.00
22 S	2-Nitrophenol-d4	0.150	0.142	5.3	93	0.00
23 C	2-Nitrophenol	0.166	0.163	1.8	96	0.00
24	2,4-Dimethylphenol	0.365	0.369	-1.1	100	0.00
25	Bis(2-Chloroethoxy)methane	0.424	0.419	1.2	98	0.00
26 S	2,4-Dichlorophenol-d3	0.295	0.295	0.0	100	0.00
27 C	2,4-Dichlorophenol	0.304	0.307	-1.0	101	0.00
28	Naphthalene	1.043	1.049	-0.6	100	0.00
29 S	4-Chloroaniline-d4	0.353	0.400	-13.3	100	0.00
30	4-Chloroaniline	0.371	0.427	-15.1	101	0.00
31 C	Hexachlorobutadiene	0.181	0.185	-2.2	102	0.00
32	Caprolactam	0.102	0.093	8.8	91	0.00
33 C	4-Chloro-3-methylphenol	0.344	0.337	2.0	97	0.00
34	2-Methylnaphthalene	0.743	0.751	-1.1	101	0.00
35	1-Methylnaphthalene	0.714	0.721	-1.0	100	0.00
36 I	Acenaphthene-d10	1.000	1.000	0.0	99	0.00
37	1,2,4,5-Tetrachlorobenzene	0.639	0.653	-2.2	100	0.00
38	Hexachlorocyclopentadiene	0.346	0.308	11.0	93	0.00
39 C	2,4,6-Trichlorophenol	0.422	0.411	2.6	97	0.00
40	2,4,5-Trichlorophenol	0.455	0.454	0.2	99	0.00
41	1,1'-Biphenyl	1.670	1.689	-1.1	99	0.00
42	2-Chloronaphthalene	1.256	1.271	-1.2	99	0.00
43	2-Nitroaniline	0.338	0.312	7.7	92	0.00
44 S	Dimethylphthalate-d6	1.563	1.530	2.1	96	0.00

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM031516\
 Data File : BM004627.D
 Acq On : 16 Mar 2016 09:24
 Operator : UM/SJ
 Sample : SSTDCCC020EC
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02050

Quant Time: Mar 16 13:03:43 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM031516.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Mar 16 07:32:20 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	Dimethylphthalate	1.587	1.580	0.4	98	0.00
46	2,6-Dinitrotoluene	0.283	0.269	4.9	93	0.00
47 S	Acenaphthylene-d8	1.917	1.948	-1.6	99	0.00
48	Acenaphthylene	2.060	2.116	-2.7	100	0.00
49	3-Nitroaniline	0.305	0.319	-4.6	95	0.00
50 C	Acenaphthene	1.387	1.416	-2.1	100	0.00
51	2,4-Dinitrophenol	0.161	0.124	23.0#	85	0.00
52 S	4-Nitrophenol-d4	0.302	0.283	6.3	91	0.00
53	4-Nitrophenol	0.245	0.237	3.3	93	0.00
54	Dibenzofuran	1.974	2.024	-2.5	101	0.00
55	2,4-Dinitrotoluene	0.426	0.422	0.9	96	0.00
56	2,3,4,6-Tetrachlorophenol	0.401	0.399	0.5	98	0.00
57	Diethylphthalate	1.617	1.570	2.9	95	0.00
58 S	Fluorene-d10	1.363	1.387	-1.8	101	0.00
59	Fluorene	1.576	1.618	-2.7	99	0.00
60	4-Chlorophenyl-phenylether	0.752	0.768	-2.1	98	0.00
61	4-Nitroaniline	0.328	0.342	-4.3	107	0.00
62 I	Phenanthrene-d10	1.000	1.000	0.0	101	0.00
63 S	4,6-Dinitro-2-methylphenol-	0.105	0.088	16.2	90	0.00
64	4,6-Dinitro-2-methylphenol	0.114	0.097	14.9	91	0.00
65	N-Nitrosodiphenylamine	0.606	0.613	-1.2	99	0.00
66	4-Bromophenyl-phenylether	0.201	0.199	1.0	98	0.00
67	Hexachlorobenzene	0.226	0.229	-1.3	101	0.00
68	Atrazine	0.214	0.213	0.5	98	0.00
69 C	Pentachlorophenol	0.147	0.137	6.8	94	0.00
70	Phenanthrene	1.143	1.162	-1.7	101	0.00
71 S	Anthracene-d10	0.921	0.937	-1.7	101	0.00
72	Anthracene	1.151	1.171	-1.7	101	0.00
73	1,2,3,4-Tetrachlorobenzene	0.275	0.277	-0.7	100	0.00
74	Pentachlorobenzene	0.266	0.271	-1.9	101	0.00
75	Carbazole	1.029	1.078	-4.8	99	0.00
76	Di-n-butylphthalate	1.459	1.165	20.2#	87	0.00
77 C	Fluoranthene	1.262	1.357	-7.5	102	0.00
78 I	Chrysene-d12	1.000	1.000	0.0	102	0.00
79 S	Pyrene-d10	0.913	0.915	-0.2	101	0.00
80	Pyrene	1.212	1.230	-1.5	102	0.00
81	Butylbenzylphthalate	0.485	0.473	2.5	99	0.00
82	3,3'-Dichlorobenzidine	0.309	0.357	-15.5	112	0.00
83	Benzo(a)anthracene	1.167	1.194	-2.3	103	0.00
84	Bis(2-ethylhexyl)phthalate	0.689	0.696	-1.0	100	0.00
85	Chrysene	1.115	1.124	-0.8	101	0.00
86 I	Perylene-d12	1.000	1.000	0.0	103	0.00
87	Di-n-octyl phthalate	1.196	1.251	-4.6	102	0.00

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM031516\
Data File : BM004627.D
Acq On : 16 Mar 2016 09:24
Operator : UM/SJ
Sample : SSTDCCC020EC
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTD02050

Quant Time: Mar 16 13:03:43 2016
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM031516.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Mar 16 07:32:20 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.193	1.182	0.9	98	0.00
89	Benzo(k)fluoranthene	1.134	1.173	-3.4	106	0.00
90 S	Benzo(a)pyrene-d12	0.893	0.912	-2.1	103	0.00
91 C	Benzo(a)pyrene	1.143	1.162	-1.7	102	0.00
92	Indeno(1,2,3-cd)pyrene	1.285	1.290	-0.4	100	0.00
93	Dibenzo(a,h)anthracene	1.063	1.073	-0.9	101	0.00
94	Benzo(g,h,i)perylene	1.088	1.080	0.7	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0