


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM031522\
 Data File : BM034240.D
 Acq On : 15 Mar 2022 15:20
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM031522.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Mar 16 02:38:26 2022






AutoFind: Scans 2471, 2472, 2473; Background Corrected with Scan 2464

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	42.0	110232	PASS
68	69	0.00	2	1.9	2281	PASS
69	198	0.00	100	44.8	117611	PASS
70	69	0.00	2	0.5	556	PASS
127	198	10	80	50.0	131360	PASS
197	198	0.00	2	0.6	1506	PASS
198	198	100	100	100.0	262613	PASS
199	198	5	9	6.9	18132	PASS
275	198	10	60	29.6	77803	PASS
365	198	1	100	4.7	12442	PASS
441	198	0.01	100	13.2	34765	PASS
442	442	50	100	100.0	223531	PASS
443	442	15	24	19.2	42851	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM031522\
Data File : BM034240.D
Acq On : 15 Mar 2022 15:20
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP

Quant Time: Mar 16 03:11:57 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM031522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 16 02:38:26 2022
Response via : Initial Calibration



TIC: BM034240.D\data.ms

(70) Pentachlorophenol (C)

16.977min (+ 0.000) 0.00 ng




response 355273

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.30	63.41
264.00	62.90	63.70
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM031522\
 Data File : BM034240.D
 Acq On : 15 Mar 2022 15:20
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP

Quant Time: Mar 16 03:11:57 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM031522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 16 02:38:26 2022
 Response via : Initial Calibration



TIC: BM034240.D\data.ms

(77) Benzidine

19.559min (+ 0.000) 1076612.98 ng

response 1514599

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.60	13.90
183.00	12.20	11.60
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
3/15/2022	BNA_M	<u>BM034240.D</u>
Compound Name	Response	Retention Time
DDT	1075989	20.8
DDD	11051	20.359
DDE	938	19.853
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
11989	1087978	1.10